

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013024\
 Data File : BN029436.D
 Acq On : 31 Jan 2024 05:45
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 31 06:12:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 30 03:38:24 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.361	9.8	90	0.00
3	n-Nitrosodimethylamine	0.400	0.361	9.8	94	0.00
4 S	2-Fluorophenol	0.400	0.374	6.5	99	0.00
5 S	Phenol-d6	0.400	0.374	6.5	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.385	3.8	104	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	100	0.01
8 S	Nitrobenzene-d5	0.400	0.383	4.3	102	0.00
9	Naphthalene	0.400	0.390	2.5	99	0.00
10	Hexachlorobutadiene	0.400	0.410	-2.5	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.390	2.5	102	0.00
12	2-Methylnaphthalene	0.400	0.383	4.3	99	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.400	0.353	11.8	93	0.00
15 S	2-Fluorobiphenyl	0.400	0.402	-0.5	99	0.00
16	Acenaphthylene	0.400	0.380	5.0	97	0.00
17	Acenaphthene	0.400	0.388	3.0	96	0.00
18	Fluorene	0.400	0.387	3.3	97	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	95	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.381	4.8	103	0.00
21	4-Bromophenyl-phenylether	0.400	0.383	4.3	96	0.00
22	Hexachlorobenzene	0.400	0.391	2.3	97	0.00
23	Atrazine	0.400	0.392	2.0	103	0.00
24	Pentachlorophenol	0.400	0.335	16.3	92	0.00
25	Phenanthrene	0.400	0.380	5.0	96	0.00
26	Anthracene	0.400	0.371	7.3	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.372	7.0	97	0.00
28	Fluoranthene	0.400	0.370	7.5	99	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	97	0.00
30	Pyrene	0.400	0.396	1.0	95	0.00
31 S	Terphenyl-d14	0.400	0.395	1.3	96	0.00
32	Benzo(a)anthracene	0.400	0.378	5.5	98	0.00
33	Chrysene	0.400	0.395	1.3	98	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.385	3.8	100	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.362	9.5	96	0.00
37	Benzo(b)fluoranthene	0.400	0.375	6.3	98	0.00
38	Benzo(k)fluoranthene	0.400	0.386	3.5	95	0.00
39 C	Benzo(a)pyrene	0.400	0.369	7.8	94	0.00
40	Dibenzo(a,h)anthracene	0.400	0.368	8.0	97	0.00
41	Benzo(g,h,i)perylene	0.400	0.366	8.5	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4