

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013025\
 Data File : BN036152.D
 Acq On : 31 Jan 2025 03:31
 Operator : RC/JU
 Sample : PB166343BS
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS343

Quant Time: Jan 31 07:37:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN013025.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 30 17:36:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	100932	20.000	ng/u1	0.00
20) Naphthalene-d8	10.587	136	376340	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.434	164	235887	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.181	188	418445	20.000	ng/u1	0.00
79) Chrysene-d12	21.427	240	355916	20.000	ng/u1	0.00
88) Perylene-d12	24.457	264	358451	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	14990	6.052	ng/uL	0.00
4) Pyridine-d5	3.681	84	217290	30.799	ng/u1	0.00
7) Phenol-d5	6.987	99	248492	29.710	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	161760	30.138	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	198121	30.256	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	190945	28.418	ng/u1	0.00
21) Nitrobenzene-d5	8.952	128	89998	29.787	ng/u1	0.00
24) 2-Nitrophenol-d4	9.669	143	102824	29.994	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	184568	29.818	ng/u1	0.00
31) 4-Chloroaniline-d4	10.722	131	212872	26.178	ng/u1	0.00
46) Dimethylphthalate-d6	13.846	166	492654	27.694	ng/u1	0.00
49) Acenaphthylene-d8	14.128	160	565722	28.561	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	82692	26.935	ng/u1	-0.01
60) Fluorene-d10	15.428	176	422696	28.387	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.545	200	82959	29.955	ng/u1	0.00
73) Anthracene-d10	17.281	188	621253	30.563	ng/u1	0.00
81) Pyrene-d10	19.575	212	639323	30.161	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.251	264	563797	29.913	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	31552	11.926	ng/uL	91
5) Pyridine	3.705	79	223007	31.122	ng/u1	97
6) Benzaldehyde	6.934	77	14993	3.002	ng/u1	95
8) Phenol	7.016	94	265702	30.771	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.216	93	207097	30.520	ng/u1	97
12) 2-Chlorophenol	7.369	128	209386	31.370	ng/u1	97
13) 2-Methylphenol	8.258	108	187015	29.331	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.316	45	310697	29.996	ng/u1	98
16) Acetophenone	8.616	105	300280	27.410	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.605	70	161010	27.914	ng/u1	98
18) 4-Methylphenol	8.581	108	200972	28.889	ng/u1	95
19) Hexachloroethane	8.875	117	92733	30.073	ng/u1	97
22) Nitrobenzene	8.993	77	249752	29.992	ng/u1	99
23) Isophorone	9.522	82	430307	29.157	ng/u1	99
25) 2-Nitrophenol	9.705	139	110457	30.113	ng/u1	99
26) 2,4-Dimethylphenol	9.775	107	218517	29.240	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.999	93	255727	29.521	ng/u1	97
29) 2,4-Dichlorophenol	10.252	162	190306	30.517	ng/u1	96
30) Naphthalene	10.640	128	601229	29.694	ng/u1	99
32) 4-Chloroaniline	10.746	127	111958	14.264	ng/u1	99
33) Hexachlorobutadiene	10.928	225	139724	30.696	ng/u1	97
34) Caprolactam	11.534	113	49432	24.493	ng/u1	90
35) 4-Chloro-3-methylphenol	11.899	107	200348	27.966	ng/u1	95
36) 2-Methylnaphthalene	12.251	142	403010	28.396	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013025\
 Data File : BN036152.D
 Acq On : 31 Jan 2025 03:31
 Operator : RC/JU
 Sample : PB166343BS
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS343

Quant Time: Jan 31 07:37:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN013025.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 30 17:36:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.475	142	401901	28.825	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.622	216	231341	30.693	ng/ul	94
40) Hexachlorocyclopentadiene	12.610	237	122304	29.071	ng/ul	90
41) 2,4,6-Trichlorophenol	12.869	196	143907	29.611	ng/ul	98
42) 2,4,5-Trichlorophenol	12.957	196	156637	29.783	ng/ul	97
43) 1,1'-Biphenyl	13.269	154	525997	30.041	ng/ul	98
44) 2-Chloronaphthalene	13.310	162	416709	30.271	ng/ul	97
45) 2-Nitroaniline	13.510	65	137682	29.333	ng/ul	98
47) Dimethylphthalate	13.893	163	499893	28.061	ng/ul	100
48) 2,6-Dinitrotoluene	14.010	165	104155	29.645	ng/ul	98
50) Acenaphthylene	14.157	152	607296	29.136	ng/ul	99
51) 3-Nitroaniline	14.340	138	91295	25.933	ng/ul	91
52) Acenaphthene	14.498	153	443651	29.567	ng/ul	99
53) 2,4-Dinitrophenol	14.540	184	60357	27.340	ng/ul	86
55) 4-Nitrophenol	14.687	109	87756	26.810	ng/ul	93
56) Dibenzofuran	14.834	168	609098	29.322	ng/ul	97
57) 2,4-Dinitrotoluene	14.798	165	149026	28.347	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.069	232	123777	28.193	ng/ul	99
59) Diethylphthalate	15.263	149	511890	27.855	ng/ul	99
61) Fluorene	15.487	166	483256	28.688	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.481	204	253027	29.051	ng/ul	95
63) 4-Nitroaniline	15.504	138	102234	29.683	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.563	198	88488	30.089	ng/ul	96
67) N-Nitrosodiphenylamine	15.692	169	403540	31.774	ng/ul	95
68) 4-Bromophenyl-phenylether	16.375	248	142313	31.744	ng/ul	90
69) Hexachlorobenzene	16.492	284	155925	31.881	ng/ul	95
70) Atrazine	16.645	200	144572	28.882	ng/ul	97
71) Pentachlorophenol	16.839	266	105056	31.467	ng/ul	98
72) Phenanthrene	17.222	178	706185	30.101	ng/ul	99
74) Anthracene	17.316	178	721297	30.552	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.234	216	224094	34.411	ng/uL	96
76) Pentachlorobenzene	14.757	250	204300	32.110	ng/uL	96
77) Carbazole	17.587	167	652557	31.509	ng/ul	98
78) Di-n-butylphthalate	18.151	149	882578	31.994	ng/ul	99
80) Fluoranthene	19.239	202	789867	31.900	ng/ul	98
82) Pyrene	19.604	202	817794	31.155	ng/ul	100
83) Butylbenzylphthalate	20.510	149	341404	30.246	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.333	252	200182	31.847	ng/ul	93
85) Benzo(a)anthracene	21.410	228	730795	29.865	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.339	149	487928	30.087	ng/ul	100
87) Chrysene	21.475	228	700345	30.186	ng/ul	98
89) Di-n-octyl phthalate	22.486	149	783912	26.613	ng/ul	100
90) Benzo(b)fluoranthene	23.492	252	685943	28.941	ng/ul	99
91) Benzo(k)fluoranthene	23.557	252	672090	28.478	ng/ul	99
93) Benzo(a)pyrene	24.316	252	619160	29.889	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.868	276	836215	37.100	ng/ul	99
95) Dibenzo(a,h)anthracene	27.933	278	679599	37.805	ng/ul	97
96) Benzo(g,h,i)perylene	28.945	276	669860	39.049	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.926 ng/uL
RT: 3.299 min Scan# 18
Delta R.T. 0.006 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion: 88 Resp: 31552
Ion Ratio Lower Upper
88 100
43 43.9 29.6 44.4
58 80.3 59.1 88.7



#5
Pyridine
Concen: 31.122 ng/uL
RT: 3.705 min Scan# 88
Delta R.T. 0.012 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion: 79 Resp: 223007
Ion Ratio Lower Upper
79 100
52 81.5 67.3 100.9
51 42.6 36.2 54.4









#6
Benzaldehyde
Concen: 3.002 ng/ul
RT: 6.934 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion: 77 Resp: 14993
Ion Ratio Lower Upper
77 100
105 87.2 74.8 112.2
106 83.9 69.8 104.8



#8
Phenol
Concen: 30.771 ng/ul
RT: 7.016 min Scan# 651
Delta R.T. -0.006 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion: 94 Resp: 265702
Ion Ratio Lower Upper
94 100
65 34.7 27.0 40.4
66 47.9 40.2 60.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 30.520 ng/ul
 RT: 7.216 min Scan# 685
 Delta R.T. -0.000 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31

Instrument :
 BNA_N
 ClientSampleId :
 SLCS343

Tgt Ion: 93 Resp: 207097




Ion	Ratio	Lower	Upper
93	100		
63	80.0	67.0	100.6
95	32.9	25.8	38.6




#12
 2-Chlorophenol
 Concen: 31.370 ng/ul
 RT: 7.369 min Scan# 711
 Delta R.T. -0.006 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31

Tgt Ion:128 Resp: 209386

Ion	Ratio	Lower	Upper
128	100		
64	53.9	41.8	62.6
130	32.9	24.9	37.3









#13
2-Methylphenol
Concen: 29.331 ng/ul
RT: 8.258 min Scan# 862
Delta R.T. -0.006 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:108 Resp: 187015
Ion Ratio Lower Upper
108 100
107 90.3 73.7 110.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 29.996 ng/ul
RT: 8.316 min Scan# 872
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion: 45 Resp: 310697
Ion Ratio Lower Upper
45 100
77 13.5 11.8 17.6
79 10.0 8.3 12.5









#16
Acetophenone
Concen: 27.410 ng/ul
RT: 8.616 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:105 Resp: 300280
Ion Ratio Lower Upper
105 100
77 85.4 68.7 103.1
51 35.8 28.9 43.3



#17
N-Nitroso-di-n-propylamine
Concen: 27.914 ng/ul
RT: 8.605 min Scan# 921
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion: 70 Resp: 161010
Ion Ratio Lower Upper
70 100
42 66.9 54.7 82.1
101 9.1 7.1 10.7
130 18.5 15.7 23.5





#18

4-Methylphenol

Concen: 28.889 ng/ul

RT: 8.581 min Scan# 918

Delta R.T. -0.006 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31

Instrument :

BNA_N

ClientSampleId :





SLCS343

Tgt Ion:108 Resp: 200972

Ion Ratio Lower Upper

108 100

107 113.3 95.4 143.2



#19

Hexachloroethane

Concen: 30.073 ng/ul

RT: 8.875 min Scan# 967

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31




Tgt Ion:117 Resp: 92733


Ion Ratio Lower Upper

117 100

201 94.7 77.5 116.3

199 59.8 45.8 68.6









#22
Nitrobenzene
Concen: 29.992 ng/ul
RT: 8.993 min Scan# 987
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion: 77 Resp: 249752
Ion Ratio Lower Upper
77 100
123 39.8 32.5 48.7
65 14.4 12.2 18.2



#23
Isophorone
Concen: 29.157 ng/ul
RT: 9.522 min Scan# 1077
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion: 82 Resp: 430307
Ion Ratio Lower Upper
82 100
95 8.2 6.6 10.0
138 15.6 12.8 19.2









#25
2-Nitrophenol
Concen: 30.113 ng/ul
RT: 9.705 min Scan# 1108
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:139 Resp: 110457
Ion Ratio Lower Upper
139 100
65 63.7 51.4 77.2
109 38.3 31.6 47.4



#26
2,4-Dimethylphenol
Concen: 29.240 ng/ul
RT: 9.775 min Scan# 1120
Delta R.T. -0.006 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:107 Resp: 218517
Ion Ratio Lower Upper
107 100
121 46.0 36.6 55.0
122 78.0 61.4 92.0





#27

Bis(2-Chloroethoxy)methane

Concen: 29.521 ng/ul

RT: 9.999 min Scan# 1158

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31

Instrument :

BNA_N

ClientSampleId :

SLCS343





Tgt Ion: 93 Resp: 255727

Ion Ratio Lower Upper

93 100

95 33.5 25.4 38.0

123 11.0 8.2 12.4



#29

2,4-Dichlorophenol

Concen: 30.517 ng/ul

RT: 10.252 min Scan# 1201

Delta R.T. -0.006 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31




Tgt Ion:162 Resp: 190306


Ion Ratio Lower Upper

162 100

164 67.0 49.7 74.5

98 38.9 30.8 46.2









#30
Naphthalene
Concen: 29.694 ng/ul
RT: 10.640 min Scan# 1267
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:128 Resp: 601229
Ion Ratio Lower Upper
128 100
129 10.7 9.0 13.4
127 13.9 11.4 17.2



#32
4-Chloroaniline
Concen: 14.264 ng/ul
RT: 10.746 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:127 Resp: 111958
Ion Ratio Lower Upper
127 100
129 31.7 25.8 38.8





#33

Hexachlorobutadiene

Concen: 30.696 ng/ul

RT: 10.928 min Scan# 11

Delta R.T. -0.006 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31

Instrument :

BNA_N

ClientSampleId :

SLCS343





Tgt Ion:225 Resp: 139724

Ion Ratio Lower Upper

225 100

223 63.7 48.5 72.7

227 60.0 49.2 73.8



#34

Caprolactam

Concen: 24.493 ng/ul

RT: 11.534 min Scan# 1419

Delta R.T. 0.006 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31




Tgt Ion:113 Resp: 49432


Ion Ratio Lower Upper

113 100

55 168.0 143.6 215.4

56 133.3 119.0 178.6









#35
4-Chloro-3-methylphenol
Concen: 27.966 ng/ul
RT: 11.899 min Scan# 1481
Delta R.T. -0.006 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:107 Resp: 200348
Ion Ratio Lower Upper
107 100
144 24.0 17.5 26.3
142 74.8 56.7 85.1



#36
2-Methylnaphthalene
Concen: 28.396 ng/ul
RT: 12.251 min Scan# 1541
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:142 Resp: 403010
Ion Ratio Lower Upper
142 100
141 90.9 73.5 110.3





#37

1-Methylnaphthalene

Concen: 28.825 ng/ul

RT: 12.475 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31

Instrument :

BNA_N

ClientSampleId :

SLCS343





Tgt Ion:142 Resp: 401901

Ion Ratio Lower Upper

142 100

141 96.2 74.7 112.1

116 4.3 3.9 5.9



#39

1,2,4,5-Tetrachlorobenzene

Concen: 30.693 ng/ul

RT: 12.622 min Scan# 1604

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31

Tgt Ion:216 Resp: 231341




Ion Ratio Lower Upper


216 100

214 72.4 63.0 94.6

179 22.1 18.6 27.8

108 19.1 16.7 25.1









#40
Hexachlorocyclopentadiene
Concen: 29.071 ng/ul
RT: 12.610 min Scan# 1602
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:237 Resp: 122304
Ion Ratio Lower Upper
237 100
235 64.3 44.7 67.1
272 12.0 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 29.611 ng/ul
RT: 12.869 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:196 Resp: 143907
Ion Ratio Lower Upper
196 100
198 100.1 81.8 122.6
200 32.7 25.0 37.4









#42
2,4,5-Trichlorophenol
Concen: 29.783 ng/ul
RT: 12.957 min Scan# 1
Delta R.T. -0.012 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:196 Resp: 156637
Ion Ratio Lower Upper
196 100
198 99.3 77.0 115.6
200 30.5 24.7 37.1



#43
1,1'-Biphenyl
Concen: 30.041 ng/ul
RT: 13.269 min Scan# 1714
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:154 Resp: 525997
Ion Ratio Lower Upper
154 100
153 40.3 33.4 50.2
76 12.9 10.0 15.0









#44
2-Chloronaphthalene
Concen: 30.271 ng/ul
RT: 13.310 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:162 Resp: 416709
Ion Ratio Lower Upper
162 100
164 34.1 27.4 41.0
127 40.2 34.5 51.7



#45
2-Nitroaniline
Concen: 29.333 ng/ul
RT: 13.510 min Scan# 1755
Delta R.T. -0.006 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion: 65 Resp: 137682
Ion Ratio Lower Upper
65 100
92 63.5 52.0 78.0
138 86.5 71.2 106.8









#47
Dimethylphthalate
Concen: 28.061 ng/ul
RT: 13.893 min Scan# 1820
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:163 Resp: 499893
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.6 8.4 12.6



#48
2,6-Dinitrotoluene
Concen: 29.645 ng/ul
RT: 14.010 min Scan# 1840
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:165 Resp: 104155
Ion Ratio Lower Upper
165 100
89 67.8 53.4 80.0
121 22.0 19.0 28.6





#50

Acenaphthylene

Concen: 29.136 ng/ul

RT: 14.157 min Scan# 1865

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31

Instrument :

BNA_N

ClientSampleId :

SLCS343





Tgt Ion:152 Resp: 607296

Ion Ratio Lower Upper

152 100

151 20.7 15.9 23.9

153 12.9 10.0 15.0



#51

3-Nitroaniline

Concen: 25.933 ng/ul

RT: 14.340 min Scan# 1896

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31




Tgt Ion:138 Resp: 91295


Ion Ratio Lower Upper

138 100

108 12.5 10.3 15.5

92 124.1 108.2 162.2









#52
Acenaphthene
Concen: 29.567 ng/ul
RT: 14.498 min Scan# 1923
Delta R.T. -0.006 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:153 Resp: 443651
Ion Ratio Lower Upper
153 100
152 47.9 38.1 57.1
154 86.7 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 27.340 ng/ul
RT: 14.540 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:184 Resp: 60357
Ion Ratio Lower Upper
184 100
63 89.5 62.0 93.0
154 76.8 53.0 79.6









#55
4-Nitrophenol
Concen: 26.810 ng/ul
RT: 14.687 min Scan# 1955
Delta R.T. -0.018 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:109 Resp: 87756
Ion Ratio Lower Upper
109 100
139 102.4 79.3 118.9
65 132.7 97.7 146.5



#56
Dibenzofuran
Concen: 29.322 ng/ul
RT: 14.834 min Scan# 1980
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:168 Resp: 609098
Ion Ratio Lower Upper
168 100
139 40.4 33.8 50.8









#57
2,4-Dinitrotoluene
Concen: 28.347 ng/ul
RT: 14.798 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:165 Resp: 149026
Ion Ratio Lower Upper
165 100
63 52.0 43.8 65.8
182 2.9 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 28.193 ng/ul
RT: 15.069 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:232 Resp: 123777
Ion Ratio Lower Upper
232 100
131 49.3 40.0 60.0
130 3.0 3.0 4.4
166 33.8 27.0 40.4









#59
Diethylphthalate
Concen: 27.855 ng/ul
RT: 15.263 min Scan# 2053
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:149 Resp: 511890
Ion Ratio Lower Upper
149 100
177 22.7 17.8 26.8
150 12.6 9.8 14.8



#61
Fluorene
Concen: 28.688 ng/ul
RT: 15.487 min Scan# 2091
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:166 Resp: 483256
Ion Ratio Lower Upper
166 100
165 102.4 79.8 119.8
167 13.3 11.2 16.8









#62
4-Chlorophenyl-phenylether
Concen: 29.051 ng/ul
RT: 15.481 min Scan# 2090
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:204 Resp: 253027
Ion Ratio Lower Upper
204 100
206 33.8 25.5 38.3
141 68.0 51.0 76.4



#63
4-Nitroaniline
Concen: 29.683 ng/ul
RT: 15.504 min Scan# 2094
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:138 Resp: 102234
Ion Ratio Lower Upper
138 100
92 57.8 47.2 70.8
108 102.3 93.0 139.6









#66
4,6-Dinitro-2-methylphenol
Concen: 30.089 ng/ul
RT: 15.563 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:198 Resp: 88488
Ion Ratio Lower Upper
198 100
182 3.7 3.5 5.3
77 28.6 24.6 36.8



#67
N-Nitrosodiphenylamine
Concen: 31.774 ng/ul
RT: 15.692 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:169 Resp: 403540
Ion Ratio Lower Upper
169 100
168 64.7 55.1 82.7
167 35.2 30.3 45.5









#68
4-Bromophenyl-phenylether
Concen: 31.744 ng/ul
RT: 16.375 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:248 Resp: 142313
Ion Ratio Lower Upper
248 100
250 102.6 73.0 109.6
141 88.8 65.8 98.8



#69
Hexachlorobenzene
Concen: 31.881 ng/ul
RT: 16.492 min Scan# 2262
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:284 Resp: 155925
Ion Ratio Lower Upper
284 100
142 38.5 32.6 48.8
249 35.2 25.4 38.2









#70
Atrazine
 Concen: 28.882 ng/ul
 RT: 16.645 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31




Instrument :
 BNA_N
ClientSampleId :
 SLCS343


Tgt Ion:	200	Resp:	144572
Ion Ratio	Lower	Upper	
200	100		
173	26.0	21.6	32.4
215	45.6	38.4	57.6



#71
Pentachlorophenol
 Concen: 31.467 ng/ul
 RT: 16.839 min Scan# 2321
 Delta R.T. -0.000 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31

Tgt Ion:	266	Resp:	105056
Ion Ratio	Lower	Upper	
266	100		
264	58.7	49.4	74.2
268	60.0	47.8	71.6









#72
Phenanthrene
Concen: 30.101 ng/u1
RT: 17.222 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:178 Resp: 706185
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.2
176 19.1 15.4 23.2



#74
Anthracene
Concen: 30.552 ng/u1
RT: 17.316 min Scan# 2402
Delta R.T. -0.000 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:178 Resp: 721297
Ion Ratio Lower Upper
178 100
179 15.7 12.4 18.6
176 19.3 15.1 22.7









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 34.411 ng/uL
 RT: 13.234 min Scan# 1
 Delta R.T. -0.000 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31

Instrument :
 BNA_N
 ClientSampleId :
 SLCS343

Tgt Ion:216 Resp: 224094




Ion	Ratio	Lower	Upper
216	100		
214	80.7	62.6	94.0
218	50.5	37.4	56.0




#76
 Pentachlorobenzene
 Concen: 32.110 ng/uL
 RT: 14.757 min Scan# 1967
 Delta R.T. -0.000 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31

Tgt Ion:250 Resp: 204300

Ion	Ratio	Lower	Upper
250	100		
248	61.3	49.1	73.7
252	63.3	46.2	69.2









#77
 Carbazole
 Concen: 31.509 ng/ul
 RT: 17.587 min Scan# 2448
 Delta R.T. -0.000 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31

Instrument :
 BNA_N
 ClientSampleId :
 SLCS343

Tgt Ion:167 Resp: 652557




Ion	Ratio	Lower	Upper
167	100		
166	22.3	16.8	25.2
139	12.6	10.5	15.7




#78
 Di-n-butylphthalate
 Concen: 31.994 ng/ul
 RT: 18.151 min Scan# 2544
 Delta R.T. -0.006 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31

Tgt Ion:149 Resp: 882578

Ion	Ratio	Lower	Upper
149	100		
150	9.8	7.2	10.8
104	6.0	4.9	7.3









#80
 Fluoranthene
 Concen: 31.900 ng/u1
 RT: 19.239 min Scan# 2729
 Delta R.T. -0.000 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31




Instrument :
 BNA_N
 ClientSampleId :
 SLCS343


Tgt Ion:202 Resp: 789867
 Ion Ratio Lower Upper
 202 100
 101 9.6 8.4 12.6
 100 8.1 7.0 10.6



#82
 Pyrene
 Concen: 31.155 ng/u1
 RT: 19.604 min Scan# 2791
 Delta R.T. -0.000 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31

Tgt Ion:202 Resp: 817794
 Ion Ratio Lower Upper
 202 100
 101 11.6 9.4 14.0
 100 10.3 8.1 12.1





#83

Butylbenzylphthalate

Concen: 30.246 ng/ul

RT: 20.510 min Scan# 2945

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31

Instrument :

BNA_N

ClientSampleId :

SLCS343





Tgt Ion:149 Resp: 341404

Ion Ratio Lower Upper

149 100

91 70.9 60.2 90.4

206 21.7 16.3 24.5



#84

3,3'-Dichlorobenzidine

Concen: 31.847 ng/ul

RT: 21.333 min Scan# 3085

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31




Tgt Ion:252 Resp: 200182


Ion Ratio Lower Upper

252 100

254 59.1 52.3 78.5

126 11.2 9.0 13.6





#85

Benzo(a)anthracene

Concen: 29.865 ng/ul

RT: 21.410 min Scan# 3098

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31

Instrument :

BNA_N

ClientSampleId :

SLCS343





Tgt Ion:228 Resp: 730795

Ion Ratio Lower Upper

228 100

229 18.1 16.2 24.4

226 27.7 21.5 32.3



#86

Bis(2-ethylhexyl)phthalate

Concen: 30.087 ng/ul

RT: 21.339 min Scan# 3086

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31




Tgt Ion:149 Resp: 487928


Ion Ratio Lower Upper

149 100

167 27.3 21.8 32.8

279 4.2 3.0 4.6





#87

Chrysene

Concen: 30.186 ng/ul

RT: 21.475 min Scan# 3109

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31

Instrument :

BNA_N

ClientSampleId :

SLCS343





Tgt Ion:228 Resp: 700345

Ion Ratio Lower Upper

228 100

226 28.2 23.4 35.2

229 18.3 15.0 22.4



#89

Di-n-octyl phthalate

Concen: 26.613 ng/ul




RT: 22.486 min Scan# 3281


Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31

Tgt Ion:149 Resp: 783912









#90
Benzo(b)fluoranthene
Concen: 28.941 ng/ul
RT: 23.492 min Scan# 3454
Delta R.T. -0.012 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31




Instrument :
BNA_N
ClientSampleId :
SLCS343


Tgt Ion:252 Resp: 685943
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 8.9 7.9 11.9



#91
Benzo(k)fluoranthene
Concen: 28.478 ng/ul
RT: 23.557 min Scan# 3463
Delta R.T. -0.006 min
Lab File: BN036152.D
Acq: 31 Jan 2025 03:31

Tgt Ion:252 Resp: 672090
Ion Ratio Lower Upper
252 100
253 21.1 17.0 25.6
125 9.5 7.3 10.9





#93

Benzo(a)pyrene

Concen: 29.889 ng/ul

RT: 24.316 min Scan# 3592

Delta R.T. -0.000 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31

Instrument :

BNA_N

ClientSampleId :

SLCS343





Tgt Ion:252 Resp: 619160

Ion Ratio Lower Upper

252 100

253 21.7 17.0 25.6

125 10.4 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 37.100 ng/ul

RT: 27.868 min Scan# 4196

Delta R.T. -0.012 min

Lab File: BN036152.D

Acq: 31 Jan 2025 03:31




Tgt Ion:276 Resp: 836215


Ion Ratio Lower Upper

276 100

138 19.3 15.8 23.6

277 23.4 19.0 28.6









#95
 Dibenzo(a,h)anthracene
 Concen: 37.805 ng/ul
 RT: 27.933 min Scan# 4379
 Delta R.T. -0.012 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31

Instrument :
 BNA_N
 ClientSampleId :
 SLCS343

Tgt Ion:278 Resp: 679599
 Ion Ratio Lower Upper
 278 100
 139 15.2 12.4 18.6
 279 23.7 20.5 30.7



#96
 Benzo(g,h,i)perylene
 Concen: 39.049 ng/ul
 RT: 28.945 min Scan# 4379
 Delta R.T. 0.012 min
 Lab File: BN036152.D
 Acq: 31 Jan 2025 03:31

Tgt Ion:276 Resp: 669860
 Ion Ratio Lower Upper
 276 100
 138 18.8 16.0 24.0
 277 23.5 17.9 26.9

