

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013025\
 Data File : BN036154.D
 Acq On : 31 Jan 2025 05:29
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020360

Quant Time: Jan 31 07:38:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN013025.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 30 17:36:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	106763	20.000	ng/u1	0.00
20) Naphthalene-d8	10.587	136	439773	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.433	164	291602	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.180	188	599298	20.000	ng/u1	0.00
79) Chrysene-d12	21.427	240	558549	20.000	ng/u1	0.00
88) Perylene-d12	24.456	264	503585	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	22038	8.411	ng/uL	0.00
4) Pyridine-d5	3.669	84	156618	20.986	ng/u1	0.00
7) Phenol-d5	6.981	99	193313	21.851	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.122	67	127166	22.399	ng/u1	0.00
11) 2-Chlorophenol-d4	7.334	132	152443	22.009	ng/u1	0.00
15) 4-Methylphenol-d8	8.516	113	162632	22.882	ng/u1	0.00
21) Nitrobenzene-d5	8.945	128	78959	22.364	ng/u1	0.00
24) 2-Nitrophenol-d4	9.669	143	90044	22.478	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.222	165	157808	21.818	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.722	131	220556	23.210	ng/u1	0.00
46) Dimethylphthalate-d6	13.845	166	499884	22.731	ng/u1	0.00
49) Acenaphthylene-d8	14.128	160	541356	22.109	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	85460	22.518	ng/u1	-0.01
60) Fluorene-d10	15.428	176	400035	21.732	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.545	200	84504	21.305	ng/u1	0.00
73) Anthracene-d10	17.280	188	645482	22.172	ng/u1	0.00
81) Pyrene-d10	19.574	212	797803	23.983	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.251	264	576363	21.766	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	24444	8.734	ng/uL	97
5) Pyridine	3.687	79	161980	21.371	ng/u1	94
6) Benzaldehyde	6.934	77	144678	27.389	ng/u1	96
8) Phenol	7.010	94	202631	22.185	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.216	93	165360	23.038	ng/u1	97
12) 2-Chlorophenol	7.369	128	156255	22.131	ng/u1	99
13) 2-Methylphenol	8.251	108	152023	22.541	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.316	45	260427	23.770	ng/u1	97
16) Acetophenone	8.610	105	267774	23.108	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.598	70	141472	23.187	ng/u1	98
18) 4-Methylphenol	8.581	108	169273	23.003	ng/u1	87
19) Hexachloroethane	8.875	117	72914	22.354	ng/u1	95
22) Nitrobenzene	8.987	77	216182	22.216	ng/u1	100
23) Isophorone	9.522	82	390077	22.618	ng/u1	99
25) 2-Nitrophenol	9.698	139	95553	22.293	ng/u1	98
26) 2,4-Dimethylphenol	9.775	107	195290	22.363	ng/u1	96
27) Bis(2-Chloroethoxy)met...	9.998	93	224512	22.179	ng/u1	96
29) 2,4-Dichlorophenol	10.251	162	155452	21.333	ng/u1	96
30) Naphthalene	10.640	128	522798	22.096	ng/u1	100
32) 4-Chloroaniline	10.745	127	212469	23.165	ng/u1	96
33) Hexachlorobutadiene	10.928	225	112805	21.207	ng/u1	96
34) Caprolactam	11.522	113	54204	22.984	ng/u1	96
35) 4-Chloro-3-methylphenol	11.898	107	188890	22.564	ng/u1	98
36) 2-Methylnaphthalene	12.251	142	371344	22.391	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013025\
 Data File : BN036154.D
 Acq On : 31 Jan 2025 05:29
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 28 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020360


Quant Time: Jan 31 07:38:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN013025.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 30 17:36:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.475	142	374840	23.006	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.622	216	212917	22.851	ng/ul	95
40) Hexachlorocyclopentadiene	12.610	237	118963	22.874	ng/ul	94
41) 2,4,6-Trichlorophenol	12.869	196	137091	22.819	ng/ul	96
42) 2,4,5-Trichlorophenol	12.957	196	147780	22.730	ng/ul	96
43) 1,1'-Biphenyl	13.269	154	496496	22.938	ng/ul	99
44) 2-Chloronaphthalene	13.310	162	389929	22.914	ng/ul	97
45) 2-Nitroaniline	13.510	65	138793	23.920	ng/ul	96
47) Dimethylphthalate	13.892	163	497343	22.584	ng/ul	99
48) 2,6-Dinitrotoluene	14.010	165	100614	23.166	ng/ul	97
50) Acenaphthylene	14.157	152	562572	21.833	ng/ul	99
51) 3-Nitroaniline	14.339	138	101662	23.360	ng/ul	95
52) Acenaphthene	14.498	153	406784	21.930	ng/ul	97
53) 2,4-Dinitrophenol	14.539	184	56599	20.740	ng/ul	91
55) 4-Nitrophenol	14.686	109	89345	22.080	ng/ul	97
56) Dibenzofuran	14.833	168	565805	22.034	ng/ul	100
57) 2,4-Dinitrotoluene	14.792	165	146631	22.562	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.069	232	118976	21.922	ng/ul#	99
59) Diethylphthalate	15.263	149	501300	22.067	ng/ul	99
61) Fluorene	15.486	166	453773	21.791	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.480	204	229805	21.344	ng/ul	98
63) 4-Nitroaniline	15.504	138	105146	24.695	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.563	198	90011	21.371	ng/ul	96
67) N-Nitrosodiphenylamine	15.692	169	393876	21.654	ng/ul	97
68) 4-Bromophenyl-phenylether	16.375	248	138323	21.543	ng/ul	98
69) Hexachlorobenzene	16.492	284	155536	22.205	ng/ul	96
70) Atrazine	16.645	200	158976	22.175	ng/ul	100
71) Pentachlorophenol	16.839	266	101405	21.208	ng/ul	97
72) Phenanthrene	17.222	178	728959	21.695	ng/ul	99
74) Anthracene	17.316	178	746132	22.067	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.234	216	210232	22.541	ng/uL	99
76) Pentachlorobenzene	14.757	250	193017	21.182	ng/uL#	89
77) Carbazole	17.586	167	660881	22.281	ng/ul	95
78) Di-n-butylphthalate	18.151	149	860393	21.777	ng/ul	99
80) Fluoranthene	19.239	202	922035	23.728	ng/ul	99
82) Pyrene	19.604	202	961527	23.342	ng/ul	97
83) Butylbenzylphthalate	20.510	149	393116	22.192	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.333	252	233603	23.682	ng/ul	99
85) Benzo(a)anthracene	21.410	228	855189	22.270	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.339	149	564157	22.167	ng/ul	99
87) Chrysene	21.474	228	808093	22.194	ng/ul	98
89) Di-n-octyl phthalate	22.486	149	935806	22.613	ng/ul	100
90) Benzo(b)fluoranthene	23.498	252	744086	22.346	ng/ul	100
91) Benzo(k)fluoranthene	23.562	252	738434	22.272	ng/ul	100
93) Benzo(a)pyrene	24.309	252	647893	22.262	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.880	276	669037	21.128	ng/ul	98
95) Dibenzo(a,h)anthracene	27.939	278	545415	21.596	ng/ul	96
96) Benzo(g,h,i)perylene	28.939	276	515780	21.401	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 31 07:38:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN013025.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 30 17:36:07 2025
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.734 ng/uL
RT: 3.293 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion: 88 Resp: 24444
Ion Ratio Lower Upper
88 100
43 38.7 29.6 44.4
58 71.6 59.1 88.7



#5
Pyridine
Concen: 21.371 ng/uL
RT: 3.687 min Scan# 85
Delta R.T. -0.006 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion: 79 Resp: 161980
Ion Ratio Lower Upper
79 100
52 91.3 67.3 100.9
51 46.4 36.2 54.4









#6
Benzaldehyde
Concen: 27.389 ng/ul
RT: 6.934 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion: 77 Resp: 144678
Ion Ratio Lower Upper
77 100
105 88.1 74.8 112.2
106 84.8 69.8 104.8



#8
Phenol
Concen: 22.185 ng/ul
RT: 7.010 min Scan# 650
Delta R.T. -0.012 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion: 94 Resp: 202631
Ion Ratio Lower Upper
94 100
65 34.7 27.0 40.4
66 50.6 40.2 60.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 23.038 ng/ul
 RT: 7.216 min Scan# 685
 Delta R.T. -0.000 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29




Instrument :
 BNA_N
 ClientSampleId :
 SST020360


Tgt Ion: 93 Resp: 165360
 Ion Ratio Lower Upper
 93 100
 63 80.9 67.0 100.6
 95 31.5 25.8 38.6



#12
 2-Chlorophenol
 Concen: 22.131 ng/ul
 RT: 7.369 min Scan# 711
 Delta R.T. -0.006 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29

Tgt Ion: 128 Resp: 156255
 Ion Ratio Lower Upper
 128 100
 64 53.2 41.8 62.6
 130 32.0 24.9 37.3









#13
2-Methylphenol
Concen: 22.541 ng/ul
RT: 8.251 min Scan# 861
Delta R.T. -0.012 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:108 Resp: 152023
Ion Ratio Lower Upper
108 100
107 90.3 73.7 110.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 23.770 ng/ul
RT: 8.316 min Scan# 872
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion: 45 Resp: 260427
Ion Ratio Lower Upper
45 100
77 13.4 11.8 17.6
79 9.7 8.3 12.5









#16
Acetophenone
Concen: 23.108 ng/ul
RT: 8.610 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:105 Resp: 267774
Ion Ratio Lower Upper
105 100
77 86.1 68.7 103.1
51 35.5 28.9 43.3



#17
N-Nitroso-di-n-propylamine
Concen: 23.187 ng/ul
RT: 8.598 min Scan# 920
Delta R.T. -0.006 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion: 70 Resp: 141472
Ion Ratio Lower Upper
70 100
42 70.0 54.7 82.1
101 8.9 7.1 10.7
130 17.4 15.7 23.5









#18
4-Methylphenol
Concen: 23.003 ng/ul
RT: 8.581 min Scan# 918
Delta R.T. -0.006 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:108 Resp: 169273
Ion Ratio Lower Upper
108 100
107 105.4 95.4 143.2



#19
Hexachloroethane
Concen: 22.354 ng/ul
RT: 8.875 min Scan# 967
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:117 Resp: 72914
Ion Ratio Lower Upper
117 100
201 101.4 77.5 116.3
199 61.8 45.8 68.6









#22
Nitrobenzene
Concen: 22.216 ng/ul
RT: 8.987 min Scan# 987
Delta R.T. -0.006 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion: 77 Resp: 216182
Ion Ratio Lower Upper
77 100
123 40.6 32.5 48.7
65 15.6 12.2 18.2



#23
Isophorone
Concen: 22.618 ng/ul
RT: 9.522 min Scan# 1077
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion: 82 Resp: 390077
Ion Ratio Lower Upper
82 100
95 8.1 6.6 10.0
138 16.3 12.8 19.2









#25
2-Nitrophenol
Concen: 22.293 ng/ul
RT: 9.698 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:139 Resp: 95553
Ion Ratio Lower Upper
139 100
65 66.2 51.4 77.2
109 39.8 31.6 47.4



#26
2,4-Dimethylphenol
Concen: 22.363 ng/ul
RT: 9.775 min Scan# 1120
Delta R.T. -0.006 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:107 Resp: 195290
Ion Ratio Lower Upper
107 100
121 47.8 36.6 55.0
122 80.6 61.4 92.0





#27

Bis(2-Chloroethoxy)methane

Concen: 22.179 ng/ul

RT: 9.998 min Scan# 1158

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Instrument :

BNA_N

ClientSampleId :

SSTD020360





Tgt Ion: 93 Resp: 224512

Ion Ratio Lower Upper

93 100

95 34.1 25.4 38.0

123 11.3 8.2 12.4



#29

2,4-Dichlorophenol

Concen: 21.333 ng/ul

RT: 10.251 min Scan# 1201

Delta R.T. -0.006 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29




Tgt Ion:162 Resp: 155452


Ion Ratio Lower Upper

162 100

164 66.3 49.7 74.5

98 40.1 30.8 46.2









#30
Naphthalene
Concen: 22.096 ng/ul
RT: 10.640 min Scan# 1267
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:	128	Resp:	522798
Ion Ratio	Lower	Upper	
128	100		
129	11.1	9.0	13.4
127	14.0	11.4	17.2



#32
4-Chloroaniline
Concen: 23.165 ng/ul
RT: 10.745 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:	127	Resp:	212469
Ion Ratio	Lower	Upper	
127	100		
129	30.1	25.8	38.8





#33

Hexachlorobutadiene

Concen: 21.207 ng/ul

RT: 10.928 min Scan# 11

Delta R.T. -0.006 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Instrument :

BNA_N

ClientSampleId :

SSTD020360





Tgt Ion:225 Resp: 112805

Ion Ratio Lower Upper

225 100

223 62.3 48.5 72.7

227 65.2 49.2 73.8



#34

Caprolactam

Concen: 22.984 ng/ul

RT: 11.522 min Scan# 1417

Delta R.T. -0.006 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29




Tgt Ion:113 Resp: 54204


Ion Ratio Lower Upper

113 100

55 186.1 143.6 215.4

56 151.5 119.0 178.6









#35
 4-Chloro-3-methylphenol
 Concen: 22.564 ng/ul
 RT: 11.898 min Scan# 1481
 Delta R.T. -0.006 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29




Instrument :
 BNA_N
 ClientSampleId :
 SST020360


Tgt Ion:107 Resp: 188890
 Ion Ratio Lower Upper
 107 100
 144 23.5 17.5 26.3
 142 72.0 56.7 85.1



#36
 2-Methylnaphthalene
 Concen: 22.391 ng/ul
 RT: 12.251 min Scan# 1541
 Delta R.T. -0.000 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29

Tgt Ion:142 Resp: 371344
 Ion Ratio Lower Upper
 142 100
 141 92.0 73.5 110.3









#37
1-Methylnaphthalene
Concen: 23.006 ng/ul
RT: 12.475 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:142 Resp: 374840
Ion Ratio Lower Upper
142 100
141 94.7 74.7 112.1
116 4.3 3.9 5.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.851 ng/ul
RT: 12.622 min Scan# 1604
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:216 Resp: 212917
Ion Ratio Lower Upper
216 100
214 75.1 63.0 94.6
179 20.6 18.6 27.8
108 17.8 16.7 25.1









#40
Hexachlorocyclopentadiene
Concen: 22.874 ng/ul
RT: 12.610 min Scan# 1602
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:237 Resp: 118963
Ion Ratio Lower Upper
237 100
235 60.6 44.7 67.1
272 12.1 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 22.819 ng/ul
RT: 12.869 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:196 Resp: 137091
Ion Ratio Lower Upper
196 100
198 97.1 81.8 122.6
200 29.7 25.0 37.4









#42
2,4,5-Trichlorophenol
Concen: 22.730 ng/ul
RT: 12.957 min Scan# 1661
Delta R.T. -0.012 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:196 Resp: 147780
Ion Ratio Lower Upper
196 100
198 92.6 77.0 115.6
200 29.4 24.7 37.1



#43
1,1'-Biphenyl
Concen: 22.938 ng/ul
RT: 13.269 min Scan# 1714
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:154 Resp: 496496
Ion Ratio Lower Upper
154 100
153 41.2 33.4 50.2
76 13.1 10.0 15.0





#44

2-Chloronaphthalene

Concen: 22.914 ng/ul

RT: 13.310 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Instrument :

BNA_N

ClientSampleId :

SSTD020360





Tgt Ion:162 Resp: 389929

Ion Ratio Lower Upper

162 100

164 32.8 27.4 41.0

127 40.7 34.5 51.7



#45

2-Nitroaniline

Concen: 23.920 ng/ul

RT: 13.510 min Scan# 1755

Delta R.T. -0.006 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29




Tgt Ion: 65 Resp: 138793


Ion Ratio Lower Upper

65 100

92 61.3 52.0 78.0

138 85.8 71.2 106.8









#47
Dimethylphthalate
Concen: 22.584 ng/ul
RT: 13.892 min Scan# 1820
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:163 Resp: 497343
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.2 8.4 12.6



#48
2,6-Dinitrotoluene
Concen: 23.166 ng/ul
RT: 14.010 min Scan# 1840
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:165 Resp: 100614
Ion Ratio Lower Upper
165 100
89 69.3 53.4 80.0
121 22.7 19.0 28.6





#50

Acenaphthylene

Concen: 21.833 ng/ul

RT: 14.157 min Scan# 1865

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Instrument :

BNA_N

ClientSampleId :

SSTD020360





Tgt Ion:152 Resp: 562572

Ion Ratio Lower Upper

152 100

151 20.2 15.9 23.9

153 12.8 10.0 15.0



#51

3-Nitroaniline

Concen: 23.360 ng/ul

RT: 14.339 min Scan# 1896

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29




Tgt Ion:138 Resp: 101662


Ion Ratio Lower Upper

138 100

108 12.7 10.3 15.5

92 128.5 108.2 162.2









#52
Acenaphthene
Concen: 21.930 ng/ul
RT: 14.498 min Scan# 1923
Delta R.T. -0.006 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:153 Resp: 406784
Ion Ratio Lower Upper
153 100
152 45.9 38.1 57.1
154 85.8 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 20.740 ng/ul
RT: 14.539 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:184 Resp: 56599
Ion Ratio Lower Upper
184 100
63 88.3 62.0 93.0
154 70.2 53.0 79.6









#55
4-Nitrophenol
Concen: 22.080 ng/ul
RT: 14.686 min Scan# 1955
Delta R.T. -0.018 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:109 Resp: 89345
Ion Ratio Lower Upper
109 100
139 101.4 79.3 118.9
65 126.5 97.7 146.5



#56
Dibenzofuran
Concen: 22.034 ng/ul
RT: 14.833 min Scan# 1980
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:168 Resp: 565805
Ion Ratio Lower Upper
168 100
139 42.1 33.8 50.8









#57
 2,4-Dinitrotoluene
 Concen: 22.562 ng/ul
 RT: 14.792 min Scan# 1973
 Delta R.T. -0.006 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29




Instrument :
 BNA_N
 ClientSampleId :
 SST020360


Tgt Ion:165 Resp: 146631
 Ion Ratio Lower Upper
 165 100
 63 57.7 43.8 65.8
 182 2.7 2.1 3.1



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 21.922 ng/ul
 RT: 15.069 min Scan# 2020
 Delta R.T. -0.000 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29

Tgt Ion:232 Resp: 118976
 Ion Ratio Lower Upper
 232 100
 131 49.5 40.0 60.0
 130 2.8 3.0 4.4#
 166 32.5 27.0 40.4





#59

Diethylphthalate

Concen: 22.067 ng/ul

RT: 15.263 min Scan# 2053

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Instrument :

BNA_N

ClientSampleId :

SSTD020360





Tgt Ion:149 Resp: 501300

Ion Ratio Lower Upper

149 100

177 22.1 17.8 26.8

150 12.7 9.8 14.8



#61

Fluorene

Concen: 21.791 ng/ul

RT: 15.486 min Scan# 2091

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29




Tgt Ion:166 Resp: 453773


Ion Ratio Lower Upper

166 100

165 99.9 79.8 119.8

167 13.6 11.2 16.8









#62
4-Chlorophenyl-phenylether
Concen: 21.344 ng/ul
RT: 15.480 min Scan# 2090
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:204 Resp: 229805
Ion Ratio Lower Upper
204 100
206 33.6 25.5 38.3
141 64.3 51.0 76.4



#63
4-Nitroaniline
Concen: 24.695 ng/ul
RT: 15.504 min Scan# 2094
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:138 Resp: 105146
Ion Ratio Lower Upper
138 100
92 59.2 47.2 70.8
108 107.7 93.0 139.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 21.371 ng/ul
 RT: 15.563 min Scan# 2104
 Delta R.T. -0.000 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020360


Tgt Ion:198 Resp: 90011
 Ion Ratio Lower Upper
 198 100
 182 4.0 3.5 5.3
 77 28.5 24.6 36.8



#67
 N-Nitrosodiphenylamine
 Concen: 21.654 ng/ul
 RT: 15.692 min Scan# 2126
 Delta R.T. -0.000 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29

Tgt Ion:169 Resp: 393876
 Ion Ratio Lower Upper
 169 100
 168 66.4 55.1 82.7
 167 36.9 30.3 45.5









#68
4-Bromophenyl-phenylether
Concen: 21.543 ng/ul
RT: 16.375 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:248 Resp: 138323
Ion Ratio Lower Upper
248 100
250 93.6 73.0 109.6
141 83.8 65.8 98.8



#69
Hexachlorobenzene
Concen: 22.205 ng/ul
RT: 16.492 min Scan# 2262
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:284 Resp: 155536
Ion Ratio Lower Upper
284 100
142 38.3 32.6 48.8
249 34.1 25.4 38.2









#70
 Atrazine
 Concen: 22.175 ng/ul
 RT: 16.645 min Scan# 2288
 Delta R.T. -0.006 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020360

Tgt Ion:200 Resp: 158976




Ion	Ratio	Lower	Upper
200	100		
173	26.5	21.6	32.4
215	47.8	38.4	57.6




#71
 Pentachlorophenol
 Concen: 21.208 ng/ul
 RT: 16.839 min Scan# 2321
 Delta R.T. -0.000 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29

Tgt Ion:266 Resp: 101405

Ion	Ratio	Lower	Upper
266	100		
264	59.7	49.4	74.2
268	57.6	47.8	71.6









#72
 Phenanthrene
 Concen: 21.695 ng/u1
 RT: 17.222 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020360

Tgt Ion:178 Resp: 728959




Ion	Ratio	Lower	Upper
178	100		
179	15.6	12.2	18.2
176	19.6	15.4	23.2




#74
 Anthracene
 Concen: 22.067 ng/u1
 RT: 17.316 min Scan# 2402
 Delta R.T. -0.000 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29

Tgt Ion:178 Resp: 746132

Ion	Ratio	Lower	Upper
178	100		
179	14.9	12.4	18.6
176	18.5	15.1	22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.541 ng/uL
RT: 13.234 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:216 Resp: 210232
Ion Ratio Lower Upper
216 100
214 78.8 62.6 94.0
218 48.1 37.4 56.0



#76
Pentachlorobenzene
Concen: 21.182 ng/uL
RT: 14.757 min Scan# 1967
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:250 Resp: 193017
Ion Ratio Lower Upper
250 100
248 65.5 49.1 73.7
252 70.3 46.2 69.2#









#77
 Carbazole
 Concen: 22.281 ng/ul
 RT: 17.586 min Scan# 2448
 Delta R.T. -0.000 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020360


Tgt Ion:167 Resp: 660881
 Ion Ratio Lower Upper
 167 100
 166 24.0 16.8 25.2
 139 14.1 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 21.777 ng/ul
 RT: 18.151 min Scan# 2544
 Delta R.T. -0.006 min
 Lab File: BN036154.D
 Acq: 31 Jan 2025 05:29

Tgt Ion:149 Resp: 860393
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.2 10.8
 104 6.0 4.9 7.3









#80
Fluoranthene
Concen: 23.728 ng/ul
RT: 19.239 min Scan# 2729
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29




Instrument :
BNA_N
ClientSampleId :
SSTD020360


Tgt Ion:202 Resp: 922035
Ion Ratio Lower Upper
202 100
101 11.1 8.4 12.6
100 8.7 7.0 10.6



#82
Pyrene
Concen: 23.342 ng/ul
RT: 19.604 min Scan# 2791
Delta R.T. -0.000 min
Lab File: BN036154.D
Acq: 31 Jan 2025 05:29

Tgt Ion:202 Resp: 961527
Ion Ratio Lower Upper
202 100
101 12.8 9.4 14.0
100 11.0 8.1 12.1





#83

Butylbenzylphthalate

Concen: 22.192 ng/ul

RT: 20.510 min Scan# 2945

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Instrument :

BNA_N

ClientSampleId :

SSTD020360





Tgt Ion:149 Resp: 393116

Ion Ratio Lower Upper

149 100

91 74.9 60.2 90.4

206 22.0 16.3 24.5



#84

3,3'-Dichlorobenzidine

Concen: 23.682 ng/ul

RT: 21.333 min Scan# 3085

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29




Tgt Ion:252 Resp: 233603


Ion Ratio Lower Upper

252 100

254 64.3 52.3 78.5

126 11.5 9.0 13.6





#85

Benzo(a)anthracene

Concen: 22.270 ng/ul

RT: 21.410 min Scan# 3098

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Instrument :

BNA_N

ClientSampleId :

SSTD020360





Tgt Ion:228 Resp: 855189

Ion Ratio Lower Upper

228 100

229 17.8 16.2 24.4

226 26.2 21.5 32.3



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.167 ng/ul

RT: 21.339 min Scan# 3086

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29




Tgt Ion:149 Resp: 564157


Ion Ratio Lower Upper

149 100

167 28.1 21.8 32.8

279 3.8 3.0 4.6





#87

Chrysene

Concen: 22.194 ng/ul

RT: 21.474 min Scan# 3109

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Instrument :

BNA_N

ClientSampleId :

SSTD020360





Tgt Ion:228 Resp: 808093

Ion Ratio Lower Upper

228 100

226 29.7 23.4 35.2

229 20.4 15.0 22.4



#89

Di-n-octyl phthalate

Concen: 22.613 ng/ul




RT: 22.486 min Scan# 3281


Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Tgt Ion:149 Resp: 935806





#90

Benzo(b)fluoranthene

Concen: 22.346 ng/ul

RT: 23.498 min Scan# 3454

Delta R.T. -0.006 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Instrument :

BNA_N

ClientSampleId :

SSTD020360





Tgt Ion:252 Resp: 744086

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 9.6 7.9 11.9



#91

Benzo(k)fluoranthene

Concen: 22.272 ng/ul

RT: 23.562 min Scan# 3464

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29




Tgt Ion:252 Resp: 738434


Ion Ratio Lower Upper

252 100

253 21.4 17.0 25.6

125 9.2 7.3 10.9





#93

Benzo(a)pyrene

Concen: 22.262 ng/ul

RT: 24.309 min Scan# 3591

Delta R.T. -0.006 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Instrument :

BNA_N

ClientSampleId :

SSTD020360





Tgt Ion:252 Resp: 647893

Ion Ratio Lower Upper

252 100

253 19.9 17.0 25.6

125 11.0 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.128 ng/ul

RT: 27.880 min Scan# 4198

Delta R.T. -0.000 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29




Tgt Ion:276 Resp: 669037


Ion Ratio Lower Upper

276 100

138 20.1 15.8 23.6

277 25.7 19.0 28.6





#95

Dibenzo(a,h)anthracene

Concen: 21.596 ng/ul

RT: 27.939 min Scan# 4378

Delta R.T. -0.006 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Instrument :

BNA_N

ClientSampleId :

SSTD020360





Tgt Ion:278 Resp: 545415

Ion Ratio Lower Upper

278 100

139 14.3 12.4 18.6

279 23.4 20.5 30.7



#96

Benzo(g,h,i)perylene

Concen: 21.401 ng/ul

RT: 28.939 min Scan# 4378

Delta R.T. 0.006 min

Lab File: BN036154.D

Acq: 31 Jan 2025 05:29

Tgt Ion:276 Resp: 515780

Ion Ratio Lower Upper

276 100

138 20.7 16.0 24.0

277 22.5 17.9 26.9

