




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013124\
Data File : BN029448.D
Acq On : 31 Jan 2024 13:47
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 01 03:24:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 03:24:38 2024
Response via : Initial Calibration



TIC: BN029448.D\data.ms

(70) Pentachlorophenol (C)




16.928min (0.000) 12358.50 ng

response 114153

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.97
264.00	61.60	66.22
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013124\
Data File : BN029448.D
Acq On : 31 Jan 2024 13:47
Operator : MA/JU
Sample : DFTTP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 01 03:24:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 03:24:38 2024
Response via : Initial Calibration



TIC: BN029448.D\data.ms

(77) Benzidine

19.551min (0.000) 0.00 ng

response 938301

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.39
183.00	13.20	10.93
0.00	0.00	0.00