

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013124\
 Data File : BN029449.D
 Acq On : 31 Jan 2024 14:25
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 01 01:39:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 30 03:38:24 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	119	0.00
2	1,4-Dioxane	0.400	0.381	4.8	113	0.00
3	n-Nitrosodimethylamine	0.400	0.365	8.8	112	0.00
4 S	2-Fluorophenol	0.400	0.398	0.5	125	0.00
5 S	Phenol-d6	0.400	0.387	3.3	126	0.00
6	bis(2-Chloroethyl)ether	0.400	0.391	2.3	125	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	115	0.00
8 S	Nitrobenzene-d5	0.400	0.394	1.5	122	0.00
9	Naphthalene	0.400	0.390	2.5	115	0.00
10	Hexachlorobutadiene	0.400	0.397	0.8	116	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.382	4.5	116	0.00
12	2-Methylnaphthalene	0.400	0.383	4.3	115	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.400	0.354	11.5	106	0.00
15 S	2-Fluorobiphenyl	0.400	0.400	0.0	111	0.00
16	Acenaphthylene	0.400	0.375	6.3	109	0.00
17	Acenaphthene	0.400	0.388	3.0	109	0.00
18	Fluorene	0.400	0.379	5.3	108	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	107	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.377	5.8	114	0.00
21	4-Bromophenyl-phenylether	0.400	0.384	4.0	108	0.00
22	Hexachlorobenzene	0.400	0.389	2.8	108	0.00
23	Atrazine	0.400	0.389	2.8	114	0.00
24	Pentachlorophenol	0.400	0.341	14.8	105	0.00
25	Phenanthrene	0.400	0.389	2.8	110	0.00
26	Anthracene	0.400	0.373	6.8	109	0.00
27 SURR	Fluoranthene-d10	0.400	0.394	1.5	114	0.00
28	Fluoranthene	0.400	0.382	4.5	114	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	114	0.00
30	Pyrene	0.400	0.393	1.8	110	0.00
31 S	Terphenyl-d14	0.400	0.393	1.8	112	0.00
32	Benzo(a)anthracene	0.400	0.388	3.0	117	0.00
33	Chrysene	0.400	0.400	0.0	116	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.369	7.8	113	0.00
35 I	Perylene-d12	0.400	0.400	0.0	125	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.366	8.5	123	0.00
37	Benzo(b)fluoranthene	0.400	0.360	10.0	121	0.00
38	Benzo(k)fluoranthene	0.400	0.368	8.0	116	0.00
39 C	Benzo(a)pyrene	0.400	0.367	8.3	119	0.00
40	Dibenzo(a,h)anthracene	0.400	0.375	6.3	126	0.00
41	Benzo(g,h,i)perylene	0.400	0.360	10.0	116	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4