

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020124\
 Data File : BN029477.D
 Acq On : 01 Feb 2024 09:12
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 01 09:47:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 01:51:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	141	0.00
2	1,4-Dioxane	0.400	0.372	7.0	131	0.00
3	n-Nitrosodimethylamine	0.400	0.356	11.0	130	0.00
4 S	2-Fluorophenol	0.400	0.396	1.0	148	0.00
5 S	Phenol-d6	0.400	0.386	3.5	150	0.00
6	bis(2-Chloroethyl)ether	0.400	0.396	1.0	151	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	140	0.00
8 S	Nitrobenzene-d5	0.400	0.378	5.5	142	0.00
9	Naphthalene	0.400	0.392	2.0	140	0.00
10	Hexachlorobutadiene	0.400	0.383	4.3	136	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.383	4.3	141	0.00
12	2-Methylnaphthalene	0.400	0.382	4.5	139	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	132	0.00
14 S	2,4,6-Tribromophenol	0.400	0.340	15.0	124	0.00
15 S	2-Fluorobiphenyl	0.400	0.403	-0.8	137	0.00
16	Acenaphthylene	0.400	0.379	5.3	134	0.00
17	Acenaphthene	0.400	0.394	1.5	135	0.00
18	Fluorene	0.400	0.376	6.0	130	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	125	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.402	-0.5	143	0.00
21	4-Bromophenyl-phenylether	0.400	0.391	2.3	129	0.00
22	Hexachlorobenzene	0.400	0.395	1.3	128	0.00
23	Atrazine	0.400	0.393	1.8	136	0.00
24	Pentachlorophenol	0.400	0.381	4.8	137	0.00
25	Phenanthrene	0.400	0.387	3.3	128	0.00
26	Anthracene	0.400	0.371	7.3	126	0.00
27 SURR	Fluoranthene-d10	0.400	0.371	7.3	126	0.00
28	Fluoranthene	0.400	0.359	10.3	126	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	118	0.00
30	Pyrene	0.400	0.413	-3.2	120	0.00
31 S	Terphenyl-d14	0.400	0.399	0.3	118	0.00
32	Benzo(a)anthracene	0.400	0.394	1.5	123	0.00
33	Chrysene	0.400	0.384	4.0	116	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.376	6.0	119	0.00
35 I	Perylene-d12	0.400	0.400	0.0	120	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.367	8.3	119	0.00
37	Benzo(b)fluoranthene	0.400	0.375	6.3	120	0.00
38	Benzo(k)fluoranthene	0.400	0.379	5.3	114	0.00
39 C	Benzo(a)pyrene	0.400	0.383	4.3	119	0.00
40	Dibenzo(a,h)anthracene	0.400	0.375	6.3	121	0.00
41	Benzo(g,h,i)perylene	0.400	0.362	9.5	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4