

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020222\
 Data File : BN018566.D
 Acq On : 03 Feb 2022 01:31
 Operator : CG/JU
 Sample : N1307-11DL 5X
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0Q34DL

Quant Time: Feb 03 03:36:05 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN020222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 02 14:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.923	152	492	0.400	ng/ul	0.00
4) Naphthalene-d8	10.728	136	1763	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.553	164	982	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.289	188	1962	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.464	240	1612	0.400	ng/ul	# 0.00
23) Perylene-d12	23.853	264	1601	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.294	96	218	0.386	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.317	152	151	0.063	ng/ul	0.00
18) Fluoranthene-d10	19.314	212	349	0.076	ng/ul	0.00
Target Compounds						
14) Pentachlorophenol	16.943	266	117	0.320	ng/ul	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020222\
Data File : BN018566.D
Acq On : 03 Feb 2022 01:31
Operator : CG/JU
Sample : N1307-11DL 5X
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0Q34DL

Quant Time: Feb 03 03:36:05 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN020222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 02 14:08:46 2022
Response via : Initial Calibration

