

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020222\
 Data File : BN018563.D
 Acq On : 02 Feb 2022 23:45
 Operator : CG/JU
 Sample : N1307-04DL 10X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0Q33DL

Quant Time: Feb 03 03:34:58 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN020222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 02 14:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.927	152	400	0.400	ng/ul	0.00
4) Naphthalene-d8	10.728	136	1411	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.553	164	775	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.289	188	1524	0.400	ng/ul #	0.00
17) Chrysene-d12	21.462	240	1285	0.400	ng/ul #	0.00
23) Perylene-d12	23.853	264	1236	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	116	0.253	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.317	152	71	0.037	ng/ul	0.00
18) Fluoranthene-d10	19.314	212	161	0.044	ng/ul	0.00
Target Compounds						
14) Pentachlorophenol	16.943	266	86	0.303	ng/ul	Qvalue 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020222\
Data File : BN018563.D
Acq On : 02 Feb 2022 23:45
Operator : CG/JU
Sample : N1307-04DL 10X
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0Q33DL

Quant Time: Feb 03 03:34:58 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN020222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 02 14:08:46 2022
Response via : Initial Calibration





#14

Pentachlorophenol

Concen: 0.303 ng/ul

RT: 16.943 min Scan# 3070

Delta R.T. -0.004 min

Lab File: BN018563.D


Acq: 02 Feb 2022 23:45

Instrument :

BNA_N

ClientSampleId :

C0Q33DL



Tgt Ion:266 Resp: 86

Ion Ratio Lower Upper

266 100

264 66.3 49.9 74.9

268 68.6 51.7 77.5

