



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020222\
 Data File : BN018540.D
 Acq On : 02 Feb 2022 09:07
 Operator : CG/JU
 Sample : DFTPP229
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP229

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN020222.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Feb 02 14:08:46 2022



AutoFind: Scans 2507, 2508, 2509; Background Corrected with Scan 2500

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.4	277995	PASS
68	69	0.00	2	1.7	4583	PASS
69	198	0.00	100	36.4	270080	PASS
70	69	0.00	2	0.5	1414	PASS
127	198	10	80	46.6	346283	PASS
197	198	0.00	2	0.3	2375	PASS
198	198	100	100	100.0	742763	PASS
199	198	5	9	6.8	50755	PASS
275	198	10	60	25.1	186539	PASS
365	198	1	100	3.3	24584	PASS
441	198	0.01	100	11.1	82805	PASS
442	198	50	100	72.1	535317	PASS
443	442	15	24	20.5	109869	PASS

m/z	Abundance
40.00	1693.0
41.00	351.0
42.80	303.0
44.00	2207.0
48.90	943.0
50.90	448.0
73.00	529.0
83.80	422.0
85.80	400.0
96.00	472.0
132.90	408.0
190.90	368.0
207.00	3480.0
208.00	763.0
209.00	385.0
281.10	603.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

m/z	Abundance
35.00	309.0
37.00	1046.0
38.00	2638.0
39.00	19016.0
40.00	2323.0
41.00	748.0
43.00	334.0
44.00	1553.0
45.00	527.0
47.90	321.0
49.00	2261.0
50.00	66528.0
51.00	262016.0
52.00	14204.0
53.00	581.0
55.00	1429.0
56.00	8419.0
57.00	16245.0
57.90	800.0
61.00	3203.0
62.00	3433.0
63.00	9463.0
63.90	1436.0
65.00	4845.0
66.10	362.0
67.00	347.0
68.00	4011.0
69.00	247552.0
70.00	1387.0
73.00	2197.0
74.00	26640.0
75.00	42440.0
76.00	14635.0
77.00	292672.0
78.00	20952.0
79.00	19088.0
80.00	14305.0
81.00	21520.0
82.00	4499.0
83.00	4334.0
84.00	1245.0
85.00	3545.0
86.00	6605.0
87.00	3107.0
88.00	1100.0
89.00	498.0
91.00	4555.0
92.00	5414.0
93.00	32952.0
93.90	2302.0
95.00	594.0
96.00	1817.0
97.00	601.0
98.00	25920.0
99.00	24104.0
100.00	1992.0
101.00	12094.0
102.00	622.0
103.00	3758.0
104.00	7348.0
105.00	6833.0
106.00	2404.0
107.00	86072.0
108.00	14692.0
109.00	2392.0
110.00	160960.0
111.00	23904.0
112.00	2908.0
112.90	963.0
115.10	469.0
116.00	4889.0
117.00	74112.0
118.00	6287.0
118.90	885.0
120.00	1242.0
120.90	499.0
122.00	5356.0
123.00	7889.0
124.00	3716.0
125.00	3644.0
126.10	1373.0
127.00	314688.0
128.00	23328.0
129.00	113648.0
130.00	9874.0
131.00	1792.0
132.00	991.0
133.00	699.0
134.00	3556.0
135.00	9020.0
136.00	3826.0
137.00	6662.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

137.90	1460.0
138.90	558.0
140.00	1644.0
141.00	14614.0
142.00	4942.0
143.00	3642.0
144.00	987.0
145.00	1271.0
145.90	3072.0
147.00	7998.0
148.00	17328.0
149.00	4561.0
150.00	1095.0
151.10	2027.0
151.90	1234.0
153.00	3763.0
154.00	3437.0
155.00	8597.0
156.00	12208.0
157.00	2404.0
157.90	2694.0
158.90	2292.0
160.00	4503.0
161.00	6987.0
162.00	1861.0
162.90	988.0
163.90	711.0
165.00	5204.0
166.00	4171.0
167.00	31392.0
168.00	23176.0
169.00	3131.0
170.00	1152.0
171.00	1227.0
172.00	2481.0
173.00	3746.0
174.00	5562.0
175.00	11853.0
176.00	3710.0
177.00	5863.0
178.00	1759.0
178.90	22792.0
180.00	17744.0
181.00	6760.0
181.90	1152.0
182.90	738.0
183.90	1796.0
185.00	11535.0
186.00	85576.0
187.00	24968.0
188.00	3255.0
188.90	5225.0
190.00	1009.0
191.00	2779.0
192.00	7974.0
193.00	7743.0
194.00	1777.0
194.90	1153.0
196.00	16992.0
197.00	2722.0
197.90	611648.0
198.90	44256.0
200.00	3323.0
201.60	3059.0
203.00	4395.0
204.00	21072.0
205.00	36152.0
206.00	155776.0
207.00	24096.0
208.00	5796.0
209.00	1712.0
210.10	2599.0
211.00	6451.0
211.80	1456.0
213.00	416.0
214.90	1796.0
216.00	3494.0
217.00	38200.0
218.00	5264.0
218.90	653.0
221.10	24744.0
221.80	5657.0
223.00	9077.0
224.00	86208.0
225.10	22728.0
226.00	2453.0
227.00	35672.0
228.00	4573.0
228.90	7292.0
229.90	929.0
231.00	2979.0
232.00	460.0
232.90	560.0
233.90	1801.0
235.00	2768.0
236.00	1631.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

237.00	3897.0
238.00	585.0
238.90	1229.0
240.00	1321.0
241.00	2140.0
242.00	4853.0
243.00	4487.0
244.00	61408.0
245.00	8929.0
246.00	13193.0
246.90	2498.0
248.00	461.0
248.90	2136.0
250.00	439.0
251.00	678.0
251.90	455.0
253.00	1405.0
254.00	2323.0
255.00	330496.0
256.00	46920.0
256.90	4282.0
258.00	19016.0
259.00	3110.0
259.80	575.0
261.00	445.0
264.10	418.0
265.00	7061.0
266.00	1183.0
269.90	316.0
271.00	539.0
272.00	1057.0
273.00	9686.0
274.00	26832.0
275.00	144640.0
276.00	20256.0
277.00	11950.0
278.00	2426.0
278.90	691.0
280.90	511.0
282.10	399.0
283.00	1752.0
284.00	1020.0
285.00	2573.0
286.00	416.0
289.10	505.0
289.90	395.0
291.90	598.0
293.00	2907.0
293.90	783.0
295.00	780.0
296.00	37096.0
297.00	5554.0
298.10	395.0
300.90	370.0
301.90	585.0
303.00	3970.0
304.00	1337.0
308.00	650.0
308.80	417.0
309.80	561.0
313.00	411.0
314.00	1827.0
315.00	4333.0
316.00	2768.0
317.00	536.0
321.10	944.0
322.00	609.0
323.00	11999.0
324.00	1976.0
325.00	350.0
326.90	2218.0
328.00	1196.0
331.90	950.0
333.00	1246.0
334.00	7830.0
334.90	2246.0
336.00	355.0
340.90	1382.0
345.90	2464.0
347.00	520.0
350.80	327.0
352.00	3401.0
353.00	2628.0
354.00	3874.0
354.80	781.0
364.90	16432.0
366.00	2440.0
370.10	319.0
370.90	1144.0
372.00	6105.0
373.00	1562.0
382.90	1832.0
383.90	465.0
390.00	838.0
390.90	633.0
392.00	324.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

400.80	372.0
401.90	2565.0
403.00	3379.0
404.00	1382.0
404.90	409.0
421.00	2625.0
422.00	3578.0
422.90	20656.0
424.00	7798.0
425.00	1107.0
441.00	49640.0
442.00	288960.0
443.00	63240.0
444.00	6427.0
445.10	363.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

m/z	Abundance
35.90	311.0
37.00	1408.0
38.00	3197.0
39.00	22064.0
40.00	1698.0
41.00	704.0
44.00	1744.0
44.90	608.0
49.00	2723.0
50.00	82592.0
51.00	320000.0
52.00	17096.0
53.00	518.0
55.00	1681.0
56.00	9130.0
57.00	21704.0
58.00	1127.0
60.00	350.0
61.00	3972.0
62.00	4018.0
63.00	12676.0
64.00	1651.0
65.00	5868.0
67.10	309.0
68.00	5462.0
69.00	312000.0
70.00	1579.0
73.10	2997.0
74.00	32256.0
75.00	48216.0
76.00	17824.0
77.00	372032.0
78.00	27368.0
79.00	23520.0
80.00	19048.0
81.00	25808.0
82.00	6682.0
83.00	6298.0
83.90	1039.0
84.90	4589.0
86.00	7750.0
87.00	3704.0
88.00	1307.0
88.90	599.0
91.00	5069.0
92.00	6397.0
93.00	42816.0
94.00	2784.0
95.10	839.0
96.00	2080.0
96.90	772.0
98.00	32208.0
99.00	27584.0
100.00	2434.0
101.00	14725.0
102.00	900.0
103.00	4553.0
104.00	9409.0
105.00	8672.0
106.00	2757.0
107.00	116296.0
108.00	17736.0
109.00	3263.0
110.00	208512.0
111.00	32688.0
112.00	3692.0
112.90	1199.0
114.00	319.0
115.00	535.0
116.00	6163.0
117.00	93760.0
118.00	7555.0
119.00	831.0
120.00	1502.0
121.00	562.0
122.00	7653.0
123.00	11997.0
124.00	5468.0
125.00	4555.0
127.00	404224.0
128.00	29656.0
129.00	146688.0
130.00	13662.0
130.90	2315.0
132.00	1030.0
133.00	838.0
134.00	4382.0
135.00	12042.0
136.00	5474.0
137.00	7354.0
138.00	1557.0
138.90	756.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

140.00	1832.0
141.00	20712.0
142.00	6504.0
143.00	4633.0
144.10	1321.0
144.90	919.0
146.00	3431.0
147.00	10673.0
147.90	23200.0
149.00	5567.0
150.00	1483.0
151.10	2459.0
151.70	1161.0
153.00	6031.0
154.00	4523.0
155.00	11611.0
156.00	17200.0
157.00	3115.0
157.90	3997.0
159.00	3336.0
160.00	6187.0
161.00	9383.0
162.00	2376.0
163.00	951.0
164.00	1056.0
165.00	8086.0
166.00	6572.0
167.00	41400.0
168.00	25272.0
169.00	4021.0
170.00	1395.0
170.90	1598.0
172.00	3630.0
173.00	4726.0
174.00	9726.0
175.00	16235.0
176.00	4856.0
176.90	7550.0
178.00	2241.0
179.00	30664.0
180.00	20832.0
181.00	9754.0
181.90	1140.0
183.00	829.0
183.90	2563.0
185.00	15325.0
186.00	117088.0
187.00	33664.0
188.00	4293.0
188.90	6773.0
190.00	1191.0
190.90	3491.0
192.00	10072.0
193.00	11768.0
194.00	2253.0
195.00	1074.0
196.00	22968.0
197.90	854464.0
199.00	57304.0
200.00	4413.0
201.50	3998.0
202.90	6222.0
204.00	30912.0
205.00	54336.0
206.00	226880.0
207.00	31648.0
208.00	7948.0
208.90	2863.0
210.20	3146.0
211.00	8733.0
211.90	1568.0
212.90	806.0
215.00	2573.0
216.00	4801.0
217.00	60248.0
218.00	8242.0
218.90	695.0
221.00	35416.0
221.80	9557.0
223.00	12745.0
224.00	120064.0
225.00	30848.0
226.00	3661.0
227.00	47456.0
228.00	7235.0
229.00	10728.0
230.00	1876.0
231.00	4231.0
232.00	800.0
233.00	955.0
234.00	2978.0
234.90	3945.0
236.00	2668.0
237.00	4850.0
238.00	756.0
239.00	1939.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

240.00	1700.0
241.00	2770.0
242.00	5849.0
243.00	6394.0
244.00	95216.0
245.00	12646.0
246.00	18808.0
246.90	3799.0
248.10	855.0
249.00	3239.0
250.00	528.0
251.00	862.0
251.90	819.0
253.00	1885.0
254.10	3768.0
255.00	482176.0
256.00	71056.0
257.00	5451.0
258.00	28376.0
258.90	5053.0
259.90	958.0
260.90	595.0
263.00	376.0
264.00	881.0
265.00	11855.0
266.00	1588.0
270.00	370.0
271.00	943.0
272.00	1700.0
273.00	14176.0
274.00	38624.0
275.00	214720.0
276.00	30624.0
277.00	19048.0
278.00	3492.0
279.00	842.0
280.90	390.0
282.00	575.0
283.00	2006.0
284.00	1566.0
285.00	3223.0
286.00	879.0
289.00	728.0
290.00	864.0
291.00	517.0
292.00	891.0
293.00	5070.0
294.00	1209.0
296.00	60872.0
297.00	9129.0
298.00	503.0
301.00	836.0
302.00	1133.0
303.00	6253.0
304.00	2279.0
305.00	371.0
308.00	925.0
308.90	582.0
310.00	672.0
311.90	340.0
312.90	554.0
314.00	2781.0
315.00	7073.0
316.00	3948.0
317.00	771.0
320.00	340.0
320.90	1678.0
322.00	973.0
323.00	20896.0
324.00	4187.0
325.00	443.0
326.90	3934.0
328.00	2493.0
328.80	368.0
329.10	355.0
332.00	1527.0
333.00	1565.0
334.00	12064.0
335.00	3307.0
336.00	587.0
339.10	388.0
341.00	2726.0
341.90	576.0
345.90	4192.0
346.90	1066.0
351.00	435.0
352.00	5744.0
353.00	4101.0
354.00	6197.0
355.00	1200.0
359.00	420.0
364.90	29480.0
365.90	4723.0
369.90	554.0
370.90	1895.0
371.90	10070.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

373.00	2776.0
373.90	406.0
383.00	2839.0
384.00	881.0
385.00	344.0
389.90	1539.0
390.90	962.0
392.00	719.0
400.90	645.0
402.00	4033.0
403.00	7212.0
403.90	3037.0
405.00	617.0
421.00	5634.0
422.00	5803.0
423.00	34568.0
424.00	10615.0
425.00	1495.0
441.00	98160.0
442.00	641152.0
443.00	134336.0
444.00	10932.0
445.00	553.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

m/z	Abundance
37.00	960.0
38.00	2841.0
39.10	17440.0
40.00	2041.0
40.90	611.0
44.00	2021.0
44.90	338.0
48.00	317.0
49.00	2742.0
50.00	65272.0
51.00	253312.0
52.00	13375.0
53.10	694.0
55.00	1331.0
56.00	7140.0
57.00	16608.0
58.00	842.0
59.00	331.0
61.00	3314.0
62.00	3358.0
63.00	10103.0
64.00	1230.0
65.00	4689.0
67.00	302.0
68.00	4277.0
69.00	250688.0
70.00	1275.0
73.00	2699.0
74.00	25408.0
75.00	40376.0
76.00	13648.0
77.00	297216.0
78.00	20272.0
79.00	18280.0
80.00	14991.0
81.00	20608.0
82.00	4726.0
83.00	4736.0
83.90	879.0
85.00	4251.0
86.00	6099.0
87.00	2835.0
87.90	959.0
88.80	490.0
91.00	4179.0
91.90	5336.0
93.00	31656.0
94.00	2049.0
95.00	593.0
96.00	1848.0
97.00	722.0
98.00	27032.0
99.00	21192.0
99.90	1892.0
101.00	11443.0
102.00	707.0
103.00	4098.0
104.00	7642.0
105.00	6534.0
106.00	2524.0
107.00	94216.0
108.00	15237.0
109.00	2656.0
110.00	169088.0
111.00	26136.0
112.00	3394.0
113.00	1371.0
114.90	504.0
116.00	5363.0
117.00	80992.0
118.00	6014.0
118.90	645.0
119.90	1136.0
122.00	5777.0
123.00	9655.0
124.00	4315.0
125.00	3984.0
127.00	319936.0
128.00	25960.0
129.00	126152.0
130.00	10599.0
131.00	1954.0
131.90	948.0
133.00	712.0
134.00	3758.0
135.00	10605.0
136.00	4733.0
137.00	5910.0
137.90	1363.0
139.00	591.0
140.00	1577.0
141.00	16944.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

142.00	5274.0
143.00	3759.0
143.90	1047.0
144.90	805.0
146.00	2694.0
147.00	8312.0
148.00	19600.0
149.00	4017.0
149.90	1328.0
151.00	2116.0
151.70	1258.0
153.00	4982.0
154.00	3885.0
155.00	9318.0
156.00	15797.0
157.00	3122.0
157.90	3111.0
158.90	2475.0
160.00	5609.0
161.00	8350.0
161.90	2621.0
163.00	601.0
164.00	975.0
165.00	6280.0
166.00	4589.0
167.00	35936.0
168.00	20464.0
169.00	3074.0
170.00	1299.0
170.90	1364.0
172.00	3399.0
173.00	4086.0
174.00	7889.0
175.00	13258.0
176.00	3885.0
177.00	6119.0
178.00	2392.0
178.90	25720.0
180.00	18632.0
181.00	9443.0
182.00	1316.0
183.00	898.0
183.90	2266.0
185.00	12750.0
186.00	103624.0
187.00	30920.0
188.00	3230.0
188.90	6326.0
190.00	1043.0
191.00	3231.0
192.00	8555.0
193.00	9474.0
194.00	2084.0
195.10	1601.0
196.00	20800.0
197.00	4402.0
198.00	762176.0
199.00	50704.0
200.00	3379.0
201.40	3559.0
203.00	4883.0
204.00	26888.0
205.00	44552.0
206.00	196416.0
207.00	28960.0
208.00	7285.0
209.00	2576.0
210.20	2980.0
211.00	8245.0
212.90	477.0
214.90	2142.0
216.00	4706.0
217.00	53480.0
218.00	6443.0
219.00	602.0
221.00	31632.0
221.80	7307.0
223.00	11396.0
224.00	112248.0
225.00	27480.0
226.00	2941.0
227.00	43312.0
228.00	6158.0
229.00	10081.0
229.90	1364.0
231.00	3781.0
231.90	582.0
232.90	1043.0
233.90	2705.0
235.00	3592.0
236.00	2008.0
237.00	4238.0
238.10	584.0
239.00	1706.0
239.90	1862.0
240.90	2301.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

242.00	6002.0
243.00	7237.0
244.00	87416.0
245.00	11845.0
246.00	17480.0
247.00	3372.0
248.00	792.0
249.00	3274.0
250.00	644.0
251.00	694.0
251.80	601.0
253.00	1702.0
255.00	432384.0
256.00	62896.0
257.00	4820.0
258.00	26792.0
259.00	3871.0
260.00	839.0
261.00	907.0
263.90	590.0
265.00	10918.0
265.90	1619.0
270.00	432.0
270.90	855.0
272.10	1416.0
273.00	13253.0
274.00	35736.0
275.00	200256.0
276.00	27608.0
277.00	18560.0
278.00	3366.0
279.00	544.0
281.00	658.0
282.00	590.0
283.00	1702.0
284.10	1422.0
285.00	2811.0
286.00	538.0
289.00	715.0
290.00	709.0
291.10	514.0
292.10	730.0
293.00	4331.0
294.00	1197.0
294.90	1354.0
296.00	57616.0
297.00	8225.0
298.00	563.0
300.90	762.0
302.00	951.0
303.00	5958.0
304.00	1818.0
304.90	358.0
308.00	640.0
308.80	627.0
309.90	1057.0
313.00	411.0
314.00	3461.0
315.00	6886.0
316.00	3789.0
317.00	632.0
320.90	1657.0
322.00	950.0
323.00	18560.0
324.00	3420.0
324.90	386.0
327.00	4084.0
328.00	2317.0
329.00	407.0
332.00	1156.0
333.00	2322.0
334.00	12922.0
335.00	3089.0
336.00	568.0
339.90	351.0
341.00	1740.0
342.00	701.0
346.00	4524.0
346.80	717.0
351.00	390.0
352.00	6240.0
353.00	3632.0
354.00	6720.0
355.00	1485.0
358.90	672.0
365.00	27840.0
365.90	3554.0
366.90	485.0
370.00	680.0
371.00	1457.0
372.00	10852.0
373.00	2699.0
374.00	477.0
377.00	452.0
383.00	3076.0
383.80	812.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

384.90	328.0
390.00	1532.0
391.00	1004.0
392.00	656.0
400.90	710.0
402.00	4983.0
403.00	6263.0
404.00	2619.0
404.90	553.0
415.10	344.0
420.90	6018.0
422.00	5886.0
423.00	42040.0
424.00	9939.0
425.00	1060.0
441.00	100616.0
442.00	675840.0
443.00	132032.0
444.00	12496.0
444.90	765.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229