

DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/2/2024	BNA_N	<u>BN029492.D</u>
Compound Name	Response	Retention Time
DDT	977035	20.786
DDD	10062	20.392
DDE	117	19.827
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
10179	987214	1.03