

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN020320\
 Data File : BN009526.D
 Acq On : 03 Feb 2020 13:04
 Operator : CG/JU
 Sample : SSTD00553
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD00553

Quant Time: Feb 03 15:02:30 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020320MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 03 14:52:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.45	152	112635	20.00	ng/ul	0.00
18) Naphthalene-d8	10.19	136	461113	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.08	164	298418	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.83	188	637591	20.00	ng/ul	0.00
77) Chrysene-d12	21.05	240	605239	20.00	ng/ul	0.00
85) Perylene-d12	23.16	264	694000	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.08	96	5605	2.06	ng/uL	0.00
5) Phenol-d5	6.64	99	43326	4.34	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.80	67	32075	4.52	ng/ul	0.00
9) 2-Chlorophenol-d4	6.99	132	35135	4.82	ng/ul	0.00
13) 4-Methylphenol-d8	8.16	113	34233	4.31	ng/ul	0.00
19) Nitrobenzene-d5	8.59	128	15986	4.77	ng/ul	0.00
22) 2-Nitrophenol-d4	9.30	143	17338	4.93	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.83	165	33481	4.80	ng/ul	0.00
29) 4-Chloroaniline-d4	10.35	131	30840	3.66	ng/ul	0.00
43) Dimethylphthalate-d6	13.50	166	109385	4.95	ng/ul	0.00
46) Acenaphthylene-d8	13.77	160	125397	4.78	ng/ul	0.00
51) 4-Nitrophenol-d4	14.32	143	14998	3.65	ng/ul	0.00
57) Fluorene-d10	15.08	176	97201	5.01	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.23	200	13069	3.38	ng/ul	0.00
70) Anthracene-d10	16.93	188	143240	4.86	ng/ul	0.00
78) Pyrene-d10	19.24	212	154645	4.75	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.03	264	162628	4.64	ng/ul	0.00

Target Compounds

					Ovalue
2) 1,4-Dioxane	3.12	88	6409	2.114	ng/uL 96
4) Benzaldehyde	6.61	77	31910	6.087	ng/ul 96
6) Phenol	6.67	94	48272	4.613	ng/ul 96
8) Bis(2-Chloroethyl)ether	6.89	93	39012	4.649	ng/ul 89
10) 2-Chlorophenol	7.02	128	35492	4.678	ng/ul 96
11) 2-Methylphenol	7.89	108	33831	4.378	ng/ul 85
12) 2,2'-oxybis(1-Chloropropan	7.98	45	104715	4.749	ng/ul 99
14) Acetophenone	8.26	105	59616	4.642	ng/ul 96
15) N-Nitroso-di-n-propylamine	8.25	70	30914	4.432	ng/ul 93
16) 4-Methylphenol	8.22	108	38131	4.506	ng/ul 93
17) Hexachloroethane	8.49	117	17407	4.953	ng/ul 88
20) Nitrobenzene	8.63	77	47305	4.796	ng/ul 95
21) Isophorone	9.15	82	82682	4.726	ng/ul 98
23) 2-Nitrophenol	9.33	139	19175	4.949	ng/ul 90
24) 2,4-Dimethylphenol	9.40	107	43954	4.944	ng/ul 99
25) Bis(2-Chloroethoxy)methane	9.63	93	54000	5.041	ng/ul 96
27) 2,4-Dichlorophenol	9.86	162	34703	5.081	ng/ul 96
28) Naphthalene	10.24	128	124780	5.029	ng/ul 98
30) 4-Chloroaniline	10.37	127	30741	3.622	ng/ul 100
31) Hexachlorobutadiene	10.53	225	23966	5.113	ng/ul 91
32) Caprolactam	11.15	113	3735	1.596	ng/ul 84
33) 4-Chloro-3-methylphenol	11.50	107	40735	4.915	ng/ul 93
34) 2-Methylnaphthalene	11.87	142	87336	5.112	ng/ul 96

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN020320\
 Data File : BN009526.D
 Acq On : 03 Feb 2020 13:04
 Operator : CG/JU
 Sample : SSTD00553
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD00553

Quant Time: Feb 03 15:02:30 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020320MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 03 14:52:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 1,2,4,5-Tetrachlorobenzene	12.25	216	42687	4.949	ng/ul	93
37) Hexachlorocyclopentadiene	12.23	237	14092	2.682	ng/ul	98
38) 2,4,6-Trichlorophenol	12.50	196	25186	4.724	ng/ul	96
39) 2,4,5-Trichlorophenol	12.57	196	29078	4.957	ng/ul	95
40) 1,1'-Biphenyl	12.90	154	118216	5.232	ng/ul	99
41) 2-Chloronaphthalene	12.94	162	92598	5.190	ng/ul	97
42) 2-Nitroaniline	13.16	65	27980	4.295	ng/ul	94
44) Dimethylphthalate	13.55	163	114654	5.043	ng/ul	99
45) 2,6-Dinitrotoluene	13.67	165	18877	4.363	ng/ul	95
47) Acenaphthylene	13.80	152	141808	5.013	ng/ul	97
48) 3-Nitroaniline	14.00	138	16289	4.033	ng/ul	89
49) Acenaphthene	14.15	153	98546	5.106	ng/ul	95
50) 2,4-Dinitrophenol	14.22	184	6820	2.715	ng/ul	90
52) 4-Nitrophenol	14.33	109	14548	3.824	ng/ul	93
53) Dibenzofuran	14.49	168	137975	5.145	ng/ul	96
54) 2,4-Dinitrotoluene	14.47	165	29034	4.499	ng/ul#	77
55) 2,3,4,6-Tetrachlorophenol	14.72	232	23935	4.775	ng/ul	97
56) Diethylphthalate	14.93	149	116442	5.008	ng/ul	97
58) Fluorene	15.14	166	115787	5.153	ng/ul	97
59) 4-Chlorophenyl-phenylether	15.14	204	56002	5.056	ng/ul	98
60) 4-Nitroaniline	15.17	138	16103	3.531	ng/ul	92
63) 4,6-Dinitro-2-methylphenol	15.24	198	14766	3.593	ng/ul#	82
64) N-Nitrosodiphenylamine	15.36	169	94494	5.005	ng/ul	96
65) 4-Bromophenyl-phenylether	16.04	248	31298	4.870	ng/ul	93
66) Hexachlorobenzene	16.15	284	33823	4.745	ng/ul	94
67) Atrazine	16.32	200	32843	4.755	ng/ul	89
68) Pentachlorophenol	16.50	266	14094	3.401	ng/ul	91
69) Phenanthrene	16.87	178	186437	5.182	ng/ul	97
71) Anthracene	16.97	178	183462	5.009	ng/ul	99
72) 1,2,3,4-Tetrachlorobenzene	12.86	216	46473	5.322	ng/uL	96
73) Pentachlorobenzene	14.40	250	46639	5.494	ng/uL	93
74) Carbazole	17.25	167	155710	4.965	ng/ul	97
75) Di-n-butylphthalate	17.83	149	187076	4.883	ng/ul	99
76) Fluoranthene	18.90	202	203953	4.945	ng/ul	99
79) Pvrene	19.27	202	223985	5.111	ng/ul	99
80) Butylbenzylphthalate	20.20	149	76678	4.534	ng/ul	93
81) 3,3'-Dichlorobenzidine	20.97	252	56082	4.396	ng/ul#	96
82) Benzo(a)anthracene	21.03	228	211748	5.138	ng/ul	99
83) Bis(2-ethylhexyl)phthalate	20.99	149	116977	4.560	ng/ul	100
84) Chrysene	21.08	228	218305	5.437	ng/ul	98
86) Di-n-octyl phthalate	21.85	149	202015	4.069	ng/ul	100
87) Benzo(b)fluoranthene	22.54	252	210168	4.790	ng/ul	99
88) Benzo(k)fluoranthene	22.58	252	202953	4.704	ng/ul	99
90) Benzo(a)pyrene	23.07	252	198112	5.030	ng/ul	97
91) Indeno(1,2,3-cd)pyrene	25.22	276	233447	4.891	ng/ul	99
92) Dibenzo(a,h)anthracene	25.23	278	192623	4.783	ng/ul	99
93) Benzo(g,h,i)perylene	25.84	276	196950	4.960	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

