

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN020320\
 Data File : BN009573.D
 Acq On : 04 Feb 2020 20:27
 Operator : CG/JU
 Sample : L1302-11
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C5AM6

Quant Time: Feb 05 01:01:34 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020320MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 03 16:39:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.44	152	137824	20.00	ng/ul	0.00
18) Naphthalene-d8	10.19	136	571636	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.07	164	353635	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.83	188	755330	20.00	ng/ul	0.00
77) Chrysene-d12	21.04	240	721225	20.00	ng/ul	0.00
85) Perylene-d12	23.15	264	790349	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.08	96	4552	1.36	ng/uL	0.00
5) Phenol-d5	6.64	99	76877	6.40	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.80	67	60138	7.45	ng/ul	0.00
9) 2-Chlorophenol-d4	6.98	132	64907	7.34	ng/ul	0.00
13) 4-Methylphenol-d8	8.16	113	44107	4.61	ng/ul	0.00
19) Nitrobenzene-d5	8.59	128	30680	7.22	ng/ul	0.00
22) 2-Nitrophenol-d4	9.30	143	32018	6.75	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.83	165	54810	6.17	ng/ul	0.00
29) 4-Chloroaniline-d4	10.35	131	62815	6.57	ng/ul	0.00
43) Dimethylphthalate-d6	13.50	166	193494	7.50	ng/ul	0.00
46) Acenaphthylene-d8	13.76	160	228401	7.32	ng/ul	0.00
51) 4-Nitrophenol-d4	14.32	143	16090	3.30	ng/ul	0.01
57) Fluorene-d10	15.08	176	179313	7.90	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.23	200	13863	2.94	ng/ul	0.00
70) Anthracene-d10	16.93	188	253881	7.40	ng/ul	0.00
78) Pyrene-d10	19.23	212	315320	8.32	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.02	264	328225	8.43	ng/ul	0.00

Target Compounds

				Ovalue	
44) Dimethylphthalate	13.55	163	88275	3.222	ng/ul 99
76) Fluoranthene	18.90	202	70595	1.436	ng/ul# 94
79) Pyrene	19.26	202	68292	1.297	ng/ul 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN020320\
Data File : BN009573.D
Acq On : 04 Feb 2020 20:27
Operator : CG/JU
Sample : L1302-11
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C5AM6

Quant Time: Feb 05 01:01:34 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020320MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 03 16:39:56 2020
Response via : Initial Calibration

