



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020425\
Data File : BN036255.D
Acq On : 04 Feb 2025 09:13
Operator : RC/JU
Sample : DFTPP263
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP263

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN012125.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Feb 03 12:27:24 2025



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2463

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.4	192048	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	43.7	184732	PASS
70	69	0.00	2	0.4	785	PASS
127	198	10	80	46.2	195397	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	423189	PASS
199	198	5	9	6.9	29072	PASS
275	198	10	60	26.3	111341	PASS
365	198	1	100	3.8	16028	PASS
441	198	0.01	100	10.9	46101	PASS
442	198	50	100	69.4	293589	PASS
443	442	15	24	19.7	57789	PASS

m/z	Abundance
39.90	761.0
41.00	538.0
41.90	191.0
43.00	581.0
44.00	1920.0
49.00	156.0
55.10	403.0
56.00	215.0
57.00	276.0
69.00	377.0
70.10	158.0
73.00	463.0
82.00	153.0
83.00	271.0
85.00	197.0
96.00	241.0
111.00	184.0
132.90	178.0
147.00	157.0
192.80	174.0
207.10	1322.0
208.00	263.0
208.80	202.0
281.10	430.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

m/z	Abundance
36.10	220.0
37.10	1149.0
38.10	3463.0
39.00	20152.0
40.00	1683.0
41.00	658.0
43.10	246.0
44.00	1604.0
45.00	504.0
46.80	173.0
48.10	157.0
49.10	2253.0
50.00	69816.0
51.00	253056.0
52.00	12335.0
53.00	458.0
54.90	1453.0
56.00	7747.0
57.00	16856.0
58.00	835.0
59.00	308.0
59.90	243.0
61.00	2984.0
62.00	3800.0
63.00	9448.0
64.00	1240.0
65.00	4465.0
66.00	372.0
67.10	440.0
69.00	241408.0
70.00	1251.0
73.00	2758.0
74.00	25192.0
75.00	37864.0
76.10	13575.0
77.00	256768.0
78.10	17464.0
79.00	18744.0
80.00	14055.0
81.00	19288.0
82.00	4529.0
83.00	3681.0
83.90	445.0
85.00	3222.0
86.00	5595.0
87.00	2553.0
88.00	981.0
88.90	333.0
91.00	4215.0
92.00	5120.0
93.00	34832.0
94.00	2213.0
95.00	513.0
96.00	1169.0
98.00	26280.0
99.00	18928.0
100.00	1408.0
101.00	9768.0
101.90	594.0
103.00	3786.0
104.00	7105.0
105.00	6554.0
107.00	79984.0
108.00	12266.0
110.00	132416.0
111.00	20600.0
112.00	2773.0
112.90	1136.0
114.00	236.0
114.90	490.0
116.00	4617.0
117.00	71200.0
118.00	5158.0
119.00	625.0
120.00	863.0
121.00	468.0
122.00	5187.0
123.00	7482.0
124.00	3528.0
125.00	3139.0
127.00	238656.0
128.00	19712.0
129.00	100672.0
130.00	8965.0
131.00	1688.0
132.00	900.0
132.80	593.0
134.00	3252.0
135.00	8030.0
135.90	3398.0
137.00	3975.0
137.80	998.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

139.00	509.0
140.00	1342.0
141.00	14745.0
142.00	4249.0
143.00	2755.0
144.00	747.0
145.00	682.0
145.90	2614.0
147.00	7417.0
148.00	17384.0
148.90	3201.0
150.00	766.0
151.00	1578.0
151.80	805.0
153.00	4650.0
154.00	3030.0
155.00	7386.0
156.00	11195.0
157.00	2361.0
158.00	2656.0
159.00	1975.0
160.00	4273.0
161.00	6558.0
162.00	1851.0
163.00	568.0
164.00	679.0
165.00	5096.0
166.00	4644.0
167.00	28968.0
168.00	13643.0
169.00	2693.0
169.90	1174.0
170.90	1135.0
172.00	2817.0
173.00	3194.0
174.00	5831.0
175.00	11194.0
176.10	2390.0
177.00	5077.0
178.00	1977.0
178.90	20320.0
180.00	14127.0
181.00	6664.0
182.00	948.0
183.00	324.0
184.00	1584.0
185.00	10622.0
186.00	73232.0
187.00	20904.0
188.00	2345.0
189.00	4969.0
190.00	743.0
191.00	2177.0
192.00	6144.0
193.00	7074.0
194.00	1479.0
196.00	15052.0
198.00	484224.0
199.00	34424.0
200.00	2907.0
201.50	1898.0
203.00	4053.0
204.00	19208.0
205.10	34584.0
206.00	141376.0
207.00	19152.0
208.00	5733.0
208.90	1928.0
210.10	2651.0
211.00	5993.0
211.80	834.0
212.90	439.0
213.80	175.0
214.90	1651.0
216.00	3561.0
217.00	39256.0
218.00	4938.0
218.90	457.0
221.00	19000.0
223.00	8062.0
224.00	78888.0
225.00	19424.0
226.20	2310.0
227.00	35992.0
228.00	4752.0
228.90	7038.0
230.00	939.0
231.00	2904.0
232.10	381.0
233.00	658.0
234.00	2127.0
235.00	2522.0
235.90	1756.0
237.00	2852.0
238.00	585.0
239.00	1565.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

240.00	1222.0
241.00	1970.0
242.00	3874.0
243.10	4282.0
244.00	53584.0
245.00	7629.0
246.00	11735.0
247.00	2171.0
247.90	485.0
249.00	2048.0
249.80	278.0
250.10	292.0
250.60	205.0
251.10	424.0
252.00	526.0
253.00	1582.0
255.00	297984.0
256.00	43584.0
257.10	3115.0
257.90	18632.0
259.00	2828.0
259.90	494.0
260.90	487.0
262.00	179.0
263.10	303.0
263.90	453.0
265.00	7789.0
266.00	1079.0
266.90	188.0
270.00	318.0
271.00	637.0
272.00	917.0
273.00	7757.0
274.00	23832.0
275.00	115728.0
276.00	15708.0
277.00	12483.0
278.00	1741.0
278.90	317.0
280.90	417.0
282.00	450.0
283.00	1272.0
284.00	747.0
285.00	1776.0
286.00	463.0
288.00	201.0
289.00	650.0
290.10	420.0
290.80	336.0
292.10	460.0
293.00	2517.0
293.90	726.0
296.00	36432.0
297.00	5477.0
298.00	368.0
300.90	343.0
302.00	635.0
303.00	3903.0
304.00	1048.0
308.00	339.0
308.80	254.0
309.90	453.0
312.80	394.0
314.00	1486.0
315.00	4383.0
316.00	2123.0
317.00	475.0
319.90	186.0
321.00	1021.0
322.00	696.0
323.10	11002.0
324.00	2206.0
325.00	214.0
326.00	230.0
326.90	1951.0
328.00	888.0
329.00	162.0
332.00	923.0
333.00	957.0
334.00	7028.0
335.00	1860.0
335.90	309.0
339.00	167.0
339.90	235.0
341.00	1609.0
341.80	353.0
345.90	2156.0
346.80	420.0
352.00	3371.0
353.00	1898.0
354.00	2964.0
355.00	685.0
358.90	271.0
365.00	14157.0
365.90	2413.0
369.80	373.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

371.00	712.0
372.00	5792.0
373.10	1368.0
373.90	185.0
383.00	1776.0
384.00	349.0
389.90	640.0
390.80	326.0
391.10	340.0
392.00	318.0
401.00	269.0
401.90	2157.0
403.00	2688.0
403.90	1124.0
421.00	2123.0
422.00	2516.0
423.00	15977.0
424.00	4175.0
424.90	510.0
439.60	358.0
441.00	39592.0
442.00	266944.0
443.00	51648.0
444.00	4925.0
445.00	345.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

m/z	Abundance
37.00	1114.0
38.00	2779.0
39.00	16880.0
40.00	1555.0
41.00	508.0
43.00	190.0
44.00	1841.0
45.00	433.0
48.00	180.0
49.20	1872.0
50.00	59992.0
51.00	219904.0
52.00	11237.0
53.00	349.0
55.00	1247.0
56.00	6481.0
57.00	14018.0
58.00	620.0
59.10	285.0
60.00	204.0
61.00	2885.0
62.00	3275.0
63.00	8482.0
64.00	1402.0
65.00	4100.0
66.10	289.0
67.00	366.0
69.00	214464.0
70.00	1014.0
73.00	2252.0
74.00	23264.0
75.00	35008.0
76.10	12928.0
77.00	226816.0
78.00	15514.0
79.00	18152.0
80.00	12432.0
81.00	18488.0
82.00	4127.0
83.00	3528.0
84.00	393.0
85.00	3013.0
86.00	5557.0
87.00	2274.0
88.00	878.0
89.00	262.0
90.10	158.0
91.00	3983.0
91.90	4568.0
93.00	31528.0
94.00	2104.0
95.00	430.0
96.00	1257.0
96.80	471.0
98.00	24720.0
99.00	17480.0
100.00	1562.0
101.00	9466.0
102.00	638.0
103.00	2982.0
104.00	6594.0
105.00	5783.0
107.00	73272.0
108.00	11444.0
110.00	122096.0
111.00	19448.0
112.00	2674.0
112.90	876.0
114.00	254.0
114.90	433.0
116.10	4549.0
117.00	68904.0
117.90	4584.0
118.90	558.0
119.90	847.0
122.00	5255.0
123.00	7458.0
124.00	3240.0
125.00	3158.0
127.00	231744.0
128.00	16576.0
129.00	95848.0
130.00	7925.0
130.90	1279.0
132.00	922.0
133.00	454.0
134.00	3294.0
135.00	8268.0
135.90	3442.0
137.00	4170.0
137.90	766.0
139.00	653.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

140.00	1421.0
141.00	13101.0
142.00	4238.0
142.90	3054.0
143.90	901.0
144.80	576.0
146.00	2702.0
147.00	6441.0
148.00	16038.0
149.00	3261.0
149.90	831.0
151.10	1525.0
151.80	1298.0
153.00	4772.0
154.00	3191.0
155.00	6976.0
156.00	10712.0
157.00	2089.0
158.00	2338.0
159.00	2146.0
160.00	4685.0
161.00	6048.0
162.00	1678.0
162.90	563.0
164.00	535.0
165.00	5822.0
166.00	4248.0
167.00	27216.0
168.00	13539.0
169.00	2234.0
170.00	988.0
170.90	1073.0
172.00	2668.0
173.00	2728.0
174.00	5873.0
175.00	11253.0
176.00	2518.0
177.00	4968.0
177.90	1833.0
179.00	21840.0
180.00	13891.0
181.00	6278.0
182.00	1014.0
183.00	549.0
184.00	1658.0
185.00	10228.0
186.00	75120.0
187.00	22032.0
188.00	2210.0
189.00	4654.0
190.00	926.0
191.00	2502.0
192.00	6002.0
193.00	7917.0
194.00	1622.0
195.00	940.0
196.00	13469.0
198.00	503872.0
199.00	34200.0
200.00	3007.0
201.50	2071.0
203.00	4594.0
204.00	20168.0
205.00	34760.0
206.10	141184.0
207.00	19224.0
208.00	5642.0
209.00	1767.0
210.10	2524.0
211.00	5601.0
213.00	459.0
214.00	215.0
215.00	1929.0
215.90	3363.0
217.00	41400.0
218.00	5557.0
219.00	419.0
221.10	19688.0
221.80	5340.0
223.00	9240.0
224.00	80824.0
225.00	20744.0
226.00	2585.0
227.00	37664.0
228.00	4896.0
229.00	7878.0
230.00	788.0
231.00	2790.0
232.00	414.0
233.00	702.0
234.00	2347.0
235.00	2383.0
235.90	1614.0
237.00	2412.0
238.00	476.0
239.00	1684.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

240.00	1167.0
241.00	1952.0
241.90	4116.0
243.10	5256.0
244.10	61368.0
245.00	7871.0
246.00	13788.0
246.90	2691.0
248.00	709.0
249.00	2135.0
249.90	415.0
251.10	684.0
252.10	644.0
253.00	1583.0
255.00	334976.0
256.00	50680.0
257.00	3570.0
258.00	20784.0
258.90	3225.0
259.90	437.0
261.00	613.0
263.00	236.0
264.10	530.0
265.00	8078.0
266.00	1141.0
266.90	174.0
270.00	633.0
271.00	654.0
272.00	1008.0
273.00	10921.0
274.10	25664.0
275.00	136896.0
276.00	17480.0
277.00	12356.0
278.00	2394.0
279.00	629.0
281.10	503.0
282.00	355.0
283.00	1477.0
284.00	1237.0
285.10	2465.0
285.90	558.0
287.80	176.0
289.00	492.0
290.00	321.0
291.00	358.0
292.00	522.0
292.90	2809.0
294.10	729.0
296.00	42496.0
297.00	6131.0
297.90	458.0
300.90	442.0
302.00	631.0
303.00	4922.0
304.00	1329.0
304.90	166.0
307.90	725.0
308.90	585.0
309.90	551.0
313.10	463.0
314.00	1865.0
314.90	4400.0
316.00	2313.0
316.90	408.0
320.10	262.0
321.00	1372.0
322.00	884.0
323.00	12676.0
324.00	2491.0
325.00	244.0
325.90	273.0
326.90	2474.0
328.00	1097.0
329.00	201.0
331.90	1093.0
333.00	1379.0
334.00	8969.0
335.00	1984.0
335.90	311.0
339.10	298.0
340.10	228.0
341.00	1781.0
342.00	504.0
346.00	3014.0
347.00	504.0
351.10	321.0
352.00	3723.0
353.00	2645.0
354.00	4207.0
355.10	617.0
359.00	304.0
360.10	153.0
365.00	20312.0
366.00	2790.0
367.00	216.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

369.80	415.0
370.90	1133.0
372.00	5938.0
373.00	1418.0
373.90	254.0
377.00	170.0
382.90	1953.0
383.80	490.0
384.90	198.0
390.00	936.0
391.00	668.0
392.00	448.0
400.90	258.0
401.90	2545.0
403.00	4125.0
404.00	1473.0
404.90	329.0
421.00	3372.0
422.00	3212.0
423.00	19752.0
424.00	4520.0
425.00	486.0
439.20	367.0
441.00	58344.0
442.00	368896.0
443.00	71904.0
444.00	7062.0
444.90	417.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

m/z	Abundance
37.00	468.0
38.10	1505.0
39.00	7909.0
40.00	1281.0
41.10	384.0
43.00	241.0
44.00	1743.0
44.90	285.0
49.00	1032.0
50.00	28248.0
51.00	103184.0
52.00	5155.0
52.90	257.0
54.90	783.0
56.00	3313.0
57.00	6413.0
57.90	409.0
61.00	1468.0
62.00	1744.0
63.00	3645.0
64.00	557.0
65.00	1838.0
65.90	168.0
69.00	99456.0
70.00	565.0
73.00	1548.0
74.00	10862.0
75.00	16576.0
76.00	5596.0
77.00	110416.0
78.00	7159.0
79.00	8288.0
80.00	6229.0
81.00	7820.0
82.00	2122.0
82.90	1602.0
83.90	356.0
84.90	1361.0
85.90	2287.0
87.00	1141.0
88.00	353.0
89.00	302.0
91.00	1730.0
92.00	2066.0
93.00	14617.0
94.00	844.0
95.00	361.0
96.00	691.0
98.00	12513.0
99.00	8726.0
100.00	628.0
101.00	4942.0
102.00	216.0
103.00	1489.0
104.00	3013.0
105.00	3109.0
105.90	965.0
107.00	37216.0
108.00	5418.0
110.00	61824.0
111.00	9416.0
111.90	1159.0
112.90	430.0
115.00	283.0
116.00	2006.0
117.00	35864.0
118.00	2734.0
118.90	311.0
120.00	436.0
120.90	207.0
122.00	2331.0
123.00	3993.0
124.00	1773.0
125.00	1527.0
127.00	115792.0
128.00	8419.0
129.00	48608.0
130.00	3868.0
130.90	995.0
132.10	472.0
133.00	362.0
133.90	1460.0
135.00	4012.0
135.90	1615.0
137.00	2172.0
138.00	412.0
138.90	260.0
140.00	723.0
141.00	6670.0
142.00	2328.0
142.90	1528.0
144.00	528.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

144.90	533.0
146.00	1150.0
147.00	3628.0
148.00	8874.0
148.90	1742.0
150.00	503.0
151.00	920.0
151.70	544.0
153.00	2615.0
153.90	1965.0
155.00	3666.0
156.00	5952.0
157.10	1030.0
158.00	1375.0
159.00	1089.0
159.90	2133.0
160.90	3346.0
162.00	929.0
162.90	251.0
164.10	318.0
165.00	2821.0
166.00	2117.0
167.00	14972.0
168.00	7555.0
168.90	1096.0
170.00	584.0
171.00	735.0
172.00	1377.0
172.90	1699.0
174.00	3185.0
175.00	5702.0
175.90	1536.0
176.90	2391.0
178.10	1061.0
179.00	11432.0
180.00	6962.0
181.00	3151.0
181.90	649.0
183.00	337.0
184.10	1065.0
185.00	5357.0
186.00	40552.0
187.00	11318.0
188.00	1207.0
189.00	3085.0
189.90	438.0
191.00	1577.0
192.00	4187.0
193.00	4216.0
194.00	1072.0
196.00	7915.0
198.00	281472.0
199.00	18592.0
200.00	1567.0
201.30	1208.0
202.00	523.0
203.00	2380.0
204.00	11815.0
205.00	20280.0
206.10	79816.0
207.00	11152.0
208.00	2757.0
209.00	1175.0
210.10	1498.0
211.00	3092.0
212.60	181.0
213.00	151.0
213.90	176.0
214.90	1238.0
217.00	22200.0
218.00	2788.0
219.00	262.0
221.00	10355.0
221.90	3036.0
223.00	5464.0
224.00	45320.0
225.00	11134.0
226.10	1314.0
227.00	21464.0
227.90	2969.0
229.00	4156.0
230.00	563.0
231.00	1770.0
232.00	268.0
233.10	518.0
234.00	1308.0
235.00	1444.0
235.90	830.0
237.00	1551.0
238.00	231.0
239.00	777.0
239.90	666.0
241.00	1191.0
242.00	2695.0
243.00	3039.0
244.00	32736.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

245.00	4414.0
246.00	7532.0
246.90	1187.0
248.10	371.0
249.00	1729.0
249.90	189.0
251.00	406.0
252.00	351.0
253.10	887.0
255.00	182592.0
256.00	24840.0
257.00	2108.0
258.00	11933.0
259.00	1961.0
260.00	372.0
260.90	388.0
264.10	332.0
265.00	5124.0
265.90	816.0
269.90	216.0
271.00	447.0
273.00	5726.0
274.00	15799.0
275.00	81400.0
276.00	10661.0
277.00	7733.0
278.00	1371.0
279.00	298.0
281.00	430.0
282.00	239.0
283.00	1093.0
284.00	622.0
285.00	1202.0
286.00	289.0
289.00	354.0
289.90	222.0
291.10	196.0
292.00	390.0
293.00	1880.0
294.00	580.0
295.00	655.0
296.00	25264.0
296.90	3456.0
297.90	252.0
301.00	313.0
301.90	504.0
303.00	2931.0
304.00	757.0
307.90	612.0
308.90	290.0
310.10	355.0
312.80	297.0
314.00	1162.0
315.00	3070.0
316.00	1453.0
317.00	421.0
319.70	172.0
320.90	756.0
322.00	519.0
323.00	7730.0
324.00	1200.0
325.00	172.0
325.90	253.0
327.00	1686.0
328.00	738.0
332.00	647.0
333.00	803.0
334.00	5716.0
334.90	1334.0
335.90	171.0
338.90	222.0
341.00	997.0
342.00	367.0
345.90	1855.0
347.00	383.0
350.90	189.0
351.90	2552.0
353.00	1759.0
354.00	2813.0
355.00	629.0
358.70	159.0
363.90	310.0
365.00	13614.0
366.00	1657.0
370.00	236.0
371.00	911.0
372.00	4548.0
373.00	1149.0
382.90	1217.0
383.90	308.0
384.80	176.0
389.80	437.0
390.80	374.0
391.80	186.0
392.10	206.0
400.90	339.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263

401.90	1830.0
403.00	2443.0
403.90	987.0
404.90	191.0
414.90	278.0
421.00	2381.0
421.90	2066.0
423.00	13368.0
424.00	3228.0
424.80	263.0
441.10	40368.0
442.00	244928.0
443.00	49816.0
443.90	4181.0
445.00	253.0

Instrument :
BNA_N
ClientSampleId :
DFTPP263