

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020521\
 Data File : BN013589.D
 Acq On : 05 Feb 2021 14:54
 Operator : CG/JU
 Sample : M1296-07
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-TT-VPB-RW6-GW-238-240

Quant Time: Feb 05 16:07:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 11:21:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.555	152	6189	0.40	ng	0.00	
7) Naphthalene-d8	10.315	136	24205	0.40	ng	# 0.00	
13) Acenaphthene-d10	14.181	164	13361	0.40	ng	0.00	
19) Phenanthrene-d10	16.933	188	27892	0.40	ng	0.00	
29) Chrysene-d12	21.141	240	26903	0.40	ng	# 0.00	
36) Perylene-d12	23.308	264	25433	0.40	ng	# 0.00	
System Monitoring Compounds							
4) 2-Fluorophenol	5.188	112	2064	0.11	ng	0.00	
5) Phenol-d6	6.734	99	1494	0.07	ng	0.00	
8) Nitrobenzene-d5	8.693	82	3762	0.23	ng	0.00	
11) 2-Methylnaphthalene-d10	11.911	152	10262	0.24	ng	0.00	
14) 2,4,6-Tribromophenol	15.678	330	1092	0.19	ng	0.00	
15) 2-Fluorobiphenyl	12.798	172	14041	0.24	ng	0.00	
27) Fluoranthene-d10	18.977	212	24532	0.28	ng	0.00	
31) Terphenyl-d14	19.588	244	21709	0.31	ng	0.00	
Target Compounds							Qvalue
2) 1,4-Dioxane	3.182	88	13545	1.77	ng		97
25) Phenanthrene	16.982	178	2226	0.02	ng	#	91
34) Bis(2-ethylhexyl)phtha...	21.077	149	47848	1.64	ng		100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020521\
 Data File : BN013589.D
 Acq On : 05 Feb 2021 14:54
 Operator : CG/JU
 Sample : M1296-07
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-TT-VPB-RW6-GW-238-240

Quant Time: Feb 05 16:07:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 11:21:14 2021
 Response via : Initial Calibration

