

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020521\
 Data File : BN013581.D
 Acq On : 05 Feb 2021 10:06
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Feb 05 11:21:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 11:21:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.555	152	6044	0.40	ng	0.00
7) Naphthalene-d8	10.315	136	22984	0.40	ng	# 0.00
13) Acenaphthene-d10	14.181	164	12697	0.40	ng	0.00
19) Phenanthrene-d10	16.934	188	26346	0.40	ng	# 0.00
29) Chrysene-d12	21.144	240	27848	0.40	ng	# 0.00
36) Perylene-d12	23.308	264	25391	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.180	112	7025	0.40	ng	0.00
5) Phenol-d6	6.734	99	8517	0.38	ng	0.00
8) Nitrobenzene-d5	8.693	82	5845	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	11.911	152	15675	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.678	330	1813	0.33	ng	0.00
15) 2-Fluorobiphenyl	12.798	172	21759	0.39	ng	0.00
27) Fluoranthene-d10	18.975	212	31681	0.38	ng	0.00
31) Terphenyl-d14	19.586	244	27099	0.38	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.190	88	2982	0.40	ng	97
3) n-Nitrosodimethylamine	3.497	42	2156	0.33	ng	# 91
6) bis(2-Chloroethyl)ether	6.992	93	6948	0.40	ng	99
9) Naphthalene	10.366	128	26599	0.39	ng	100
10) Hexachlorobutadiene	10.657	225	4547	0.38	ng	# 100
12) 2-Methylnaphthalene	11.986	142	18195	0.38	ng	98
16) Acenaphthylene	13.902	152	22701	0.37	ng	100
17) Acenaphthene	14.245	154	16366	0.38	ng	98
18) Fluorene	15.234	166	20873	0.38	ng	100
20) 4,6-Dinitro-2-methylph...	15.323	198	1242	0.38	ng	98
21) 4-Bromophenyl-phenylether	16.143	248	6300	0.38	ng	# 77
22) Hexachlorobenzene	16.252	284	7004	0.38	ng	99
23) Atrazine	16.410	200	3747	0.36	ng	95
24) Pentachlorophenol	16.593	266	1563	0.29	ng	94
25) Phenanthrene	16.970	178	34493	0.39	ng	100
26) Anthracene	17.068	178	28177	0.38	ng	100
28) Fluoranthene	19.005	202	38012	0.38	ng	99
30) Pyrene	19.372	202	38194	0.37	ng	100
32) Benzo(a)anthracene	21.123	228	37050	0.38	ng	98
33) Chrysene	21.176	228	39785	0.37	ng	98
34) Bis(2-ethylhexyl)phtha...	21.080	149	11678	0.39	ng	98
35) Indeno(1,2,3-cd)pyrene	25.464	276	43792	0.39	ng	98
37) Benzo(b)fluoranthene	22.664	252	39941	0.38	ng	95
38) Benzo(k)fluoranthene	22.705	252	41270	0.40	ng	# 95
39) Benzo(a)pyrene	23.215	252	36204	0.42	ng	# 95
40) Dibenzo(a,h)anthracene	25.478	278	37062	0.38	ng	98
41) Benzo(g,h,i)perylene	26.115	276	37987	0.38	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020521\
Data File : BN013581.D
Acq On : 05 Feb 2021 10:06
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Feb 05 11:21:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 05 11:21:14 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.555 min Scan# 593
 Delta R.T. 0.000 min
 Lab File: BN013581.D
 Acq: 05 Feb 2021 10:06




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:152 Resp: 6044
 Ion Ratio Lower Upper
 152 100
 150 152.7 121.8 182.8
 115 57.6 46.6 69.8



#2
 1,4-Dioxane
 Concen: 0.40 ng
 RT: 3.190 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: BN013581.D
 Acq: 05 Feb 2021 10:06

Tgt Ion: 88 Resp: 2982
 Ion Ratio Lower Upper
 88 100
 58 78.9 64.6 96.8
 43 31.2 26.6 40.0









#3
n-Nitrosodimethylamine
Concen: 0.33 ng
RT: 3.497 min Scan# 89
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 2156
Ion Ratio Lower Upper
42 100
74 128.5 95.0 142.6
44 3.8 5.8 8.8#



#4
2-Fluorophenol
Concen: 0.40 ng
RT: 5.180 min Scan# 298
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion: 112 Resp: 7025
Ion Ratio Lower Upper
112 100
64 55.6 44.3 66.5
63 28.4 23.0 34.6









#5
Phenol-d6
Concen: 0.38 ng
RT: 6.734 min Scan# 491
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	99	Resp:	8517
Ion	Ratio	Lower	Upper
99	100		
42	18.9	15.2	22.8
71	31.3	26.1	39.1



#6
bis(2-Chloroethyl)ether
Concen: 0.40 ng
RT: 6.992 min Scan# 523
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:	93	Resp:	6948
Ion	Ratio	Lower	Upper
93	100		
63	77.2	60.8	91.2
95	32.9	26.0	39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.315 min Scan# 845
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	22984
Ion Ratio	Lower	Upper	
136	100		
137	11.3	9.0	13.6
54	5.8	6.4	9.6#
68	4.8	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.693 min Scan# 717
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:	82	Resp:	5845
Ion Ratio	Lower	Upper	
82	100		
128	51.2	39.2	58.8
54	46.3	40.1	60.1









#9
Naphthalene
Concen: 0.39 ng
RT: 10.366 min Scan# 849
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 26599
Ion Ratio Lower Upper
128 100
129 11.4 9.0 13.4
127 13.1 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.38 ng
RT: 10.657 min Scan# 872
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:225 Resp: 4547
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 0.38 ng
RT: 11.911 min Scan# 1030
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 15675
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.38 ng
RT: 11.986 min Scan# 1047
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:142 Resp: 18195
Ion Ratio Lower Upper
142 100
141 87.0 70.5 105.7
115 31.7 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.181 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 12697
Ion Ratio Lower Upper
164 100
162 101.3 79.6 119.4
160 46.0 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 0.33 ng
RT: 15.678 min Scan# 1414
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:330 Resp: 1813
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 36.8 26.8 40.2









#15
2-Fluorobiphenyl
Concen: 0.39 ng
RT: 12.798 min Scan# 1187
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172	Resp:	21759
Ion Ratio	Lower	Upper
172	100	
171	35.2	28.6 42.8
170	23.8	18.9 28.3



#16
Acenaphthylene
Concen: 0.37 ng
RT: 13.902 min Scan# 1274
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:152	Resp:	22701
Ion Ratio	Lower	Upper
152	100	
151	19.4	15.7 23.5
153	13.1	10.5 15.7









#17
Acenaphthene
Concen: 0.38 ng
RT: 14.245 min Scan# 1301
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 16366
Ion Ratio Lower Upper
154 100
153 112.6 88.4 132.6
152 55.2 43.0 64.4



#18
Fluorene
Concen: 0.38 ng
RT: 15.234 min Scan# 1379
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:166 Resp: 20873
Ion Ratio Lower Upper
166 100
165 97.6 77.8 116.8
167 13.4 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.934 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 26346
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.8 9.0 13.6#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.38 ng
RT: 15.323 min Scan# 1386
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:198 Resp: 1242
Ion Ratio Lower Upper
198 100
51 79.6 61.7 92.5
105 41.9 33.0 49.6









#21
4-Bromophenyl-phenylether
Concen: 0.38 ng
RT: 16.143 min Scan# 1451
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 6300
Ion Ratio Lower Upper
248 100
250 103.7 75.5 113.3
141 33.0 53.8 80.8#



#22
Hexachlorobenzene
Concen: 0.38 ng
RT: 16.252 min Scan# 1460
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:284 Resp: 7004
Ion Ratio Lower Upper
284 100
142 35.0 27.9 41.9
249 30.8 24.2 36.2









#23
Atrazine
Concen: 0.36 ng
RT: 16.410 min Scan# 1473
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 3747
Ion Ratio Lower Upper
200 100
173 27.4 18.6 28.0
215 48.8 41.0 61.6



#24
Pentachlorophenol
Concen: 0.29 ng
RT: 16.593 min Scan# 1488
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:266 Resp: 1563
Ion Ratio Lower Upper
266 100
264 61.0 51.6 77.4
268 62.1 54.6 82.0









#25
Phenanthrene
Concen: 0.39 ng
RT: 16.970 min Scan# 1519
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 34493
Ion Ratio Lower Upper
178 100
176 18.5 15.0 22.4
179 15.2 12.3 18.5



#26
Anthracene
Concen: 0.38 ng
RT: 17.068 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:178 Resp: 28177
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 18.975 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 31681
Ion Ratio Lower Upper
212 100
106 13.0 9.7 14.5
104 7.1 5.5 8.3



#28
Fluoranthene
Concen: 0.38 ng
RT: 19.005 min Scan# 1804
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:202 Resp: 38012
Ion Ratio Lower Upper
202 100
101 10.9 8.1 12.1
203 17.1 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.144 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 27848
Ion Ratio Lower Upper
240 100
120 6.4 10.2 15.4#
236 25.2 21.3 31.9



#30
Pyrene
Concen: 0.37 ng
RT: 19.372 min Scan# 1878
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:202 Resp: 38194
Ion Ratio Lower Upper
202 100
200 20.1 16.3 24.5
203 17.4 14.0 21.0









#31
Terphenyl-d14
Concen: 0.38 ng
RT: 19.586 min Scan# 1921
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 27099
Ion Ratio Lower Upper
244 100
212 7.3 6.1 9.1
122 13.1 9.8 14.8



#32
Benzo(a)anthracene
Concen: 0.38 ng
RT: 21.123 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:228 Resp: 37050
Ion Ratio Lower Upper
228 100
226 26.8 20.5 30.7
229 19.4 16.2 24.4









#33
Chrysene
Concen: 0.37 ng
RT: 21.176 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228 Resp: 39785
Ion Ratio Lower Upper
228 100
226 29.1 24.3 36.5
229 19.5 15.1 22.7



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.39 ng
RT: 21.080 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:149 Resp: 11678
Ion Ratio Lower Upper
149 100
167 28.7 22.1 33.1
279 4.3 3.5 5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.39 ng
RT: 25.464 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:276 Resp: 43792
Ion Ratio Lower Upper
276 100
138 26.3 19.8 29.6
277 25.2 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.308 min Scan# 2495
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:264 Resp: 25391
Ion Ratio Lower Upper
264 100
260 24.3 19.3 28.9
265 32.8 66.2 99.4#









#37
Benzo(b)fluoranthene
Concen: 0.38 ng
RT: 22.664 min Scan# 2307
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 39941
Ion Ratio Lower Upper
252 100
253 23.8 20.4 30.6
125 11.6 11.6 17.4



#38
Benzo(k)fluoranthene
Concen: 0.40 ng
RT: 22.705 min Scan# 2319
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Tgt Ion:252 Resp: 41270
Ion Ratio Lower Upper
252 100
253 23.8 20.5 30.7
125 11.9 12.3 18.5#









#39
Benzo(a)pyrene
Concen: 0.42 ng
RT: 23.215 min Scan# 2468
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 36204
Ion Ratio Lower Upper
252 100
253 25.4 21.8 32.6
125 14.5 14.6 21.8#



#40
Dibenzo(a,h)anthracene
Concen: 0.38 ng
RT: 25.478 min Scan# 3129
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06


Tgt Ion:278 Resp: 37062
Ion Ratio Lower Upper
278 100
139 16.6 13.1 19.7
279 25.1 21.4 32.0





#41
Benzo(g,h,i)perylene
Concen: 0.38 ng
RT: 26.115 min Scan# 3315
Delta R.T. 0.000 min
Lab File: BN013581.D
Acq: 05 Feb 2021 10:06

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion: 276 Resp: 37987
Ion Ratio Lower Upper
276 100
277 24.4 20.0 30.0
138 21.9 17.8 26.8

