

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020524\
 Data File : BN029517.D
 Acq On : 05 Feb 2024 20:34
 Operator : MA/JU
 Sample : P1343-05
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-195-GW-680-682

Quant Time: Feb 06 04:22:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 01:51:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.890	152	5134	0.400	ng	0.00
7) Naphthalene-d8	10.690	136	13776	0.400	ng	0.00
13) Acenaphthene-d10	14.532	164	6021	0.400	ng	0.00
19) Phenanthrene-d10	17.280	188	11499	0.400	ng	0.00
29) Chrysene-d12	21.483	240	8784	0.400	ng	# 0.00
35) Perylene-d12	23.858	264	8967	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.464	112	1036	0.084	ng	0.00
5) Phenol-d6	7.060	99	851	0.060	ng	0.00
8) Nitrobenzene-d5	9.057	82	3580	0.312	ng	0.00
11) 2-Methylnaphthalene-d10	12.280	152	6064	0.323	ng	0.00
14) 2,4,6-Tribromophenol	16.026	330	344	0.178	ng	0.00
15) 2-Fluorobiphenyl	13.164	172	8649	0.335	ng	0.00
27) Fluoranthene-d10	19.317	212	9412	0.338	ng	0.00
31) Terphenyl-d14	19.926	244	10067	0.462	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.411	149	2579	0.126	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020524\
Data File : BN029517.D
Acq On : 05 Feb 2024 20:34
Operator : MA/JU
Sample : P1343-05
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-195-GW-680-682

Quant Time: Feb 06 04:22:22 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 01:51:59 2024
Response via : Initial Calibration

