

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020525\
 Data File : BN036306.D
 Acq On : 05 Feb 2025 20:35
 Operator : RC/JU
 Sample : Q1173-06RE
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 VPB192-HYD-20250122RE

Quant Time: Feb 06 01:06:05 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 06 01:04:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	2111	0.400	ng	0.00
7) Naphthalene-d8	10.562	136	4569	0.400	ng	# 0.00
13) Acenaphthene-d10	14.409	164	2296	0.400	ng	0.00
19) Phenanthrene-d10	17.161	188	4463	0.400	ng	# 0.00
29) Chrysene-d12	21.340	240	3799	0.400	ng	# 0.00
35) Perylene-d12	23.628	264	3362	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.370	112	104	0.019	ng	0.00
5) Phenol-d6	6.959	99	26	0.004	ng	0.00
8) Nitrobenzene-d5	8.929	82	1612	0.374	ng	0.00
11) 2-Methylnaphthalene-d10	12.162	152	198	0.032	ng	0.00
14) 2,4,6-Tribromophenol	15.907	330	104	0.071	ng	0.00
15) 2-Fluorobiphenyl	13.041	172	3370	0.329	ng	0.00
27) Fluoranthene-d10	19.188	212	4073	0.352	ng	0.00
31) Terphenyl-d14	19.787	244	4262	0.540	ng	0.00
Target Compounds						
						Qvalue
9) Naphthalene	10.616	128	377	0.028	ng	# 53
17) Acenaphthene	14.473	154	235	0.032	ng	97
18) Fluorene	15.467	166	430	0.046	ng	# 98
25) Phenanthrene	17.198	178	1709	0.127	ng	100
28) Fluoranthene	19.220	202	646	0.041	ng	# 90
34) Bis(2-ethylhexyl)phtha...	21.259	149	1330	0.176	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020525\
Data File : BN036306.D
Acq On : 05 Feb 2025 20:35
Operator : RC/JU
Sample : Q1173-06RE
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
VPB192-HYD-20250122RE

Quant Time: Feb 06 01:06:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 06 01:04:09 2025
Response via : Initial Calibration

