

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020525\
 Data File : BN036316.D
 Acq On : 06 Feb 2025 02:34
 Operator : RC/JU
 Sample : Q1283-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 VPB192-HYD-20250131

Quant Time: Feb 06 03:28:39 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 06 01:04:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	2355	0.400	ng	0.00
7) Naphthalene-d8	10.562	136	5280	0.400	ng	# 0.00
13) Acenaphthene-d10	14.409	164	2768	0.400	ng	0.00
19) Phenanthrene-d10	17.161	188	5326	0.400	ng	# 0.00
29) Chrysene-d12	21.339	240	5009	0.400	ng	0.00
35) Perylene-d12	23.622	264	4931	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.370	112	760	0.124	ng	0.00
5) Phenol-d6	6.952	99	16	0.002	ng	0.00
8) Nitrobenzene-d5	8.929	82	1971	0.395	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	444	0.062	ng	0.02
14) 2,4,6-Tribromophenol	15.907	330	442	0.249	ng	0.00
15) 2-Fluorobiphenyl	13.040	172	4239	0.343	ng	0.00
27) Fluoranthene-d10	19.187	212	5656	0.410	ng	0.00
31) Terphenyl-d14	19.787	244	4494	0.432	ng	0.00
Target Compounds						
24) Pentachlorophenol	16.813	266	722	0.334	ng	95
34) Bis(2-ethylhexyl)phtha...	21.259	149	1745	0.175	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020525\
Data File : BN036316.D
Acq On : 06 Feb 2025 02:34
Operator : RC/JU
Sample : Q1283-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
VPB192-HYD-20250131

Quant Time: Feb 06 03:28:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 06 01:04:09 2025
Response via : Initial Calibration

