

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020723\
 Data File : BN023973.D
 Acq On : 07 Feb 2023 16:33
 Operator : CG/JU
 Sample : PB150670BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB150670BL

Quant Time: Feb 08 02:12:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN020723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 08 01:53:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	4138	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	9588	0.400	ng	0.00
13) Acenaphthene-d10	14.485	164	5408	0.400	ng	0.00
19) Phenanthrene-d10	17.241	188	12694	0.400	ng	0.00
29) Chrysene-d12	21.428	240	11138	0.400	ng	0.00
35) Perylene-d12	23.814	264	8929	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.399	112	3836	0.446	ng	0.00
5) Phenol-d6	7.009	99	4275	0.422	ng	0.00
8) Nitrobenzene-d5	9.003	82	2873	0.373	ng	0.00
11) 2-Methylnaphthalene-d10	12.238	152	4942	0.371	ng	0.00
14) 2,4,6-Tribromophenol	15.987	330	1082	0.389	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	8442	0.395	ng	0.00
27) Fluoranthene-d10	19.275	212	11989	0.389	ng	0.00
31) Terphenyl-d14	19.869	244	9568	0.460	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020723\
Data File : BN023973.D
Acq On : 07 Feb 2023 16:33
Operator : CG/JU
Sample : PB150670BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB150670BL

Quant Time: Feb 08 02:12:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN020723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 08 01:53:47 2023
Response via : Initial Calibration

