

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020723\
 Data File : BN023971.D
 Acq On : 07 Feb 2023 15:07
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN020723

Quant Time: Feb 08 01:55:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN020723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 08 01:53:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	91	0.00
2	1,4-Dioxane	0.400	0.388	3.0	91	0.00
3	n-Nitrosodimethylamine	0.400	0.348	13.0	88	0.00
4 S	2-Fluorophenol	0.400	0.372	7.0	90	0.00
5 S	Phenol-d6	0.400	0.356	11.0	88	0.00
6	bis(2-Chloroethyl)ether	0.400	0.372	7.0	89	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	88	0.00
8 S	Nitrobenzene-d5	0.400	0.348	13.0	88	0.00
9	Naphthalene	0.400	0.360	10.0	87	0.01
10	Hexachlorobutadiene	0.400	0.367	8.3	88	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.346	13.5	85	0.00
12	2-Methylnaphthalene	0.400	0.347	13.3	85	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	81	0.01
14 S	2,4,6-Tribromophenol	0.400	0.317	20.8	74	0.00
15 S	2-Fluorobiphenyl	0.400	0.371	7.3	84	0.00
16	Acenaphthylene	0.400	0.335	16.3	80	0.00
17	Acenaphthene	0.400	0.358	10.5	81	0.00
18	Fluorene	0.400	0.355	11.3	78	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	74	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.398	0.5	71	0.00
21	4-Bromophenyl-phenylether	0.400	0.352	12.0	78	0.00
22	Hexachlorobenzene	0.400	0.362	9.5	78	0.00
23	Atrazine	0.400	0.319	20.3	70	0.00
24	Pentachlorophenol	0.400	0.385	3.8	69	0.00
25	Phenanthrene	0.400	0.344	14.0	73	0.00
26	Anthracene	0.400	0.331	17.3	72	0.01
27 SURR	Fluoranthene-d10	0.400	0.338	15.5	68	0.00
28	Fluoranthene	0.400	0.336	16.0	68	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	64	0.00
30	Pyrene	0.400	0.347	13.3	68	0.00
31 S	Terphenyl-d14	0.400	0.356	11.0	68	0.00
32	Benzo(a)anthracene	0.400	0.346	13.5	64	0.00
33	Chrysene	0.400	0.357	10.8	66	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.316	21.0	60	0.00
35 I	Perylene-d12	0.400	0.400	0.0	65	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.353	11.8	71	0.00
37	Benzo(b)fluoranthene	0.400	0.348	13.0	67	0.00
38	Benzo(k)fluoranthene	0.400	0.342	14.5	65	0.00
39 C	Benzo(a)pyrene	0.400	0.339	15.3	66	0.00
40	Dibenzo(a,h)anthracene	0.400	0.352	12.0	71	0.00
41	Benzo(g,h,i)perylene	0.400	0.362	9.5	73	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN020723