

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020723\
 Data File : BN023979.D
 Acq On : 07 Feb 2023 20:15
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 08 02:13:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN020723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 08 01:53:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	78	0.00
2	1,4-Dioxane	0.400	0.398	0.5	80	0.00
3	n-Nitrosodimethylamine	0.400	0.352	12.0	76	0.00
4 S	2-Fluorophenol	0.400	0.372	7.0	78	0.00
5 S	Phenol-d6	0.400	0.363	9.3	78	0.00
6	bis(2-Chloroethyl)ether	0.400	0.375	6.3	78	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	79	0.00
8 S	Nitrobenzene-d5	0.400	0.346	13.5	79	0.00
9	Naphthalene	0.400	0.366	8.5	80	0.00
10	Hexachlorobutadiene	0.400	0.362	9.5	78	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.343	14.2	75	0.00
12	2-Methylnaphthalene	0.400	0.354	11.5	78	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	75	0.00
14 S	2,4,6-Tribromophenol	0.400	0.339	15.3	74	0.00
15 S	2-Fluorobiphenyl	0.400	0.361	9.8	76	0.00
16	Acenaphthylene	0.400	0.332	17.0	74	0.00
17	Acenaphthene	0.400	0.356	11.0	75	0.00
18	Fluorene	0.400	0.360	10.0	74	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	76	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.401	-0.3	74	0.00
21	4-Bromophenyl-phenylether	0.400	0.322	19.5	73	0.00
22	Hexachlorobenzene	0.400	0.331	17.3	74	0.00
23	Atrazine	0.400	0.312	22.0	70	0.01
24	Pentachlorophenol	0.400	0.387	3.3	71	0.00
25	Phenanthrene	0.400	0.347	13.3	75	0.00
26	Anthracene	0.400	0.330	17.5	74	0.01
27 SURR	Fluoranthene-d10	0.400	0.370	7.5	76	0.00
28	Fluoranthene	0.400	0.369	7.8	77	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	78	0.00
30	Pyrene	0.400	0.328	18.0	77	0.00
31 S	Terphenyl-d14	0.400	0.328	18.0	75	0.00
32	Benzo(a)anthracene	0.400	0.343	14.2	77	0.00
33	Chrysene	0.400	0.358	10.5	80	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.310	22.5	71	0.00
35 I	Perylene-d12	0.400	0.400	0.0	79	0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.333	16.8	81	0.03
37	Benzo(b)fluoranthene	0.400	0.348	13.0	82	0.01
38	Benzo(k)fluoranthene	0.400	0.344	14.0	80	0.00
39 C	Benzo(a)pyrene	0.400	0.341	14.8	81	0.02
40	Dibenzo(a,h)anthracene	0.400	0.327	18.3	80	0.03
41	Benzo(g,h,i)perylene	0.400	0.327	18.3	79	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4