

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020723\
 Data File : BN023987.D
 Acq On : 08 Feb 2023 03:00
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 08 03:40:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN020723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 08 02:37:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2	1,4-Dioxane	0.400	0.377	5.8	92	0.00
3	n-Nitrosodimethylamine	0.400	0.371	7.3	98	0.00
4 S	2-Fluorophenol	0.400	0.356	11.0	90	0.00
5 S	Phenol-d6	0.400	0.357	10.8	92	0.00
6	bis(2-Chloroethyl)ether	0.400	0.385	3.8	97	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	97	0.00
8 S	Nitrobenzene-d5	0.400	0.351	12.3	98	0.00
9	Naphthalene	0.400	0.362	9.5	97	0.00
10	Hexachlorobutadiene	0.400	0.374	6.5	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.358	10.5	97	0.00
12	2-Methylnaphthalene	0.400	0.362	9.5	98	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.400	0.316	21.0	89	0.00
15 S	2-Fluorobiphenyl	0.400	0.361	9.8	98	0.00
16	Acenaphthylene	0.400	0.326	18.5	94	0.00
17	Acenaphthene	0.400	0.360	10.0	99	0.00
18	Fluorene	0.400	0.357	10.8	95	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	92	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.384	4.0	83	0.00
21	4-Bromophenyl-phenylether	0.400	0.343	14.2	94	0.00
22	Hexachlorobenzene	0.400	0.351	12.3	95	0.00
23	Atrazine	0.400	0.316	21.0	87	0.00
24	Pentachlorophenol	0.400	0.372	7.0	81	0.00
25	Phenanthrene	0.400	0.339	15.3	89	0.00
26	Anthracene	0.400	0.326	18.5	88	0.00
27 SURR	Fluoranthene-d10	0.400	0.328	18.0	82	0.00
28	Fluoranthene	0.400	0.327	18.3	82	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	67	0.00
30	Pyrene	0.400	0.395	1.3	80	0.00
31 S	Terphenyl-d14	0.400	0.384	4.0	77	0.00
32	Benzo(a)anthracene	0.400	0.328	18.0	64	0.00
33	Chrysene	0.400	0.359	10.3	69	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.282	29.5#	56	0.00
35 I	Perylene-d12	0.400	0.400	0.0	58	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.347	13.3	62	0.00
37	Benzo(b)fluoranthene	0.400	0.365	8.8	62	0.00
38	Benzo(k)fluoranthene	0.400	0.364	9.0	61	0.00
39 C	Benzo(a)pyrene	0.400	0.342	14.5	59	0.00
40	Dibenzo(a,h)anthracene	0.400	0.351	12.3	62	0.00
41	Benzo(g,h,i)perylene	0.400	0.361	9.8	64	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4