

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020723\
 Data File : BN023964.D
 Acq On : 07 Feb 2023 10:50
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Feb 08 01:30:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN020723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 08 01:29:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	4527	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	11295	0.400	ng	0.00
13) Acenaphthene-d10	14.495	164	6849	0.400	ng	0.01
19) Phenanthrene-d10	17.241	188	14773	0.400	ng	0.00
29) Chrysene-d12	21.432	240	11703	0.400	ng	0.00
35) Perylene-d12	23.826	264	9505	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.399	112	984	0.105	ng	0.00
5) Phenol-d6	7.009	99	1133	0.102	ng	0.00
8) Nitrobenzene-d5	9.003	82	844	0.093	ng	0.00
11) 2-Methylnaphthalene-d10	12.238	152	1518	0.097	ng	0.00
14) 2,4,6-Tribromophenol	15.987	330	316	0.090	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	2669	0.099	ng	0.00
27) Fluoranthene-d10	19.275	212	3295	0.092	ng	0.00
31) Terphenyl-d14	19.873	244	2335	0.107	ng	0.00
Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.608	42	462	0.096	ng	# 90
6) bis(2-Chloroethyl)ether	7.262	93	1161	0.105	ng	99
9) Naphthalene	10.690	128	2757	0.098	ng	# 93
10) Hexachlorobutadiene	10.978	225	710	0.096	ng	# 96
12) 2-Methylnaphthalene	12.314	142	1733	0.092	ng	99
16) Acenaphthylene	14.206	152	2432	0.091	ng	99
17) Acenaphthene	14.549	154	1777	0.098	ng	98
18) Fluorene	15.543	166	2394	0.097	ng	99
21) 4-Bromophenyl-phenylether	16.434	248	926	0.096	ng	# 95
22) Hexachlorobenzene	16.545	284	1142	0.099	ng	97
23) Atrazine	16.707	200	553	0.088	ng	# 88
25) Phenanthrene	17.278	178	4164	0.102	ng	98
26) Anthracene	17.377	178	3047	0.091	ng	98
28) Fluoranthene	19.304	202	4284	0.092	ng	99
30) Pyrene	19.667	202	4349	0.107	ng	98
32) Benzo(a)anthracene	21.423	228	3805	0.102	ng	96
33) Chrysene	21.468	228	4319	0.107	ng	97
34) Bis(2-ethylhexyl)phtha...	21.343	149	2110	0.123	ng	96
36) Indeno(1,2,3-cd)pyrene	26.309	276	4148	0.096	ng	98
37) Benzo(b)fluoranthene	23.098	252	4043	0.105	ng	# 68
38) Benzo(k)fluoranthene	23.145	252	3957	0.104	ng	# 68
39) Benzo(a)pyrene	23.727	252	2890	0.099	ng	# 49
40) Dibenzo(a,h)anthracene	26.323	278	3303	0.096	ng	# 75
41) Benzo(g,h,i)perylene	27.075	276	3625	0.099	ng	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020723\
Data File : BN023964.D
Acq On : 07 Feb 2023 10:50
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Feb 08 01:30:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN020723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 08 01:29:21 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.840 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 4527
Ion Ratio Lower Upper
152 100
150 152.2 121.8 182.6
115 58.9 47.7 71.5



#3
n-Nitrosodimethylamine
Concen: 0.096 ng
RT: 3.608 min Scan# 73
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion: 42 Resp: 462
Ion Ratio Lower Upper
42 100
74 114.1 99.3 148.9
44 19.3 4.2 6.4









#4
2-Fluorophenol
Concen: 0.105 ng
RT: 5.399 min Scan# 321
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	112	Resp:	984
Ion	Ratio	Lower	Upper
112	100		
64	53.3	44.8	67.2
63	30.3	25.1	37.7



#5
Phenol-d6
Concen: 0.102 ng
RT: 7.009 min Scan# 544
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:	99	Resp:	1133
Ion	Ratio	Lower	Upper
99	100		
42	22.9	17.0	25.6
71	36.7	29.0	43.4









#6
bis(2-Chloroethyl)ether
Concen: 0.105 ng
RT: 7.262 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 93 Resp: 1161
Ion Ratio Lower Upper
93 100
63 72.0 58.4 87.6
95 31.2 25.8 38.6



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.648 min Scan# 943
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion: 136 Resp: 11295
Ion Ratio Lower Upper
136 100
137 11.6 9.1 13.7
54 6.9 5.8 8.8
68 5.6 5.0 7.4









#8
Nitrobenzene-d5
Concen: 0.093 ng
RT: 9.003 min Scan# 789
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 82 Resp: 844
Ion Ratio Lower Upper
82 100
128 43.8 31.5 47.3
54 55.3 42.8 64.2



#9
Naphthalene
Concen: 0.098 ng
RT: 10.690 min Scan# 947
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50


Tgt Ion: 128 Resp: 2757
Ion Ratio Lower Upper
128 100
129 14.1 9.3 13.9#
127 16.5 11.0 16.4#








#10
Hexachlorobutadiene
Concen: 0.096 ng
RT: 10.978 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1







Tgt Ion:225 Resp: 710
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 67.3 51.1 76.7



#11
2-Methylnaphthalene-d10
Concen: 0.097 ng
RT: 12.238 min Scan# 1216
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:152 Resp: 1518
Ion Ratio Lower Upper
152 100
151 21.2 16.6 24.8









#12
2-Methylnaphthalene
Concen: 0.092 ng
RT: 12.314 min Scan# 11
Delta R.T. 0.004 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:142 Resp: 1733
Ion Ratio Lower Upper
142 100
141 90.5 72.9 109.3
115 43.1 33.1 49.7



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.495 min Scan# 1473
Delta R.T. 0.010 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:164 Resp: 6849
Ion Ratio Lower Upper
164 100
162 101.3 84.8 127.2
160 48.2 42.0 63.0









#14
2,4,6-Tribromophenol
Concen: 0.090 ng
RT: 15.987 min Scan# 1609
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:330 Resp: 316
Ion Ratio Lower Upper
330 100
332 91.8 77.7 116.5
141 40.2 26.6 39.8#



#15
2-Fluorobiphenyl
Concen: 0.099 ng
RT: 13.116 min Scan# 1344
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:172 Resp: 2669
Ion Ratio Lower Upper
172 100
171 36.7 27.4 41.2
170 25.0 18.2 27.4









#16
Acenaphthylene
Concen: 0.091 ng
RT: 14.206 min Scan# 1446
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 2432
Ion Ratio Lower Upper
152 100
151 19.2 15.9 23.9
153 13.0 10.5 15.7



#17
Acenaphthene
Concen: 0.098 ng
RT: 14.549 min Scan# 1478
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:154 Resp: 1777
Ion Ratio Lower Upper
154 100
153 116.9 92.6 139.0
152 60.4 46.5 69.7





#18

Fluorene

Concen: 0.097 ng

RT: 15.543 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN023964.D

Acq: 07 Feb 2023 10:50

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:166 Resp: 2394

Ion Ratio Lower Upper

166 100

165 100.6 79.3 118.9

167 13.7 11.0 16.6



#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.241 min Scan# 1710

Delta R.T. -0.000 min

Lab File: BN023964.D

Acq: 07 Feb 2023 10:50




Tgt Ion:188 Resp: 14773


Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 8.8 6.1 9.1









#21
4-Bromophenyl-phenylether
Concen: 0.096 ng
RT: 16.434 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:248 Resp: 926
Ion Ratio Lower Upper
248 100
250 102.2 80.8 121.2
141 46.3 29.2 43.8#



#22
Hexachlorobenzene
Concen: 0.099 ng
RT: 16.545 min Scan# 1654
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:284 Resp: 1142
Ion Ratio Lower Upper
284 100
142 33.4 26.2 39.2
249 33.4 25.0 37.4





#23

Atrazine

Concen: 0.088 ng

RT: 16.707 min Scan# 1667

Delta R.T. -0.000 min

Lab File: BN023964.D

Acq: 07 Feb 2023 10:50

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:200 Resp: 553

Ion Ratio Lower Upper

200 100

173 34.4 19.5 29.3#

215 53.8 38.6 58.0



#25

Phenanthrene

Concen: 0.102 ng

RT: 17.278 min Scan# 1713

Delta R.T. -0.000 min

Lab File: BN023964.D

Acq: 07 Feb 2023 10:50




Tgt Ion:178 Resp: 4164


Ion Ratio Lower Upper

178 100

176 20.7 15.6 23.4

179 16.2 12.6 19.0









#26
 Anthracene
 Concen: 0.091 ng
 RT: 17.377 min Scan# 1721
 Delta R.T. 0.012 min
 Lab File: BN023964.D
 Acq: 07 Feb 2023 10:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	178	Resp:	3047
Ion Ratio	Lower	Upper	
178	100		
176	19.5	14.9	22.3
179	15.8	12.2	18.4



#27
 Fluoranthene-d10
 Concen: 0.092 ng
 RT: 19.275 min Scan# 2072
 Delta R.T. 0.004 min
 Lab File: BN023964.D
 Acq: 07 Feb 2023 10:50

Tgt Ion:	212	Resp:	3295
Ion Ratio	Lower	Upper	
212	100		
106	11.6	9.2	13.8
104	7.3	5.7	8.5









#28
Fluoranthene
Concen: 0.092 ng
RT: 19.304 min Scan# 2079
Delta R.T. -0.000 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	202	Resp:	4284
Ion	Ratio	Lower	Upper
202	100		
101	9.7	7.7	11.5
203	17.4	13.7	20.5



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.432 min Scan# 2405
Delta R.T. 0.004 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:	240	Resp:	11703
Ion	Ratio	Lower	Upper
240	100		
120	14.5	11.4	17.0
236	30.0	23.2	34.8





#30

Pyrene

Concen: 0.107 ng

RT: 19.667 min Scan# 2165

Delta R.T. -0.000 min

Lab File: BN023964.D

Acq: 07 Feb 2023 10:50

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:202 Resp: 4349

Ion Ratio Lower Upper

202 100

200 21.6 16.7 25.1

203 18.8 14.4 21.6



#31

Terphenyl-d14

Concen: 0.107 ng

RT: 19.873 min Scan# 2214

Delta R.T. 0.004 min

Lab File: BN023964.D

Acq: 07 Feb 2023 10:50




Tgt Ion:244 Resp: 2335


Ion Ratio Lower Upper

244 100

212 12.1 7.7 11.5#

122 0.0 0.0 0.0









#32
Benzo(a)anthracene
Concen: 0.102 ng
RT: 21.423 min Scan# 2403
Delta R.T. 0.004 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:228 Resp: 3805
Ion Ratio Lower Upper
228 100
226 29.1 21.8 32.8
229 22.0 15.9 23.9



#33
Chrysene
Concen: 0.107 ng
RT: 21.468 min Scan# 2409
Delta R.T. 0.004 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:228 Resp: 4319
Ion Ratio Lower Upper
228 100
226 32.5 25.2 37.8
229 21.6 15.7 23.5









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.123 ng
RT: 21.343 min Scan# 2395
Delta R.T. -0.005 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:149 Resp: 2110
Ion Ratio Lower Upper
149 100
167 24.4 21.1 31.7
279 3.7 3.1 4.7



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.826 min Scan# 2999
Delta R.T. 0.007 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:264 Resp: 9505
Ion Ratio Lower Upper
264 100
260 27.5 21.4 32.0
265 72.0 41.4 62.0#









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.096 ng
RT: 26.309 min Scan# 3844
Delta R.T. 0.013 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:276 Resp: 4148
Ion Ratio Lower Upper
276 100
138 23.6 18.0 27.0
277 25.3 19.9 29.9



#37
Benzo(b)fluoranthene
Concen: 0.105 ng
RT: 23.098 min Scan# 2750
Delta R.T. 0.007 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:252 Resp: 4043
Ion Ratio Lower Upper
252 100
253 40.1 19.8 29.6#
125 25.3 9.4 14.0#









#38
Benzo(k)fluoranthene
Concen: 0.104 ng
RT: 23.145 min Scan# 21
Delta R.T. 0.007 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 3957
Ion Ratio Lower Upper
252 100
253 41.1 20.2 30.2#
125 24.1 9.0 13.4#



#39
Benzo(a)pyrene
Concen: 0.099 ng
RT: 23.727 min Scan# 2965
Delta R.T. 0.007 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:252 Resp: 2890
Ion Ratio Lower Upper
252 100
253 52.2 21.9 32.9#
125 39.2 12.4 18.6#









#40
Dibenzo(a,h)anthracene
Concen: 0.096 ng
RT: 26.323 min Scan# 3851
Delta R.T. 0.007 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Tgt Ion:278 Resp: 3303
Ion Ratio Lower Upper
278 100
139 21.4 13.1 19.7#
279 44.5 21.7 32.5#



#41
Benzo(g,h,i)perylene
Concen: 0.099 ng
RT: 27.075 min Scan# 4110
Delta R.T. 0.007 min
Lab File: BN023964.D
Acq: 07 Feb 2023 10:50

Tgt Ion:276 Resp: 3625
Ion Ratio Lower Upper
276 100
277 29.3 19.6 29.4
138 23.9 16.2 24.4

