

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020723\
 Data File : BN023970.D
 Acq On : 07 Feb 2023 14:30
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Feb 08 01:32:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN020723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 08 01:29:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	5106	0.400	ng	0.00
7) Naphthalene-d8	10.647	136	11206	0.400	ng	0.00
13) Acenaphthene-d10	14.495	164	6783	0.400	ng	0.01
19) Phenanthrene-d10	17.240	188	13560	0.400	ng	0.00
29) Chrysene-d12	21.432	240	12996	0.400	ng	# 0.00
35) Perylene-d12	23.809	264	10245	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.399	112	52187	4.918	ng	0.00
5) Phenol-d6	7.009	99	64395	5.150	ng	0.00
8) Nitrobenzene-d5	9.003	82	49747	5.530	ng	0.00
11) 2-Methylnaphthalene-d10	12.238	152	82565	5.302	ng	0.00
14) 2,4,6-Tribromophenol	15.987	330	20643	5.910	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	138112	5.147	ng	0.00
27) Fluoranthene-d10	19.274	212	181076	5.501	ng	0.00
31) Terphenyl-d14	19.869	244	123472	5.091	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.268	88	23135	4.776	ng	99
3) n-Nitrosodimethylamine	3.593	42	28234	5.204	ng	99
6) bis(2-Chloroethyl)ether	7.269	93	59487	4.758	ng	99
9) Naphthalene	10.701	128	143925	5.150	ng	98
10) Hexachlorobutadiene	10.978	225	37535	5.119	ng	# 100
12) 2-Methylnaphthalene	12.314	142	98121	5.277	ng	97
16) Acenaphthylene	14.206	152	152241	5.764	ng	100
17) Acenaphthene	14.548	154	94615	5.264	ng	98
18) Fluorene	15.543	166	125411	5.145	ng	99
21) 4-Bromophenyl-phenylether	16.433	248	49190	5.581	ng	99
22) Hexachlorobenzene	16.545	284	56552	5.336	ng	99
23) Atrazine	16.707	200	34275	5.942	ng	97
25) Phenanthrene	17.278	178	198405	5.304	ng	99
26) Anthracene	17.377	178	178656	5.803	ng	100
28) Fluoranthene	19.304	202	233589	5.447	ng	100
30) Pyrene	19.666	202	232984	5.169	ng	100
32) Benzo(a)anthracene	21.414	228	225578	5.425	ng	99
33) Chrysene	21.468	228	226696	5.051	ng	98
34) Bis(2-ethylhexyl)phtha...	21.342	149	106202	5.587	ng	99
36) Indeno(1,2,3-cd)pyrene	26.267	276	258124	5.545	ng	100
37) Benzo(b)fluoranthene	23.084	252	218851	5.278	ng	# 94
38) Benzo(k)fluoranthene	23.133	252	215653	5.262	ng	# 93
39) Benzo(a)pyrene	23.703	252	174364	5.544	ng	# 89
40) Dibenzo(a,h)anthracene	26.282	278	207396	5.566	ng	94
41) Benzo(g,h,i)perylene	27.022	276	216168	5.470	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020723\
Data File : BN023970.D
Acq On : 07 Feb 2023 14:30
Operator : CG/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Feb 08 01:32:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN020723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 08 01:29:21 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.840 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 5106
Ion Ratio Lower Upper
152 100
150 155.1 121.8 182.6
115 61.0 47.7 71.5



#2
1,4-Dioxane
Concen: 4.776 ng
RT: 3.268 min Scan# 26
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30

Tgt Ion: 88 Resp: 23135
Ion Ratio Lower Upper
88 100
43 28.7 23.7 35.5
58 67.4 53.4 80.2









#3
n-Nitrosodimethylamine
Concen: 5.204 ng
RT: 3.593 min Scan# 71
Delta R.T. -0.015 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 28234
Ion Ratio Lower Upper
42 100
74 123.4 99.3 148.9
44 5.6 4.2 6.4



#4
2-Fluorophenol
Concen: 4.918 ng
RT: 5.399 min Scan# 321
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30

Tgt Ion: 112 Resp: 52187
Ion Ratio Lower Upper
112 100
64 56.5 44.8 67.2
63 31.1 25.1 37.7









#5
Phenol-d6
Concen: 5.150 ng
RT: 7.009 min Scan# 54
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 64395
Ion Ratio Lower Upper
99 100
42 21.6 17.0 25.6
71 36.7 29.0 43.4



#6
bis(2-Chloroethyl)ether
Concen: 4.758 ng
RT: 7.269 min Scan# 580
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30

Tgt Ion: 93 Resp: 59487
Ion Ratio Lower Upper
93 100
63 74.5 58.4 87.6
95 32.7 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.647 min Scan# 94
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	11206
Ion Ratio	Lower	Upper	
136	100		
137	11.7	9.1	13.7
54	7.8	5.8	8.8
68	6.6	5.0	7.4



#8
Nitrobenzene-d5
Concen: 5.530 ng
RT: 9.003 min Scan# 789
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30

Tgt Ion:	82	Resp:	49747
Ion Ratio	Lower	Upper	
82	100		
128	37.3	31.5	47.3
54	51.7	42.8	64.2









#9
Naphthalene
Concen: 5.150 ng
RT: 10.701 min Scan# 948
Delta R.T. 0.010 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 143925
Ion Ratio Lower Upper
128 100
129 11.0 9.3 13.9
127 12.9 11.0 16.4



#10
Hexachlorobutadiene
Concen: 5.119 ng
RT: 10.978 min Scan# 974
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30

Tgt Ion:225 Resp: 37535
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 51.1 76.7









#11
2-Methylnaphthalene-d10
Concen: 5.302 ng
RT: 12.238 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 82565
Ion Ratio Lower Upper
152 100
151 20.6 16.6 24.8



#12
2-Methylnaphthalene
Concen: 5.277 ng
RT: 12.314 min Scan# 1236
Delta R.T. 0.003 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30

Tgt Ion:142 Resp: 98121
Ion Ratio Lower Upper
142 100
141 88.3 72.9 109.3
115 38.7 33.1 49.7









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.495 min Scan# 14
 Delta R.T. 0.010 min
 Lab File: BN023970.D
 Acq: 07 Feb 2023 14:30




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:164 Resp: 6783
 Ion Ratio Lower Upper
 164 100
 162 99.8 84.8 127.2
 160 47.5 42.0 63.0



#14
 2,4,6-Tribromophenol
 Concen: 5.910 ng
 RT: 15.987 min Scan# 1609
 Delta R.T. -0.000 min
 Lab File: BN023970.D
 Acq: 07 Feb 2023 14:30

Tgt Ion:330 Resp: 20643
 Ion Ratio Lower Upper
 330 100
 332 97.4 77.7 116.5
 141 35.0 26.6 39.8









#15
2-Fluorobiphenyl
Concen: 5.147 ng
RT: 13.116 min Scan# 1344
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 138112
Ion Ratio Lower Upper
172 100
171 34.5 27.4 41.2
170 23.0 18.2 27.4



#16
Acenaphthylene
Concen: 5.764 ng
RT: 14.206 min Scan# 1446
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30

Tgt Ion:152 Resp: 152241
Ion Ratio Lower Upper
152 100
151 19.6 15.9 23.9
153 13.0 10.5 15.7





#17

Acenaphthene

Concen: 5.264 ng

RT: 14.548 min Scan# 1478

Delta R.T. -0.000 min

Lab File: BN023970.D

Acq: 07 Feb 2023 14:30

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:154 Resp: 94615

Ion Ratio Lower Upper

154 100

153 113.5 92.6 139.0

152 56.5 46.5 69.7



#18

Fluorene

Concen: 5.145 ng

RT: 15.543 min Scan# 1571

Delta R.T. -0.000 min

Lab File: BN023970.D

Acq: 07 Feb 2023 14:30




Tgt Ion:166 Resp: 125411


Ion Ratio Lower Upper

166 100

165 98.1 79.3 118.9

167 13.4 11.0 16.6









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.240 min Scan# 1710
Delta R.T. -0.001 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188 Resp: 13560
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.2 6.1 9.1



#21
4-Bromophenyl-phenylether
Concen: 5.581 ng
RT: 16.433 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30

Tgt Ion:248 Resp: 49190
Ion Ratio Lower Upper
248 100
250 99.6 80.8 121.2
141 35.4 29.2 43.8









#22
Hexachlorobenzene
Concen: 5.336 ng
RT: 16.545 min Scan# 1654
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:284 Resp: 56552
Ion Ratio Lower Upper
284 100
142 32.4 26.2 39.2
249 30.9 25.0 37.4



#23
Atrazine
Concen: 5.942 ng
RT: 16.707 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30

Tgt Ion:200 Resp: 34275
Ion Ratio Lower Upper
200 100
173 22.3 19.5 29.3
215 47.0 38.6 58.0









#25
Phenanthrene
Concen: 5.304 ng
RT: 17.278 min Scan# 1713
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 198405
Ion Ratio Lower Upper
178 100
176 19.5 15.6 23.4
179 15.2 12.6 19.0



#26
Anthracene
Concen: 5.803 ng
RT: 17.377 min Scan# 1721
Delta R.T. 0.012 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30

Tgt Ion:178 Resp: 178656
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.2 12.2 18.4





#27

Fluoranthene-d10

Concen: 5.501 ng

RT: 19.274 min Scan# 2071

Delta R.T. 0.004 min

Lab File: BN023970.D

Acq: 07 Feb 2023 14:30

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:212 Resp: 181076

Ion Ratio Lower Upper

212 100

106 11.7 9.2 13.8

104 6.7 5.7 8.5



#28

Fluoranthene

Concen: 5.447 ng

RT: 19.304 min Scan# 2079

Delta R.T. -0.000 min

Lab File: BN023970.D

Acq: 07 Feb 2023 14:30




Tgt Ion:202 Resp: 233589


Ion Ratio Lower Upper

202 100

101 9.5 7.7 11.5

203 17.3 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.432 min Scan# 2404
Delta R.T. 0.004 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:240 Resp: 12996
Ion Ratio Lower Upper
240 100
120 9.4 11.4 17.0#
236 28.3 23.2 34.8



#30
Pyrene
Concen: 5.169 ng
RT: 19.666 min Scan# 2165
Delta R.T. -0.000 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30

Tgt Ion:202 Resp: 232984
Ion Ratio Lower Upper
202 100
200 21.1 16.7 25.1
203 17.8 14.4 21.6









#31
 Terphenyl-d14
 Concen: 5.091 ng
 RT: 19.869 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN023970.D
 Acq: 07 Feb 2023 14:30




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:244 Resp: 123472
 Ion Ratio Lower Upper
 244 100
 212 8.9 7.7 11.5
 122 0.0 0.0 0.0



#32
 Benzo(a)anthracene
 Concen: 5.425 ng
 RT: 21.414 min Scan# 2403
 Delta R.T. -0.005 min
 Lab File: BN023970.D
 Acq: 07 Feb 2023 14:30

Tgt Ion:228 Resp: 225578
 Ion Ratio Lower Upper
 228 100
 226 27.6 21.8 32.8
 229 19.6 15.9 23.9





#33

Chrysene

Concen: 5.051 ng

RT: 21.468 min Scan# 2408

Delta R.T. 0.004 min

Lab File: BN023970.D

Acq: 07 Feb 2023 14:30

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:228 Resp: 226696

Ion Ratio Lower Upper

228 100

226 30.1 25.2 37.8

229 19.5 15.7 23.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 5.587 ng

RT: 21.342 min Scan# 2395

Delta R.T. -0.005 min

Lab File: BN023970.D

Acq: 07 Feb 2023 14:30




Tgt Ion:149 Resp: 106202


Ion Ratio Lower Upper

149 100

167 27.0 21.1 31.7

279 3.4 3.1 4.7









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.809 min Scan# 2993
 Delta R.T. -0.011 min
 Lab File: BN023970.D
 Acq: 07 Feb 2023 14:30




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:264 Resp: 10245
 Ion Ratio Lower Upper
 264 100
 260 27.6 21.4 32.0
 265 72.9 41.4 62.0#



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 5.545 ng
 RT: 26.267 min Scan# 3834
 Delta R.T. -0.028 min
 Lab File: BN023970.D
 Acq: 07 Feb 2023 14:30

Tgt Ion:276 Resp: 258124
 Ion Ratio Lower Upper
 276 100
 138 22.3 18.0 27.0
 277 24.9 19.9 29.9





#37

Benzo(b)fluoranthene

Concen: 5.278 ng

RT: 23.084 min Scan# 21

Delta R.T. -0.008 min

Lab File: BN023970.D

Acq: 07 Feb 2023 14:30

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:252 Resp: 218851

Ion Ratio Lower Upper

252 100

253 21.9 19.8 29.6

125 8.8 9.4 14.0#



#38

Benzo(k)fluoranthene

Concen: 5.262 ng

RT: 23.133 min Scan# 2762

Delta R.T. -0.005 min

Lab File: BN023970.D

Acq: 07 Feb 2023 14:30




Tgt Ion:252 Resp: 215653


Ion Ratio Lower Upper

252 100

253 21.9 20.2 30.2

125 8.7 9.0 13.4#









#39
Benzo(a)pyrene
Concen: 5.544 ng
RT: 23.703 min Scan# 2963
Delta R.T. -0.017 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 174364
Ion Ratio Lower Upper
252 100
253 22.0 21.9 32.9
125 9.9 12.4 18.6#



#40
Dibenzo(a,h)anthracene
Concen: 5.566 ng
RT: 26.282 min Scan# 3839
Delta R.T. -0.034 min
Lab File: BN023970.D
Acq: 07 Feb 2023 14:30


Tgt Ion:278 Resp: 207396
Ion Ratio Lower Upper
278 100
139 15.0 13.1 19.7
279 23.0 21.7 32.5





#41
 Benzo(g,h,i)perylene
 Concen: 5.470 ng
 RT: 27.022 min Scan# 40
 Delta R.T. -0.046 min
 Lab File: BN023970.D
 Acq: 07 Feb 2023 14:30

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0



Tgt Ion:276 Resp: 216168

Ion	Ratio	Lower	Upper
276	100		
277	23.3	19.6	29.4
138	19.3	16.2	24.4

