

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020824\
 Data File : BN029589.D
 Acq On : 08 Feb 2024 22:35
 Operator : MA/JU
 Sample : P1303-17
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0FH6

Quant Time: Feb 08 23:59:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN020624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 07 02:17:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.889	152	5743	0.400	ng/u1	0.00
4) Naphthalene-d8	10.683	136	15783	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.525	164	7231	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.276	188	13838	0.400	ng/u1	0.00
17) Chrysene-d12	21.476	240	10127	0.400	ng/u1	0.00
23) Perylene-d12	23.843	264	13255	0.400	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.332	96	50417	7.041	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.278	152	8755	0.436	ng/u1	0.00
18) Fluoranthene-d10	19.309	212	14013	0.421	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020824\
Data File : BN029589.D
Acq On : 08 Feb 2024 22:35
Operator : MA/JU
Sample : P1303-17
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0FH6

Quant Time: Feb 08 23:59:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN020624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 07 02:17:13 2024
Response via : Initial Calibration

