

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
 Data File : BN029636.D
 Acq On : 10 Feb 2024 07:21
 Operator : MA/JU
 Sample : P1399-01
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AH4

Quant Time: Feb 11 22:09:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 08 00:25:29 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	220603	20.000	ng/ul	0.00
20) Naphthalene-d8	10.675	136	895644	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.522	164	431174	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.275	188	776313	20.000	ng/ul	0.00
79) Chrysene-d12	21.469	240	633408	20.000	ng/ul	-0.01
88) Perylene-d12	23.845	264	725224	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.323	96	26268	3.917	ng/uL	0.00
4) Pyridine-d5	3.740	84	98311	5.496	ng/ul	0.00
7) Phenol-d5	7.046	99	154771	7.501	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	485945	35.287	ng/ul	0.00
11) 2-Chlorophenol-d4	7.405	132	427098	28.563	ng/ul	-0.01
15) 4-Methylphenol-d8	8.587	113	276340	17.957	ng/ul	-0.01
21) Nitrobenzene-d5	9.046	128	264629	38.069	ng/ul	0.00
24) 2-Nitrophenol-d4	9.763	143	231510	38.240	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.299	165	400089	33.481	ng/ul	0.00
31) 4-Chloroaniline-d4	10.822	131	279055	13.688	ng/ul	-0.01
46) Dimethylphthalate-d6	13.946	166	1228903	38.282	ng/ul	0.00
49) Acenaphthylene-d8	14.216	160	1468605	38.523	ng/ul	0.00
54) 4-Nitrophenol-d4	14.728	143	27348	4.489	ng/ul	0.00
60) Fluorene-d10	15.516	176	1022104	37.427	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.640	200	117621	31.095	ng/ul	0.00
73) Anthracene-d10	17.375	188	1484753	40.809	ng/ul	0.00
81) Pyrene-d10	19.669	212	1558982	41.014	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.692	264	1517166	41.807	ng/ul	-0.01
Target Compounds						
48) 2,6-Dinitrotoluene	13.951	165	8308	1.413	ng/ul#	Qvalue 34


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
Data File : BN029636.D
Acq On : 10 Feb 2024 07:21
Operator : MA/JU
Sample : P1399-01
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AH4


Quant Time: Feb 11 22:09:39 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 08 00:25:29 2024
Response via : Initial Calibration





#48
 2,6-Dinitrotoluene
 Concen: 1.413 ng/ul
 RT: 13.951 min Scan# 11
 Delta R.T. -0.170 min
 Lab File: BN029636.D
 Acq: 10 Feb 2024 07:21

Instrument :
 BNA_N
 ClientSampleId :
 C0AH4



Tgt Ion:165 Resp: 8308
 Ion Ratio Lower Upper
 165 100
 89 0.0 55.8 83.8#
 121 29.9 19.8 29.6#

