

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
 Data File : BN029644.D
 Acq On : 10 Feb 2024 12:10
 Operator : MA/JU
 Sample : PB158921BS
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS921

Quant Time: Feb 11 22:11:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 08 00:25:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	254672	20.000	ng/u1	0.00
20) Naphthalene-d8	10.675	136	1024243	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.522	164	482428	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.274	188	836344	20.000	ng/u1	0.00
79) Chrysene-d12	21.474	240	642382	20.000	ng/u1	0.00
88) Perylene-d12	23.845	264	713779	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.328	96	45226	5.841	ng/uL	0.00
4) Pyridine-d5	3.734	84	619085	29.982	ng/u1	0.00
7) Phenol-d5	7.046	99	736063	30.900	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	488268	30.713	ng/u1	0.00
11) 2-Chlorophenol-d4	7.410	132	569126	32.970	ng/u1	0.00
15) 4-Methylphenol-d8	8.593	113	543675	30.603	ng/u1	0.00
21) Nitrobenzene-d5	9.045	128	270223	33.993	ng/u1	0.00
24) 2-Nitrophenol-d4	9.763	143	256088	36.989	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.298	165	461461	33.768	ng/u1	0.00
31) 4-Chloroaniline-d4	10.822	131	653683	28.038	ng/u1	-0.01
46) Dimethylphthalate-d6	13.945	166	1174184	32.691	ng/u1	0.00
49) Acenaphthylene-d8	14.216	160	1455538	34.124	ng/u1	0.00
54) 4-Nitrophenol-d4	14.728	143	203021	29.781	ng/u1	0.00
60) Fluorene-d10	15.516	176	964068	31.552	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.639	200	135051	33.140	ng/u1	0.00
73) Anthracene-d10	17.374	188	1288640	32.877	ng/u1	0.00
81) Pyrene-d10	19.668	212	1306489	33.891	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.692	264	1215292	34.026	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.363	88	96513	12.035	ng/uL	97
5) Pyridine	3.752	79	643317	30.115	ng/u1	98
6) Benzaldehyde	7.028	77	396607	37.143	ng/u1	96
8) Phenol	7.075	94	784933	32.152	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.316	93	662587	32.276	ng/u1	98
12) 2-Chlorophenol	7.440	128	606607	33.495	ng/u1	99
13) 2-Methylphenol	8.322	108	561400	31.866	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.404	45	528955	19.729	ng/u1	99
16) Acetophenone	8.710	105	897665	31.748	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.698	70	485928	33.288	ng/u1	99
18) 4-Methylphenol	8.657	108	602147	32.345	ng/u1	98
19) Hexachloroethane	8.951	117	267654	31.847	ng/u1	97
22) Nitrobenzene	9.093	77	738750	32.798	ng/u1	99
23) Isophorone	9.616	82	1290780	32.900	ng/u1	99
25) 2-Nitrophenol	9.798	139	293019	37.062	ng/u1	99
26) 2,4-Dimethylphenol	9.857	107	549320	28.991	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.104	93	816524	32.748	ng/u1	99
29) 2,4-Dichlorophenol	10.328	162	474319	33.590	ng/u1	97
30) Naphthalene	10.728	128	1870748	31.954	ng/u1	99
32) 4-Chloroaniline	10.845	127	561582	24.401	ng/u1	100
33) Hexachlorobutadiene	11.004	225	279977	32.127	ng/u1	98
34) Caprolactam	11.622	113	85536	18.204	ng/u1	94
35) 4-Chloro-3-methylphenol	11.969	107	534355	32.762	ng/u1	98
36) 2-Methylnaphthalene	12.345	142	1159490	32.309	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
 Data File : BN029644.D
 Acq On : 10 Feb 2024 12:10
 Operator : MA/JU
 Sample : PB158921BS
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS921

Quant Time: Feb 11 22:11:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 08 00:25:29 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.563	142	1140539	32.056	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.704	216	505015	35.316	ng/ul	98
40) Hexachlorocyclopentadiene	12.681	237	273098	29.217	ng/ul	95
41) 2,4,6-Trichlorophenol	12.951	196	301856	36.623	ng/ul	95
42) 2,4,5-Trichlorophenol	13.022	196	338559	36.556	ng/ul	98
43) 1,1'-Biphenyl	13.357	154	1385948	34.065	ng/ul	99
44) 2-Chloronaphthalene	13.398	162	1091874	34.531	ng/ul	100
45) 2-Nitroaniline	13.610	65	334386	35.969	ng/ul	97
47) Dimethylphthalate	13.992	163	1215014	33.047	ng/ul	99
48) 2,6-Dinitrotoluene	14.116	165	240980	36.629	ng/ul	99
50) Acenaphthylene	14.245	152	1737365	33.939	ng/ul	99
51) 3-Nitroaniline	14.439	138	249068	32.528	ng/ul	98
52) Acenaphthene	14.586	153	1131105	33.449	ng/ul	92
53) 2,4-Dinitrophenol	14.645	184	96958	29.040	ng/ul	94
55) 4-Nitrophenol	14.745	109	181296	29.084	ng/ul	95
56) Dibenzofuran	14.922	168	1472307	32.972	ng/ul	99
57) 2,4-Dinitrotoluene	14.898	165	322545	34.822	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.145	232	216911	33.001	ng/ul	99
59) Diethylphthalate	15.363	149	1243353	32.922	ng/ul	99
61) Fluorene	15.575	166	1155006	32.723	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.575	204	523677	33.172	ng/ul	96
63) 4-Nitroaniline	15.604	138	251718	36.424	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.657	198	155799	34.678	ng/ul	98
67) N-Nitrosodiphenylamine	15.792	169	938495	36.010	ng/ul	99
68) 4-Bromophenyl-phenylether	16.469	248	295754	37.361	ng/ul	89
69) Hexachlorobenzene	16.569	284	320103	35.876	ng/ul	97
70) Atrazine	16.745	200	163339	20.060	ng/ul	97
71) Pentachlorophenol	16.916	266	160726	31.018	ng/ul	97
72) Phenanthrene	17.316	178	1650216	34.110	ng/ul	98
74) Anthracene	17.404	178	1668726	34.369	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.316	216	493817	40.237	ng/ul	95
76) Pentachlorobenzene	14.833	250	420291	38.246	ng/ul	99
77) Carbazole	17.680	167	1443596	32.552	ng/ul	98
78) Di-n-butylphthalate	18.263	149	1867196	35.014	ng/ul	99
80) Fluoranthene	19.339	202	1665026	35.556	ng/ul	99
82) Pyrene	19.698	202	1757016	35.467	ng/ul	99
83) Butylbenzylphthalate	20.621	149	732249	36.268	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.398	252	438381	35.233	ng/ul	97
85) Benzo(a)anthracene	21.457	228	1626377	34.465	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.404	149	1109941	35.541	ng/ul	95
87) Chrysene	21.510	228	1529221	34.579	ng/ul	96
89) Di-n-octyl phthalate	22.333	149	1829270	33.062	ng/ul	100
90) Benzo(b)fluoranthene	23.121	252	1657321	36.863	ng/ul	99
91) Benzo(k)fluoranthene	23.174	252	1564300	33.994	ng/ul	97
93) Benzo(a)pyrene	23.739	252	1551706	35.990	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.303	276	1905115	36.286	ng/ul	97
95) Dibenzo(a,h)anthracene	26.333	278	1584456	36.966	ng/ul	97
96) Benzo(g,h,i)perylene	27.045	276	787987	17.945	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
 Data File : BN029644.D
 Acq On : 10 Feb 2024 12:10
 Operator : MA/JU
 Sample : PB158921BS
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS921

Quant Time: Feb 11 22:11:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 08 00:25:29 2024
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 12.035 ng/uL
RT: 3.363 min Scan# 41
Delta R.T. -0.000 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion: 88 Resp: 96513
Ion Ratio Lower Upper
88 100
43 27.9 22.1 33.1
58 66.8 56.5 84.7



#5
Pyridine
Concen: 30.115 ng/uL
RT: 3.752 min Scan# 113
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion: 79 Resp: 643317
Ion Ratio Lower Upper
79 100
52 67.5 55.4 83.0
51 32.1 26.9 40.3









#6
Benzaldehyde
Concen: 37.143 ng/ul
RT: 7.028 min Scan# 61
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion: 77 Resp: 396607
Ion Ratio Lower Upper
77 100
105 93.6 71.0 106.4
106 89.5 69.4 104.0



#8
Phenol
Concen: 32.152 ng/ul
RT: 7.075 min Scan# 678
Delta R.T. -0.000 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion: 94 Resp: 784933
Ion Ratio Lower Upper
94 100
65 27.8 23.9 35.9
66 37.6 32.2 48.2





#10

Bis(2-Chloroethyl)ether

Concen: 32.276 ng/ul

RT: 7.316 min Scan# 718

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion: 93 Resp: 662587

Ion Ratio Lower Upper

93 100

63 74.2 61.4 92.2

95 31.2 24.9 37.3



#12

2-Chlorophenol

Concen: 33.495 ng/ul

RT: 7.440 min Scan# 740

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10




Tgt Ion:128 Resp: 606607


Ion Ratio Lower Upper

128 100

64 50.3 39.8 59.6

130 32.5 25.6 38.4









#13
2-Methylphenol
Concen: 31.866 ng/ul
RT: 8.322 min Scan# 890
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:108 Resp: 561400
Ion Ratio Lower Upper
108 100
107 91.2 71.2 106.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.729 ng/ul
RT: 8.404 min Scan# 904
Delta R.T. -0.012 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion: 45 Resp: 528955
Ion Ratio Lower Upper
45 100
77 15.7 12.7 19.1
79 11.1 9.4 14.2









#16
Acetophenone
Concen: 31.748 ng/ul
RT: 8.710 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:105 Resp: 897665
Ion Ratio Lower Upper
105 100
77 81.1 67.4 101.0
51 24.9 20.7 31.1



#17
N-Nitroso-di-n-propylamine
Concen: 33.288 ng/ul
RT: 8.698 min Scan# 954
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion: 70 Resp: 485928
Ion Ratio Lower Upper
70 100
42 54.4 43.0 64.4
101 9.3 7.4 11.2
130 19.3 15.1 22.7









#18
4-Methylphenol
Concen: 32.345 ng/ul
RT: 8.657 min Scan# 945
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:108 Resp: 602147
Ion Ratio Lower Upper
108 100
107 109.0 88.7 133.1



#19
Hexachloroethane
Concen: 31.847 ng/ul
RT: 8.951 min Scan# 997
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:117 Resp: 267654
Ion Ratio Lower Upper
117 100
201 82.4 64.4 96.6
199 50.6 38.3 57.5









#22
Nitrobenzene
Concen: 32.798 ng/ul
RT: 9.093 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion: 77 Resp: 738750
Ion Ratio Lower Upper
77 100
123 42.3 33.6 50.4
65 12.9 10.6 16.0



#23
Isophorone
Concen: 32.900 ng/ul
RT: 9.616 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion: 82 Resp: 1290780
Ion Ratio Lower Upper
82 100
95 7.0 5.8 8.6
138 16.3 12.6 19.0









#25
2-Nitrophenol
Concen: 37.062 ng/ul
RT: 9.798 min Scan# 1140
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:139 Resp: 293019
Ion Ratio Lower Upper
139 100
65 57.0 46.6 70.0
109 33.2 27.0 40.4



#26
2,4-Dimethylphenol
Concen: 28.991 ng/ul
RT: 9.857 min Scan# 1151
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:107 Resp: 549320
Ion Ratio Lower Upper
107 100
121 49.7 40.0 60.0
122 83.4 67.6 101.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.748 ng/ul
 RT: 10.104 min Scan# 1193
 Delta R.T. -0.006 min
 Lab File: BN029644.D
 Acq: 10 Feb 2024 12:10




Instrument :
 BNA_N
 ClientSampleId :
 SLCS921


Tgt Ion: 93 Resp: 816524
 Ion Ratio Lower Upper
 93 100
 95 32.8 25.7 38.5
 123 10.6 8.2 12.2



#29
 2,4-Dichlorophenol
 Concen: 33.590 ng/ul
 RT: 10.328 min Scan# 1231
 Delta R.T. -0.006 min
 Lab File: BN029644.D
 Acq: 10 Feb 2024 12:10

Tgt Ion: 162 Resp: 474319
 Ion Ratio Lower Upper
 162 100
 164 62.3 53.3 79.9
 98 42.5 33.8 50.8









#30
Naphthalene
Concen: 31.954 ng/ul
RT: 10.728 min Scan# 1298
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:128 Resp: 1870748
Ion Ratio Lower Upper
128 100
129 10.6 9.0 13.6
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 24.401 ng/ul
RT: 10.845 min Scan# 1319
Delta R.T. -0.012 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:127 Resp: 561582
Ion Ratio Lower Upper
127 100
129 32.9 26.4 39.6









#33
Hexachlorobutadiene
Concen: 32.127 ng/ul
RT: 11.004 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:225 Resp: 279977
Ion Ratio Lower Upper
225 100
223 65.4 51.6 77.4
227 67.2 52.2 78.2



#34
Caprolactam
Concen: 18.204 ng/ul
RT: 11.622 min Scan# 1451
Delta R.T. -0.000 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:113 Resp: 85536
Ion Ratio Lower Upper
113 100
55 162.8 125.3 187.9
56 120.8 103.4 155.0









#35
4-Chloro-3-methylphenol
Concen: 32.762 ng/ul
RT: 11.969 min Scan# 1510
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:107 Resp: 534355
Ion Ratio Lower Upper
107 100
144 24.3 18.8 28.2
142 74.0 60.6 90.8



#36
2-Methylnaphthalene
Concen: 32.309 ng/ul
RT: 12.345 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:142 Resp: 1159490
Ion Ratio Lower Upper
142 100
141 89.7 70.4 105.6









#37
1-Methylnaphthalene
Concen: 32.056 ng/ul
RT: 12.563 min Scan# 1610
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:142 Resp: 1140539
Ion Ratio Lower Upper
142 100
141 91.7 73.8 110.8
116 3.9 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 35.316 ng/ul
RT: 12.704 min Scan# 1635
Delta R.T. -0.012 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:216 Resp: 505015
Ion Ratio Lower Upper
216 100
214 77.3 60.4 90.6
179 22.4 18.6 28.0
108 23.6 19.0 28.6









#40
Hexachlorocyclopentadiene
Concen: 29.217 ng/ul
RT: 12.681 min Scan# 1631
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:237 Resp: 273098
Ion Ratio Lower Upper
237 100
235 57.9 50.2 75.4
272 12.1 9.5 14.3



#41
2,4,6-Trichlorophenol
Concen: 36.623 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:196 Resp: 301856
Ion Ratio Lower Upper
196 100
198 90.4 76.8 115.2
200 29.5 24.6 37.0





#42

2,4,5-Trichlorophenol

Concen: 36.556 ng/ul

RT: 13.022 min Scan# 1

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion:196 Resp: 338559

Ion Ratio Lower Upper

196 100

198 92.2 75.3 112.9

200 29.3 24.7 37.1



#43

1,1'-Biphenyl

Concen: 34.065 ng/ul

RT: 13.357 min Scan# 1746

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10




Tgt Ion:154 Resp: 1385948


Ion Ratio Lower Upper

154 100

153 40.6 32.8 49.2

76 15.8 12.9 19.3





#44

2-Chloronaphthalene

Concen: 34.531 ng/ul

RT: 13.398 min Scan# 1753

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion:162 Resp: 1091874

Ion Ratio Lower Upper

162 100

164 32.1 25.5 38.3

127 44.5 35.7 53.5



#45

2-Nitroaniline

Concen: 35.969 ng/ul

RT: 13.610 min Scan# 1789

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10




Tgt Ion: 65 Resp: 334386


Ion Ratio Lower Upper

65 100

92 67.8 54.3 81.5

138 94.4 71.2 106.8









#47
Dimethylphthalate
Concen: 33.047 ng/ul
RT: 13.992 min Scan# 1854
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:163 Resp: 1215014
Ion Ratio Lower Upper
163 100
194 3.7 2.7 4.1
164 9.8 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 36.629 ng/ul
RT: 14.116 min Scan# 1875
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:165 Resp: 240980
Ion Ratio Lower Upper
165 100
89 70.4 55.8 83.8
121 24.8 19.8 29.6





#50

Acenaphthylene

Concen: 33.939 ng/ul

RT: 14.245 min Scan# 1897

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion:152 Resp: 1737365

Ion Ratio Lower Upper

152 100

151 19.5 15.4 23.2

153 12.7 10.5 15.7



#51

3-Nitroaniline

Concen: 32.528 ng/ul

RT: 14.439 min Scan# 1930

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10




Tgt Ion:138 Resp: 249068


Ion Ratio Lower Upper

138 100

108 11.7 9.4 14.2

92 125.8 102.7 154.1









#52
 Acenaphthene
 Concen: 33.449 ng/ul
 RT: 14.586 min Scan# 1955
 Delta R.T. -0.006 min
 Lab File: BN029644.D
 Acq: 10 Feb 2024 12:10

Instrument :
 BNA_N
 ClientSampleId :
 SLCS921

Tgt Ion:153 Resp: 1131105




Ion	Ratio	Lower	Upper
153	100		
152	46.6	39.2	58.8
154	83.6	74.6	111.8




#53
 2,4-Dinitrophenol
 Concen: 29.040 ng/ul
 RT: 14.645 min Scan# 1965
 Delta R.T. -0.006 min
 Lab File: BN029644.D
 Acq: 10 Feb 2024 12:10

Tgt Ion:184 Resp: 96958

Ion	Ratio	Lower	Upper
184	100		
63	82.2	71.2	106.8
154	69.1	58.8	88.2









#55
4-Nitrophenol
Concen: 29.084 ng/ul
RT: 14.745 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:109 Resp: 181296
Ion Ratio Lower Upper
109 100
139 130.0 94.6 142.0
65 135.5 108.5 162.7



#56
Dibenzofuran
Concen: 32.972 ng/ul
RT: 14.922 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:168 Resp: 1472307
Ion Ratio Lower Upper
168 100
139 40.3 32.8 49.2





#57

2,4-Dinitrotoluene

Concen: 34.822 ng/ul

RT: 14.898 min Scan# 2008

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion:165 Resp: 322545

Ion Ratio Lower Upper

165 100

63 53.1 46.7 70.1

182 2.7 1.9 2.9



#58

2,3,4,6-Tetrachlorophenol

Concen: 33.001 ng/ul

RT: 15.145 min Scan# 2050

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Tgt Ion:232 Resp: 216911




Ion Ratio Lower Upper


232 100

131 58.3 46.9 70.3

130 3.3 2.6 4.0

166 35.9 29.4 44.2





#59

Diethylphthalate

Concen: 32.922 ng/ul

RT: 15.363 min Scan# 2086

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion:149 Resp: 1243353

Ion Ratio Lower Upper

149 100

177 20.5 15.8 23.6

150 12.0 9.6 14.4



#61

Fluorene

Concen: 32.723 ng/ul

RT: 15.575 min Scan# 2123

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10




Tgt Ion:166 Resp: 1155006


Ion Ratio Lower Upper

166 100

165 98.4 77.4 116.0

167 13.1 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 33.172 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:204 Resp: 523677
Ion Ratio Lower Upper
204 100
206 33.0 27.5 41.3
141 71.6 60.5 90.7



#63
4-Nitroaniline
Concen: 36.424 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:138 Resp: 251718
Ion Ratio Lower Upper
138 100
92 55.8 47.4 71.2
108 88.1 70.9 106.3









#66
4,6-Dinitro-2-methylphenol
Concen: 34.678 ng/ul
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:198 Resp: 155799
Ion Ratio Lower Upper
198 100
182 4.3 3.9 5.9
77 36.3 29.8 44.8



#67
N-Nitrosodiphenylamine
Concen: 36.010 ng/ul
RT: 15.792 min Scan# 2160
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:169 Resp: 938495
Ion Ratio Lower Upper
169 100
168 67.3 54.5 81.7
167 36.5 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 37.361 ng/ul
RT: 16.469 min Scan# 2275
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:248 Resp: 295754
Ion Ratio Lower Upper
248 100
250 97.1 85.6 128.4
141 91.8 84.6 126.8



#69
Hexachlorobenzene
Concen: 35.876 ng/ul
RT: 16.569 min Scan# 2292
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:284 Resp: 320103
Ion Ratio Lower Upper
284 100
142 46.4 39.2 58.8
249 33.5 25.7 38.5





#70

Atrazine

Concen: 20.060 ng/ul

RT: 16.745 min Scan# 21

Delta R.T. -0.012 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion:200 Resp: 163339

Ion Ratio Lower Upper

200 100

173 28.4 22.0 33.0

215 47.1 36.1 54.1



#71

Pentachlorophenol

Concen: 31.018 ng/ul

RT: 16.916 min Scan# 2351

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10




Tgt Ion:266 Resp: 160726


Ion Ratio Lower Upper

266 100

264 58.9 45.8 68.6

268 62.2 51.6 77.4









#72
Phenanthrene
Concen: 34.110 ng/ul
RT: 17.316 min Scan# 2419
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:178 Resp: 1650216
Ion Ratio Lower Upper
178 100
179 14.9 12.5 18.7
176 18.4 15.7 23.5



#74
Anthracene
Concen: 34.369 ng/ul
RT: 17.404 min Scan# 2434
Delta R.T. -0.012 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:178 Resp: 1668726
Ion Ratio Lower Upper
178 100
179 14.7 11.6 17.4
176 18.0 14.2 21.2









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 40.237 ng/uL
 RT: 13.316 min Scan# 1739
 Delta R.T. -0.006 min
 Lab File: BN029644.D
 Acq: 10 Feb 2024 12:10

Instrument :
 BNA_N
 ClientSampleId :
 SLCS921

Tgt Ion:216 Resp: 493817




Ion	Ratio	Lower	Upper
216	100		
214	81.0	60.1	90.1
218	48.9	38.5	57.7




#76
 Pentachlorobenzene
 Concen: 38.246 ng/uL
 RT: 14.833 min Scan# 1997
 Delta R.T. -0.006 min
 Lab File: BN029644.D
 Acq: 10 Feb 2024 12:10

Tgt Ion:250 Resp: 420291

Ion	Ratio	Lower	Upper
250	100		
248	60.7	49.8	74.8
252	63.5	50.5	75.7









#77
 Carbazole
 Concen: 32.552 ng/ul
 RT: 17.680 min Scan# 2482
 Delta R.T. -0.012 min
 Lab File: BN029644.D
 Acq: 10 Feb 2024 12:10




Instrument :
 BNA_N
 ClientSampleId :
 SLCS921


Tgt Ion:167 Resp: 1443596
 Ion Ratio Lower Upper
 167 100
 166 22.4 16.6 25.0
 139 14.7 11.4 17.0



#78
 Di-n-butylphthalate
 Concen: 35.014 ng/ul
 RT: 18.263 min Scan# 2580
 Delta R.T. -0.006 min
 Lab File: BN029644.D
 Acq: 10 Feb 2024 12:10

Tgt Ion:149 Resp: 1867196
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.0 10.4
 104 5.4 4.7 7.1









#80
Fluoranthene
Concen: 35.556 ng/ul
RT: 19.339 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:202 Resp: 1665026
Ion Ratio Lower Upper
202 100
101 14.0 11.1 16.7
100 10.6 8.9 13.3



#82
Pyrene
Concen: 35.467 ng/ul
RT: 19.698 min Scan# 2824
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:202 Resp: 1757016
Ion Ratio Lower Upper
202 100
101 18.1 14.2 21.4
100 14.7 11.4 17.2





#83

Butylbenzylphthalate

Concen: 36.268 ng/ul

RT: 20.621 min Scan# 2981

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion:149 Resp: 732249

Ion Ratio Lower Upper

149 100

91 82.6 70.0 105.0

206 16.9 14.1 21.1



#84

3,3'-Dichlorobenzidine

Concen: 35.233 ng/ul

RT: 21.398 min Scan# 3113

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10




Tgt Ion:252 Resp: 438381


Ion Ratio Lower Upper

252 100

254 60.1 49.8 74.6

126 13.9 12.2 18.4





#85

Benzo(a)anthracene

Concen: 34.465 ng/ul

RT: 21.457 min Scan# 3123

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion:228 Resp: 1626377

Ion Ratio Lower Upper

228 100

229 18.4 15.0 22.6

226 26.6 21.0 31.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 35.541 ng/ul

RT: 21.404 min Scan# 3114

Delta R.T. -0.012 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10




Tgt Ion:149 Resp: 1109941


Ion Ratio Lower Upper

149 100

167 23.0 20.6 30.8

279 3.0 2.7 4.1





#87

Chrysene

Concen: 34.579 ng/ul

RT: 21.510 min Scan# 3133

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion:228 Resp: 1529221

Ion Ratio Lower Upper

228 100

226 27.5 24.6 37.0

229 19.5 16.2 24.4



#89

Di-n-octyl phthalate

Concen: 33.062 ng/ul




RT: 22.333 min Scan# 3272


Delta R.T. -0.012 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Tgt Ion:149 Resp: 1829270









#90
Benzo(b)fluoranthene
Concen: 36.863 ng/ul
RT: 23.121 min Scan# 3407
Delta R.T. -0.012 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10




Instrument :
BNA_N
ClientSampleId :
SLCS921


Tgt Ion:252 Resp: 1657321
Ion Ratio Lower Upper
252 100
253 21.5 17.5 26.3
125 13.2 11.1 16.7



#91
Benzo(k)fluoranthene
Concen: 33.994 ng/ul
RT: 23.174 min Scan# 3415
Delta R.T. -0.006 min
Lab File: BN029644.D
Acq: 10 Feb 2024 12:10

Tgt Ion:252 Resp: 1564300
Ion Ratio Lower Upper
252 100
253 20.5 18.2 27.4
125 13.7 11.2 16.8





#93

Benzo(a)pyrene

Concen: 35.990 ng/ul

RT: 23.739 min Scan# 3511

Delta R.T. -0.012 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion:252 Resp: 1551706

Ion Ratio Lower Upper

252 100

253 20.4 18.8 28.2

125 14.9 12.2 18.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 36.286 ng/ul

RT: 26.303 min Scan# 3947

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10




Tgt Ion:276 Resp: 1905115


Ion Ratio Lower Upper

276 100

138 28.8 21.7 32.5

277 23.8 17.8 26.8





#95

Dibenzo(a,h)anthracene

Concen: 36.966 ng/ul

RT: 26.333 min Scan# 3951

Delta R.T. -0.006 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Instrument :

BNA_N

ClientSampleId :

SLCS921





Tgt Ion:278 Resp: 1584456

Ion Ratio Lower Upper

278 100

139 21.0 17.8 26.6

279 23.0 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 17.945 ng/ul

RT: 27.045 min Scan# 4073

Delta R.T. -0.024 min

Lab File: BN029644.D

Acq: 10 Feb 2024 12:10

Tgt Ion:276 Resp: 787987

Ion Ratio Lower Upper

276 100

138 28.3 23.9 35.9

277 23.9 18.6 28.0

