

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
 Data File : BN029646.D
 Acq On : 10 Feb 2024 13:59
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020300

Quant Time: Feb 11 22:12:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 08 00:25:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	252578	20.000	ng/u1	0.00
20) Naphthalene-d8	10.675	136	1051720	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.522	164	512713	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.274	188	888628	20.000	ng/u1	0.00
79) Chrysene-d12	21.474	240	703960	20.000	ng/u1	0.00
88) Perylene-d12	23.839	264	803279	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.328	96	56310	7.333	ng/uL	0.00
4) Pyridine-d5	3.734	84	376797	18.399	ng/u1	0.00
7) Phenol-d5	7.045	99	446093	18.882	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	296901	18.830	ng/u1	0.00
11) 2-Chlorophenol-d4	7.404	132	349057	20.389	ng/u1	-0.01
15) 4-Methylphenol-d8	8.592	113	331599	18.820	ng/u1	0.00
21) Nitrobenzene-d5	9.045	128	161138	19.741	ng/u1	0.00
24) 2-Nitrophenol-d4	9.763	143	149983	21.097	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.298	165	273818	19.514	ng/u1	0.00
31) 4-Chloroaniline-d4	10.822	131	420209	17.553	ng/u1	-0.01
46) Dimethylphthalate-d6	13.939	166	685925	17.969	ng/u1	-0.01
49) Acenaphthylene-d8	14.210	160	852160	18.798	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.727	143	113611	15.681	ng/u1	0.00
60) Fluorene-d10	15.516	176	562671	17.327	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.639	200	73377	16.947	ng/u1	0.00
73) Anthracene-d10	17.368	188	766969	18.416	ng/u1	-0.01
81) Pyrene-d10	19.668	212	795608	18.833	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.686	264	763027	18.983	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.363	88	61179	7.692	ng/uL	93
5) Pyridine	3.752	79	391540	18.481	ng/u1	97
6) Benzaldehyde	7.022	77	259103	24.467	ng/u1	98
8) Phenol	7.069	94	452746	18.699	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.316	93	384763	18.898	ng/u1	98
12) 2-Chlorophenol	7.440	128	358136	19.939	ng/u1	98
13) 2-Methylphenol	8.322	108	326399	18.681	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.404	45	494613	18.601	ng/u1	99
16) Acetophenone	8.710	105	537462	19.166	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.692	70	283439	19.578	ng/u1	99
18) 4-Methylphenol	8.651	108	349825	18.947	ng/u1	97
19) Hexachloroethane	8.945	117	160452	19.250	ng/u1	96
22) Nitrobenzene	9.087	77	434819	18.800	ng/u1	99
23) Isophorone	9.610	82	738348	18.328	ng/u1	99
25) 2-Nitrophenol	9.792	139	169522	20.882	ng/u1	98
26) 2,4-Dimethylphenol	9.857	107	361163	18.563	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.104	93	470587	18.381	ng/u1	98
29) 2,4-Dichlorophenol	10.322	162	277237	19.120	ng/u1	97
30) Naphthalene	10.728	128	1106400	18.405	ng/u1	100
32) 4-Chloroaniline	10.845	127	407215	17.232	ng/u1	99
33) Hexachlorobutadiene	10.998	225	172671	19.296	ng/u1	99
34) Caprolactam	11.616	113	69736	14.453	ng/u1	93
35) 4-Chloro-3-methylphenol	11.969	107	305838	18.262	ng/u1	98
36) 2-Methylnaphthalene	12.339	142	678081	18.401	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
 Data File : BN029646.D
 Acq On : 10 Feb 2024 13:59
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020300

Quant Time: Feb 11 22:12:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 08 00:25:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.557	142	668213	18.290	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.704	216	297113	19.550	ng/ul	97
40) Hexachlorocyclopentadiene	12.680	237	172092	17.324	ng/ul	100
41) 2,4,6-Trichlorophenol	12.951	196	176183	20.113	ng/ul	99
42) 2,4,5-Trichlorophenol	13.022	196	193029	19.611	ng/ul	98
43) 1,1'-Biphenyl	13.357	154	824067	19.058	ng/ul	98
44) 2-Chloronaphthalene	13.392	162	633288	18.845	ng/ul	97
45) 2-Nitroaniline	13.610	65	180547	18.274	ng/ul	93
47) Dimethylphthalate	13.992	163	687118	17.585	ng/ul	99
48) 2,6-Dinitrotoluene	14.116	165	131822	18.853	ng/ul	97
50) Acenaphthylene	14.239	152	1024581	18.833	ng/ul	100
51) 3-Nitroaniline	14.439	138	134316	16.505	ng/ul	99
52) Acenaphthene	14.586	153	661984	18.420	ng/ul	94
53) 2,4-Dinitrophenol	14.645	184	48943	13.793	ng/ul	93
55) 4-Nitrophenol	14.739	109	102762	15.512	ng/ul	97
56) Dibenzofuran	14.922	168	862601	18.176	ng/ul	99
57) 2,4-Dinitrotoluene	14.898	165	177396	18.021	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.145	232	131916	18.884	ng/ul	98
59) Diethylphthalate	15.357	149	690357	17.200	ng/ul	97
61) Fluorene	15.569	166	679090	18.103	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.569	204	299595	17.856	ng/ul	100
63) 4-Nitroaniline	15.598	138	124725	16.982	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.651	198	83354	17.462	ng/ul	98
67) N-Nitrosodiphenylamine	15.786	169	533793	19.277	ng/ul	97
68) 4-Bromophenyl-phenylether	16.469	248	168053	19.980	ng/ul	93
69) Hexachlorobenzene	16.569	284	185572	19.574	ng/ul	93
70) Atrazine	16.745	200	156864	18.131	ng/ul	98
71) Pentachlorophenol	16.916	266	99673	18.104	ng/ul	96
72) Phenanthrene	17.316	178	967271	18.817	ng/ul	99
74) Anthracene	17.404	178	944866	18.315	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.316	216	284540	21.821	ng/uL	98
76) Pentachlorobenzene	14.833	250	240914	20.633	ng/uL	95
77) Carbazole	17.680	167	846450	17.964	ng/ul	99
78) Di-n-butylphthalate	18.257	149	1013892	17.894	ng/ul	100
80) Fluoranthene	19.333	202	978198	19.062	ng/ul	99
82) Pyrene	19.698	202	1044737	19.244	ng/ul	98
83) Butylbenzylphthalate	20.621	149	406418	18.369	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.398	252	258961	18.992	ng/ul	94
85) Benzo(a)anthracene	21.456	228	955951	18.486	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.404	149	621099	18.148	ng/ul	98
87) Chrysene	21.509	228	935274	19.298	ng/ul	96
89) Di-n-octyl phthalate	22.333	149	997246	16.016	ng/ul	100
90) Benzo(b)fluoranthene	23.121	252	955198	18.879	ng/ul	97
91) Benzo(k)fluoranthene	23.174	252	970731	18.745	ng/ul	95
93) Benzo(a)pyrene	23.739	252	916533	18.889	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.292	276	1116669	18.899	ng/ul	96
95) Dibenzo(a,h)anthracene	26.315	278	943160	19.553	ng/ul	98
96) Benzo(g,h,i)perylene	27.050	276	933980	18.899	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.692 ng/uL
RT: 3.363 min Scan# 41
Delta R.T. -0.000 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion: 88 Resp: 61179
Ion Ratio Lower Upper
88 100
43 25.8 22.1 33.1
58 63.8 56.5 84.7



#5
Pyridine
Concen: 18.481 ng/uL
RT: 3.752 min Scan# 113
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion: 79 Resp: 391540
Ion Ratio Lower Upper
79 100
52 67.4 55.4 83.0
51 31.7 26.9 40.3









#6
Benzaldehyde
Concen: 24.467 ng/ul
RT: 7.022 min Scan# 669
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion: 77 Resp: 259103
Ion Ratio Lower Upper
77 100
105 90.2 71.0 106.4
106 88.0 69.4 104.0



#8
Phenol
Concen: 18.699 ng/ul
RT: 7.069 min Scan# 677
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion: 94 Resp: 452746
Ion Ratio Lower Upper
94 100
65 29.4 23.9 35.9
66 38.6 32.2 48.2









#10
Bis(2-Chloroethyl)ether
Concen: 18.898 ng/ul
RT: 7.316 min Scan# 718
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion: 93 Resp: 384763
Ion Ratio Lower Upper
93 100
63 75.1 61.4 92.2
95 32.4 24.9 37.3



#12
2-Chlorophenol
Concen: 19.939 ng/ul
RT: 7.440 min Scan# 740
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion: 128 Resp: 358136
Ion Ratio Lower Upper
128 100
64 48.4 39.8 59.6
130 32.7 25.6 38.4





#13

2-Methylphenol

Concen: 18.681 ng/ul

RT: 8.322 min Scan# 890

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :





SSTD020300

Tgt Ion:108 Resp: 326399

Ion Ratio Lower Upper

108 100

107 90.7 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 18.601 ng/ul

RT: 8.404 min Scan# 904

Delta R.T. -0.012 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59




Tgt Ion: 45 Resp: 494613


Ion Ratio Lower Upper

45 100

77 16.2 12.7 19.1

79 11.4 9.4 14.2









#16
Acetophenone
Concen: 19.166 ng/ul
RT: 8.710 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:105 Resp: 537462
Ion Ratio Lower Upper
105 100
77 80.9 67.4 101.0
51 24.9 20.7 31.1



#17
N-Nitroso-di-n-propylamine
Concen: 19.578 ng/ul
RT: 8.692 min Scan# 953
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion: 70 Resp: 283439
Ion Ratio Lower Upper
70 100
42 53.7 43.0 64.4
101 8.4 7.4 11.2
130 17.7 15.1 22.7









#18
4-Methylphenol
Concen: 18.947 ng/ul
RT: 8.651 min Scan# 945
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:108 Resp: 349825
Ion Ratio Lower Upper
108 100
107 113.7 88.7 133.1



#19
Hexachloroethane
Concen: 19.250 ng/ul
RT: 8.945 min Scan# 996
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:117 Resp: 160452
Ion Ratio Lower Upper
117 100
201 76.6 64.4 96.6
199 49.9 38.3 57.5









#22
Nitrobenzene
Concen: 18.800 ng/ul
RT: 9.087 min Scan# 1020
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion: 77 Resp: 434819
Ion Ratio Lower Upper
77 100
123 42.8 33.6 50.4
65 12.7 10.6 16.0



#23
Isophorone
Concen: 18.328 ng/ul
RT: 9.610 min Scan# 1109
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion: 82 Resp: 738348
Ion Ratio Lower Upper
82 100
95 7.0 5.8 8.6
138 15.4 12.6 19.0









#25
2-Nitrophenol
Concen: 20.882 ng/ul
RT: 9.792 min Scan# 1140
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:139 Resp: 169522
Ion Ratio Lower Upper
139 100
65 58.0 46.6 70.0
109 31.1 27.0 40.4



#26
2,4-Dimethylphenol
Concen: 18.563 ng/ul
RT: 9.857 min Scan# 1151
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:107 Resp: 361163
Ion Ratio Lower Upper
107 100
121 50.3 40.0 60.0
122 83.4 67.6 101.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.381 ng/ul
 RT: 10.104 min Scan# 1193
 Delta R.T. -0.006 min
 Lab File: BN029646.D
 Acq: 10 Feb 2024 13:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020300


Tgt Ion: 93 Resp: 470587
 Ion Ratio Lower Upper
 93 100
 95 32.7 25.7 38.5
 123 11.7 8.2 12.2



#29
 2,4-Dichlorophenol
 Concen: 19.120 ng/ul
 RT: 10.322 min Scan# 1230
 Delta R.T. -0.012 min
 Lab File: BN029646.D
 Acq: 10 Feb 2024 13:59

Tgt Ion: 162 Resp: 277237
 Ion Ratio Lower Upper
 162 100
 164 68.9 53.3 79.9
 98 44.8 33.8 50.8





#30

Naphthalene

Concen: 18.405 ng/ul

RT: 10.728 min Scan# 1298

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :

SSTD020300





Tgt Ion:128 Resp: 1106400

Ion Ratio Lower Upper

128 100

129 11.2 9.0 13.6

127 13.1 10.4 15.6



#32

4-Chloroaniline

Concen: 17.232 ng/ul

RT: 10.845 min Scan# 1319

Delta R.T. -0.012 min

Lab File: BN029646.D




Acq: 10 Feb 2024 13:59


Tgt Ion:127 Resp: 407215

Ion Ratio Lower Upper

127 100

129 33.5 26.4 39.6





#33

Hexachlorobutadiene

Concen: 19.296 ng/ul

RT: 10.998 min Scan# 1345

Delta R.T. -0.012 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :

SSTD020300





Tgt Ion:225 Resp: 172671

Ion Ratio Lower Upper

225 100

223 63.6 51.6 77.4

227 63.8 52.2 78.2



#34

Caprolactam

Concen: 14.453 ng/ul

RT: 11.616 min Scan# 1450

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59




Tgt Ion:113 Resp: 69736


Ion Ratio Lower Upper

113 100

55 164.4 125.3 187.9

56 118.7 103.4 155.0









#35
 4-Chloro-3-methylphenol
 Concen: 18.262 ng/ul
 RT: 11.969 min Scan# 1510
 Delta R.T. -0.006 min
 Lab File: BN029646.D
 Acq: 10 Feb 2024 13:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020300


Tgt Ion:107 Resp: 305838
 Ion Ratio Lower Upper
 107 100
 144 24.6 18.8 28.2
 142 77.5 60.6 90.8



#36
 2-Methylnaphthalene
 Concen: 18.401 ng/ul
 RT: 12.339 min Scan# 1573
 Delta R.T. -0.012 min
 Lab File: BN029646.D
 Acq: 10 Feb 2024 13:59

Tgt Ion:142 Resp: 678081
 Ion Ratio Lower Upper
 142 100
 141 88.6 70.4 105.6









#37
1-Methylnaphthalene
Concen: 18.290 ng/ul
RT: 12.557 min Scan# 1610
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:142 Resp: 668213
Ion Ratio Lower Upper
142 100
141 90.2 73.8 110.8
116 4.4 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.550 ng/ul
RT: 12.704 min Scan# 1635
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:216 Resp: 297113
Ion Ratio Lower Upper
216 100
214 78.9 60.4 90.6
179 21.6 18.6 28.0
108 23.0 19.0 28.6









#40
Hexachlorocyclopentadiene
Concen: 17.324 ng/ul
RT: 12.680 min Scan# 1631
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:237 Resp: 172092
Ion Ratio Lower Upper
237 100
235 62.9 50.2 75.4
272 12.5 9.5 14.3



#41
2,4,6-Trichlorophenol
Concen: 20.113 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:196 Resp: 176183
Ion Ratio Lower Upper
196 100
198 94.4 76.8 115.2
200 30.6 24.6 37.0









#42
2,4,5-Trichlorophenol
Concen: 19.611 ng/ul
RT: 13.022 min Scan# 1689
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:196 Resp: 193029
Ion Ratio Lower Upper
196 100
198 96.6 75.3 112.9
200 30.5 24.7 37.1



#43
1,1'-Biphenyl
Concen: 19.058 ng/ul
RT: 13.357 min Scan# 1746
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:154 Resp: 824067
Ion Ratio Lower Upper
154 100
153 39.6 32.8 49.2
76 15.2 12.9 19.3









#44
2-Chloronaphthalene
Concen: 18.845 ng/ul
RT: 13.392 min Scan# 1752
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:162 Resp: 633288
Ion Ratio Lower Upper
162 100
164 31.0 25.5 38.3
127 41.8 35.7 53.5



#45
2-Nitroaniline
Concen: 18.274 ng/ul
RT: 13.610 min Scan# 1789
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion: 65 Resp: 180547
Ion Ratio Lower Upper
65 100
92 70.4 54.3 81.5
138 99.1 71.2 106.8









#47
Dimethylphthalate
Concen: 17.585 ng/ul
RT: 13.992 min Scan# 1854
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:163 Resp: 687118
Ion Ratio Lower Upper
163 100
194 3.9 2.7 4.1
164 10.0 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 18.853 ng/ul
RT: 14.116 min Scan# 1875
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:165 Resp: 131822
Ion Ratio Lower Upper
165 100
89 66.6 55.8 83.8
121 24.4 19.8 29.6









#50
Acenaphthylene
Concen: 18.833 ng/ul
RT: 14.239 min Scan# 1896
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:152 Resp: 1024581
Ion Ratio Lower Upper
152 100
151 19.1 15.4 23.2
153 13.2 10.5 15.7



#51
3-Nitroaniline
Concen: 16.505 ng/ul
RT: 14.439 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:138 Resp: 134316
Ion Ratio Lower Upper
138 100
108 12.6 9.4 14.2
92 127.5 102.7 154.1





#52

Acenaphthene

Concen: 18.420 ng/ul

RT: 14.586 min Scan# 1955

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :

SSTD020300





Tgt Ion:153 Resp: 661984

Ion Ratio Lower Upper

153 100

152 48.4 39.2 58.8

154 84.9 74.6 111.8



#53

2,4-Dinitrophenol

Concen: 13.793 ng/ul

RT: 14.645 min Scan# 1965

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59




Tgt Ion:184 Resp: 48943


Ion Ratio Lower Upper

184 100

63 83.7 71.2 106.8

154 65.9 58.8 88.2









#55
4-Nitrophenol
Concen: 15.512 ng/ul
RT: 14.739 min Scan# 1981
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:109 Resp: 102762
Ion Ratio Lower Upper
109 100
139 119.9 94.6 142.0
65 130.5 108.5 162.7



#56
Dibenzofuran
Concen: 18.176 ng/ul
RT: 14.922 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:168 Resp: 862601
Ion Ratio Lower Upper
168 100
139 40.4 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 18.021 ng/ul
RT: 14.898 min Scan# 2008
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:165 Resp: 177396
Ion Ratio Lower Upper
165 100
63 54.0 46.7 70.1
182 2.6 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.884 ng/ul
RT: 15.145 min Scan# 2050
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:232 Resp: 131916
Ion Ratio Lower Upper
232 100
131 59.8 46.9 70.3
130 3.4 2.6 4.0
166 37.9 29.4 44.2





#59

Diethylphthalate

Concen: 17.200 ng/ul

RT: 15.357 min Scan# 2086

Delta R.T. -0.012 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :

SSTD020300





Tgt Ion:149 Resp: 690357

Ion Ratio Lower Upper

149 100

177 21.1 15.8 23.6

150 12.9 9.6 14.4



#61

Fluorene

Concen: 18.103 ng/ul

RT: 15.569 min Scan# 2122

Delta R.T. -0.012 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59




Tgt Ion:166 Resp: 679090


Ion Ratio Lower Upper

166 100

165 99.5 77.4 116.0

167 13.7 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 17.856 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:204 Resp: 299595
Ion Ratio Lower Upper
204 100
206 34.8 27.5 41.3
141 75.7 60.5 90.7



#63
4-Nitroaniline
Concen: 16.982 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:138 Resp: 124725
Ion Ratio Lower Upper
138 100
92 59.9 47.4 71.2
108 90.6 70.9 106.3









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.462 ng/ul
 RT: 15.651 min Scan# 2136
 Delta R.T. -0.012 min
 Lab File: BN029646.D
 Acq: 10 Feb 2024 13:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020300


Tgt Ion:198 Resp: 83354
 Ion Ratio Lower Upper
 198 100
 182 5.4 3.9 5.9
 77 38.5 29.8 44.8



#67
 N-Nitrosodiphenylamine
 Concen: 19.277 ng/ul
 RT: 15.786 min Scan# 2159
 Delta R.T. -0.012 min
 Lab File: BN029646.D
 Acq: 10 Feb 2024 13:59

Tgt Ion:169 Resp: 533793
 Ion Ratio Lower Upper
 169 100
 168 65.4 54.5 81.7
 167 35.7 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 19.980 ng/ul
RT: 16.469 min Scan# 2175
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:248 Resp: 168053
Ion Ratio Lower Upper
248 100
250 103.3 85.6 128.4
141 95.6 84.6 126.8



#69
Hexachlorobenzene
Concen: 19.574 ng/ul
RT: 16.569 min Scan# 2292
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:284 Resp: 185572
Ion Ratio Lower Upper
284 100
142 42.4 39.2 58.8
249 30.3 25.7 38.5





#70

Atrazine

Concen: 18.131 ng/ul

RT: 16.745 min Scan# 21

Delta R.T. -0.012 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :

SSTD020300





Tgt Ion:200 Resp: 156864

Ion Ratio Lower Upper

200 100

173 29.5 22.0 33.0

215 44.1 36.1 54.1



#71

Pentachlorophenol

Concen: 18.104 ng/ul

RT: 16.916 min Scan# 2351

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59




Tgt Ion:266 Resp: 99673


Ion Ratio Lower Upper

266 100

264 60.7 45.8 68.6

268 67.2 51.6 77.4





#72

Phenanthrene

Concen: 18.817 ng/ul

RT: 17.316 min Scan# 2419

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :

SSTD020300





Tgt Ion:178 Resp: 967271

Ion Ratio Lower Upper

178 100

179 14.9 12.5 18.7

176 19.0 15.7 23.5



#74

Anthracene

Concen: 18.315 ng/ul

RT: 17.404 min Scan# 2434

Delta R.T. -0.012 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59




Tgt Ion:178 Resp: 944866


Ion Ratio Lower Upper

178 100

179 15.5 11.6 17.4

176 19.2 14.2 21.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.821 ng/uL

RT: 13.316 min Scan# 1739

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :

SSTD020300





Tgt Ion:216 Resp: 284540

Ion Ratio Lower Upper

216 100

214 76.7 60.1 90.1

218 46.2 38.5 57.7



#76

Pentachlorobenzene

Concen: 20.633 ng/uL

RT: 14.833 min Scan# 1997

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59




Tgt Ion:250 Resp: 240914


Ion Ratio Lower Upper

250 100

248 66.8 49.8 74.8

252 66.5 50.5 75.7









#77
 Carbazole
 Concen: 17.964 ng/ul
 RT: 17.680 min Scan# 2482
 Delta R.T. -0.012 min
 Lab File: BN029646.D
 Acq: 10 Feb 2024 13:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020300


Tgt Ion:167 Resp: 846450
 Ion Ratio Lower Upper
 167 100
 166 21.6 16.6 25.0
 139 14.0 11.4 17.0



#78
 Di-n-butylphthalate
 Concen: 17.894 ng/ul
 RT: 18.257 min Scan# 2579
 Delta R.T. -0.012 min
 Lab File: BN029646.D
 Acq: 10 Feb 2024 13:59

Tgt Ion:149 Resp: 1013892
 Ion Ratio Lower Upper
 149 100
 150 8.8 7.0 10.4
 104 6.0 4.7 7.1









#80
Fluoranthene
Concen: 19.062 ng/ul
RT: 19.333 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:202 Resp: 978198
Ion Ratio Lower Upper
202 100
101 14.9 11.1 16.7
100 11.1 8.9 13.3



#82
Pyrene
Concen: 19.244 ng/ul
RT: 19.698 min Scan# 2824
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:202 Resp: 1044737
Ion Ratio Lower Upper
202 100
101 16.7 14.2 21.4
100 13.4 11.4 17.2





#83

Butylbenzylphthalate

Concen: 18.369 ng/ul

RT: 20.621 min Scan# 2981

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :

SSTD020300





Tgt Ion:149 Resp: 406418

Ion Ratio Lower Upper

149 100

91 78.8 70.0 105.0

206 18.6 14.1 21.1



#84

3,3'-Dichlorobenzidine

Concen: 18.992 ng/ul

RT: 21.398 min Scan# 3113

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59




Tgt Ion:252 Resp: 258961


Ion Ratio Lower Upper

252 100

254 67.4 49.8 74.6

126 13.8 12.2 18.4





#85

Benzo(a)anthracene

Concen: 18.486 ng/ul

RT: 21.456 min Scan# 3114

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :

SSTD020300





Tgt Ion:228 Resp: 955951

Ion Ratio Lower Upper

228 100

229 18.8 15.0 22.6

226 27.8 21.0 31.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.148 ng/ul

RT: 21.404 min Scan# 3114

Delta R.T. -0.012 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59




Tgt Ion:149 Resp: 621099


Ion Ratio Lower Upper

149 100

167 24.7 20.6 30.8

279 3.0 2.7 4.1





#87

Chrysene

Concen: 19.298 ng/ul

RT: 21.509 min Scan# 3133

Delta R.T. -0.006 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :

SSTD020300





Tgt Ion:228 Resp: 935274

Ion Ratio Lower Upper

228 100

226 28.7 24.6 37.0

229 18.6 16.2 24.4



#89

Di-n-octyl phthalate

Concen: 16.016 ng/ul




RT: 22.333 min Scan# 3272


Delta R.T. -0.012 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Tgt Ion:149 Resp: 997246









#90
Benzo(b)fluoranthene
Concen: 18.879 ng/ul
RT: 23.121 min Scan# 3406
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:252 Resp: 955198
Ion Ratio Lower Upper
252 100
253 20.8 17.5 26.3
125 12.3 11.1 16.7



#91
Benzo(k)fluoranthene
Concen: 18.745 ng/ul
RT: 23.174 min Scan# 3415
Delta R.T. -0.006 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:252 Resp: 970731
Ion Ratio Lower Upper
252 100
253 20.5 18.2 27.4
125 12.2 11.2 16.8









#93
Benzo(a)pyrene
Concen: 18.889 ng/ul
RT: 23.739 min Scan# 3511
Delta R.T. -0.012 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59




Instrument :
BNA_N
ClientSampleId :
SSTD020300


Tgt Ion:252 Resp: 916533
Ion Ratio Lower Upper
252 100
253 21.1 18.8 28.2
125 13.9 12.2 18.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 18.899 ng/ul
RT: 26.292 min Scan# 3945
Delta R.T. -0.018 min
Lab File: BN029646.D
Acq: 10 Feb 2024 13:59

Tgt Ion:276 Resp: 1116669
Ion Ratio Lower Upper
276 100
138 28.2 21.7 32.5
277 25.5 17.8 26.8





#95

Dibenzo(a,h)anthracene

Concen: 19.553 ng/ul

RT: 26.315 min Scan# 3949

Delta R.T. -0.024 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Instrument :

BNA_N

ClientSampleId :

SSTD020300





Tgt Ion:278 Resp: 943160

Ion Ratio Lower Upper

278 100

139 21.0 17.8 26.6

279 23.4 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 18.899 ng/ul

RT: 27.050 min Scan# 4074

Delta R.T. -0.018 min

Lab File: BN029646.D

Acq: 10 Feb 2024 13:59

Tgt Ion:276 Resp: 933980

Ion Ratio Lower Upper

276 100

138 27.1 23.9 35.9

277 22.5 18.6 28.0

