

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
 Data File : BN029649.D
 Acq On : 10 Feb 2024 15:48
 Operator : MA/JU
 Sample : PB158929BL
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK929

Quant Time: Feb 11 22:58:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 08 00:25:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	262520	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.675	136	1072594	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.522	164	522908	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.269	188	919535	20.000	ng/u1	-0.01
79) Chrysene-d12	21.469	240	663961	20.000	ng/u1	-0.01
88) Perylene-d12	23.839	264	757656	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.328	96	43022	5.390	ng/uL	0.00
4) Pyridine-d5	3.734	84	577012	27.109	ng/u1	0.00
7) Phenol-d5	7.046	99	674979	27.488	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	465846	28.427	ng/u1	0.00
11) 2-Chlorophenol-d4	7.405	132	524020	29.450	ng/u1	-0.01
15) 4-Methylphenol-d8	8.587	113	510925	27.900	ng/u1	-0.01
21) Nitrobenzene-d5	9.046	128	249912	30.020	ng/u1	0.00
24) 2-Nitrophenol-d4	9.763	143	233954	32.269	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.293	165	405474	28.334	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.822	131	667482	27.340	ng/u1	-0.01
46) Dimethylphthalate-d6	13.940	166	1177637	30.249	ng/u1	-0.01
49) Acenaphthylene-d8	14.210	160	1376047	29.763	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.722	143	144933	19.615	ng/u1	-0.01
60) Fluorene-d10	15.516	176	953695	28.796	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.640	200	100193	22.362	ng/u1	0.00
73) Anthracene-d10	17.369	188	1291890	29.978	ng/u1	-0.01
81) Pyrene-d10	19.669	212	1277179	32.054	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.686	264	1197385	31.583	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
Data File : BN029649.D
Acq On : 10 Feb 2024 15:48
Operator : MA/JU
Sample : PB158929BL
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK929

Quant Time: Feb 11 22:58:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 08 00:25:29 2024
Response via : Initial Calibration

