




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
Data File : BN029641.D
Acq On : 10 Feb 2024 10:22
Operator : MA/JU
Sample : P1399-06
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AH9

Manual IntegrationsAPPROVED

Quant Time: Feb 11 22:10:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 08 00:25:29 2024
Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 02/12/2024
Supervised By :mohammad ahmed 02/13/2024



TIC: BN029641.D\data.ms

(54) 4-Nitrophenol-d4 (S)

14.728min (-0.006) 4.57 ng/u1

response 23872

Ion	Exp%	Act%
143.00	100.00	100.00
113.00	87.40	81.35
41.00	45.90	45.97
42.00	28.40	27.68




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
Data File : BN029641.D
Acq On : 10 Feb 2024 10:22
Operator : MA/JU
Sample : P1399-06
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AH9

Manual IntegrationsAPPROVED

Quant Time: Feb 11 22:10:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 08 00:25:29 2024
Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 02/12/2024
Supervised By :mohammad ahmed 02/13/2024



TIC: BN029641.D\data.ms

(54) 4-Nitrophenol-d4 (S)

14.728min (-0.006) 4.87 ng/u1 m

response 25466

Ion	Exp%	Act%
143.00	100.00	100.00
113.00	87.40	81.35
41.00	45.90	45.97
42.00	28.40	27.68

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
 Data File : BN029641.D
 Acq On : 10 Feb 2024 10:22
 Operator : MA/JU
 Sample : P1399-06
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AH9

Manual IntegrationsAPPROVED

Reviewed By :Jagrut Upadhyay 02/12/2024
 Supervised By :mohammad ahmed 02/13/2024

Quant Time: Feb 11 22:36:06 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 08 00:25:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	191526	20.000	ng/ul	0.00
20) Naphthalene-d8	10.675	136	750113	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.522	164	369783	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.269	188	669820	20.000	ng/ul	-0.01
79) Chrysene-d12	21.474	240	533706	20.000	ng/ul	0.00
88) Perylene-d12	23.845	264	624848	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.322	96	24174	4.151	ng/uL	0.00
4) Pyridine-d5	3.740	84	137828	8.876	ng/ul	0.00
7) Phenol-d5	7.046	99	144817	8.084	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	433037	36.219	ng/ul	0.00
11) 2-Chlorophenol-d4	7.404	132	390343	30.069	ng/ul	-0.01
15) 4-Methylphenol-d8	8.587	113	247309	18.511	ng/ul	-0.01
21) Nitrobenzene-d5	9.046	128	230594	39.608	ng/ul	0.00
24) 2-Nitrophenol-d4	9.763	143	201604	39.761	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.298	165	345894	34.561	ng/ul	0.00
31) 4-Chloroaniline-d4	10.822	131	440482	25.798	ng/ul	-0.01
46) Dimethylphthalate-d6	13.945	166	1107089	40.213	ng/ul	0.00
49) Acenaphthylene-d8	14.216	160	1283303	39.251	ng/ul	0.00
54) 4-Nitrophenol-d4	14.728	143	25466m	4.874	ng/ul	0.00
60) Fluorene-d10	15.516	176	935982	39.964	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.639	200	103441	31.694	ng/ul	0.00
73) Anthracene-d10	17.375	188	1356952	43.226	ng/ul	0.00
81) Pyrene-d10	19.669	212	1506294	47.030	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.692	264	1447879	46.307	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020924\
Data File : BN029641.D
Acq On : 10 Feb 2024 10:22
Operator : MA/JU
Sample : P1399-06
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AH9

Manual IntegrationsAPPROVED

Quant Time: Feb 11 22:36:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 08 00:25:29 2024
Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 02/12/2024
Supervised By :mohammad ahmed 02/13/2024

