

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021023\
 Data File : BN024000.D
 Acq On : 10 Feb 2023 10:35
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020253

Quant Time: Feb 10 21:57:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020623.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 10 21:55:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	164116	20.000	ng/u1	0.00
20) Naphthalene-d8	10.640	136	725771	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.487	164	470022	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.239	188	1063021	20.000	ng/u1	0.00
79) Chrysene-d12	21.433	240	987022	20.000	ng/u1	0.00
88) Perylene-d12	23.816	264	959788	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.223	96	32774	8.075	ng/uL	0.00
4) Pyridine-d5	3.646	84	220985	19.629	ng/u1	0.00
7) Phenol-d5	6.999	99	283964	20.240	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.163	67	188947	21.342	ng/u1	0.00
11) 2-Chlorophenol-d4	7.358	132	228824	20.970	ng/u1	0.00
15) 4-Methylphenol-d8	8.552	113	226428	20.659	ng/u1	0.00
21) Nitrobenzene-d5	9.005	128	112875	19.481	ng/u1	0.00
24) 2-Nitrophenol-d4	9.722	143	124633	19.157	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.263	165	239457	19.751	ng/u1	0.00
31) 4-Chloroaniline-d4	10.787	131	323934	19.770	ng/u1	0.00
46) Dimethylphthalate-d6	13.904	166	716981	19.736	ng/u1	0.00
49) Acenaphthylene-d8	14.181	160	811045	20.153	ng/u1	0.00
54) 4-Nitrophenol-d4	14.698	143	128412	20.549	ng/u1	0.00
60) Fluorene-d10	15.481	176	628382	20.544	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.610	200	138585	18.869	ng/u1	0.00
73) Anthracene-d10	17.339	188	978904	20.215	ng/u1	0.00
81) Pyrene-d10	19.633	212	1164672	19.589	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.668	264	965500	19.663	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.264	88	34281	7.986	ng/uL#	92
5) Pyridine	3.670	79	227415	20.093	ng/u1	96
6) Benzaldehyde	6.969	77	127887	21.039	ng/u1	98
8) Phenol	7.028	94	295427	20.861	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.258	93	239940	20.845	ng/u1	98
12) 2-Chlorophenol	7.393	128	232599	20.965	ng/u1	98
13) 2-Methylphenol	8.281	108	219906	20.728	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.363	45	204835	14.285	ng/u1	99
16) Acetophenone	8.663	105	372038	21.759	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.652	70	202879	22.094	ng/u1	97
18) 4-Methylphenol	8.616	108	245563	21.486	ng/u1	93
19) Hexachloroethane	8.910	117	100176	21.609	ng/u1	92
22) Nitrobenzene	9.046	77	300930	20.472	ng/u1	99
23) Isophorone	9.569	82	545200	20.155	ng/u1	98
25) 2-Nitrophenol	9.757	139	136542	20.062	ng/u1	96
26) 2,4-Dimethylphenol	9.816	107	274850	20.268	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.057	93	335593	19.973	ng/u1	98
29) 2,4-Dichlorophenol	10.287	162	238859	20.316	ng/u1	98
30) Naphthalene	10.693	128	774162	19.891	ng/u1	99
32) 4-Chloroaniline	10.810	127	329238	20.128	ng/u1	100
33) Hexachlorobutadiene	10.969	225	168504	19.986	ng/u1	100
34) Caprolactam	11.599	113	38145	11.409	ng/u1	89
35) 4-Chloro-3-methylphenol	11.940	107	255216	20.659	ng/u1	100
36) 2-Methylnaphthalene	12.304	142	530132	20.231	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021023\
 Data File : BN024000.D
 Acq On : 10 Feb 2023 10:35
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020253


Quant Time: Feb 10 21:57:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020623.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 10 21:55:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.528	142	540956	20.292	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	12.675	216	321117	19.501	ng/ul	95
40) Hexachlorocyclopentadiene	12.651	237	212291	17.792	ng/ul	99
41) 2,4,6-Trichlorophenol	12.916	196	205201	19.637	ng/ul	95
42) 2,4,5-Trichlorophenol	12.993	196	223289	19.732	ng/ul	95
43) 1,1'-Biphenyl	13.322	154	728514	19.467	ng/ul	99
44) 2-Chloronaphthalene	13.363	162	570751	19.619	ng/ul	98
45) 2-Nitroaniline	13.575	65	173849	20.937	ng/ul	99
47) Dimethylphthalate	13.951	163	724736	20.028	ng/ul	99
48) 2,6-Dinitrotoluene	14.075	165	144783	20.470	ng/ul	93
50) Acenaphthylene	14.210	152	884109	20.635	ng/ul	98
51) 3-Nitroaniline	14.404	138	145898	23.423	ng/ul	97
52) Acenaphthene	14.551	153	610712	20.405	ng/ul	98
53) 2,4-Dinitrophenol	14.610	184	88195	18.545	ng/ul	96
55) 4-Nitrophenol	14.716	109	120784	21.838	ng/ul	97
56) Dibenzofuran	14.887	168	875081	20.666	ng/ul	99
57) 2,4-Dinitrotoluene	14.863	165	210513	21.445	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.116	232	202367	20.560	ng/ul	99
59) Diethylphthalate	15.316	149	729921	20.659	ng/ul	99
61) Fluorene	15.540	166	711772	20.510	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.534	204	374931	20.535	ng/ul	99
63) 4-Nitroaniline	15.569	138	137604	25.783	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.622	198	137453	19.244	ng/ul	94
67) N-Nitrosodiphenylamine	15.751	169	596204	19.462	ng/ul	99
68) 4-Bromophenyl-phenylether	16.428	248	236174	18.961	ng/ul	96
69) Hexachlorobenzene	16.539	284	280336	19.698	ng/ul	95
70) Atrazine	16.710	200	227054	18.972	ng/ul	98
71) Pentachlorophenol	16.887	266	167323	18.509	ng/ul	95
72) Phenanthrene	17.281	178	1132553	20.015	ng/ul	98
74) Anthracene	17.375	178	1149608	20.290	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.281	216	314003	18.415	ng/ul	99
76) Pentachlorobenzene	14.804	250	318077	18.795	ng/ul	96
77) Carbazole	17.651	167	1004019	20.561	ng/ul	99
78) Di-n-butylphthalate	18.210	149	1199338	14.706	ng/ul	99
80) Fluoranthene	19.298	202	1379943	19.506	ng/ul	99
82) Pyrene	19.663	202	1448702	20.118	ng/ul	99
83) Butylbenzylphthalate	20.569	149	508485	18.024	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.357	252	400186	17.236	ng/ul	98
85) Benzo(a)anthracene	21.416	228	1378634	20.157	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.345	149	721285	17.229	ng/ul	97
87) Chrysene	21.469	228	1289866	20.379	ng/ul	99
89) Di-n-octyl phthalate	22.274	149	1150152	17.333	ng/ul	100
90) Benzo(b)fluoranthene	23.086	252	1314192	20.782	ng/ul	96
91) Benzo(k)fluoranthene	23.139	252	1231779	20.444	ng/ul	96
93) Benzo(a)pyrene	23.716	252	1106533	20.611	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.298	276	1326151	19.138	ng/ul	98
95) Dibenzo(a,h)anthracene	26.315	278	1104746	19.213	ng/ul	98
96) Benzo(g,h,i)perylene	27.068	276	1053167	18.793	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 10 21:57:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020623.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Feb 10 21:55:50 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.986 ng/uL
 RT: 3.264 min Scan# 30
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020253


Tgt Ion: 88 Resp: 34281
 Ion Ratio Lower Upper
 88 100
 43 29.4 33.4 50.2#
 58 66.9 52.9 79.3



#5
 Pyridine
 Concen: 20.093 ng/uL
 RT: 3.670 min Scan# 99
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Tgt Ion: 79 Resp: 227415
 Ion Ratio Lower Upper
 79 100
 52 68.4 52.1 78.1
 51 34.7 26.0 39.0









#6
Benzaldehyde
Concen: 21.039 ng/ul
RT: 6.969 min Scan# 660
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion: 77 Resp: 127887
Ion Ratio Lower Upper
77 100
105 92.4 71.3 106.9
106 88.0 71.1 106.7



#8
Phenol
Concen: 20.861 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion: 94 Resp: 295427
Ion Ratio Lower Upper
94 100
65 30.0 24.7 37.1
66 41.0 31.4 47.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.845 ng/ul
 RT: 7.258 min Scan# 709
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35




Instrument :
 BNA_N
 ClientSampleId :
 SST020253


Tgt Ion: 93 Resp: 239940
 Ion Ratio Lower Upper
 93 100
 63 76.6 60.2 90.2
 95 31.9 26.2 39.4



#12
 2-Chlorophenol
 Concen: 20.965 ng/ul
 RT: 7.393 min Scan# 732
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Tgt Ion: 128 Resp: 232599
 Ion Ratio Lower Upper
 128 100
 64 48.4 37.8 56.6
 130 31.4 25.6 38.4





#13

2-Methylphenol

Concen: 20.728 ng/ul

RT: 8.281 min Scan# 883

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35

Instrument :

BNA_N

ClientSampleId :





SSTD020253

Tgt Ion:108 Resp: 219906

Ion Ratio Lower Upper

108 100

107 90.4 69.3 103.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 14.285 ng/ul

RT: 8.363 min Scan# 897

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35




Tgt Ion: 45 Resp: 204835


Ion Ratio Lower Upper

45 100

77 16.6 12.8 19.2

79 12.5 9.9 14.9









#16
Acetophenone
Concen: 21.759 ng/ul
RT: 8.663 min Scan# 948
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:105 Resp: 372038
Ion Ratio Lower Upper
105 100
77 81.7 66.4 99.6
51 27.0 22.2 33.2



#17
N-Nitroso-di-n-propylamine
Concen: 22.094 ng/ul
RT: 8.652 min Scan# 946
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion: 70 Resp: 202879
Ion Ratio Lower Upper
70 100
42 59.1 49.6 74.4
101 8.8 7.4 11.2
130 20.5 16.4 24.6









#18
4-Methylphenol
Concen: 21.486 ng/ul
RT: 8.616 min Scan# 940
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:108 Resp: 245563
Ion Ratio Lower Upper
108 100
107 106.4 90.8 136.2



#19
Hexachloroethane
Concen: 21.609 ng/ul
RT: 8.910 min Scan# 990
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:117 Resp: 100176
Ion Ratio Lower Upper
117 100
201 109.3 92.7 139.1
199 71.9 50.4 75.6









#22
Nitrobenzene
Concen: 20.472 ng/ul
RT: 9.046 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion: 77 Resp: 300930
Ion Ratio Lower Upper
77 100
123 41.2 33.8 50.6
65 13.1 10.6 15.8



#23
Isophorone
Concen: 20.155 ng/ul
RT: 9.569 min Scan# 1102
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion: 82 Resp: 545200
Ion Ratio Lower Upper
82 100
95 7.3 6.0 9.0
138 16.0 13.6 20.4









#25
2-Nitrophenol
Concen: 20.062 ng/ul
RT: 9.757 min Scan# 1134
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:139 Resp: 136542
Ion Ratio Lower Upper
139 100
65 54.0 46.9 70.3
109 33.4 27.1 40.7



#26
2,4-Dimethylphenol
Concen: 20.268 ng/ul
RT: 9.816 min Scan# 1144
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:107 Resp: 274850
Ion Ratio Lower Upper
107 100
121 49.2 38.9 58.3
122 83.1 65.9 98.9









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.973 ng/ul
 RT: 10.057 min Scan# 1185
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020253


Tgt Ion: 93 Resp: 335593
 Ion Ratio Lower Upper
 93 100
 95 33.5 25.9 38.9
 123 11.3 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 20.316 ng/ul
 RT: 10.287 min Scan# 1224
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Tgt Ion: 162 Resp: 238859
 Ion Ratio Lower Upper
 162 100
 164 63.8 52.2 78.4
 98 35.2 29.4 44.0









#30
Naphthalene
Concen: 19.891 ng/ul
RT: 10.693 min Scan# 1293
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:	128	Resp:	774162
Ion	Ratio	Lower	Upper
128	100		
129	10.5	9.0	13.4
127	12.7	10.5	15.7



#32
4-Chloroaniline
Concen: 20.128 ng/ul
RT: 10.810 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:	127	Resp:	329238
Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.8	38.6









#33
Hexachlorobutadiene
Concen: 19.986 ng/ul
RT: 10.969 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:225 Resp: 168504
Ion Ratio Lower Upper
225 100
223 61.5 49.0 73.4
227 62.2 50.1 75.1



#34
Caprolactam
Concen: 11.409 ng/ul
RT: 11.599 min Scan# 1447
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:113 Resp: 38145
Ion Ratio Lower Upper
113 100
55 150.7 132.3 198.5
56 115.3 103.3 154.9









#35
4-Chloro-3-methylphenol
Concen: 20.659 ng/ul
RT: 11.940 min Scan# 1505
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:107 Resp: 255216
Ion Ratio Lower Upper
107 100
144 25.8 20.6 30.8
142 79.4 63.8 95.8



#36
2-Methylnaphthalene
Concen: 20.231 ng/ul
RT: 12.304 min Scan# 1567
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:142 Resp: 530132
Ion Ratio Lower Upper
142 100
141 90.5 70.6 106.0









#37
1-Methylnaphthalene
Concen: 20.292 ng/ul
RT: 12.528 min Scan# 1605
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:142 Resp: 540956
Ion Ratio Lower Upper
142 100
141 89.6 75.4 113.0
116 4.2 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.501 ng/ul
RT: 12.675 min Scan# 1630
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:216 Resp: 321117
Ion Ratio Lower Upper
216 100
214 81.1 60.5 90.7
179 20.7 16.0 24.0
108 17.2 13.8 20.8









#40
Hexachlorocyclopentadiene
Concen: 17.792 ng/ul
RT: 12.651 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:237 Resp: 212291
Ion Ratio Lower Upper
237 100
235 65.6 52.0 78.0
272 12.8 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 19.637 ng/ul
RT: 12.916 min Scan# 1671
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:196 Resp: 205201
Ion Ratio Lower Upper
196 100
198 96.0 81.3 121.9
200 30.3 26.2 39.2









#42
2,4,5-Trichlorophenol
Concen: 19.732 ng/ul
RT: 12.993 min Scan# 1684
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:196 Resp: 223289
Ion Ratio Lower Upper
196 100
198 100.6 76.0 114.0
200 30.7 24.0 36.0



#43
1,1'-Biphenyl
Concen: 19.467 ng/ul
RT: 13.322 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:154 Resp: 728514
Ion Ratio Lower Upper
154 100
153 39.3 31.9 47.9
76 14.1 10.6 16.0









#44
2-Chloronaphthalene
Concen: 19.619 ng/ul
RT: 13.363 min Scan# 1747
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:162 Resp: 570751
Ion Ratio Lower Upper
162 100
164 32.6 26.4 39.6
127 40.4 30.6 45.8



#45
2-Nitroaniline
Concen: 20.937 ng/ul
RT: 13.575 min Scan# 1783
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion: 65 Resp: 173849
Ion Ratio Lower Upper
65 100
92 67.0 52.0 78.0
138 98.2 79.0 118.6









#47
 Dimethylphthalate
 Concen: 20.028 ng/ul
 RT: 13.951 min Scan# 1847
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020253

Tgt Ion:163 Resp: 724736




Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.5	5.3
164	10.4	7.8	11.8




#48
 2,6-Dinitrotoluene
 Concen: 20.470 ng/ul
 RT: 14.075 min Scan# 1868
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Tgt Ion:165 Resp: 144783

Ion	Ratio	Lower	Upper
165	100		
89	66.6	48.1	72.1
121	21.6	16.8	25.2









#50
 Acenaphthylene
 Concen: 20.635 ng/ul
 RT: 14.210 min Scan# 1891
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020253


Tgt Ion:	152	Resp:	884109
Ion Ratio	Lower	Upper	
152	100		
151	20.3	15.3	22.9
153	12.9	10.6	16.0



#51
 3-Nitroaniline
 Concen: 23.423 ng/ul
 RT: 14.404 min Scan# 1924
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Tgt Ion:	138	Resp:	145898
Ion Ratio	Lower	Upper	
138	100		
108	10.2	8.2	12.4
92	119.8	92.6	138.8





#52

Acenaphthene

Concen: 20.405 ng/ul

RT: 14.551 min Scan# 1949

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35

Instrument :

BNA_N

ClientSampleId :

SSTD020253





Tgt Ion:153 Resp: 610712

Ion Ratio Lower Upper

153 100

152 47.9 37.4 56.2

154 87.0 71.5 107.3



#53

2,4-Dinitrophenol

Concen: 18.545 ng/ul

RT: 14.610 min Scan# 1959

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35




Tgt Ion:184 Resp: 88195


Ion Ratio Lower Upper

184 100

63 76.2 58.3 87.5

154 68.6 52.0 78.0









#55
4-Nitrophenol
Concen: 21.838 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:109 Resp: 120784
Ion Ratio Lower Upper
109 100
139 114.8 96.2 144.2
65 122.6 99.5 149.3



#56
Dibenzofuran
Concen: 20.666 ng/ul
RT: 14.887 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:168 Resp: 875081
Ion Ratio Lower Upper
168 100
139 38.4 30.2 45.2









#57
2,4-Dinitrotoluene
Concen: 21.445 ng/ul
RT: 14.863 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:165 Resp: 210513
Ion Ratio Lower Upper
165 100
63 48.2 37.8 56.8
182 2.9 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.560 ng/ul
RT: 15.116 min Scan# 2045
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:232 Resp: 202367
Ion Ratio Lower Upper
232 100
131 41.2 33.0 49.6
130 2.3 2.1 3.1
166 27.6 23.2 34.8









#59
 Diethylphthalate
 Concen: 20.659 ng/ul
 RT: 15.316 min Scan# 2079
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020253

Tgt Ion:149 Resp: 729921




Ion	Ratio	Lower	Upper
149	100		
177	22.5	17.8	26.6
150	12.0	10.0	15.0




#61
 Fluorene
 Concen: 20.510 ng/ul
 RT: 15.540 min Scan# 2117
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Tgt Ion:166 Resp: 711772

Ion	Ratio	Lower	Upper
166	100		
165	96.1	78.4	117.6
167	12.5	10.4	15.6









#62
4-Chlorophenyl-phenylether
Concen: 20.535 ng/ul
RT: 15.534 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:204 Resp: 374931
Ion Ratio Lower Upper
204 100
206 32.6 26.9 40.3
141 53.9 43.2 64.8



#63
4-Nitroaniline
Concen: 25.783 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:138 Resp: 137604
Ion Ratio Lower Upper
138 100
92 57.1 42.5 63.7
108 99.0 70.6 105.8









#66
4,6-Dinitro-2-methylphenol
Concen: 19.244 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:198 Resp: 137453
Ion Ratio Lower Upper
198 100
182 4.3 3.4 5.0
77 30.3 21.4 32.0



#67
N-Nitrosodiphenylamine
Concen: 19.462 ng/ul
RT: 15.751 min Scan# 2153
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:169 Resp: 596204
Ion Ratio Lower Upper
169 100
168 67.3 54.2 81.2
167 37.4 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 18.961 ng/ul
RT: 16.428 min Scan# 2268
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:248 Resp: 236174
Ion Ratio Lower Upper
248 100
250 97.3 82.0 123.0
141 71.1 58.8 88.2



#69
Hexachlorobenzene
Concen: 19.698 ng/ul
RT: 16.539 min Scan# 2287
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:284 Resp: 280336
Ion Ratio Lower Upper
284 100
142 36.1 24.7 37.1
249 31.4 25.0 37.6









#70
 Atrazine
 Concen: 18.972 ng/ul
 RT: 16.710 min Scan# 2316
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020253

Tgt Ion: 200 Resp: 227054




Ion	Ratio	Lower	Upper
200	100		
173	25.2	20.3	30.5
215	49.4	41.0	61.6




#71
 Pentachlorophenol
 Concen: 18.509 ng/ul
 RT: 16.887 min Scan# 2346
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Tgt Ion: 266 Resp: 167323

Ion	Ratio	Lower	Upper
266	100		
264	61.8	47.4	71.0
268	66.0	48.8	73.2





#72

Phenanthrene

Concen: 20.015 ng/u1

RT: 17.281 min Scan# 2413

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35

Instrument :





BNA_N

ClientSampleId :

SSTD020253

Tgt Ion:178 Resp: 1132553

Ion	Ratio	Lower	Upper
178	100		
179	15.7	13.6	20.4
176	20.3	15.8	23.6



#74

Anthracene

Concen: 20.290 ng/u1

RT: 17.375 min Scan# 2429




Delta R.T. 0.000 min


Lab File: BN024000.D

Acq: 10 Feb 2023 10:35

Tgt Ion:178 Resp: 1149608

Ion	Ratio	Lower	Upper
178	100		
179	15.5	12.1	18.1
176	18.0	14.4	21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.415 ng/uL
RT: 13.281 min Scan# 1733
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:216 Resp: 314003
Ion Ratio Lower Upper
216 100
214 77.3 61.2 91.8
218 49.0 39.8 59.6



#76
Pentachlorobenzene
Concen: 18.795 ng/uL
RT: 14.804 min Scan# 1992
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:250 Resp: 318077
Ion Ratio Lower Upper
250 100
248 68.7 50.7 76.1
252 64.9 51.5 77.3





#77

Carbazole

Concen: 20.561 ng/ul

RT: 17.651 min Scan# 2476

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35

Instrument :





BNA_N

ClientSampleId :

SSTD020253

Tgt Ion:167 Resp: 1004019

Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.1	25.7
139	12.9	10.3	15.5



#78

Di-n-butylphthalate

Concen: 14.706 ng/ul

RT: 18.210 min Scan# 2571




Delta R.T. 0.000 min


Lab File: BN024000.D

Acq: 10 Feb 2023 10:35

Tgt Ion:149 Resp: 1199338

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.0	10.6
104	6.1	4.2	6.2









#80
Fluoranthene
Concen: 19.506 ng/u1
RT: 19.298 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:202 Resp: 1379943
Ion Ratio Lower Upper
202 100
101 9.6 7.8 11.6
100 7.7 5.8 8.8



#82
Pyrene
Concen: 20.118 ng/u1
RT: 19.663 min Scan# 2818
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:202 Resp: 1448702
Ion Ratio Lower Upper
202 100
101 11.3 9.3 13.9
100 9.4 7.6 11.4





#83

Butylbenzylphthalate

Concen: 18.024 ng/ul

RT: 20.569 min Scan# 2972

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35

Instrument :

BNA_N

ClientSampleId :

SSTD020253





Tgt Ion:149 Resp: 508485

Ion Ratio Lower Upper

149 100

91 76.7 59.5 89.3

206 22.1 18.8 28.2



#84

3,3'-Dichlorobenzidine

Concen: 17.236 ng/ul

RT: 21.357 min Scan# 3106

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35




Tgt Ion:252 Resp: 400186


Ion Ratio Lower Upper

252 100

254 58.4 48.4 72.6

126 9.2 7.4 11.0





#85

Benzo(a)anthracene

Concen: 20.157 ng/ul

RT: 21.416 min Scan# 3116

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35

Instrument :

BNA_N

ClientSampleId :

SSTD020253





Tgt Ion:228 Resp: 1378634

Ion Ratio Lower Upper

228 100

229 19.4 15.8 23.6

226 27.3 21.3 31.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.229 ng/ul

RT: 21.345 min Scan# 3104

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35




Tgt Ion:149 Resp: 721285


Ion Ratio Lower Upper

149 100

167 26.7 22.7 34.1

279 4.3 3.9 5.9









#87
Chrysene
Concen: 20.379 ng/ul
RT: 21.469 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35




Instrument :
BNA_N
ClientSampleId :
SSTD020253


Tgt Ion:228 Resp: 1289866
Ion Ratio Lower Upper
228 100
226 31.3 24.5 36.7
229 19.6 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 17.333 ng/ul
RT: 22.274 min Scan# 3262
Delta R.T. 0.000 min
Lab File: BN024000.D
Acq: 10 Feb 2023 10:35

Tgt Ion:149 Resp: 1150152





#90

Benzo(b)fluoranthene

Concen: 20.782 ng/ul

RT: 23.086 min Scan# 3400

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35

Instrument :

BNA_N

ClientSampleId :

SSTD020253





Tgt Ion:252 Resp: 1314192

Ion Ratio Lower Upper

252 100

253 22.6 16.4 24.6

125 9.0 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 20.444 ng/ul

RT: 23.139 min Scan# 3409

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35




Tgt Ion:252 Resp: 1231779


Ion Ratio Lower Upper

252 100

253 21.3 19.3 28.9

125 9.1 7.2 10.8





#93

Benzo(a)pyrene

Concen: 20.611 ng/ul

RT: 23.716 min Scan# 3507

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35

Instrument :

BNA_N

ClientSampleId :

SSTD020253





Tgt Ion:252 Resp: 1106533

Ion Ratio Lower Upper

252 100

253 20.7 17.6 26.4

125 10.5 7.1 10.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.138 ng/ul

RT: 26.298 min Scan# 3946

Delta R.T. 0.000 min

Lab File: BN024000.D

Acq: 10 Feb 2023 10:35




Tgt Ion:276 Resp: 1326151


Ion Ratio Lower Upper

276 100

138 21.9 16.4 24.6

277 26.0 20.6 30.8









#95
 Dibenzo(a,h)anthracene
 Concen: 19.213 ng/ul
 RT: 26.315 min Scan# 3949
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020253

Tgt Ion:278 Resp: 1104746

Ion	Ratio	Lower	Upper
278	100		
139	13.5	11.6	17.4
279	24.3	19.9	29.9



#96
 Benzo(g,h,i)perylene
 Concen: 18.793 ng/ul
 RT: 27.068 min Scan# 4077
 Delta R.T. 0.000 min
 Lab File: BN024000.D
 Acq: 10 Feb 2023 10:35

Tgt Ion:276 Resp: 1053167

Ion	Ratio	Lower	Upper
276	100		
138	18.4	15.4	23.2
277	24.6	17.8	26.8

