



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021023\
 Data File : BN023999.D
 Acq On : 10 Feb 2023 09:54
 Operator : CG/JU
 Sample : DFTPP276
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP276

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020623.M
 Title : SVOA CALIBRATION
 Last Update : Fri Feb 10 21:55:50 2023



AutoFind: Scans 2479, 2480, 2481; Background Corrected with Scan 2474

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.7	104915	PASS
68	69	0.00	2	1.9	2099	PASS
69	198	0.00	100	44.8	112688	PASS
70	69	0.00	2	0.6	718	PASS
127	198	10	80	47.5	119491	PASS
197	198	0.00	2	0.6	1574	PASS
198	198	100	100	100.0	251584	PASS
199	198	5	9	7.0	17653	PASS
275	198	10	60	27.9	70264	PASS
365	198	1	100	3.9	9909	PASS
441	198	0.01	100	12.0	30126	PASS
442	198	50	100	71.1	178997	PASS
443	442	15	24	19.8	35525	PASS

m/z	Abundance
38.90	316.0
40.00	742.0
40.90	488.0
43.00	256.0
44.00	457.0
44.90	193.0
49.00	193.0
51.00	237.0
55.00	273.0
56.90	189.0
67.00	203.0
68.90	173.0
73.10	1131.0
76.90	353.0
77.90	222.0
78.90	247.0
83.00	174.0
83.90	223.0
85.90	180.0
91.00	307.0
95.90	195.0
96.80	235.0
105.10	162.0
115.00	228.0
132.90	253.0
147.00	271.0
163.00	159.0
191.00	274.0
192.90	388.0
207.00	2523.0
208.00	471.0
209.00	267.0
220.90	194.0
267.00	161.0
281.00	557.0
282.10	169.0
355.00	152.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

m/z	Abundance
35.90	163.0
37.00	664.0
38.00	1393.0
39.00	7203.0
40.00	837.0
41.10	713.0
43.00	308.0
44.00	383.0
44.90	189.0
49.00	756.0
50.00	22320.0
51.00	84600.0
52.00	4470.0
53.00	312.0
55.00	955.0
55.90	2694.0
57.00	5616.0
58.00	288.0
58.90	170.0
60.90	1246.0
62.00	1270.0
63.00	3244.0
64.00	662.0
65.00	1635.0
66.00	154.0
67.00	263.0
68.00	1903.0
69.00	90152.0
69.90	563.0
71.00	270.0
72.90	1511.0
74.00	10143.0
75.00	15352.0
76.00	5235.0
77.00	106240.0
78.00	7073.0
79.00	6688.0
80.00	5469.0
81.00	7215.0
81.90	1829.0
82.90	1684.0
84.00	391.0
85.00	1205.0
86.00	2574.0
86.90	866.0
87.90	339.0
89.00	204.0
91.00	1507.0
91.90	1910.0
93.00	12520.0
93.90	894.0
95.00	263.0
96.10	671.0
96.90	314.0
98.00	11061.0
99.00	8992.0
100.00	792.0
101.00	3958.0
101.90	277.0
102.90	1226.0
104.00	2342.0
104.90	2428.0
105.90	719.0
107.00	29424.0
108.00	4513.0
109.00	1026.0
110.00	46792.0
111.00	8899.0
112.00	1020.0
112.90	402.0
115.10	222.0
116.00	1561.0
117.00	33360.0
118.00	2608.0
119.00	358.0
119.90	490.0
120.90	297.0
122.00	1916.0
123.00	3064.0
123.90	1400.0
125.00	1313.0
127.00	89648.0
128.00	7205.0
129.00	36968.0
130.00	3642.0
131.00	794.0
131.90	307.0
133.00	382.0
133.90	908.0
135.00	2910.0
136.00	1205.0
137.00	2096.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

137.90	377.0
138.90	284.0
140.00	464.0
141.00	4981.0
141.90	1551.0
143.00	1119.0
143.90	320.0
144.90	416.0
146.00	1064.0
147.00	3246.0
148.00	7298.0
149.00	1882.0
150.00	329.0
151.00	736.0
152.00	429.0
153.00	1798.0
154.00	1222.0
155.00	2689.0
156.00	4320.0
157.00	806.0
158.00	948.0
159.00	675.0
160.00	1532.0
161.00	2589.0
161.90	478.0
162.90	387.0
163.90	329.0
165.00	2099.0
166.00	1457.0
167.00	12573.0
168.00	10155.0
169.00	993.0
170.10	272.0
170.90	539.0
172.00	800.0
173.00	1251.0
174.00	2195.0
175.00	3875.0
176.00	1046.0
176.90	1791.0
177.90	649.0
178.90	7466.0
180.00	6505.0
181.00	2585.0
182.00	451.0
182.90	342.0
184.00	739.0
185.00	3728.0
186.00	25664.0
187.00	8169.0
188.00	1012.0
188.90	1980.0
189.90	305.0
191.00	1051.0
192.00	2554.0
193.00	3347.0
193.90	602.0
195.10	241.0
196.00	4530.0
197.00	1114.0
197.90	165312.0
198.90	12638.0
199.90	958.0
201.40	1071.0
202.90	1396.0
204.00	6785.0
205.00	12284.0
206.00	48288.0
207.00	9103.0
208.00	2121.0
209.00	1127.0
210.00	844.0
211.00	2271.0
211.90	413.0
212.90	152.0
215.00	731.0
216.00	1345.0
216.90	13839.0
218.00	1691.0
219.10	266.0
221.00	6999.0
221.90	1923.0
223.00	3082.0
224.00	27464.0
225.10	7524.0
226.00	959.0
227.00	12598.0
227.90	1651.0
229.00	2026.0
230.00	408.0
230.90	870.0
232.00	203.0
233.10	277.0
233.90	682.0
235.00	861.0
236.00	750.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

236.90	1817.0
237.90	209.0
239.00	366.0
239.90	583.0
241.10	543.0
242.00	1426.0
243.00	1688.0
244.00	21232.0
245.00	3066.0
246.00	4281.0
247.00	784.0
248.00	212.0
248.90	728.0
249.90	208.0
250.90	247.0
252.00	236.0
253.00	710.0
255.00	109240.0
256.00	14598.0
257.00	1868.0
258.00	5594.0
259.00	1229.0
260.00	228.0
260.80	196.0
264.00	240.0
264.90	2761.0
265.90	444.0
271.00	394.0
271.90	429.0
273.00	3302.0
274.00	8282.0
275.00	45312.0
276.00	6295.0
276.90	4096.0
277.90	727.0
279.00	238.0
281.00	706.0
281.90	191.0
283.00	552.0
283.90	262.0
285.00	782.0
290.90	158.0
292.00	192.0
293.00	914.0
293.90	451.0
295.00	309.0
296.00	12625.0
297.00	2178.0
297.90	170.0
300.90	178.0
302.10	208.0
303.00	1202.0
304.00	312.0
308.80	219.0
309.90	232.0
313.00	177.0
314.00	621.0
314.90	1421.0
315.90	919.0
321.00	481.0
322.00	274.0
323.00	3783.0
324.00	670.0
324.80	205.0
326.90	901.0
327.90	498.0
332.00	458.0
333.00	395.0
333.90	2720.0
334.90	707.0
340.90	457.0
345.90	843.0
346.90	236.0
351.90	1216.0
352.90	689.0
354.00	1083.0
355.00	231.0
365.00	5662.0
366.00	865.0
371.00	219.0
371.90	1621.0
373.00	536.0
382.90	488.0
389.90	234.0
391.00	172.0
391.80	202.0
400.80	199.0
402.00	755.0
402.90	793.0
403.90	560.0
404.80	221.0
420.90	679.0
422.10	740.0
423.00	6076.0
424.00	2497.0
425.00	329.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

441.00	15475.0
442.00	84064.0
443.00	17952.0
444.00	1700.0
445.00	168.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

m/z	Abundance
35.00	167.0
37.00	570.0
38.00	1731.0
39.00	9119.0
39.00	928.0
41.00	742.0
42.00	192.0
43.00	425.0
44.00	561.0
44.00	255.0
46.00	193.0
49.00	1128.0
50.00	33264.0
51.00	124760.0
52.00	6286.0
53.00	325.0
54.00	910.0
56.00	3355.0
57.00	7869.0
58.00	395.0
59.00	178.0
61.00	1662.0
62.00	1602.0
63.00	5525.0
64.00	820.0
65.00	2298.0
66.00	259.0
68.00	2567.0
69.00	131648.0
69.00	789.0
71.00	310.0
73.00	2241.0
74.00	15663.0
75.00	22760.0
76.00	8156.0
77.00	154560.0
78.00	11827.0
79.00	10919.0
80.00	8196.0
81.00	10827.0
82.00	2940.0
83.00	2560.0
84.00	421.0
85.00	2140.0
86.00	3471.0
87.00	1349.0
88.00	515.0
89.00	324.0
89.00	186.0
91.00	2845.0
92.00	2729.0
93.00	19880.0
94.00	1056.0
95.00	457.0
96.00	983.0
97.10	489.0
98.00	15541.0
99.00	12761.0
100.00	1007.0
101.00	5575.0
102.00	324.0
102.00	1571.0
104.00	3922.0
105.00	3799.0
106.00	1089.0
107.00	45856.0
108.00	6966.0
109.00	1126.0
110.00	72704.0
111.00	12292.0
112.00	1463.0
113.00	571.0
113.00	156.0
114.00	273.0
116.00	2769.0
117.00	53232.0
118.00	4092.0
118.00	412.0
120.00	609.0
120.00	311.0
122.00	3025.0
123.00	4505.0
124.00	2326.0
125.00	2127.0
127.00	140608.0
128.00	11364.0
129.00	57928.0
130.00	4768.0
130.00	1093.0
131.00	469.0
132.00	453.0
134.00	1713.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

135.00	4493.0
135.90	2107.0
137.00	3273.0
138.00	609.0
138.90	476.0
140.00	789.0
141.00	7941.0
142.00	2411.0
143.00	2228.0
143.90	567.0
144.90	528.0
145.90	1828.0
147.00	4698.0
148.00	11391.0
149.00	2784.0
150.00	798.0
151.10	995.0
151.70	606.0
153.00	2908.0
154.00	2057.0
155.00	4835.0
156.00	6644.0
157.10	1300.0
158.00	1397.0
158.90	1002.0
159.90	2515.0
161.00	3570.0
161.90	1239.0
162.90	463.0
164.00	361.0
165.00	2893.0
166.00	2390.0
167.00	21168.0
168.00	14097.0
169.00	1993.0
170.00	722.0
170.90	766.0
171.90	1221.0
172.90	2170.0
174.00	3580.0
175.00	7173.0
176.00	1856.0
177.00	2885.0
177.90	962.0
179.00	12050.0
180.00	9489.0
181.00	4181.0
181.90	498.0
182.20	481.0
182.80	305.0
184.00	1213.0
185.00	6308.0
186.00	44648.0
187.00	12524.0
188.00	1760.0
188.90	2949.0
189.90	619.0
191.00	1636.0
191.90	4287.0
193.00	5362.0
193.90	1079.0
195.00	609.0
196.00	8276.0
196.90	1629.0
198.00	314304.0
199.00	19968.0
200.00	1802.0
201.40	1467.0
203.00	2402.0
204.00	12524.0
205.00	20856.0
206.00	91040.0
207.00	14487.0
208.00	3567.0
209.00	1379.0
209.90	1452.0
211.00	3626.0
212.80	293.0
214.10	196.0
214.90	1091.0
216.00	2302.0
217.00	23632.0
218.00	3195.0
219.00	293.0
221.00	12425.0
221.90	3170.0
223.00	5476.0
224.00	49280.0
225.00	13758.0
226.00	1747.0
227.00	20744.0
228.00	3539.0
229.00	4530.0
230.00	684.0
231.00	1659.0
232.00	230.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

233.00	470.0
233.90	1145.0
235.00	1518.0
235.90	1214.0
237.00	2550.0
238.00	351.0
239.00	1039.0
240.00	914.0
241.00	1121.0
242.00	2658.0
243.00	2625.0
244.00	38976.0
245.00	4461.0
246.00	8423.0
247.00	1436.0
248.00	432.0
248.90	1341.0
250.00	356.0
251.00	416.0
252.00	272.0
253.00	801.0
254.10	1547.0
255.00	179328.0
256.00	27728.0
257.00	2450.0
258.00	11888.0
258.90	2153.0
260.00	422.0
260.90	438.0
264.10	270.0
264.90	5064.0
266.00	651.0
270.10	196.0
270.90	416.0
272.10	602.0
273.00	6006.0
274.00	16147.0
275.00	81600.0
276.00	10381.0
277.00	7497.0
278.00	1600.0
278.80	194.0
281.00	601.0
282.00	274.0
283.00	866.0
284.00	804.0
285.00	1313.0
286.00	181.0
288.90	267.0
290.00	365.0
290.90	200.0
291.90	465.0
293.00	1662.0
294.00	653.0
295.10	437.0
296.00	24256.0
296.90	3043.0
297.90	225.0
301.00	373.0
301.90	485.0
303.10	2735.0
304.00	804.0
305.00	153.0
307.90	409.0
308.90	267.0
309.90	318.0
313.00	190.0
314.00	1357.0
315.00	2962.0
316.00	1728.0
317.00	222.0
321.00	694.0
322.00	458.0
323.00	7314.0
324.00	1031.0
325.90	229.0
326.90	1401.0
327.90	695.0
332.00	517.0
333.00	874.0
334.00	4903.0
335.00	1226.0
335.90	258.0
341.00	935.0
341.90	217.0
345.90	1743.0
347.00	445.0
352.00	2393.0
353.10	1707.0
354.00	2104.0
355.00	683.0
364.00	155.0
364.90	9988.0
365.90	1682.0
366.80	155.0
369.90	262.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

371.00	471.0
372.00	4445.0
372.90	896.0
376.90	172.0
382.90	1285.0
384.00	294.0
384.80	174.0
389.90	605.0
390.80	198.0
392.10	228.0
400.90	282.0
402.00	1487.0
402.90	2614.0
404.00	920.0
404.90	258.0
420.90	1945.0
421.90	1753.0
423.00	10739.0
424.00	4668.0
424.90	533.0
441.00	31920.0
442.00	176896.0
443.00	39672.0
443.90	3279.0
444.90	215.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

m/z	Abundance
37.00	540.0
38.00	1760.0
39.00	7958.0
40.00	732.0
41.00	774.0
42.10	199.0
43.00	524.0
43.90	539.0
44.80	271.0
49.00	930.0
50.00	28824.0
51.00	106096.0
52.00	5650.0
53.00	309.0
54.90	952.0
56.00	2919.0
57.00	7189.0
57.90	266.0
59.00	198.0
61.00	1540.0
62.00	1675.0
63.00	4198.0
64.00	771.0
65.00	1917.0
65.90	242.0
66.90	310.0
68.00	1828.0
68.90	116784.0
70.00	802.0
71.00	289.0
73.00	1735.0
74.00	12998.0
75.00	19488.0
76.00	7148.0
77.00	137984.0
78.00	9981.0
79.00	9235.0
80.00	6656.0
81.00	9591.0
82.00	2467.0
83.00	2250.0
84.10	302.0
85.00	1768.0
86.00	2893.0
86.90	1300.0
87.90	562.0
88.90	255.0
91.00	2514.0
92.00	2407.0
93.00	18032.0
94.00	1202.0
95.00	366.0
96.00	797.0
97.00	514.0
98.00	13673.0
99.00	11121.0
99.90	934.0
101.00	5495.0
101.90	399.0
102.90	1646.0
104.00	3186.0
105.00	3569.0
106.00	1095.0
107.00	39560.0
108.00	6284.0
109.10	1179.0
110.00	66032.0
111.00	10944.0
112.00	1534.0
113.00	470.0
115.00	434.0
116.00	2546.0
116.90	48096.0
117.90	3254.0
118.90	454.0
120.00	556.0
121.00	313.0
122.00	2814.0
123.00	4367.0
124.00	2125.0
124.90	1732.0
126.10	737.0
127.00	128216.0
128.00	9960.0
129.00	54184.0
130.00	4767.0
130.90	926.0
132.10	468.0
132.90	587.0
134.00	1751.0
135.00	4604.0
136.00	1665.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

137.00	3225.0
137.80	593.0
138.80	279.0
139.90	669.0
141.00	7779.0
142.00	2347.0
142.90	1767.0
144.00	504.0
145.00	378.0
146.00	1128.0
147.00	4273.0
148.00	11064.0
149.00	2485.0
149.90	559.0
151.10	909.0
151.90	422.0
152.90	2438.0
153.90	1878.0
155.00	4392.0
156.00	6243.0
157.10	1447.0
157.90	1485.0
159.00	1196.0
160.00	2683.0
161.00	3504.0
161.90	1086.0
162.90	321.0
164.00	402.0
165.00	3022.0
166.00	2192.0
167.00	18168.0
168.00	11713.0
169.00	1667.0
170.10	463.0
170.90	710.0
172.00	1664.0
173.00	1834.0
174.00	3420.0
175.00	6705.0
176.10	1308.0
176.90	2940.0
178.00	964.0
178.90	12040.0
180.00	8494.0
181.00	3787.0
182.00	849.0
183.00	315.0
184.00	1012.0
185.00	6110.0
186.00	43816.0
187.00	12273.0
188.00	1524.0
189.00	3074.0
189.90	441.0
190.90	1433.0
192.00	4832.0
193.00	5298.0
194.00	931.0
195.00	406.0
196.00	8660.0
197.00	1980.0
198.00	275136.0
199.00	20352.0
200.00	2027.0
201.40	1501.0
203.00	2507.0
204.00	12437.0
205.00	19768.0
206.00	87792.0
207.00	11798.0
208.00	3565.0
209.00	1582.0
210.10	1611.0
211.00	3610.0
211.80	624.0
212.90	359.0
214.10	175.0
215.00	1059.0
215.90	2204.0
217.00	23304.0
218.00	3566.0
219.00	359.0
221.00	13294.0
221.80	3324.0
223.00	5435.0
224.00	50568.0
225.00	13303.0
226.00	1513.0
227.00	21408.0
227.90	3157.0
229.00	4642.0
229.90	610.0
231.00	2074.0
232.00	411.0
233.00	285.0
234.00	1627.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

235.00	1834.0
235.90	1101.0
236.90	2438.0
237.90	366.0
239.00	1013.0
239.90	713.0
240.90	1369.0
242.00	3070.0
242.90	2878.0
244.00	39264.0
245.00	4951.0
246.00	8707.0
246.90	1441.0
247.90	280.0
249.00	1574.0
249.90	387.0
251.00	453.0
251.80	414.0
253.00	956.0
254.10	1677.0
255.00	202432.0
256.00	29392.0
257.00	2504.0
258.00	12508.0
259.00	2076.0
259.90	311.0
260.90	418.0
264.00	268.0
265.00	4960.0
265.90	746.0
267.10	183.0
269.90	283.0
271.00	713.0
272.10	621.0
273.00	6042.0
274.00	15766.0
275.00	83880.0
276.00	12421.0
276.90	8210.0
278.00	1968.0
278.80	426.0
281.00	695.0
281.90	447.0
283.00	1052.0
284.00	560.0
285.00	1453.0
286.00	251.0
288.90	347.0
289.90	301.0
290.90	213.0
291.90	386.0
293.00	1677.0
293.90	641.0
296.00	26904.0
297.00	3929.0
297.90	196.0
300.90	419.0
302.00	677.0
303.00	3118.0
304.00	911.0
304.90	239.0
308.10	261.0
308.90	245.0
309.90	321.0
313.00	240.0
314.00	1236.0
314.90	2851.0
315.90	1711.0
316.80	257.0
320.90	921.0
322.00	369.0
323.00	7860.0
324.00	1431.0
324.90	191.0
327.00	1837.0
328.00	928.0
331.90	689.0
333.00	951.0
334.00	5456.0
334.90	1307.0
335.90	177.0
340.90	943.0
341.80	159.0
345.90	1906.0
346.90	420.0
352.10	2537.0
353.00	1632.0
354.00	3010.0
355.00	616.0
359.00	230.0
365.00	14078.0
366.00	1948.0
369.90	221.0
371.00	747.0
372.00	4762.0
373.00	904.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276

373.90	159.0
382.90	1275.0
384.00	329.0
385.10	187.0
389.80	594.0
390.90	436.0
391.80	249.0
401.00	293.0
401.90	1844.0
403.00	2381.0
404.10	673.0
404.90	326.0
420.90	1891.0
422.00	2286.0
423.00	15019.0
424.00	4994.0
424.90	626.0
441.00	42984.0
442.00	276032.0
443.00	48952.0
444.00	3555.0
444.90	262.0

Instrument :
BNA_N
ClientSampleId :
DFTPP276