


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021122\
 Data File : BN018588.D
 Acq On : 11 Feb 2022 09:32
 Operator : CG/JU
 Sample : DFTPP232
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP232

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021122.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Feb 11 15:26:34 2022



AutoFind: Scans 651, 652, 653; Background Corrected with Scan 648

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.9	59727	PASS
68	69	0.00	2	1.9	1171	PASS
69	198	0.00	100	42.4	61841	PASS
70	69	0.00	2	0.5	281	PASS
127	198	10	80	50.6	73858	PASS
197	198	0.00	2	0.7	950	PASS
198	198	100	100	100.0	145886	PASS
199	198	5	9	6.8	9992	PASS
275	198	10	60	23.0	33624	PASS
365	198	1	100	3.0	4399	PASS
441	198	0.01	100	10.5	15319	PASS
442	198	50	100	70.9	103424	PASS
443	442	15	24	18.3	18894	PASS

m/z	Abundance
39.10	299.0
40.00	678.0
41.10	1014.0
42.20	199.0
43.10	1095.0
44.00	798.0
49.00	215.0
51.00	151.0
53.10	177.0
55.10	1340.0
56.10	407.0
57.10	1244.0
67.10	646.0
68.00	166.0
69.10	894.0
70.10	356.0
71.10	729.0
73.10	627.0
77.00	238.0
79.10	278.0
81.10	659.0
82.10	610.0
83.10	1141.0
84.00	427.0
85.10	572.0
91.00	316.0
93.10	161.0
95.10	666.0
96.10	356.0
97.10	758.0
105.00	202.0
107.10	152.0
109.10	414.0
110.10	178.0
111.10	415.0
119.10	196.0
121.00	171.0
123.10	243.0
125.10	233.0
133.00	304.0
135.10	249.0
137.10	159.0
147.10	273.0
149.10	173.0
167.10	263.0
177.00	208.0
191.00	331.0
193.00	205.0
207.10	2508.0
208.00	557.0
209.10	356.0
281.00	434.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

m/z	Abundance
39.10	677.0
40.00	792.0
41.10	1073.0
43.10	1180.0
44.00	780.0
49.00	256.0
50.00	1366.0
51.10	4918.0
52.00	307.0
53.00	171.0
54.10	154.0
55.10	1255.0
56.10	499.0
57.10	1444.0
63.00	228.0
65.00	192.0
67.00	548.0
68.00	253.0
69.00	4619.0
70.10	376.0
71.10	795.0
73.10	3091.0
74.00	661.0
75.00	817.0
76.00	286.0
77.00	4067.0
78.10	449.0
79.00	499.0
80.00	350.0
81.10	939.0
82.10	232.0
83.00	657.0
84.10	373.0
85.10	663.0
86.00	208.0
91.10	376.0
92.10	196.0
93.00	538.0
95.10	634.0
96.10	464.0
97.20	520.0
98.00	412.0
99.00	531.0
101.00	164.0
104.00	161.0
105.00	295.0
107.00	890.0
108.10	180.0
109.20	350.0
110.00	1433.0
111.10	542.0
113.10	163.0
117.00	658.0
119.00	286.0
123.10	265.0
125.10	214.0
127.00	2116.0
128.10	217.0
129.00	793.0
133.00	363.0
135.00	298.0
137.10	220.0
141.10	151.0
147.10	1417.0
148.10	394.0
149.00	321.0
163.10	158.0
167.00	194.0
168.10	200.0
177.10	187.0
179.00	151.0
186.00	231.0
191.00	375.0
193.10	291.0
198.00	1243.0
206.00	342.0
207.00	3188.0
208.10	621.0
209.10	438.0
221.10	880.0
222.10	229.0
223.10	235.0
224.00	193.0
255.00	271.0
267.00	209.0
281.10	989.0
282.10	309.0
283.10	151.0
341.00	209.0
355.10	288.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

m/z	Abundance
36.10	161.0
37.10	864.0
38.10	2057.0
39.10	12178.0
40.00	1143.0
41.10	1234.0
42.10	202.0
43.10	1089.0
44.00	593.0
45.10	376.0
49.10	1479.0
50.10	44624.0
51.10	172352.0
52.10	8473.0
53.10	501.0
54.10	168.0
55.10	2090.0
56.00	5500.0
57.00	12785.0
58.00	594.0
59.00	190.0
61.10	2270.0
62.00	2166.0
63.10	7035.0
63.90	467.0
65.10	3308.0
66.10	330.0
67.10	707.0
68.10	3480.0
69.00	179776.0
70.10	1159.0
71.10	912.0
72.10	162.0
73.00	2486.0
74.10	18336.0
75.00	28592.0
76.10	10426.0
77.10	203840.0
78.10	14373.0
79.00	13975.0
80.00	10666.0
81.00	15102.0
82.00	3667.0
83.00	3652.0
84.00	623.0
85.00	3181.0
86.00	4449.0
87.00	1820.0
87.90	585.0
89.00	366.0
91.00	3549.0
92.00	3826.0
93.00	24344.0
94.00	1790.0
95.10	897.0
96.10	1375.0
97.10	1043.0
98.00	19384.0
99.00	14877.0
100.00	1543.0
101.00	8676.0
102.00	407.0
103.00	2665.0
104.00	5406.0
105.00	5072.0
106.10	1635.0
107.00	64552.0
108.00	9664.0
109.10	2346.0
110.00	110920.0
111.00	17512.0
112.10	2177.0
113.00	848.0
115.10	387.0
116.00	3735.0
117.00	55064.0
118.00	4191.0
119.00	644.0
120.10	732.0
121.00	405.0
122.00	4335.0
123.00	6359.0
124.00	2861.0
125.00	2846.0
127.00	210752.0
128.00	15787.0
129.00	82032.0
130.00	7012.0
131.00	1442.0
132.10	870.0
133.00	513.0
134.00	2425.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

135.00	6784.0
136.00	2789.0
137.00	3785.0
138.00	853.0
139.00	545.0
140.00	1093.0
141.00	11208.0
142.00	3580.0
143.00	2555.0
144.00	521.0
145.10	614.0
146.00	2128.0
147.00	6613.0
148.00	13584.0
149.00	2910.0
150.10	960.0
151.10	1518.0
152.00	792.0
153.00	3263.0
154.00	2497.0
155.00	6102.0
156.10	8871.0
157.10	1825.0
158.00	1953.0
159.00	1734.0
160.00	3436.0
161.00	4890.0
162.00	1515.0
163.00	480.0
164.00	596.0
165.00	3971.0
166.00	3061.0
167.00	20928.0
168.00	11883.0
169.00	2123.0
170.10	623.0
171.00	835.0
172.00	2042.0
173.00	2128.0
174.00	4471.0
175.00	8116.0
176.10	2134.0
177.00	3825.0
178.10	1251.0
179.00	15304.0
180.00	10970.0
181.00	4964.0
182.10	853.0
183.10	408.0
184.00	1391.0
185.00	7353.0
186.00	57560.0
187.00	16936.0
188.00	2060.0
189.00	3805.0
190.00	619.0
191.00	1656.0
192.00	4987.0
193.00	5685.0
194.10	1097.0
195.00	593.0
196.00	10675.0
197.10	2851.0
198.00	387968.0
199.00	26992.0
200.00	2128.0
201.50	1769.0
203.00	2905.0
204.00	13785.0
205.00	23864.0
206.10	100896.0
207.10	15784.0
208.00	4212.0
209.00	1608.0
210.10	1650.0
211.00	3968.0
213.00	304.0
215.00	1247.0
216.00	2052.0
217.00	26336.0
218.00	3095.0
219.10	386.0
221.00	13056.0
221.80	3521.0
223.00	5843.0
224.00	51848.0
225.00	13626.0
226.10	1689.0
227.00	21776.0
228.00	3297.0
229.00	4644.0
230.10	638.0
231.10	1719.0
232.10	365.0
233.00	373.0
234.00	1349.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

235.00	1530.0
236.00	1029.0
237.00	2291.0
238.20	322.0
239.00	808.0
240.00	735.0
241.00	1065.0
242.00	2574.0
243.10	2724.0
244.00	35624.0
245.10	4960.0
246.00	7552.0
247.00	1378.0
248.10	393.0
249.00	1245.0
250.10	201.0
251.00	352.0
252.00	350.0
253.00	824.0
255.00	186880.0
256.00	27904.0
257.00	2484.0
258.00	10433.0
259.00	1684.0
260.10	284.0
261.00	299.0
264.10	270.0
265.00	4042.0
266.00	567.0
267.00	274.0
270.10	208.0
271.00	308.0
272.00	481.0
273.00	4954.0
274.00	12506.0
275.00	71792.0
276.00	9622.0
277.00	6572.0
278.00	1247.0
279.00	204.0
281.00	690.0
282.00	335.0
283.00	852.0
284.10	465.0
285.00	1100.0
286.10	257.0
289.10	237.0
292.00	257.0
293.00	1292.0
294.00	456.0
295.10	481.0
296.00	18152.0
297.00	2666.0
298.00	206.0
301.00	273.0
302.00	386.0
303.00	2140.0
304.00	639.0
308.00	232.0
310.10	302.0
313.10	157.0
314.00	739.0
315.00	1863.0
316.00	984.0
317.10	177.0
321.00	511.0
322.10	239.0
323.10	4987.0
324.00	945.0
325.00	174.0
327.00	1105.0
328.00	455.0
332.10	382.0
333.00	426.0
334.00	2986.0
335.00	740.0
341.00	693.0
342.00	209.0
346.00	846.0
347.00	211.0
352.00	1148.0
353.00	931.0
354.10	1353.0
355.10	737.0
356.10	189.0
365.00	4981.0
366.00	869.0
371.00	271.0
372.00	1937.0
373.00	446.0
383.00	456.0
384.00	183.0
390.00	265.0
401.00	243.0
402.00	647.0
403.00	823.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

404.00	370.0
421.00	488.0
422.00	514.0
423.00	3867.0
424.00	1804.0
425.00	315.0
441.00	7292.0
442.10	49024.0
443.10	10555.0
444.10	1006.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

m/z	Abundance
39.10	416.0
40.00	652.0
41.10	1015.0
42.10	240.0
43.10	1038.0
44.00	836.0
49.10	275.0
50.10	603.0
51.00	2364.0
52.00	169.0
53.00	167.0
55.10	1393.0
56.10	418.0
57.10	1486.0
63.00	166.0
65.00	183.0
67.10	529.0
68.10	279.0
69.00	3811.0
70.10	375.0
71.10	736.0
73.10	827.0
74.00	474.0
75.00	695.0
76.10	245.0
77.10	3821.0
78.10	458.0
79.00	647.0
80.00	293.0
81.10	1081.0
82.00	369.0
83.10	717.0
84.00	329.0
85.10	612.0
86.00	291.0
91.10	409.0
92.00	216.0
93.00	836.0
94.10	150.0
95.10	779.0
96.10	535.0
97.10	805.0
98.00	556.0
99.00	701.0
101.00	219.0
103.00	184.0
104.00	204.0
105.00	369.0
107.00	2012.0
108.00	415.0
109.10	435.0
110.00	3838.0
111.10	881.0
112.10	175.0
114.90	157.0
117.00	2022.0
118.00	206.0
119.10	225.0
121.10	180.0
122.10	219.0
123.00	425.0
124.00	182.0
125.10	351.0
127.00	8706.0
128.00	649.0
129.00	3395.0
130.10	294.0
131.00	172.0
133.00	374.0
134.10	216.0
135.10	469.0
136.10	190.0
137.00	353.0
141.00	652.0
142.00	231.0
143.00	174.0
147.10	762.0
148.00	851.0
149.00	400.0
151.20	274.0
153.00	276.0
154.00	231.0
155.00	486.0
156.00	540.0
160.00	289.0
161.00	364.0
163.00	220.0
165.10	531.0
166.10	303.0
167.10	1787.0
168.00	1186.0
169.10	223.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

172.00	170.0
173.10	271.0
174.00	468.0
175.00	792.0
176.00	252.0
177.00	505.0
178.00	180.0
179.00	1480.0
180.00	953.0
181.00	556.0
185.00	821.0
186.00	5943.0
187.00	1654.0
188.00	207.0
189.00	464.0
191.00	518.0
192.10	580.0
193.00	846.0
194.00	200.0
196.00	1311.0
198.00	48448.0
199.00	2985.0
200.00	258.0
201.40	243.0
203.00	482.0
204.00	1896.0
205.00	3362.0
206.00	13579.0
207.10	4245.0
208.00	1012.0
209.10	581.0
210.10	326.0
211.00	668.0
214.80	199.0
216.00	371.0
217.00	4240.0
218.00	453.0
221.00	2645.0
221.90	663.0
223.00	1126.0
224.00	9440.0
225.10	2374.0
226.00	269.0
227.00	4394.0
228.00	609.0
229.00	947.0
231.10	398.0
234.00	290.0
235.00	344.0
236.00	306.0
237.10	582.0
238.90	229.0
240.00	187.0
241.00	308.0
242.10	575.0
243.00	713.0
244.00	9127.0
245.00	1115.0
246.00	1960.0
247.10	424.0
249.00	396.0
253.10	183.0
255.00	55272.0
256.00	7673.0
257.00	629.0
258.00	3466.0
259.00	506.0
265.00	1509.0
266.00	249.0
271.00	174.0
272.00	163.0
273.00	1722.0
274.00	5143.0
275.00	29080.0
276.00	3790.0
277.00	2499.0
278.10	451.0
281.10	496.0
282.00	178.0
283.10	404.0
284.10	184.0
285.10	484.0
292.00	188.0
293.00	688.0
294.00	188.0
295.00	218.0
296.00	10374.0
297.10	1451.0
301.00	171.0
302.00	213.0
303.00	1278.0
304.10	302.0
308.00	181.0
310.00	162.0
314.00	625.0
315.00	1410.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

316.00	837.0
321.10	447.0
322.00	195.0
323.00	4052.0
324.00	710.0
327.00	860.0
328.00	485.0
332.00	375.0
333.10	439.0
334.00	2883.0
335.10	756.0
341.00	614.0
346.00	991.0
347.00	171.0
352.00	1623.0
353.00	1155.0
354.00	1602.0
355.00	404.0
365.00	8216.0
366.00	1139.0
370.00	165.0
371.00	549.0
372.00	3265.0
373.10	780.0
383.00	956.0
384.10	328.0
390.00	531.0
391.10	294.0
392.00	241.0
401.00	312.0
402.00	1629.0
403.00	2317.0
404.00	815.0
405.00	162.0
415.10	153.0
421.00	2101.0
422.10	2060.0
423.10	14892.0
424.00	3522.0
425.10	406.0
441.10	38664.0
442.10	261248.0
443.00	46128.0
444.00	4184.0
445.00	153.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232