

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021124\
 Data File : BN029669.D
 Acq On : 11 Feb 2024 05:04
 Operator : MA/JU
 Sample : P1399-09
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AJ2

Quant Time: Feb 11 23:24:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN020624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Feb 11 23:18:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.880	152	6288	0.400	ng/u1	0.00
4) Naphthalene-d8	10.678	136	17323	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.520	164	7848	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.272	188	14348	0.400	ng/u1	0.00
17) Chrysene-d12	21.470	240	9872	0.400	ng/u1	0.00
23) Perylene-d12	23.837	264	13692	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.328	96	37130	4.736	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.273	152	9057	0.411	ng/u1	0.00
18) Fluoranthene-d10	19.304	212	14371	0.443	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021124\
Data File : BN029669.D
Acq On : 11 Feb 2024 05:04
Operator : MA/JU
Sample : P1399-09
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AJ2

Quant Time: Feb 11 23:24:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN020624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Feb 11 23:18:17 2024
Response via : Initial Calibration

