

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021124\
 Data File : BN029695.D
 Acq On : 11 Feb 2024 21:56
 Operator : MA/JU
 Sample : PB158924BL
 Misc :
 ALS Vial : 89 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK924

Quant Time: Feb 11 23:29:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN020624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Feb 11 23:18:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.880	152	5240	0.400	ng/u1	0.00
4) Naphthalene-d8	10.672	136	14323	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.520	164	6340	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.272	188	11544	0.400	ng/u1	0.00
17) Chrysene-d12	21.471	240	7733	0.400	ng/u1	0.00
23) Perylene-d12	23.836	264	10155	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.328	96	45181	6.915	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.272	152	6123	0.336	ng/u1	0.00
18) Fluoranthene-d10	19.304	212	8866	0.349	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021124\
Data File : BN029695.D
Acq On : 11 Feb 2024 21:56
Operator : MA/JU
Sample : PB158924BL
Misc :
ALS Vial : 89 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK924

Quant Time: Feb 11 23:29:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN020624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Feb 11 23:18:17 2024
Response via : Initial Calibration

