

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021125\
 Data File : BN036430.D
 Acq On : 11 Feb 2025 19:04
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 12 00:14:34 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.438	0.435	0.7	87	0.00
3	n-Nitrosodimethylamine	0.760	0.760	0.0	86	0.00
4 S	2-Fluorophenol	0.945	0.920	2.6	85	0.00
5 S	Phenol-d6	1.109	1.038	6.4	87	0.00
6	bis(2-Chloroethyl)ether	1.160	1.142	1.6	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.395	0.375	5.1	89	0.00
9	Naphthalene	1.154	1.113	3.6	86	0.00
10	Hexachlorobutadiene	0.281	0.285	-1.4	87	0.00
11 SURR	2-Methylnaphthalene-d10	0.615	0.595	3.3	85	0.00
12	2-Methylnaphthalene	0.757	0.730	3.6	85	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
14 S	2,4,6-Tribromophenol	0.198	0.166	16.2	80	0.00
15 S	2-Fluorobiphenyl	1.504	1.374	8.6	90	0.00
16	Acenaphthylene	1.767	1.606	9.1	85	0.00
17	Acenaphthene	1.180	1.104	6.4	86	0.00
18	Fluorene	1.680	1.558	7.3	84	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
20	4,6-Dinitro-2-methylphenol	0.078	0.066	15.4	80	0.00
21	4-Bromophenyl-phenylether	0.239	0.230	3.8	83	0.00
22	Hexachlorobenzene	0.295	0.295	0.0	86	-0.01
23	Atrazine	0.199	0.178	10.6	79	-0.01
24	Pentachlorophenol	0.140	0.114	18.6	78	0.00
25	Phenanthrene	1.156	1.107	4.2	84	0.00
26	Anthracene	1.020	0.964	5.5	83	0.00
27 SURR	Fluoranthene-d10	1.112	1.049	5.7	82	0.00
28	Fluoranthene	1.421	1.346	5.3	83	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
30	Pyrene	1.541	1.552	-0.7	83	0.00
31 S	Terphenyl-d14	0.854	0.835	2.2	80	0.00
32	Benzo(a)anthracene	1.316	1.320	-0.3	83	0.00
33	Chrysene	1.425	1.459	-2.4	87	0.00
34	Bis(2-ethylhexyl)phthalate	0.820	0.731	10.9	77	0.00
35 I	Perylene-d12	1.000	1.000	0.0	85	0.00
36	Indeno(1,2,3-cd)pyrene	1.398	1.465	-4.8	90	0.00
37	Benzo(b)fluoranthene	1.317	1.319	-0.2	89	0.00
38	Benzo(k)fluoranthene	1.356	1.401	-3.3	87	0.00
39 C	Benzo(a)pyrene	1.150	1.171	-1.8	90	0.00
40	Dibenzo(a,h)anthracene	1.103	1.151	-4.4	91	0.00
41	Benzo(g,h,i)perylene	1.250	1.307	-4.6	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4