



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021125\
Data File : BN036418.D
Acq On : 11 Feb 2025 09:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Feb 11 01:17:14 2025



AutoFind: Scans 2464, 2465, 2466; Background Corrected with Scan 2457

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.1	112736	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	47.9	105747	PASS
70	69	0.00	2	0.5	549	PASS
127	198	10	80	48.1	106163	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	220608	PASS
199	198	5	9	7.0	15371	PASS
275	198	10	60	25.0	55243	PASS
365	198	1	100	3.5	7745	PASS
441	198	0.01	100	8.8	19491	PASS
442	442	50	100	100.0	119517	PASS
443	442	15	24	19.6	23394	PASS

DDT Breakdown




Date	Instrument Name	DFTPP Data File
2/11/2025	BNA_N	<u>BN036418.D</u>
Compound Name	Response	Retention Time
DDT	843322	20.627
DDD	8015	20.233
DDE	0	19.728
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8015	851337	0.94

Instrument :
BNA_N
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021125\
Data File : BN036418.D
Acq On : 11 Feb 2025 09:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Feb 11 17:47:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN036418.D\data.ms

(70) Pentachlorophenol (C)

16.786min (-0.029) 20524.85 ng




response 244188

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.46
264.00	61.60	67.34
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021125\
Data File : BN036418.D
Acq On : 11 Feb 2025 09:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Feb 11 17:47:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN036418.D\data.ms

(77) Benzidine

19.386min (-0.018) 0.00 ng

response 1568249

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.52
183.00	13.20	12.85
0.00	0.00	0.00