


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN021220\
 Data File : BN009710.D
 Acq On : 12 Feb 2020 13:00
 Operator : CG/JU
 Sample : DFTPP59
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP59

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Feb 13 11:04:48 2020



AutoFind: Scans 2430, 2431, 2432; Background Corrected with Scan 2426

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	67.3	61002	PASS
68	69	0.00	2	2.0	888	PASS
69	198	0.00	100	48.1	43600	PASS
70	69	0.00	2	0.6	275	PASS
127	198	10	80	52.0	47189	PASS
197	198	0.00	2	0.2	209	PASS
198	198	100	100	100.0	90685	PASS
199	198	5	9	7.1	6443	PASS
275	198	10	60	23.5	21327	PASS
365	198	1	100	3.1	2844	PASS
441	198	0.01	100	8.8	7971	PASS
442	198	50	100	55.0	49917	PASS
443	442	15	24	20.1	10011	PASS

Scan 2426 (17.257 min): BN009710.D
DFTPP59

m/z	Abundance
39.90	989.0
40.90	379.0
44.00	2458.0
48.90	311.0
73.00	538.0
207.00	1664.0
207.90	561.0
281.10	343.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

m/z	Abundance
38.00	746.0
39.00	4123.0
40.00	1143.0
41.10	377.0
44.00	2717.0
49.00	739.0
50.00	12527.0
51.00	49576.0
52.00	2505.0
54.90	429.0
55.90	1582.0
57.00	3460.0
60.90	477.0
62.00	427.0
63.00	1937.0
65.00	968.0
67.90	626.0
68.00	628.0
69.00	34096.0
73.00	885.0
74.00	4291.0
75.00	5912.0
76.00	1742.0
77.00	38912.0
78.10	3121.0
79.10	3635.0
80.00	2225.0
81.00	2908.0
82.00	738.0
83.10	430.0
84.90	594.0
86.10	941.0
86.90	520.0
90.90	528.0
92.00	816.0
93.00	4665.0
95.80	327.0
98.00	4200.0
99.00	3321.0
99.90	471.0
101.00	1122.0
103.10	336.0
104.00	1022.0
104.90	859.0
107.00	11213.0
108.00	1725.0
108.80	315.0
110.00	17504.0
111.00	2885.0
111.90	527.0
116.10	814.0
117.00	10185.0
118.00	904.0
122.10	643.0
123.00	1068.0
123.90	307.0
124.80	474.0
127.00	34368.0
128.00	2808.0
129.00	13265.0
130.00	1177.0
131.00	420.0
133.90	594.0
134.90	833.0
135.90	457.0
137.00	945.0
137.70	325.0
141.00	1636.0
142.00	485.0
142.90	476.0
147.00	1255.0
148.00	2364.0
148.90	566.0
153.00	649.0
154.00	420.0
155.00	934.0
156.00	1517.0
156.90	377.0
159.90	593.0
161.00	1001.0
161.90	309.0
164.90	600.0
166.00	595.0
166.90	2920.0
168.00	3393.0
169.10	772.0
172.90	680.0
173.90	745.0
174.90	1311.0
177.00	531.0
178.90	2767.0
179.90	1914.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

180.90	961.0
183.90	360.0
185.00	1436.0
186.00	9181.0
187.00	2823.0
188.00	412.0
189.00	646.0
191.00	486.0
191.90	871.0
193.00	1098.0
195.90	1813.0
196.90	627.0
197.90	58152.0
199.00	4470.0
199.90	394.0
201.70	355.0
202.90	677.0
204.00	2849.0
205.00	4282.0
206.00	17096.0
207.00	4036.0
207.90	1079.0
210.00	425.0
211.00	728.0
214.90	318.0
215.90	568.0
216.90	4645.0
217.90	729.0
221.00	2248.0
222.10	510.0
223.00	790.0
224.00	8747.0
225.00	1972.0
226.90	3794.0
227.90	741.0
229.00	673.0
231.00	427.0
233.80	370.0
236.90	596.0
240.90	356.0
241.90	414.0
242.90	599.0
244.00	6149.0
244.90	690.0
245.90	957.0
247.00	339.0
249.00	415.0
254.10	312.0
254.90	32560.0
256.00	4983.0
257.00	485.0
257.90	2074.0
259.10	473.0
265.00	711.0
273.00	1152.0
273.90	2475.0
275.00	12183.0
275.90	1912.0
277.00	1309.0
278.00	337.0
285.00	441.0
296.00	2993.0
296.90	469.0
302.90	430.0
314.90	455.0
323.00	1071.0
327.00	304.0
333.90	654.0
351.90	321.0
364.90	1400.0
372.00	579.0
402.90	305.0
421.00	392.0
421.90	304.0
422.90	1380.0
423.80	567.0
441.00	3738.0
442.00	21520.0
442.90	5316.0
443.90	398.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

m/z	Abundance
37.00	342.0
37.90	922.0
39.10	6058.0
40.00	1041.0
41.00	620.0
44.00	2752.0
44.90	397.0
48.90	618.0
50.00	17408.0
51.00	73552.0
52.00	3491.0
55.00	431.0
56.00	2398.0
57.00	5007.0
61.00	977.0
62.00	855.0
63.00	2407.0
64.00	422.0
65.00	1370.0
68.10	986.0
69.00	53512.0
70.00	406.0
73.00	1025.0
74.00	5818.0
75.00	9716.0
76.10	2860.0
77.00	59024.0
78.00	4203.0
79.00	4298.0
80.00	3426.0
81.00	4441.0
82.00	1208.0
83.00	1082.0
83.90	325.0
85.00	858.0
86.00	1426.0
87.00	464.0
87.90	309.0
88.90	327.0
91.00	1283.0
92.00	1144.0
93.00	7876.0
93.90	372.0
96.10	582.0
97.90	5738.0
99.00	4973.0
101.00	2407.0
103.00	813.0
104.00	1364.0
104.90	1525.0
105.90	350.0
107.00	17504.0
107.90	2703.0
108.90	526.0
110.00	29528.0
110.90	5151.0
111.80	652.0
115.90	840.0
117.00	16856.0
117.90	1010.0
121.90	1081.0
123.00	1800.0
123.90	847.0
125.00	857.0
126.20	632.0
127.00	56608.0
128.00	3832.0
129.00	22096.0
130.00	2032.0
131.00	399.0
134.00	584.0
135.00	2080.0
135.90	683.0
136.90	1114.0
141.00	2836.0
141.90	810.0
143.00	645.0
146.00	460.0
147.00	1781.0
148.00	3888.0
148.90	1195.0
150.90	504.0
153.00	796.0
154.00	725.0
154.90	1417.0
156.00	2297.0
156.90	518.0
157.90	626.0
158.90	486.0
159.90	723.0
160.90	1287.0
162.00	452.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

164.90	1021.0
166.10	1140.0
167.00	6252.0
168.00	4593.0
168.80	561.0
170.90	399.0
171.90	413.0
173.10	956.0
174.00	1620.0
174.90	2048.0
175.90	528.0
176.90	1126.0
178.20	458.0
178.90	4502.0
180.00	3310.0
181.00	1653.0
185.00	2220.0
186.00	16960.0
187.00	4700.0
187.90	584.0
188.90	1163.0
190.80	593.0
192.00	1602.0
193.00	1705.0
194.10	377.0
196.00	2543.0
197.90	111016.0
198.90	7821.0
200.00	758.0
201.30	685.0
203.00	775.0
204.00	3800.0
205.00	7715.0
206.00	29424.0
207.00	6000.0
207.90	1552.0
209.10	592.0
209.90	792.0
210.40	391.0
211.00	1257.0
212.10	354.0
214.80	382.0
216.00	979.0
217.00	7648.0
218.00	938.0
220.90	3840.0
222.00	949.0
222.90	2087.0
224.00	16023.0
225.00	3855.0
226.00	517.0
227.00	7551.0
227.90	1128.0
228.90	1694.0
229.90	347.0
231.00	911.0
234.00	604.0
234.90	479.0
237.00	1109.0
241.90	989.0
242.90	794.0
244.00	11227.0
245.00	1562.0
246.00	2578.0
246.90	643.0
248.90	632.0
253.00	464.0
255.00	62944.0
256.00	8929.0
257.00	616.0
258.00	3078.0
258.90	569.0
265.00	1485.0
273.00	2049.0
274.00	5125.0
275.00	25832.0
276.00	3325.0
276.90	1974.0
278.00	534.0
281.00	524.0
282.90	488.0
285.00	341.0
292.80	606.0
296.00	8478.0
296.90	1075.0
302.90	649.0
313.80	554.0
315.00	746.0
315.90	383.0
320.90	451.0
323.00	1912.0
324.00	508.0
326.90	478.0
334.00	1698.0
345.80	602.0
351.80	656.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

352.90	561.0
353.90	648.0
365.00	3402.0
366.00	584.0
371.90	1343.0
382.90	316.0
402.10	459.0
402.90	509.0
420.80	445.0
421.90	607.0
422.90	3180.0
424.00	1242.0
441.00	8348.0
442.00	57728.0
443.00	9683.0
444.00	1220.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

m/z	Abundance
37.10	345.0
38.10	722.0
39.10	4185.0
40.00	1246.0
40.90	384.0
43.00	317.0
44.00	2255.0
49.10	411.0
50.00	16808.0
51.00	59880.0
52.10	2901.0
54.90	465.0
56.00	1718.0
57.00	4068.0
61.00	869.0
62.00	999.0
63.00	2360.0
64.00	373.0
64.90	735.0
68.00	1051.0
69.00	43192.0
70.00	419.0
73.00	1109.0
74.00	5012.0
75.00	7249.0
76.00	2873.0
77.00	49160.0
78.10	3142.0
79.00	4144.0
79.90	2371.0
80.90	3572.0
82.00	931.0
83.00	708.0
83.90	425.0
84.90	610.0
85.80	1074.0
87.00	565.0
91.00	810.0
92.00	1410.0
92.90	6495.0
94.10	486.0
95.90	327.0
98.00	5550.0
99.00	4510.0
100.10	310.0
101.00	2055.0
103.00	722.0
104.00	943.0
105.00	1168.0
106.10	435.0
107.00	15135.0
108.00	2402.0
108.90	502.0
110.00	26112.0
111.00	3641.0
111.90	583.0
116.00	826.0
117.00	14918.0
117.90	879.0
122.00	1235.0
122.90	1854.0
123.90	682.0
124.90	1290.0
126.00	324.0
127.00	50592.0
128.00	3369.0
129.00	19688.0
130.00	1772.0
130.90	368.0
133.90	619.0
135.00	1976.0
135.90	695.0
137.00	831.0
140.90	2275.0
141.90	539.0
143.00	555.0
143.90	397.0
146.10	443.0
147.00	1683.0
148.00	3519.0
149.00	887.0
151.00	332.0
153.00	1139.0
154.00	591.0
155.00	1590.0
156.00	2704.0
157.10	433.0
157.90	388.0
158.90	513.0
160.10	758.0
161.00	1581.0
161.90	353.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

164.90	985.0
166.00	936.0
167.00	6464.0
168.00	3965.0
168.90	318.0
172.00	591.0
172.90	731.0
174.00	1147.0
175.00	1742.0
176.00	613.0
177.00	954.0
177.90	462.0
178.90	3540.0
179.90	2637.0
180.90	1366.0
182.00	350.0
183.90	403.0
185.00	2093.0
186.00	14750.0
187.00	4382.0
188.90	1176.0
191.00	585.0
192.00	1737.0
193.00	1235.0
193.80	361.0
196.00	2743.0
197.90	102888.0
198.90	7038.0
199.90	568.0
201.30	445.0
203.00	800.0
204.00	4417.0
205.00	6386.0
206.00	27688.0
207.00	5764.0
208.00	1242.0
209.00	729.0
211.10	1197.0
215.90	749.0
216.90	7926.0
218.10	1010.0
221.00	3875.0
223.10	1660.0
224.00	16592.0
225.00	4359.0
226.10	493.0
226.90	7612.0
227.90	780.0
228.80	1201.0
231.10	738.0
233.90	394.0
235.00	607.0
235.90	316.0
236.90	1014.0
241.00	378.0
242.00	1072.0
242.90	1033.0
244.00	12096.0
244.90	1393.0
246.00	2230.0
247.00	484.0
249.00	519.0
253.00	328.0
255.00	61624.0
255.90	7875.0
257.00	984.0
258.00	3480.0
258.90	917.0
264.90	1997.0
271.90	315.0
273.00	2849.0
274.00	5194.0
274.90	25968.0
276.00	3819.0
277.00	2809.0
277.80	446.0
280.90	310.0
282.90	413.0
283.90	309.0
284.90	311.0
292.90	452.0
296.00	8127.0
296.90	1090.0
303.00	809.0
314.00	385.0
314.90	914.0
315.90	488.0
323.00	2249.0
326.90	607.0
333.90	1366.0
335.10	403.0
341.00	334.0
345.90	642.0
351.90	767.0
353.00	463.0
353.90	746.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

354.90	324.0
364.90	3732.0
366.00	518.0
372.00	1549.0
372.90	379.0
382.90	437.0
402.00	609.0
403.00	954.0
403.90	390.0
420.90	667.0
422.00	553.0
423.00	4606.0
423.90	1614.0
441.00	11829.0
442.00	70504.0
443.00	15036.0
444.00	1395.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59