

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021225\
 Data File : BN036442.D
 Acq On : 12 Feb 2025 16:24
 Operator : RC/JU
 Sample : PB166675BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB166675BL

Quant Time: Feb 12 17:37:10 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2366	0.400	ng	0.00
7) Naphthalene-d8	10.562	136	5102	0.400	ng	0.02
13) Acenaphthene-d10	14.398	164	2942	0.400	ng	0.01
19) Phenanthrene-d10	17.149	188	6604	0.400	ng	0.01
29) Chrysene-d12	21.331	240	5195	0.400	ng	0.00
35) Perylene-d12	23.595	264	4640	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2308	0.413	ng	0.00
5) Phenol-d6	6.937	99	2230	0.340	ng	0.00
8) Nitrobenzene-d5	8.918	82	1785	0.355	ng	0.01
11) 2-Methylnaphthalene-d10	12.162	152	2747	0.350	ng	0.02
14) 2,4,6-Tribromophenol	15.907	330	414	0.284	ng	0.02
15) 2-Fluorobiphenyl	13.041	172	3962	0.358	ng	0.02
27) Fluoranthene-d10	19.174	212	7210	0.393	ng	0.00
31) Terphenyl-d14	19.773	244	4909	0.443	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021225\
Data File : BN036442.D
Acq On : 12 Feb 2025 16:24
Operator : RC/JU
Sample : PB166675BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB166675BL

Quant Time: Feb 12 17:37:10 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration

