

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021225\
 Data File : BN036450.D
 Acq On : 12 Feb 2025 21:12
 Operator : RC/JU
 Sample : Q1327-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-192-GW-540-542

Quant Time: Feb 12 23:17:24 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Anahy Claudio 02/13/2025
 Supervised By :Jagrut Upadhyay 02/13/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2852	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	7637	0.400	ng	# 0.00
13) Acenaphthene-d10	14.388	164	4410	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	11189	0.400	ng	# 0.00
29) Chrysene-d12	21.322	240	8988	0.400	ng	# 0.00
35) Perylene-d12	23.587	264	2010	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	1447	0.215	ng	0.00
5) Phenol-d6	6.930	99	1369	0.173	ng	0.00
8) Nitrobenzene-d5	8.897	82	1845	0.245	ng	-0.01
11) 2-Methylnaphthalene-d10	12.131	152	3049	0.260	ng	-0.01
14) 2,4,6-Tribromophenol	15.883	330	784	0.359	ng	0.00
15) 2-Fluorobiphenyl	13.009	172	5775	0.348	ng	-0.01
27) Fluoranthene-d10	19.164	212	9345	0.300	ng	0.00
31) Terphenyl-d14	19.764	244	8624	0.450	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.268	88	3525m	1.130	ng	Qvalue
34) Bis(2-ethylhexyl)phtha...	21.232	149	4166	0.226	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021225\
Data File : BN036450.D
Acq On : 12 Feb 2025 21:12
Operator : RC/JU
Sample : Q1327-05
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-540-542

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 02/13/2025
Supervised By :Jagrut Upadhyay 02/13/2025

Quant Time: Feb 12 23:17:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration

