

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021323\
 Data File : BN024014.D
 Acq On : 13 Feb 2023 15:26
 Operator : CG/JU
 Sample : 01422-13
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0B71

Quant Time: Feb 14 01:15:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN013023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 14 01:13:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	3910	0.400	ng/u1	0.00
4) Naphthalene-d8	10.639	136	10596	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.489	164	6425	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.239	188	15005	0.400	ng/u1	# 0.00
17) Chrysene-d12	21.430	240	14089	0.400	ng/u1	# 0.00
23) Perylene-d12	23.806	264	11710	0.400	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.227	96	20477	4.886	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.234	152	4478	0.270	ng/u1	0.00
18) Fluoranthene-d10	19.268	212	11987	0.317	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021323\
Data File : BN024014.D
Acq On : 13 Feb 2023 15:26
Operator : CG/JU
Sample : 01422-13
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0B71

Quant Time: Feb 14 01:15:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN013023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 14 01:13:07 2023
Response via : Initial Calibration

