

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021423\
 Data File : BN024021.D
 Acq On : 14 Feb 2023 13:10
 Operator : CG/JU
 Sample : 01512-10
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0B84

Quant Time: Feb 14 14:54:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN013023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 14 01:13:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.829	152	4895	0.400	ng/u1	0.00
4) Naphthalene-d8	10.632	136	13100	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.486	164	7766	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.237	188	17353	0.400	ng/u1	# 0.00
17) Chrysene-d12	21.438	240	13227	0.400	ng/u1	# 0.00
23) Perylene-d12	23.828	264	10005	0.400	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.226	96	20531	3.913	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.233	152	5007	0.244	ng/u1	0.00
18) Fluoranthene-d10	19.270	212	11840	0.333	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021423\
Data File : BN024021.D
Acq On : 14 Feb 2023 13:10
Operator : CG/JU
Sample : 01512-10
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0B84

Quant Time: Feb 14 14:54:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN013023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 14 01:13:07 2023
Response via : Initial Calibration

