

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021424\
 Data File : BN029780.D
 Acq On : 14 Feb 2024 21:28
 Operator : MA/JU
 Sample : P1305-02RE
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0FP3

Quant Time: Feb 14 22:26:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 13 22:50:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	198952	20.000	ng/ul	0.00
20) Naphthalene-d8	10.663	136	806691	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.510	164	426133	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.263	188	849547	20.000	ng/ul	0.00
79) Chrysene-d12	21.463	240	815460	20.000	ng/ul	0.00
88) Perylene-d12	23.827	264	920318	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.322	96	30754	5.084	ng/uL	0.00
4) Pyridine-d5	3.734	84	212735	13.188	ng/ul	0.00
7) Phenol-d5	7.034	99	521712	28.035	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	340554	27.421	ng/ul	0.00
11) 2-Chlorophenol-d4	7.399	132	405630	30.080	ng/ul	0.00
15) 4-Methylphenol-d8	8.581	113	398289	28.699	ng/ul	0.00
21) Nitrobenzene-d5	9.034	128	197789	31.591	ng/ul	0.00
24) 2-Nitrophenol-d4	9.752	143	195520	35.857	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	340007	31.590	ng/ul	0.00
31) 4-Chloroaniline-d4	10.810	131	484790	26.402	ng/ul	0.00
46) Dimethylphthalate-d6	13.934	166	1045411	32.951	ng/ul	0.00
49) Acenaphthylene-d8	14.204	160	1214685	32.240	ng/ul	0.00
54) 4-Nitrophenol-d4	14.722	143	159814	26.540	ng/ul	0.00
60) Fluorene-d10	15.510	176	895459	33.178	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.634	200	108692	26.257	ng/ul	0.00
73) Anthracene-d10	17.363	188	1260046	31.648	ng/ul	0.00
81) Pyrene-d10	19.663	212	1515970	30.978	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.680	264	1449885	31.484	ng/ul	0.00
Target Compounds						
48) 2,6-Dinitrotoluene	13.940	165	7978	1.373	ng/ul#	39


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021424\
Data File : BN029780.D
Acq On : 14 Feb 2024 21:28
Operator : MA/JU
Sample : P1305-02RE
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0FP3

Quant Time: Feb 14 22:26:33 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 13 22:50:45 2024
Response via : Initial Calibration





#48
 2,6-Dinitrotoluene
 Concen: 1.373 ng/ul
 RT: 13.940 min Scan# 11
 Delta R.T. -0.171 min
 Lab File: BN029780.D
 Acq: 14 Feb 2024 21:28

Instrument :
 BNA_N
 ClientSampleId :
 C0FP3

Tgt Ion	Ratio	Lower	Upper
165	100		
89	3.0	55.8	83.8#
121	26.6	19.8	29.6

