

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021424\
 Data File : BN029823.D
 Acq On : 16 Feb 2024 11:34
 Operator : MA/JU
 Sample : PB158982BS
 Misc :
 ALS Vial : 104 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS982

Quant Time: Feb 17 04:45:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 13 22:50:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.863	152	152795	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	622342	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.510	164	338670	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.263	188	716782	20.000	ng/u1	0.00
79) Chrysene-d12	21.469	240	621703	20.000	ng/u1	0.00
88) Perylene-d12	23.833	264	713598	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.317	96	30478	6.561	ng/uL	0.00
4) Pyridine-d5	3.728	84	417630	33.711	ng/u1	0.00
7) Phenol-d5	7.034	99	534935	37.430	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	334234	35.042	ng/u1	0.00
11) 2-Chlorophenol-d4	7.393	132	417038	40.268	ng/u1	0.00
15) 4-Methylphenol-d8	8.581	113	411229	38.582	ng/u1	0.00
21) Nitrobenzene-d5	9.034	128	200210	41.450	ng/u1	0.00
24) 2-Nitrophenol-d4	9.751	143	209591	49.823	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	371286	44.715	ng/u1	0.00
31) 4-Chloroaniline-d4	10.810	131	506079	35.726	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	1067134	42.323	ng/u1	0.00
49) Acenaphthylene-d8	14.204	160	1232291	41.153	ng/u1	0.00
54) 4-Nitrophenol-d4	14.728	143	215184	44.965	ng/u1	0.00
60) Fluorene-d10	15.510	176	900943	42.002	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.639	200	162686	46.580	ng/u1	0.00
73) Anthracene-d10	17.363	188	1343746	40.001	ng/u1	0.00
81) Pyrene-d10	19.663	212	1547810	41.486	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.680	264	1556749	43.597	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.352	88	65868	13.690	ng/uL	94
5) Pyridine	3.746	79	441558	34.452	ng/u1	96
6) Benzaldehyde	7.016	77	283789	44.298	ng/u1	96
8) Phenol	7.063	94	576258	39.342	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.305	93	463051	37.596	ng/u1	97
12) 2-Chlorophenol	7.428	128	452496	41.644	ng/u1	97
13) 2-Methylphenol	8.310	108	418874	39.629	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.399	45	568770	35.358	ng/u1	96
16) Acetophenone	8.699	105	673502	39.703	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.687	70	358612	40.946	ng/u1	97
18) 4-Methylphenol	8.646	108	448594	40.164	ng/u1	98
19) Hexachloroethane	8.934	117	197896	39.247	ng/u1	94
22) Nitrobenzene	9.075	77	560208	40.933	ng/u1	99
23) Isophorone	9.604	82	998383	41.881	ng/u1	100
25) 2-Nitrophenol	9.787	139	236819	49.298	ng/u1	99
26) 2,4-Dimethylphenol	9.846	107	435714	37.846	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.093	93	603038	39.805	ng/u1	99
29) 2,4-Dichlorophenol	10.316	162	382545	44.586	ng/u1	97
30) Naphthalene	10.716	128	1432784	40.278	ng/u1	99
32) 4-Chloroaniline	10.834	127	436645	31.225	ng/u1	99
33) Hexachlorobutadiene	10.993	225	230630	43.554	ng/u1	96
34) Caprolactam	11.622	113	91347	31.995	ng/u1	94
35) 4-Chloro-3-methylphenol	11.963	107	458619	46.277	ng/u1	99
36) 2-Methylnaphthalene	12.328	142	909021	41.687	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021424\
 Data File : BN029823.D
 Acq On : 16 Feb 2024 11:34
 Operator : MA/JU
 Sample : PB158982BS
 Misc :
 ALS Vial : 104 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS982

Quant Time: Feb 17 04:45:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 13 22:50:45 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.551	142	903879	41.810	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.693	216	417443	41.583	ng/ul	97
40) Hexachlorocyclopentadiene	12.663	237	214031	32.618	ng/ul	96
41) 2,4,6-Trichlorophenol	12.940	196	279097	48.235	ng/ul	97
42) 2,4,5-Trichlorophenol	13.016	196	307401	47.281	ng/ul	98
43) 1,1'-Biphenyl	13.345	154	1161745	40.675	ng/ul	99
44) 2-Chloronaphthalene	13.387	162	904976	40.769	ng/ul	98
45) 2-Nitroaniline	13.604	65	316828	48.547	ng/ul	95
47) Dimethylphthalate	13.981	163	1150072	44.558	ng/ul	99
48) 2,6-Dinitrotoluene	14.104	165	232140	50.263	ng/ul	96
50) Acenaphthylene	14.234	152	1502043	41.797	ng/ul	99
51) 3-Nitroaniline	14.434	138	249248	46.369	ng/ul	98
52) Acenaphthene	14.575	153	978749	41.230	ng/ul	96
53) 2,4-Dinitrophenol	14.639	184	114541	48.869	ng/ul	97
55) 4-Nitrophenol	14.739	109	208533	47.654	ng/ul	95
56) Dibenzofuran	14.910	168	1301272	41.511	ng/ul	99
57) 2,4-Dinitrotoluene	14.892	165	350298	53.872	ng/ul	92
58) 2,3,4,6-Tetrachlorophenol	15.139	232	241780	52.398	ng/ul	99
59) Diethylphthalate	15.351	149	1232662	46.493	ng/ul	99
61) Fluorene	15.563	166	1095072	44.195	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.563	204	496867	44.833	ng/ul	96
63) 4-Nitroaniline	15.598	138	274457	56.572	ng/ul	90
66) 4,6-Dinitro-2-methylph...	15.651	198	184821	48.000	ng/ul	95
67) N-Nitrosodiphenylamine	15.781	169	941503	42.151	ng/ul	98
68) 4-Bromophenyl-phenylether	16.457	248	309968	45.688	ng/ul	91
69) Hexachlorobenzene	16.557	284	339340	44.375	ng/ul	99
70) Atrazine	16.739	200	201566	28.884	ng/ul	98
71) Pentachlorophenol	16.910	266	205357	46.242	ng/ul	92
72) Phenanthrene	17.310	178	1716600	41.401	ng/ul	99
74) Anthracene	17.398	178	1728217	41.532	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.304	216	413435	39.306	ng/ul	98
76) Pentachlorobenzene	14.828	250	382829	40.648	ng/ul	92
77) Carbazole	17.675	167	1621220	42.656	ng/ul	98
78) Di-n-butylphthalate	18.251	149	2180704	47.713	ng/ul	99
80) Fluoranthene	19.327	202	1965395	43.366	ng/ul	98
82) Pyrene	19.692	202	2051885	42.797	ng/ul	94
83) Butylbenzylphthalate	20.616	149	1025688	52.492	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.392	252	564391	46.869	ng/ul	95
85) Benzo(a)anthracene	21.451	228	1991610	43.608	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.398	149	1563656	51.734	ng/ul	100
87) Chrysene	21.504	228	1893665	44.244	ng/ul	98
89) Di-n-octyl phthalate	22.327	149	2753215	49.774	ng/ul	100
90) Benzo(b)fluoranthene	23.115	252	1937350	43.102	ng/ul	98
91) Benzo(k)fluoranthene	23.163	252	1986471	43.179	ng/ul	96
93) Benzo(a)pyrene	23.733	252	1827447	42.396	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.292	276	2260856	43.073	ng/ul	99
95) Dibenzo(a,h)anthracene	26.315	278	1837921	42.890	ng/ul	95
96) Benzo(g,h,i)perylene	27.039	276	1852422	42.195	ng/ul	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021424\
 Data File : BN029823.D
 Acq On : 16 Feb 2024 11:34
 Operator : MA/JU
 Sample : PB158982BS
 Misc :
 ALS Vial : 104 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS982

Quant Time: Feb 17 04:45:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 13 22:50:45 2024
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 13.690 ng/uL
RT: 3.352 min Scan# 45
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion: 88 Resp: 65868
Ion Ratio Lower Upper
88 100
43 30.5 22.1 33.1
58 65.3 56.5 84.7



#5
Pyridine
Concen: 34.452 ng/uL
RT: 3.746 min Scan# 112
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion: 79 Resp: 441558
Ion Ratio Lower Upper
79 100
52 65.0 55.4 83.0
51 31.8 26.9 40.3









#6
Benzaldehyde
Concen: 44.298 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion: 77 Resp: 283789
Ion Ratio Lower Upper
77 100
105 92.6 71.0 106.4
106 89.5 69.4 104.0



#8
Phenol
Concen: 39.342 ng/ul
RT: 7.063 min Scan# 676
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion: 94 Resp: 576258
Ion Ratio Lower Upper
94 100
65 29.9 23.9 35.9
66 40.5 32.2 48.2









#10
Bis(2-Chloroethyl)ether
Concen: 37.596 ng/ul
RT: 7.305 min Scan# 717
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion: 93 Resp: 463051
Ion Ratio Lower Upper
93 100
63 73.6 61.4 92.2
95 32.3 24.9 37.3



#12
2-Chlorophenol
Concen: 41.644 ng/ul
RT: 7.428 min Scan# 738
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:128 Resp: 452496
Ion Ratio Lower Upper
128 100
64 46.4 39.8 59.6
130 31.5 25.6 38.4





#13

2-Methylphenol

Concen: 39.629 ng/ul

RT: 8.310 min Scan# 888

Delta R.T. -0.006 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Instrument :

BNA_N

ClientSampleId :





SLCS982

Tgt Ion:108 Resp: 418874

Ion Ratio Lower Upper

108 100

107 89.4 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 35.358 ng/ul

RT: 8.399 min Scan# 903

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34




Tgt Ion: 45 Resp: 568770


Ion Ratio Lower Upper

45 100

77 17.9 12.7 19.1

79 13.1 9.4 14.2









#16
Acetophenone
Concen: 39.703 ng/ul
RT: 8.699 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:105 Resp: 673502
Ion Ratio Lower Upper
105 100
77 82.9 67.4 101.0
51 25.8 20.7 31.1



#17
N-Nitroso-di-n-propylamine
Concen: 40.946 ng/ul
RT: 8.687 min Scan# 952
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion: 70 Resp: 358612
Ion Ratio Lower Upper
70 100
42 56.6 43.0 64.4
101 9.3 7.4 11.2
130 20.3 15.1 22.7









#18
4-Methylphenol
Concen: 40.164 ng/ul
RT: 8.646 min Scan# 945
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:108 Resp: 448594
Ion Ratio Lower Upper
108 100
107 113.3 88.7 133.1



#19
Hexachloroethane
Concen: 39.247 ng/ul
RT: 8.934 min Scan# 994
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:117 Resp: 197896
Ion Ratio Lower Upper
117 100
201 85.1 64.4 96.6
199 52.5 38.3 57.5









#22
Nitrobenzene
Concen: 40.933 ng/ul
RT: 9.075 min Scan# 1018
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion: 77 Resp: 560208
Ion Ratio Lower Upper
77 100
123 41.6 33.6 50.4
65 13.1 10.6 16.0



#23
Isophorone
Concen: 41.881 ng/ul
RT: 9.604 min Scan# 1108
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion: 82 Resp: 998383
Ion Ratio Lower Upper
82 100
95 7.4 5.8 8.6
138 15.9 12.6 19.0









#25
2-Nitrophenol
Concen: 49.298 ng/ul
RT: 9.787 min Scan# 1139
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:139 Resp: 236819
Ion Ratio Lower Upper
139 100
65 57.5 46.6 70.0
109 34.5 27.0 40.4



#26
2,4-Dimethylphenol
Concen: 37.846 ng/ul
RT: 9.846 min Scan# 1149
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:107 Resp: 435714
Ion Ratio Lower Upper
107 100
121 48.0 40.0 60.0
122 80.8 67.6 101.4





#27

Bis(2-Chloroethoxy)methane

Concen: 39.805 ng/ul

RT: 10.093 min Scan# 1191

Delta R.T. -0.006 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Instrument :

BNA_N

ClientSampleId :

SLCS982





Tgt Ion: 93 Resp: 603038

Ion Ratio Lower Upper

93 100

95 31.9 25.7 38.5

123 10.7 8.2 12.2



#29

2,4-Dichlorophenol

Concen: 44.586 ng/ul

RT: 10.316 min Scan# 1229

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34




Tgt Ion:162 Resp: 382545


Ion Ratio Lower Upper

162 100

164 63.5 53.3 79.9

98 40.5 33.8 50.8









#30
Naphthalene
Concen: 40.278 ng/ul
RT: 10.716 min Scan# 1297
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Instrument :
BNA_N
ClientSampleId :
SLCS982

Tgt Ion:128 Resp: 1432784




Ion	Ratio	Lower	Upper
128	100		
129	10.8	9.0	13.6
127	13.2	10.4	15.6




#32
4-Chloroaniline
Concen: 31.225 ng/ul
RT: 10.834 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:127 Resp: 436645

Ion	Ratio	Lower	Upper
127	100		
129	32.4	26.4	39.6









#33
Hexachlorobutadiene
Concen: 43.554 ng/ul
RT: 10.993 min Scan# 1451
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:225 Resp: 230630
Ion Ratio Lower Upper
225 100
223 59.2 51.6 77.4
227 66.7 52.2 78.2



#34
Caprolactam
Concen: 31.995 ng/ul
RT: 11.622 min Scan# 1451
Delta R.T. 0.012 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:113 Resp: 91347
Ion Ratio Lower Upper
113 100
55 153.7 125.3 187.9
56 117.9 103.4 155.0









#35
4-Chloro-3-methylphenol
Concen: 46.277 ng/ul
RT: 11.963 min Scan# 1509
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:107 Resp: 458619
Ion Ratio Lower Upper
107 100
144 24.1 18.8 28.2
142 77.0 60.6 90.8



#36
2-Methylnaphthalene
Concen: 41.687 ng/ul
RT: 12.328 min Scan# 1571
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:142 Resp: 909021
Ion Ratio Lower Upper
142 100
141 90.2 70.4 105.6









#37
1-Methylnaphthalene
Concen: 41.810 ng/ul
RT: 12.551 min Scan# 1609
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:142 Resp: 903879
Ion Ratio Lower Upper
142 100
141 90.2 73.8 110.8
116 4.2 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 41.583 ng/ul
RT: 12.693 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:216 Resp: 417443
Ion Ratio Lower Upper
216 100
214 78.8 60.4 90.6
179 24.8 18.6 28.0
108 23.9 19.0 28.6









#40
Hexachlorocyclopentadiene
Concen: 32.618 ng/ul
RT: 12.663 min Scan# 1628
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:237 Resp: 214031
Ion Ratio Lower Upper
237 100
235 59.4 50.2 75.4
272 13.1 9.5 14.3



#41
2,4,6-Trichlorophenol
Concen: 48.235 ng/ul
RT: 12.940 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:196 Resp: 279097
Ion Ratio Lower Upper
196 100
198 93.9 76.8 115.2
200 28.0 24.6 37.0





#42

2,4,5-Trichlorophenol

Concen: 47.281 ng/ul

RT: 13.016 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Instrument :

BNA_N

ClientSampleId :

SLCS982





Tgt Ion:196 Resp: 307401

Ion Ratio Lower Upper

196 100

198 91.5 75.3 112.9

200 30.1 24.7 37.1



#43

1,1'-Biphenyl

Concen: 40.675 ng/ul

RT: 13.345 min Scan# 1744

Delta R.T. -0.006 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34




Tgt Ion:154 Resp: 1161745


Ion Ratio Lower Upper

154 100

153 41.2 32.8 49.2

76 14.8 12.9 19.3









#44
2-Chloronaphthalene
Concen: 40.769 ng/ul
RT: 13.387 min Scan# 1751
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:162 Resp: 904976
Ion Ratio Lower Upper
162 100
164 32.5 25.5 38.3
127 42.9 35.7 53.5



#45
2-Nitroaniline
Concen: 48.547 ng/ul
RT: 13.604 min Scan# 1788
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion: 65 Resp: 316828
Ion Ratio Lower Upper
65 100
92 66.3 54.3 81.5
138 96.3 71.2 106.8





#47

Dimethylphthalate

Concen: 44.558 ng/ul

RT: 13.981 min Scan# 1852

Delta R.T. -0.006 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Instrument :

BNA_N

ClientSampleId :

SLCS982





Tgt Ion:163 Resp: 1150072

Ion Ratio Lower Upper

163 100

194 3.9 2.7 4.1

164 10.1 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 50.263 ng/ul

RT: 14.104 min Scan# 1873

Delta R.T. -0.006 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34




Tgt Ion:165 Resp: 232140


Ion Ratio Lower Upper

165 100

89 74.0 55.8 83.8

121 24.3 19.8 29.6









#50
Acenaphthylene
Concen: 41.797 ng/ul
RT: 14.234 min Scan# 1895
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:152 Resp: 1502043
Ion Ratio Lower Upper
152 100
151 19.8 15.4 23.2
153 12.9 10.5 15.7



#51
3-Nitroaniline
Concen: 46.369 ng/ul
RT: 14.434 min Scan# 1929
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:138 Resp: 249248
Ion Ratio Lower Upper
138 100
108 11.3 9.4 14.2
92 125.6 102.7 154.1









#52
Acenaphthene
Concen: 41.230 ng/ul
RT: 14.575 min Scan# 1953
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:153 Resp: 978749
Ion Ratio Lower Upper
153 100
152 48.4 39.2 58.8
154 88.0 74.6 111.8



#53
2,4-Dinitrophenol
Concen: 48.869 ng/ul
RT: 14.639 min Scan# 1964
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:184 Resp: 114541
Ion Ratio Lower Upper
184 100
63 85.1 71.2 106.8
154 75.5 58.8 88.2









#55
4-Nitrophenol
Concen: 47.654 ng/ul
RT: 14.739 min Scan# 1981
Delta R.T. 0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:109 Resp: 208533
Ion Ratio Lower Upper
109 100
139 112.9 94.6 142.0
65 130.2 108.5 162.7



#56
Dibenzofuran
Concen: 41.511 ng/ul
RT: 14.910 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:168 Resp: 1301272
Ion Ratio Lower Upper
168 100
139 41.7 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 53.872 ng/ul
RT: 14.892 min Scan# 2007
Delta R.T. 0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:165 Resp: 350298
Ion Ratio Lower Upper
165 100
63 52.3 46.7 70.1
182 2.7 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 52.398 ng/ul
RT: 15.139 min Scan# 2049
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:232 Resp: 241780
Ion Ratio Lower Upper
232 100
131 57.6 46.9 70.3
130 3.1 2.6 4.0
166 36.2 29.4 44.2









#59
 Diethylphthalate
 Concen: 46.493 ng/ul
 RT: 15.351 min Scan# 2085
 Delta R.T. -0.000 min
 Lab File: BN029823.D
 Acq: 16 Feb 2024 11:34

Instrument :
 BNA_N
 ClientSampleId :
 SLCS982

Tgt Ion:149 Resp: 1232662




Ion	Ratio	Lower	Upper
149	100		
177	19.3	15.8	23.6
150	13.0	9.6	14.4




#61
 Fluorene
 Concen: 44.195 ng/ul
 RT: 15.563 min Scan# 2121
 Delta R.T. -0.006 min
 Lab File: BN029823.D
 Acq: 16 Feb 2024 11:34

Tgt Ion:166 Resp: 1095072

Ion	Ratio	Lower	Upper
166	100		
165	96.7	77.4	116.0
167	13.4	10.9	16.3









#62
4-Chlorophenyl-phenylether
Concen: 44.833 ng/ul
RT: 15.563 min Scan# 2121
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:204 Resp: 496867
Ion Ratio Lower Upper
204 100
206 32.9 27.5 41.3
141 71.2 60.5 90.7



#63
4-Nitroaniline
Concen: 56.572 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. 0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:138 Resp: 274457
Ion Ratio Lower Upper
138 100
92 57.3 47.4 71.2
108 102.1 70.9 106.3









#66
4,6-Dinitro-2-methylphenol
Concen: 48.000 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:198 Resp: 184821
Ion Ratio Lower Upper
198 100
182 4.8 3.9 5.9
77 33.8 29.8 44.8



#67
N-Nitrosodiphenylamine
Concen: 42.151 ng/ul
RT: 15.781 min Scan# 2158
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:169 Resp: 941503
Ion Ratio Lower Upper
169 100
168 66.0 54.5 81.7
167 36.1 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 45.688 ng/ul
RT: 16.457 min Scan# 2273
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:248 Resp: 309968
Ion Ratio Lower Upper
248 100
250 94.7 85.6 128.4
141 98.8 84.6 126.8



#69
Hexachlorobenzene
Concen: 44.375 ng/ul
RT: 16.557 min Scan# 2290
Delta R.T. -0.006 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:284 Resp: 339340
Ion Ratio Lower Upper
284 100
142 49.0 39.2 58.8
249 31.3 25.7 38.5









#70
 Atrazine
 Concen: 28.884 ng/ul
 RT: 16.739 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN029823.D
 Acq: 16 Feb 2024 11:34




Instrument :
 BNA_N
 ClientSampleId :
 SLCS982


Tgt Ion:200 Resp: 201566
 Ion Ratio Lower Upper
 200 100
 173 28.3 22.0 33.0
 215 46.2 36.1 54.1



#71
 Pentachlorophenol
 Concen: 46.242 ng/ul
 RT: 16.910 min Scan# 2350
 Delta R.T. -0.000 min
 Lab File: BN029823.D
 Acq: 16 Feb 2024 11:34

Tgt Ion:266 Resp: 205357
 Ion Ratio Lower Upper
 266 100
 264 65.8 45.8 68.6
 268 68.4 51.6 77.4





#72

Phenanthrene

Concen: 41.401 ng/u1

RT: 17.310 min Scan# 2417

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Instrument :

BNA_N

ClientSampleId :

SLCS982





Tgt Ion:178 Resp: 1716600

Ion Ratio Lower Upper

178 100

179 15.3 12.5 18.7

176 19.3 15.7 23.5



#74

Anthracene

Concen: 41.532 ng/u1

RT: 17.398 min Scan# 2433

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34




Tgt Ion:178 Resp: 1728217


Ion Ratio Lower Upper

178 100

179 15.9 11.6 17.4

176 19.0 14.2 21.2









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 39.306 ng/uL
 RT: 13.304 min Scan# 1737
 Delta R.T. -0.006 min
 Lab File: BN029823.D
 Acq: 16 Feb 2024 11:34




Instrument :
 BNA_N
 ClientSampleId :
 SLCS982


Tgt Ion:	216	Resp:	413435
Ion Ratio	Lower	Upper	
216	100		
214	77.9	60.1	90.1
218	47.6	38.5	57.7



#76
 Pentachlorobenzene
 Concen: 40.648 ng/uL
 RT: 14.828 min Scan# 1996
 Delta R.T. -0.000 min
 Lab File: BN029823.D
 Acq: 16 Feb 2024 11:34

Tgt Ion:	250	Resp:	382829
Ion Ratio	Lower	Upper	
250	100		
248	66.6	49.8	74.8
252	71.0	50.5	75.7









#77
 Carbazole
 Concen: 42.656 ng/ul
 RT: 17.675 min Scan# 24
 Delta R.T. -0.000 min
 Lab File: BN029823.D
 Acq: 16 Feb 2024 11:34

Instrument :
 BNA_N
 ClientSampleId :
 SLCS982

Tgt Ion:167 Resp: 1621220




Ion	Ratio	Lower	Upper
167	100		
166	21.9	16.6	25.0
139	14.5	11.4	17.0




#78
 Di-n-butylphthalate
 Concen: 47.713 ng/ul
 RT: 18.251 min Scan# 2578
 Delta R.T. -0.000 min
 Lab File: BN029823.D
 Acq: 16 Feb 2024 11:34

Tgt Ion:149 Resp: 2180704

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.0	10.4
104	6.0	4.7	7.1









#80
Fluoranthene
Concen: 43.366 ng/ul
RT: 19.327 min Scan# 2761
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34




Instrument :
BNA_N
ClientSampleId :
SLCS982


Tgt Ion:202 Resp: 1965395
Ion Ratio Lower Upper
202 100
101 13.4 11.1 16.7
100 9.9 8.9 13.3



#82
Pyrene
Concen: 42.797 ng/ul
RT: 19.692 min Scan# 2823
Delta R.T. -0.000 min
Lab File: BN029823.D
Acq: 16 Feb 2024 11:34

Tgt Ion:202 Resp: 2051885
Ion Ratio Lower Upper
202 100
101 14.9 14.2 21.4
100 12.5 11.4 17.2





#83

Butylbenzylphthalate

Concen: 52.492 ng/ul

RT: 20.616 min Scan# 2980

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Instrument :

BNA_N

ClientSampleId :

SLCS982





Tgt Ion:149 Resp: 1025688

Ion Ratio Lower Upper

149 100

91 78.9 70.0 105.0

206 18.7 14.1 21.1



#84

3,3'-Dichlorobenzidine

Concen: 46.869 ng/ul

RT: 21.392 min Scan# 3112

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34




Tgt Ion:252 Resp: 564391


Ion Ratio Lower Upper

252 100

254 66.1 49.8 74.6

126 13.2 12.2 18.4





#85

Benzo(a)anthracene

Concen: 43.608 ng/ul

RT: 21.451 min Scan# 3122

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Instrument :

BNA_N

ClientSampleId :

SLCS982





Tgt Ion:228 Resp: 1991610

Ion Ratio Lower Upper

228 100

229 18.3 15.0 22.6

226 26.6 21.0 31.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 51.734 ng/ul

RT: 21.398 min Scan# 3113

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34




Tgt Ion:149 Resp: 1563656


Ion Ratio Lower Upper

149 100

167 25.8 20.6 30.8

279 3.2 2.7 4.1





#87

Chrysene

Concen: 44.244 ng/ul

RT: 21.504 min Scan# 3131

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Instrument :

BNA_N

ClientSampleId :

SLCS982





Tgt Ion:228 Resp: 1893665

Ion Ratio Lower Upper

228 100

226 29.5 24.6 37.0

229 19.4 16.2 24.4



#89

Di-n-octyl phthalate

Concen: 49.774 ng/ul




RT: 22.327 min Scan# 3271


Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Tgt Ion:149 Resp: 2753215





#90

Benzo(b)fluoranthene

Concen: 43.102 ng/ul

RT: 23.115 min Scan# 3405

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Instrument :

BNA_N

ClientSampleId :

SLCS982





Tgt Ion:252 Resp: 1937350

Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 12.2 11.1 16.7



#91

Benzo(k)fluoranthene

Concen: 43.179 ng/ul

RT: 23.163 min Scan# 3413

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34




Tgt Ion:252 Resp: 1986471


Ion Ratio Lower Upper

252 100

253 21.1 18.2 27.4

125 11.9 11.2 16.8





#93

Benzo(a)pyrene

Concen: 42.396 ng/ul

RT: 23.733 min Scan# 3510

Delta R.T. 0.006 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Instrument :

BNA_N

ClientSampleId :

SLCS982





Tgt Ion:252 Resp: 1827447

Ion Ratio Lower Upper

252 100

253 21.2 18.8 28.2

125 13.8 12.2 18.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 43.073 ng/ul

RT: 26.292 min Scan# 3945

Delta R.T. 0.012 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34




Tgt Ion:276 Resp: 2260856


Ion Ratio Lower Upper

276 100

138 27.3 21.7 32.5

277 23.2 17.8 26.8





#95

Dibenzo(a,h)anthracene

Concen: 42.890 ng/ul

RT: 26.315 min Scan# 3949

Delta R.T. -0.000 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Instrument :

BNA_N

ClientSampleId :

SLCS982





Tgt Ion:278 Resp: 1837921

Ion Ratio Lower Upper

278 100

139 18.0 17.8 26.6

279 23.5 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 42.195 ng/ul

RT: 27.039 min Scan# 4072

Delta R.T. 0.006 min

Lab File: BN029823.D

Acq: 16 Feb 2024 11:34

Tgt Ion:276 Resp: 1852422

Ion Ratio Lower Upper

276 100

138 26.6 23.9 35.9

277 25.2 18.6 28.0

