

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021424\
 Data File : BN029757.D
 Acq On : 14 Feb 2024 07:04
 Operator : MA/JU
 Sample : PB158805BL
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK805

Quant Time: Feb 14 07:35:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 13 22:50:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	191312	20.000	ng/u1	0.00
20) Naphthalene-d8	10.669	136	787699	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.510	164	416499	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.263	188	823815	20.000	ng/u1	0.00
79) Chrysene-d12	21.463	240	661528	20.000	ng/u1	0.00
88) Perylene-d12	23.833	264	782963	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.322	96	49058	8.434	ng/uL	0.00
4) Pyridine-d5	3.728	84	657676	42.399	ng/u1	0.00
7) Phenol-d5	7.040	99	779081	43.537	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	518546	43.420	ng/u1	0.00
11) 2-Chlorophenol-d4	7.399	132	601331	46.373	ng/u1	0.00
15) 4-Methylphenol-d8	8.581	113	595823	44.646	ng/u1	0.00
21) Nitrobenzene-d5	9.040	128	295242	48.293	ng/u1	0.00
24) 2-Nitrophenol-d4	9.757	143	302296	56.775	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	498510	47.434	ng/u1	0.00
31) 4-Chloroaniline-d4	10.816	131	798159	44.516	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	1469603	47.393	ng/u1	0.00
49) Acenaphthylene-d8	14.204	160	1738078	47.198	ng/u1	0.00
54) 4-Nitrophenol-d4	14.722	143	243185	41.320	ng/u1	0.00
60) Fluorene-d10	15.510	176	1247193	47.279	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.634	200	171603	42.750	ng/u1	0.00
73) Anthracene-d10	17.363	188	1822042	47.192	ng/u1	0.00
81) Pyrene-d10	19.663	212	1935743	48.761	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.686	264	1919495	48.994	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021424\
Data File : BN029757.D
Acq On : 14 Feb 2024 07:04
Operator : MA/JU
Sample : PB158805BL
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK805

Quant Time: Feb 14 07:35:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 13 22:50:45 2024
Response via : Initial Calibration

