

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021424\
 Data File : BN029824.D
 Acq On : 16 Feb 2024 12:10
 Operator : MA/JU
 Sample : PB158885TB 10X
 Misc :
 ALS Vial : 105 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLEB885

Quant Time: Feb 17 04:57:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 13 22:50:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.863	152	148501	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	603508	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.510	164	330592	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.263	188	659045	20.000	ng/u1	0.00
79) Chrysene-d12	21.463	240	590099	20.000	ng/u1	0.00
88) Perylene-d12	23.827	264	663291	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.322	96	23692	5.248	ng/uL	0.00
4) Pyridine-d5	3.728	84	309202	25.680	ng/u1	0.00
7) Phenol-d5	7.034	99	378630	27.259	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	248516	26.808	ng/u1	0.00
11) 2-Chlorophenol-d4	7.393	132	302074	30.011	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	292593	28.245	ng/u1	-0.01
21) Nitrobenzene-d5	9.034	128	144620	30.875	ng/u1	0.00
24) 2-Nitrophenol-d4	9.751	143	145169	35.586	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	254726	31.635	ng/u1	0.00
31) 4-Chloroaniline-d4	10.810	131	397139	28.910	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	760814	30.911	ng/u1	0.00
49) Acenaphthylene-d8	14.204	160	880851	30.136	ng/u1	0.00
54) 4-Nitrophenol-d4	14.722	143	109750	23.494	ng/u1	0.00
60) Fluorene-d10	15.504	176	620142	29.617	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.628	200	73656	22.937	ng/u1	0.00
73) Anthracene-d10	17.363	188	945799	30.622	ng/u1	0.00
81) Pyrene-d10	19.663	212	1107557	31.276	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.680	264	1078764	32.502	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021424\
Data File : BN029824.D
Acq On : 16 Feb 2024 12:10
Operator : MA/JU
Sample : PB158885TB 10X
Misc :
ALS Vial : 105 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLEB885

Quant Time: Feb 17 04:57:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 13 22:50:45 2024
Response via : Initial Calibration

