

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021521\
 Data File : BN013621.D
 Acq On : 15 Feb 2021 13:22
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 15 14:06:06 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN020821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 08 15:13:18 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	113	0.00
2	1,4-Dioxane	0.400	0.379	5.3	116	0.00
3	n-Nitrosodimethylamine	0.400	0.398	0.5	118	0.00
4 S	2-Fluorophenol	0.400	0.425	-6.2	108	0.00
5 S	Phenol-d6	0.400	0.426	-6.5	109	0.00
6	bis(2-Chloroethyl)ether	0.400	0.409	-2.2	117	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	120	0.00
8 S	Nitrobenzene-d5	0.400	0.374	6.5	113	-0.01
9	Naphthalene	0.400	0.388	3.0	118	-0.01
10	Hexachlorobutadiene	0.400	0.393	1.8	118	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	118	0.00
12	2-Methylnaphthalene	0.400	0.391	2.3	120	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	122	0.00
14 S	2,4,6-Tribromophenol	0.400	0.349	12.8	114	0.00
15 S	2-Fluorobiphenyl	0.400	0.387	3.3	112	0.00
16	Acenaphthylene	0.400	0.362	9.5	121	0.00
17	Acenaphthene	0.400	0.389	2.8	123	0.00
18	Fluorene	0.400	0.386	3.5	122	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	125	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.390	2.5	127	0.00
21	4-Bromophenyl-phenylether	0.400	0.375	6.3	119	-0.01
22	Hexachlorobenzene	0.400	0.380	5.0	121	-0.01
23	Atrazine	0.400	0.346	13.5	122	0.00
24	Pentachlorophenol	0.400	0.475	-18.7	146	0.00
25	Phenanthrene	0.400	0.381	4.8	123	0.00
26	Anthracene	0.400	0.370	7.5	126	0.00
27 SURR	Fluoranthene-d10	0.400	0.381	4.8	126	0.00
28	Fluoranthene	0.400	0.379	5.3	126	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	126	0.00
30	Pyrene	0.400	0.405	-1.3	127	0.00
31 S	Terphenyl-d14	0.400	0.405	-1.3	120	0.00
32	Benzo(a)anthracene	0.400	0.373	6.8	124	0.00
33	Chrysene	0.400	0.405	-1.3	131	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.362	9.5	124	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.407	-1.7	134	0.00
36 I	Perylene-d12	0.400	0.400	0.0	133	0.00
37	Benzo(b)fluoranthene	0.400	0.377	5.8	128	0.00
38	Benzo(k)fluoranthene	0.400	0.384	4.0	133	0.00
39 C	Benzo(a)pyrene	0.400	0.375	6.3	132	0.00
40	Dibenzo(a,h)anthracene	0.400	0.376	6.0	132	0.00
41	Benzo(g,h,i)perylene	0.400	0.376	6.0	127	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4