


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021621\
 Data File : BN013630.D
 Acq On : 16 Feb 2021 08:59
 Operator : CG/JU
 Sample : DFTPP21
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP21

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN020421.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Feb 16 16:31:21 2021



AutoFind: Scans 2433, 2434, 2435; Background Corrected with Scan 2426

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.8	175237	PASS
68	69	0.00	2	1.9	3316	PASS
69	198	0.00	100	34.6	174360	PASS
70	69	0.00	2	0.6	1059	PASS
127	198	10	80	47.4	238784	PASS
197	198	0.00	2	0.3	1686	PASS
198	198	100	100	100.0	503680	PASS
199	198	5	9	6.7	33554	PASS
275	198	10	60	25.8	130186	PASS
365	198	1	100	3.2	15902	PASS
441	198	0.01	100	13.7	68978	PASS
442	198	50	100	91.7	461781	PASS
443	442	15	24	18.7	86565	PASS

m/z	Abundance
39.90	2212.0
41.00	363.0
42.90	439.0
44.00	2485.0
48.90	1028.0
50.90	536.0
73.10	1143.0
83.80	778.0
85.90	471.0
133.00	387.0
135.10	302.0
147.10	576.0
190.90	515.0
193.00	463.0
207.00	2948.0
208.10	1372.0
209.00	574.0
280.90	624.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

m/z	Abundance
36.00	415.0
37.00	952.0
38.00	2235.0
39.00	15382.0
39.90	2593.0
41.00	941.0
43.00	364.0
43.90	2437.0
44.90	336.0
47.90	366.0
49.00	2351.0
50.00	52472.0
51.00	216512.0
52.00	11047.0
52.90	574.0
55.00	1481.0
56.00	6442.0
57.00	15341.0
57.90	652.0
60.00	401.0
61.00	2837.0
62.00	2961.0
63.00	8390.0
64.00	957.0
65.00	4035.0
65.90	341.0
67.00	457.0
68.00	3935.0
69.00	210816.0
70.00	1263.0
71.00	375.0
73.00	2168.0
74.00	20776.0
75.00	34368.0
76.10	10408.0
77.00	244224.0
78.00	16760.0
79.00	15463.0
80.00	12249.0
81.00	16270.0
81.90	3933.0
83.00	4304.0
84.00	1205.0
85.00	3160.0
86.00	5448.0
87.00	2468.0
87.90	986.0
89.00	333.0
90.90	4313.0
92.00	4979.0
93.00	28424.0
94.00	2201.0
95.00	851.0
96.10	1492.0
97.00	496.0
98.00	21728.0
99.00	18000.0
100.00	1629.0
101.00	9471.0
102.00	529.0
102.90	3790.0
104.00	6489.0
105.00	6097.0
106.00	1778.0
107.00	74584.0
108.00	11431.0
109.00	1919.0
110.00	142016.0
111.00	20272.0
111.90	2388.0
112.90	1083.0
114.10	305.0
115.10	348.0
116.00	4187.0
117.00	58552.0
118.00	5448.0
118.90	802.0
120.10	938.0
120.90	304.0
122.00	4940.0
123.00	7053.0
124.00	3727.0
125.00	2848.0
126.10	1160.0
127.00	277504.0
128.00	21376.0
129.00	103864.0
130.00	9318.0
130.90	1737.0
132.00	998.0
132.90	968.0
134.00	2571.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

135.00	8413.0
136.00	3478.0
137.00	5511.0
137.80	1194.0
138.90	666.0
139.90	1223.0
141.00	13544.0
142.00	4061.0
142.90	2690.0
143.90	967.0
144.90	946.0
145.90	2456.0
146.90	7048.0
148.00	14657.0
149.00	3563.0
150.00	892.0
151.10	1460.0
151.80	937.0
152.20	614.0
153.00	3634.0
153.90	3239.0
155.00	6833.0
156.00	11739.0
157.00	2233.0
158.00	2208.0
159.00	2012.0
160.00	4589.0
161.00	6224.0
162.00	2180.0
163.00	775.0
164.10	743.0
165.00	4688.0
166.00	3859.0
167.00	24224.0
168.00	18368.0
169.00	2671.0
170.10	719.0
171.00	1240.0
172.00	2285.0
173.00	3188.0
174.00	5498.0
175.00	10244.0
176.00	3261.0
177.00	5589.0
178.00	1736.0
179.00	18496.0
180.00	15375.0
181.00	6263.0
182.00	834.0
182.90	453.0
183.00	438.0
184.10	1334.0
185.00	8430.0
186.00	71352.0
187.00	21256.0
188.00	2903.0
189.00	4660.0
189.90	776.0
191.00	2530.0
192.00	6433.0
193.00	7173.0
194.00	1469.0
195.00	1061.0
196.00	14726.0
197.10	2972.0
198.00	529600.0
198.90	37456.0
200.00	2975.0
200.90	1058.0
201.50	2448.0
203.00	3548.0
204.00	18360.0
205.00	30576.0
206.00	130960.0
207.00	22792.0
208.00	5317.0
209.00	2323.0
210.20	2160.0
211.00	4232.0
211.90	1129.0
212.90	376.0
214.90	1660.0
216.00	3075.0
217.00	31944.0
218.00	4138.0
219.10	484.0
221.00	22824.0
223.00	8158.0
224.00	71592.0
225.00	17744.0
226.00	2736.0
227.00	27448.0
228.00	4317.0
228.90	5504.0
230.00	644.0
231.00	2425.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

231.80	398.0
233.00	648.0
234.00	1768.0
235.00	1826.0
236.10	1058.0
237.00	4178.0
238.10	573.0
239.00	1013.0
240.00	891.0
240.90	1509.0
242.00	4117.0
243.00	3604.0
244.00	56528.0
245.00	7431.0
246.00	10236.0
247.00	2455.0
247.90	478.0
249.00	2242.0
249.90	400.0
252.00	582.0
253.00	1069.0
255.00	293824.0
256.00	43560.0
257.00	3552.0
258.00	15659.0
259.00	2656.0
260.00	504.0
261.10	374.0
264.10	496.0
265.10	5948.0
265.90	862.0
270.10	340.0
271.00	854.0
272.00	624.0
273.00	7495.0
274.00	21720.0
275.00	129520.0
276.00	18912.0
277.00	10866.0
278.00	1927.0
278.90	403.0
281.00	558.0
282.90	1200.0
284.00	927.0
285.00	1844.0
285.90	457.0
289.00	576.0
289.90	325.0
290.90	509.0
291.90	404.0
292.90	2465.0
294.00	976.0
296.00	29640.0
297.00	4595.0
298.00	457.0
300.90	480.0
301.90	545.0
303.00	4484.0
303.90	937.0
304.90	330.0
307.90	551.0
308.90	377.0
310.00	573.0
314.00	1735.0
315.00	3373.0
316.00	1955.0
316.90	563.0
320.90	1195.0
322.00	442.0
323.10	10361.0
324.00	1841.0
327.00	1993.0
328.00	1163.0
331.90	750.0
332.90	930.0
334.00	7210.0
335.00	1626.0
335.90	317.0
341.00	1532.0
341.90	403.0
345.90	2613.0
347.00	651.0
352.00	3494.0
353.00	2360.0
354.00	3444.0
354.90	925.0
365.00	14131.0
366.00	1962.0
370.90	967.0
372.00	5998.0
373.00	1426.0
383.00	1614.0
383.80	500.0
385.00	355.0
389.90	739.0
390.80	565.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

401.00	756.0
402.00	2214.0
403.00	3118.0
403.90	1400.0
405.00	434.0
420.90	2747.0
422.00	3001.0
423.00	20880.0
424.00	8256.0
425.00	1224.0
441.00	56576.0
442.00	363776.0
443.00	69464.0
444.10	7052.0
445.00	534.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

m/z	Abundance
35.90	436.0
36.90	1081.0
38.00	2408.0
39.00	15376.0
40.00	2591.0
41.10	596.0
43.00	462.0
44.00	2429.0
44.90	491.0
49.00	2069.0
50.00	52120.0
51.00	204416.0
52.00	10845.0
53.00	509.0
55.00	1159.0
56.00	6691.0
57.00	14112.0
57.90	646.0
61.00	2436.0
62.00	2665.0
63.00	7807.0
64.00	1068.0
65.00	4235.0
66.90	512.0
68.10	4146.0
69.00	204416.0
69.90	1226.0
73.00	2673.0
74.00	20752.0
75.00	33336.0
76.00	11021.0
77.00	241344.0
78.00	16025.0
79.00	14785.0
80.00	12929.0
81.00	16383.0
82.00	4036.0
83.00	3828.0
83.90	1213.0
85.00	2646.0
86.00	5265.0
87.10	1675.0
88.00	1130.0
89.00	508.0
91.00	3979.0
92.00	4071.0
93.00	28944.0
94.00	1913.0
95.10	629.0
96.00	1389.0
96.80	438.0
98.00	21400.0
99.00	17880.0
100.00	1324.0
101.00	9331.0
101.90	660.0
103.00	2783.0
104.00	7198.0
105.00	6239.0
106.10	1812.0
107.00	73248.0
108.00	11560.0
109.00	2822.0
110.00	142656.0
111.00	20384.0
112.00	2691.0
113.00	639.0
114.90	319.0
116.00	4532.0
117.00	60048.0
118.00	4820.0
119.00	1006.0
120.00	893.0
121.00	375.0
122.00	4912.0
123.00	8226.0
123.90	3795.0
125.00	3175.0
126.20	1704.0
127.00	279616.0
128.00	21488.0
129.00	104728.0
130.00	8749.0
131.00	1627.0
131.90	1365.0
133.00	782.0
133.90	3105.0
135.00	8792.0
136.00	3744.0
137.00	5138.0
138.00	907.0
139.00	621.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

140.00	1378.0
141.00	14746.0
142.00	5088.0
142.90	3270.0
143.90	756.0
145.00	810.0
146.00	2599.0
147.00	7269.0
148.00	16115.0
149.00	3979.0
149.90	1238.0
151.10	1754.0
151.70	1289.0
152.90	4945.0
154.00	3602.0
155.00	8309.0
156.00	11220.0
157.00	2318.0
158.00	2790.0
158.90	1854.0
160.00	4194.0
161.00	6167.0
161.90	2231.0
163.00	605.0
164.10	1270.0
165.00	5162.0
166.00	4323.0
167.00	28264.0
168.00	18784.0
169.00	3049.0
170.00	808.0
171.00	1205.0
172.00	2694.0
173.00	3051.0
174.00	5480.0
175.00	11874.0
175.90	2777.0
177.00	6048.0
177.90	1685.0
178.90	21136.0
180.00	16608.0
181.00	6616.0
182.00	1331.0
183.00	706.0
184.00	1579.0
185.00	9371.0
186.00	79440.0
187.00	22256.0
188.00	2597.0
189.00	5268.0
189.90	934.0
190.90	2501.0
192.00	6847.0
193.00	7712.0
194.00	1964.0
194.90	773.0
196.00	16384.0
198.00	618496.0
199.00	37848.0
200.00	3114.0
201.40	2921.0
203.00	5029.0
204.00	18392.0
205.00	35040.0
206.00	146560.0
207.00	22744.0
208.00	5541.0
209.00	1881.0
210.20	2401.0
211.00	5805.0
212.90	476.0
214.90	1461.0
216.00	3286.0
217.00	37720.0
218.00	5481.0
219.10	435.0
221.00	25576.0
221.80	5500.0
223.00	8149.0
224.00	79000.0
225.00	22664.0
226.00	2481.0
227.00	33712.0
228.00	5554.0
229.00	6545.0
230.00	863.0
231.00	2569.0
232.00	636.0
232.80	518.0
234.00	1899.0
235.00	2162.0
236.00	1647.0
237.00	3941.0
237.90	483.0
239.00	1133.0
240.00	1109.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

240.90	2022.0
242.00	4871.0
243.10	5123.0
244.00	67376.0
245.00	8642.0
246.00	12516.0
247.00	1988.0
247.80	463.0
248.90	2118.0
249.90	416.0
251.10	586.0
252.00	583.0
253.00	1199.0
255.00	341888.0
256.00	50400.0
257.00	4300.0
258.00	16800.0
259.00	2806.0
259.90	487.0
260.10	484.0
261.00	443.0
263.80	439.0
265.00	7568.0
266.00	1254.0
270.00	426.0
271.00	763.0
271.90	849.0
273.00	8927.0
274.00	27000.0
275.00	159360.0
276.00	21312.0
277.00	12189.0
278.00	2483.0
279.00	542.0
280.90	429.0
281.90	569.0
283.00	1496.0
284.00	872.0
285.00	2009.0
286.00	350.0
288.90	575.0
290.10	452.0
290.90	381.0
291.90	621.0
293.00	2932.0
294.00	1026.0
296.00	41560.0
297.00	5960.0
298.00	348.0
300.90	531.0
302.00	692.0
303.00	4861.0
304.00	1237.0
307.90	563.0
310.10	575.0
312.90	391.0
314.00	2302.0
315.00	4523.0
316.00	3014.0
316.90	667.0
321.10	1554.0
322.00	791.0
323.00	13698.0
324.00	2474.0
327.00	2699.0
328.10	1641.0
331.90	1286.0
333.00	1488.0
334.00	9514.0
335.00	2434.0
335.80	321.0
340.10	314.0
341.10	1478.0
342.10	478.0
345.90	3012.0
347.00	892.0
351.00	355.0
352.00	4186.0
352.90	3041.0
354.10	5286.0
355.10	824.0
365.00	18488.0
366.00	3304.0
369.90	456.0
371.00	1576.0
372.00	8931.0
373.00	1984.0
374.00	326.0
382.90	2238.0
383.90	924.0
389.90	1000.0
391.00	821.0
392.00	486.0
401.00	451.0
402.00	3437.0
403.00	4917.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

404.00	1787.0
405.00	323.0
421.00	4467.0
422.00	4493.0
423.00	30008.0
424.10	8426.0
425.00	1357.0
440.20	758.0
441.00	85552.0
442.00	551104.0
443.00	107120.0
444.00	9711.0
444.80	513.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

m/z	Abundance
37.00	453.0
38.00	1250.0
39.00	8526.0
39.90	1910.0
41.00	638.0
43.10	554.0
44.00	2624.0
45.00	402.0
49.00	1599.0
50.00	26360.0
51.00	106392.0
52.00	4657.0
53.00	440.0
55.00	1021.0
55.90	3732.0
57.00	7351.0
58.00	597.0
61.00	1166.0
61.90	1656.0
63.00	4265.0
64.00	554.0
65.00	1661.0
67.00	320.0
68.00	1867.0
69.00	107848.0
69.90	688.0
73.10	2063.0
74.00	11635.0
75.00	17800.0
76.00	6589.0
77.00	128736.0
78.00	8616.0
79.00	8399.0
80.00	6323.0
80.90	9029.0
82.00	2651.0
83.00	2283.0
83.90	688.0
84.90	1388.0
86.00	2593.0
87.00	1356.0
88.10	509.0
88.90	386.0
91.00	2429.0
92.00	3062.0
93.00	14409.0
93.90	1149.0
95.00	432.0
96.10	860.0
97.00	446.0
98.00	12423.0
99.00	10611.0
100.00	915.0
101.00	5660.0
101.90	501.0
103.00	2123.0
104.00	3571.0
105.00	3646.0
106.10	1321.0
107.00	42232.0
108.00	6491.0
109.10	1418.0
110.00	81464.0
111.00	11371.0
112.00	1623.0
112.90	557.0
114.90	516.0
115.90	2569.0
117.00	35824.0
117.90	2608.0
119.00	466.0
120.10	892.0
122.00	3212.0
123.00	5101.0
123.90	2148.0
124.90	1606.0
127.00	159232.0
128.00	11669.0
129.00	58016.0
130.00	4749.0
130.90	1338.0
132.10	591.0
133.00	606.0
134.00	2039.0
135.00	5434.0
136.00	2429.0
137.00	2770.0
138.00	542.0
138.80	424.0
139.90	782.0
141.00	7913.0
141.90	3036.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

142.90	2245.0
143.90	530.0
144.90	531.0
146.00	1380.0
146.90	4433.0
148.00	9037.0
149.00	2467.0
150.00	625.0
151.10	1413.0
151.90	685.0
152.90	2791.0
153.90	2005.0
155.00	4503.0
156.00	6839.0
157.00	1545.0
157.90	1440.0
158.90	1105.0
159.90	2305.0
161.00	4269.0
161.90	1286.0
163.00	788.0
163.80	479.0
164.90	3397.0
166.00	2421.0
167.00	15515.0
168.00	10613.0
168.90	1463.0
170.10	749.0
171.10	846.0
171.90	1331.0
173.00	2000.0
174.00	3666.0
175.00	6747.0
176.10	1623.0
176.90	2999.0
177.90	1242.0
178.90	13309.0
180.00	9362.0
181.00	4663.0
182.00	739.0
183.10	476.0
183.80	845.0
185.00	5692.0
186.00	48096.0
187.00	14011.0
188.00	1436.0
189.00	3124.0
189.80	484.0
191.00	1457.0
191.90	4089.0
193.10	4565.0
194.00	1006.0
194.90	728.0
196.00	11044.0
196.80	1873.0
197.10	2088.0
198.00	362944.0
199.00	25360.0
200.00	2015.0
201.30	1045.0
201.80	1182.0
203.00	2549.0
204.00	12166.0
205.00	20896.0
206.00	93008.0
207.00	14861.0
208.00	3468.0
209.10	1664.0
210.00	1439.0
211.00	4743.0
212.00	511.0
212.80	308.0
214.90	1000.0
216.00	2126.0
216.90	23800.0
217.90	2840.0
219.10	311.0
221.00	16348.0
223.00	5409.0
224.00	50336.0
225.00	12795.0
226.10	1550.0
227.00	19608.0
228.00	3506.0
229.00	4349.0
230.00	683.0
231.00	2411.0
232.00	362.0
233.00	359.0
234.00	1572.0
235.00	1361.0
236.00	1007.0
237.00	2920.0
237.90	424.0
239.00	837.0
240.00	883.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

241.00	992.0
242.00	3214.0
243.00	3576.0
244.00	45008.0
245.00	5737.0
246.00	8109.0
247.00	1885.0
248.00	400.0
248.90	1361.0
250.00	446.0
252.00	443.0
253.00	1011.0
254.00	1791.0
255.00	232960.0
256.00	32704.0
257.00	3023.0
257.90	11808.0
258.90	1725.0
260.10	374.0
261.00	393.0
263.90	352.0
265.00	5446.0
266.00	606.0
270.00	347.0
270.90	431.0
271.90	624.0
273.00	6695.0
274.00	17056.0
275.00	101680.0
276.00	13488.0
277.00	7952.0
278.00	1769.0
279.00	560.0
280.90	565.0
281.90	349.0
283.00	1306.0
283.90	497.0
285.00	1936.0
289.00	402.0
290.90	447.0
292.00	342.0
293.00	1955.0
294.00	449.0
295.00	650.0
296.00	27784.0
297.00	3830.0
301.10	352.0
301.90	446.0
303.00	3354.0
304.00	821.0
308.00	415.0
309.90	419.0
313.00	320.0
314.00	1416.0
315.00	3293.0
316.00	2282.0
317.00	347.0
321.00	958.0
321.90	383.0
323.00	9970.0
324.00	1483.0
324.80	309.0
326.00	440.0
327.00	2088.0
328.00	885.0
332.00	902.0
333.00	722.0
334.00	5910.0
335.10	1646.0
341.00	1174.0
342.00	487.0
345.90	2356.0
346.90	496.0
351.00	360.0
352.00	3593.0
353.00	1866.0
354.00	3266.0
355.10	716.0
365.00	15087.0
366.00	2020.0
370.90	621.0
372.10	5691.0
373.00	1555.0
383.10	1716.0
383.90	590.0
389.90	928.0
390.80	643.0
391.80	448.0
402.10	2355.0
403.00	3510.0
404.00	1283.0
404.90	509.0
420.90	3345.0
422.00	3976.0
423.00	23960.0
424.00	7591.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

425.00	756.0
441.00	64808.0
442.00	470464.0
443.00	83112.0
444.00	8054.0
444.90	658.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21