

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021622\
 Data File : BN018618.D
 Acq On : 16 Feb 2022 13:02
 Operator : CG/JU
 Sample : SSTD00531
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD005231

Quant Time: Feb 16 15:20:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 16 15:18:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	254621	20.000	ng/u1	0.00
20) Naphthalene-d8	10.722	136	1009235	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.545	164	631269	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.286	188	1239644	20.000	ng/u1	0.00
79) Chrysene-d12	21.462	240	1055917	20.000	ng/u1	0.00
88) Perylene-d12	23.856	264	1015085	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.287	96	12570	2.175	ng/uL	0.00
4) Pyridine-d5	3.710	84	78610	4.015	ng/u1	0.00
7) Phenol-d5	7.063	99	90286	4.375	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.234	67	60146	4.800	ng/u1	0.00
11) 2-Chlorophenol-d4	7.440	132	71590	4.369	ng/u1	0.00
15) 4-Methylphenol-d8	8.616	113	68142	4.283	ng/u1	0.00
21) Nitrobenzene-d5	9.069	128	31188	4.113	ng/u1	0.00
24) 2-Nitrophenol-d4	9.798	143	31982	3.907	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.340	165	69026	4.384	ng/u1	0.00
31) 4-Chloroaniline-d4	10.845	131	93680	4.548	ng/u1	0.00
46) Dimethylphthalate-d6	13.951	166	206166	4.565	ng/u1	0.00
49) Acenaphthylene-d8	14.239	160	266243	4.305	ng/u1	0.00
54) 4-Nitrophenol-d4	14.722	143	27297	3.588	ng/u1	0.00
60) Fluorene-d10	15.533	176	187816	4.918	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	23804	3.182	ng/u1	0.00
73) Anthracene-d10	17.386	188	283683	4.689	ng/u1	0.00
81) Pyrene-d10	19.674	212	292488	5.668	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.704	264	222102	4.065	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.322	88	12826	1.729	ng/uL	90
5) Pyridine	3.728	79	83448	4.342	ng/u1#	92
6) Benzaldehyde	7.040	77	59070	5.110	ng/u1	90
8) Phenol	7.093	94	98801	4.726	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.328	93	82060	4.762	ng/u1	96
12) 2-Chlorophenol	7.475	128	78765	4.662	ng/u1	93
13) 2-Methylphenol	8.351	108	66798	4.306	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.440	45	112875	4.885	ng/u1	99
16) Acetophenone	8.734	105	118847	4.973	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.716	70	54829	4.509	ng/u1	93
18) 4-Methylphenol	8.675	108	76811	4.578	ng/u1	96
19) Hexachloroethane	9.004	117	33334	4.708	ng/u1	91
22) Nitrobenzene	9.110	77	84527	4.414	ng/u1	96
23) Isophorone	9.640	82	145242	4.428	ng/u1#	94
25) 2-Nitrophenol	9.828	139	38131	4.221	ng/u1	95
26) 2,4-Dimethylphenol	9.887	107	87166	4.559	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.122	93	103652	4.701	ng/u1	99
29) 2,4-Dichlorophenol	10.363	162	71020	4.336	ng/u1	96
30) Naphthalene	10.769	128	261789	4.707	ng/u1	99
32) 4-Chloroaniline	10.875	127	100170	4.716	ng/u1	98
33) Hexachlorobutadiene	11.069	225	52844	4.605	ng/u1	92
34) Caprolactam	11.610	113	13846	3.236	ng/u1	81
35) 4-Chloro-3-methylphenol	11.992	107	70154	4.317	ng/u1	99
36) 2-Methylnaphthalene	12.381	142	175560	4.960	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021622\
 Data File : BN018618.D
 Acq On : 16 Feb 2022 13:02
 Operator : CG/JU
 Sample : SST00531
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SST005231


Quant Time: Feb 16 15:20:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 16 15:18:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.598	142	177661	4.751	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.751	216	96771	4.259	ng/ul	96
40) Hexachlorocyclopentadiene	12.734	237	55362	3.966	ng/ul	98
41) 2,4,6-Trichlorophenol	12.981	196	49820	3.970	ng/ul	100
42) 2,4,5-Trichlorophenol	13.051	196	55826	4.146	ng/ul	97
43) 1,1'-Biphenyl	13.386	154	242155	4.498	ng/ul	96
44) 2-Chloronaphthalene	13.428	162	183400	4.407	ng/ul	94
45) 2-Nitroaniline	13.622	65	36732	3.846	ng/ul	94
47) Dimethylphthalate	13.998	163	216839	4.857	ng/ul	99
48) 2,6-Dinitrotoluene	14.110	165	33143	4.066	ng/ul#	91
50) Acenaphthylene	14.269	152	285306	4.739	ng/ul	99
51) 3-Nitroaniline	14.439	138	34323	4.235	ng/ul	89
52) Acenaphthene	14.610	153	190677	4.898	ng/ul	100
53) 2,4-Dinitrophenol	14.645	184	13029	2.505	ng/ul	85
55) 4-Nitrophenol	14.733	109	25386	3.659	ng/ul	89
56) Dibenzofuran	14.945	168	272305	4.836	ng/ul	95
57) 2,4-Dinitrotoluene	14.898	165	49721	4.132	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.169	232	42820	4.269	ng/ul	94
59) Diethylphthalate	15.357	149	204370	4.501	ng/ul	98
61) Fluorene	15.592	166	218033	4.729	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.586	204	109882	4.621	ng/ul	93
63) 4-Nitroaniline	15.592	138	32485	4.953	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.657	198	25449	3.128	ng/ul#	92
67) N-Nitrosodiphenylamine	15.792	169	178174	4.105	ng/ul	97
68) 4-Bromophenyl-phenylether	16.480	248	61846	4.678	ng/ul	93
69) Hexachlorobenzene	16.598	284	72247	4.083	ng/ul#	88
70) Atrazine	16.739	200	53521	3.545	ng/ul	99
71) Pentachlorophenol	16.939	266	28110	2.785	ng/ul	95
72) Phenanthrene	17.327	178	337850	4.507	ng/ul	99
74) Anthracene	17.416	178	336950	4.550	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.351	216	103371	4.067	ng/ul	98
76) Pentachlorobenzene	14.869	250	92020	4.136	ng/ul	97
77) Carbazole	17.680	167	271982	4.202	ng/ul	98
78) Di-n-butylphthalate	18.257	149	276610	4.005	ng/ul	99
80) Fluoranthene	19.339	202	350914	5.706	ng/ul	99
82) Pyrene	19.704	202	372544	5.508	ng/ul	99
83) Butylbenzylphthalate	20.598	149	92111	5.512	ng/ul	91
84) 3,3'-Dichlorobenzidine	21.380	252	70963	3.802	ng/ul	99
85) Benzo(a)anthracene	21.451	228	322242	4.971	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.380	149	137765	3.576	ng/ul	99
87) Chrysene	21.498	228	337397	6.156	ng/ul	97
89) Di-n-octyl phthalate	22.298	149	212523	2.976	ng/ul	100
90) Benzo(b)fluoranthene	23.127	252	301221	4.114	ng/ul	98
91) Benzo(k)fluoranthene	23.174	252	297988	4.599	ng/ul	98
93) Benzo(a)pyrene	23.751	252	297022	4.296	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.339	276	320348	4.080	ng/ul	96
95) Dibenzo(a,h)anthracene	26.356	278	276254	4.153	ng/ul	97
96) Benzo(g,h,i)perylene	27.109	276	269728	4.029	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 16 15:20:40 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 16 15:18:48 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.729 ng/uL
RT: 3.322 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion: 88 Resp: 12826
Ion Ratio Lower Upper
88 100
43 32.2 21.9 32.9
58 78.4 56.3 84.5



#5
Pyridine
Concen: 4.342 ng/uL
RT: 3.728 min Scan# 126
Delta R.T. 0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion: 79 Resp: 83448
Ion Ratio Lower Upper
79 100
52 83.4 62.9 94.3
51 42.3 27.8 41.8#









#6
Benzaldehyde
Concen: 5.110 ng/ul
RT: 7.040 min Scan# 61
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion: 77 Resp: 59070
Ion Ratio Lower Upper
77 100
105 104.4 73.5 110.3
106 92.9 69.9 104.9



#8
Phenol
Concen: 4.726 ng/ul
RT: 7.093 min Scan# 698
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion: 94 Resp: 98801
Ion Ratio Lower Upper
94 100
65 27.5 21.8 32.6
66 37.0 30.2 45.4









#10
Bis(2-Chloroethyl)ether
Concen: 4.762 ng/ul
RT: 7.328 min Scan# 738
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion: 93 Resp: 82060
Ion Ratio Lower Upper
93 100
63 75.5 63.4 95.0
95 31.4 25.9 38.9



#12
2-Chlorophenol
Concen: 4.662 ng/ul
RT: 7.475 min Scan# 763
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion: 128 Resp: 78765
Ion Ratio Lower Upper
128 100
64 46.9 44.2 66.2
130 31.4 25.0 37.6









#13
2-Methylphenol
Concen: 4.306 ng/ul
RT: 8.351 min Scan# 912
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:108 Resp: 66798
Ion Ratio Lower Upper
108 100
107 92.7 69.5 104.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.885 ng/ul
RT: 8.440 min Scan# 927
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion: 45 Resp: 112875
Ion Ratio Lower Upper
45 100
77 14.8 11.9 17.9
79 10.2 8.6 12.8









#16
Acetophenone
Concen: 4.973 ng/ul
RT: 8.734 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:105 Resp: 118847
Ion Ratio Lower Upper
105 100
77 80.8 64.8 97.2
51 29.0 22.8 34.2



#17
N-Nitroso-di-n-propylamine
Concen: 4.509 ng/ul
RT: 8.716 min Scan# 974
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion: 70 Resp: 54829
Ion Ratio Lower Upper
70 100
42 58.1 42.4 63.6
101 9.8 6.6 10.0
130 19.8 13.6 20.4









#18
4-Methylphenol
Concen: 4.578 ng/ul
RT: 8.675 min Scan# 90
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:108 Resp: 76811
Ion Ratio Lower Upper
108 100
107 107.3 89.4 134.0



#19
Hexachloroethane
Concen: 4.708 ng/ul
RT: 9.004 min Scan# 1023
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:117 Resp: 33334
Ion Ratio Lower Upper
117 100
201 95.8 85.9 128.9
199 63.6 54.5 81.7





#22

Nitrobenzene

Concen: 4.414 ng/ul

RT: 9.110 min Scan# 1042

Delta R.T. -0.006 min

Lab File: BN018618.D

Acq: 16 Feb 2022 13:02

Instrument :

BNA_N

ClientSampleId :

SSTD005231





Tgt Ion: 77 Resp: 84527

Ion Ratio Lower Upper

77 100

123 42.8 32.0 48.0

65 13.8 10.7 16.1



#23

Isophorone

Concen: 4.428 ng/ul

RT: 9.640 min Scan# 1131

Delta R.T. -0.000 min

Lab File: BN018618.D

Acq: 16 Feb 2022 13:02




Tgt Ion: 82 Resp: 145242


Ion Ratio Lower Upper

82 100

95 7.6 5.7 8.5

138 18.3 11.9 17.9#





#25

2-Nitrophenol

Concen: 4.221 ng/ul

RT: 9.828 min Scan# 1163

Delta R.T. -0.000 min

Lab File: BN018618.D

Acq: 16 Feb 2022 13:02

Instrument :

BNA_N

ClientSampleId :

SSTD005231





Tgt Ion:139 Resp: 38131

Ion Ratio Lower Upper

139 100

65 53.5 47.7 71.5

109 33.7 27.0 40.4



#26

2,4-Dimethylphenol

Concen: 4.559 ng/ul

RT: 9.887 min Scan# 1173

Delta R.T. -0.000 min

Lab File: BN018618.D

Acq: 16 Feb 2022 13:02




Tgt Ion:107 Resp: 87166


Ion Ratio Lower Upper

107 100

121 49.8 38.9 58.3

122 84.4 68.6 103.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 4.701 ng/ul
 RT: 10.122 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BN018618.D
 Acq: 16 Feb 2022 13:02




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005231


Tgt Ion: 93 Resp: 103652
 Ion Ratio Lower Upper
 93 100
 95 31.1 25.7 38.5
 123 10.5 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 4.336 ng/ul
 RT: 10.363 min Scan# 1254
 Delta R.T. -0.000 min
 Lab File: BN018618.D
 Acq: 16 Feb 2022 13:02

Tgt Ion: 162 Resp: 71020
 Ion Ratio Lower Upper
 162 100
 164 67.3 51.8 77.6
 98 36.1 31.0 46.6









#30
Naphthalene
Concen: 4.707 ng/ul
RT: 10.769 min Scan# 1323
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:128 Resp: 261789
Ion Ratio Lower Upper
128 100
129 11.1 8.6 12.8
127 13.5 10.6 16.0



#32
4-Chloroaniline
Concen: 4.716 ng/ul
RT: 10.875 min Scan# 1341
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:127 Resp: 100170
Ion Ratio Lower Upper
127 100
129 32.7 25.0 37.6









#33
Hexachlorobutadiene
Concen: 4.605 ng/ul
RT: 11.069 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:225 Resp: 52844
Ion Ratio Lower Upper
225 100
223 65.9 49.7 74.5
227 70.7 49.5 74.3



#34
Caprolactam
Concen: 3.236 ng/ul
RT: 11.610 min Scan# 1466
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:113 Resp: 13846
Ion Ratio Lower Upper
113 100
55 151.8 147.4 221.2
56 118.0 107.6 161.4









#35
4-Chloro-3-methylphenol
Concen: 4.317 ng/ul
RT: 11.992 min Scan# 1531
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:107 Resp: 70154
Ion Ratio Lower Upper
107 100
144 25.5 19.4 29.0
142 78.7 62.3 93.5



#36
2-Methylnaphthalene
Concen: 4.960 ng/ul
RT: 12.381 min Scan# 1597
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:142 Resp: 175560
Ion Ratio Lower Upper
142 100
141 87.1 70.2 105.4









#37
1-Methylnaphthalene
Concen: 4.751 ng/ul
RT: 12.598 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:142 Resp: 177661
Ion Ratio Lower Upper
142 100
141 91.5 72.2 108.4
116 4.0 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.259 ng/ul
RT: 12.751 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:216 Resp: 96771
Ion Ratio Lower Upper
216 100
214 77.5 64.4 96.6
179 22.3 15.1 22.7
108 17.1 12.0 18.0









#40
Hexachlorocyclopentadiene
Concen: 3.966 ng/ul
RT: 12.734 min Scan# 1698
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:237 Resp: 55362
Ion Ratio Lower Upper
237 100
235 62.8 49.1 73.7
272 12.6 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 3.970 ng/ul
RT: 12.981 min Scan# 1699
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:196 Resp: 49820
Ion Ratio Lower Upper
196 100
198 93.1 74.5 111.7
200 30.0 24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 4.146 ng/ul
RT: 13.051 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:196 Resp: 55826
Ion Ratio Lower Upper
196 100
198 97.2 75.3 112.9
200 32.4 24.2 36.2



#43
1,1'-Biphenyl
Concen: 4.498 ng/ul
RT: 13.386 min Scan# 1768
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:154 Resp: 242155
Ion Ratio Lower Upper
154 100
153 38.6 32.4 48.6
76 12.6 12.0 18.0









#44
2-Chloronaphthalene
Concen: 4.407 ng/ul
RT: 13.428 min Scan# 1775
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:162 Resp: 183400
Ion Ratio Lower Upper
162 100
164 32.2 26.8 40.2
127 41.4 28.8 43.2



#45
2-Nitroaniline
Concen: 3.846 ng/ul
RT: 13.622 min Scan# 1808
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion: 65 Resp: 36732
Ion Ratio Lower Upper
65 100
92 69.3 53.0 79.6
138 99.0 73.8 110.8









#47
Dimethylphthalate
Concen: 4.857 ng/ul
RT: 13.998 min Scan# 1872
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:163 Resp: 216839
Ion Ratio Lower Upper
163 100
194 4.4 3.6 5.4
164 9.8 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 4.066 ng/ul
RT: 14.110 min Scan# 1891
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:165 Resp: 33143
Ion Ratio Lower Upper
165 100
89 68.1 49.3 73.9
121 25.9 16.8 25.2#









#50
Acenaphthylene
Concen: 4.739 ng/ul
RT: 14.269 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:152 Resp: 285306
Ion Ratio Lower Upper
152 100
151 19.5 16.2 24.2
153 13.4 10.6 15.8



#51
3-Nitroaniline
Concen: 4.235 ng/ul
RT: 14.439 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:138 Resp: 34323
Ion Ratio Lower Upper
138 100
108 12.2 10.6 15.8
92 120.7 107.9 161.9









#52
Acenaphthene
Concen: 4.898 ng/ul
RT: 14.610 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:153 Resp: 190677
Ion Ratio Lower Upper
153 100
152 46.7 37.5 56.3
154 86.7 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 2.505 ng/ul
RT: 14.645 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:184 Resp: 13029
Ion Ratio Lower Upper
184 100
63 59.7 58.6 87.8
154 67.1 46.2 69.2









#55
4-Nitrophenol
Concen: 3.659 ng/ul
RT: 14.733 min Scan# 1997
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:109 Resp: 25386
Ion Ratio Lower Upper
109 100
139 120.8 104.0 156.0
65 130.7 117.5 176.3



#56
Dibenzofuran
Concen: 4.836 ng/ul
RT: 14.945 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:168 Resp: 272305
Ion Ratio Lower Upper
168 100
139 37.6 27.8 41.6









#57
2,4-Dinitrotoluene
Concen: 4.132 ng/ul
RT: 14.898 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:165 Resp: 49721
Ion Ratio Lower Upper
165 100
63 46.8 36.0 54.0
182 2.5 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.269 ng/ul
RT: 15.169 min Scan# 2071
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:232 Resp: 42820
Ion Ratio Lower Upper
232 100
131 42.6 31.4 47.0
130 2.3 1.8 2.6
166 30.6 21.8 32.8









#59
Diethylphthalate
Concen: 4.501 ng/ul
RT: 15.357 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:149 Resp: 204370
Ion Ratio Lower Upper
149 100
177 22.1 18.6 28.0
150 12.4 10.0 15.0



#61
Fluorene
Concen: 4.729 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:166 Resp: 218033
Ion Ratio Lower Upper
166 100
165 96.7 79.0 118.4
167 13.0 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 4.621 ng/ul
RT: 15.586 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:204 Resp: 109882
Ion Ratio Lower Upper
204 100
206 32.5 26.8 40.2
141 61.6 43.1 64.7



#63
4-Nitroaniline
Concen: 4.953 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:138 Resp: 32485
Ion Ratio Lower Upper
138 100
92 51.3 45.6 68.4
108 83.3 68.0 102.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 3.128 ng/ul
 RT: 15.657 min Scan# 2155
 Delta R.T. -0.006 min
 Lab File: BN018618.D
 Acq: 16 Feb 2022 13:02




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005231


Tgt Ion:198 Resp: 25449
 Ion Ratio Lower Upper
 198 100
 182 6.4 3.4 5.2#
 77 30.5 21.1 31.7



#67
 N-Nitrosodiphenylamine
 Concen: 4.105 ng/ul
 RT: 15.792 min Scan# 2177
 Delta R.T. -0.006 min
 Lab File: BN018618.D
 Acq: 16 Feb 2022 13:02

Tgt Ion:169 Resp: 178174
 Ion Ratio Lower Upper
 169 100
 168 67.3 52.2 78.4
 167 36.5 27.4 41.2









#68
4-Bromophenyl-phenylether
Concen: 4.678 ng/ul
RT: 16.480 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:248 Resp: 61846
Ion Ratio Lower Upper
248 100
250 92.7 72.3 108.5
141 78.5 53.7 80.5



#69
Hexachlorobenzene
Concen: 4.083 ng/ul
RT: 16.598 min Scan# 2314
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:284 Resp: 72247
Ion Ratio Lower Upper
284 100
142 37.8 24.3 36.5#
249 34.8 23.8 35.6









#70
 Atrazine
 Concen: 3.545 ng/ul
 RT: 16.739 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN018618.D
 Acq: 16 Feb 2022 13:02




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005231


Tgt Ion:200 Resp: 53521
 Ion Ratio Lower Upper
 200 100
 173 25.5 20.2 30.2
 215 50.6 39.7 59.5



#71
 Pentachlorophenol
 Concen: 2.785 ng/ul
 RT: 16.939 min Scan# 2372
 Delta R.T. -0.000 min
 Lab File: BN018618.D
 Acq: 16 Feb 2022 13:02

Tgt Ion:266 Resp: 28110
 Ion Ratio Lower Upper
 266 100
 264 67.8 52.0 78.0
 268 62.1 53.5 80.3









#72
Phenanthrene
Concen: 4.507 ng/u1
RT: 17.327 min Scan# 2438
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:178 Resp: 337850
Ion Ratio Lower Upper
178 100
179 14.3 12.2 18.4
176 18.1 14.6 21.8



#74
Anthracene
Concen: 4.550 ng/u1
RT: 17.416 min Scan# 2453
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02


Tgt Ion:178 Resp: 336950
Ion Ratio Lower Upper
178 100
179 15.5 12.3 18.5
176 18.5 14.7 22.1








#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.067 ng/uL
RT: 13.351 min Scan# 1762
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Instrument :
BNA_N
ClientSampleId :
SSTD005231







Tgt Ion:216 Resp: 103371
Ion Ratio Lower Upper
216 100
214 76.8 63.1 94.7
218 48.6 39.8 59.6



#76
Pentachlorobenzene
Concen: 4.136 ng/uL
RT: 14.869 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:250 Resp: 92020
Ion Ratio Lower Upper
250 100
248 64.6 49.5 74.3
252 64.6 52.7 79.1









#77
 Carbazole
 Concen: 4.202 ng/ul
 RT: 17.680 min Scan# 2499
 Delta R.T. -0.006 min
 Lab File: BN018618.D
 Acq: 16 Feb 2022 13:02




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005231


Tgt Ion:167 Resp: 271982
 Ion Ratio Lower Upper
 167 100
 166 21.6 16.9 25.3
 139 13.3 9.6 14.4



#78
 Di-n-butylphthalate
 Concen: 4.005 ng/ul
 RT: 18.257 min Scan# 2596
 Delta R.T. -0.000 min
 Lab File: BN018618.D
 Acq: 16 Feb 2022 13:02

Tgt Ion:149 Resp: 276610
 Ion Ratio Lower Upper
 149 100
 150 8.6 7.3 10.9
 104 4.8 4.2 6.2









#80
Fluoranthene
Concen: 5.706 ng/u1
RT: 19.339 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:202 Resp: 350914
Ion Ratio Lower Upper
202 100
101 10.9 8.5 12.7
100 8.8 6.4 9.6



#82
Pyrene
Concen: 5.508 ng/u1
RT: 19.704 min Scan# 2842
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:202 Resp: 372544
Ion Ratio Lower Upper
202 100
101 12.5 9.6 14.4
100 10.0 7.8 11.8









#83
Butylbenzylphthalate
Concen: 5.512 ng/ul
RT: 20.598 min Scan# 2994
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:149 Resp: 92111
Ion Ratio Lower Upper
149 100
91 71.4 63.7 95.5
206 20.4 19.1 28.7



#84
3,3'-Dichlorobenzidine
Concen: 3.802 ng/ul
RT: 21.380 min Scan# 3127
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:252 Resp: 70963
Ion Ratio Lower Upper
252 100
254 63.5 51.5 77.3
126 10.3 7.9 11.9





#85

Benzo(a)anthracene

Concen: 4.971 ng/ul

RT: 21.451 min Scan# 3139

Delta R.T. -0.000 min

Lab File: BN018618.D

Acq: 16 Feb 2022 13:02

Instrument :

BNA_N

ClientSampleId :

SSTD005231





Tgt Ion:228 Resp: 322242

Ion Ratio Lower Upper

228 100

229 19.1 15.4 23.2

226 26.6 21.0 31.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 3.576 ng/ul

RT: 21.380 min Scan# 3127

Delta R.T. -0.000 min

Lab File: BN018618.D

Acq: 16 Feb 2022 13:02




Tgt Ion:149 Resp: 137765


Ion Ratio Lower Upper

149 100

167 29.0 23.4 35.0

279 4.4 4.1 6.1









#87
Chrysene
Concen: 6.156 ng/ul
RT: 21.498 min Scan# 3147
Delta R.T. -0.000 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:228 Resp: 337397
Ion Ratio Lower Upper
228 100
226 31.2 23.5 35.3
229 19.8 15.0 22.4



#89
Di-n-octyl phthalate
Concen: 2.976 ng/ul
RT: 22.298 min Scan# 3283
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:149 Resp: 212523









#90
Benzo(b)fluoranthene
Concen: 4.114 ng/ul
RT: 23.127 min Scan# 3425
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:252 Resp: 301221
Ion Ratio Lower Upper
252 100
253 21.0 17.8 26.6
125 9.6 7.2 10.8



#91
Benzo(k)fluoranthene
Concen: 4.599 ng/ul
RT: 23.174 min Scan# 3432
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:252 Resp: 297988
Ion Ratio Lower Upper
252 100
253 22.5 17.3 25.9
125 9.1 6.9 10.3









#93
Benzo(a)pyrene
Concen: 4.296 ng/ul
RT: 23.751 min Scan# 31
Delta R.T. -0.006 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02




Instrument :
BNA_N
ClientSampleId :
SSTD005231


Tgt Ion:252 Resp: 297022
Ion Ratio Lower Upper
252 100
253 21.5 19.3 28.9
125 10.9 8.0 12.0



#94
Indeno(1,2,3-cd)pyrene
Concen: 4.080 ng/ul
RT: 26.339 min Scan# 3970
Delta R.T. -0.018 min
Lab File: BN018618.D
Acq: 16 Feb 2022 13:02

Tgt Ion:276 Resp: 320348
Ion Ratio Lower Upper
276 100
138 24.3 16.6 24.8
277 24.3 19.1 28.7





#95

Dibenzo(a,h)anthracene

Concen: 4.153 ng/ul

RT: 26.356 min Scan# 3975

Delta R.T. -0.012 min

Lab File: BN018618.D

Acq: 16 Feb 2022 13:02

Instrument :

BNA_N

ClientSampleId :

SSTD005231





Tgt Ion:278 Resp: 276254

Ion Ratio Lower Upper

278 100

139 16.8 11.7 17.5

279 23.3 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 4.029 ng/ul

RT: 27.109 min Scan# 4101

Delta R.T. -0.006 min

Lab File: BN018618.D

Acq: 16 Feb 2022 13:02

Tgt Ion:276 Resp: 269728

Ion Ratio Lower Upper

276 100

138 22.8 15.6 23.4

277 23.2 17.8 26.6

