

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021622\
 Data File : BN018619.D
 Acq On : 16 Feb 2022 13:38
 Operator : CG/JU
 Sample : SSTD01032
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010232

Quant Time: Feb 16 15:21:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 16 15:18:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	182195	20.000	ng/u1	0.00
20) Naphthalene-d8	10.722	136	691350	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.545	164	393011	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.286	188	746345	20.000	ng/u1	0.00
79) Chrysene-d12	21.463	240	694573	20.000	ng/u1	0.00
88) Perylene-d12	23.857	264	702639	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.287	96	21586	5.221	ng/uL	0.00
4) Pyridine-d5	3.705	84	139896	9.984	ng/u1	0.00
7) Phenol-d5	7.063	99	156125	10.574	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.234	67	101546	11.325	ng/u1	0.00
11) 2-Chlorophenol-d4	7.440	132	126129	10.756	ng/u1	0.00
15) 4-Methylphenol-d8	8.616	113	117035	10.281	ng/u1	0.00
21) Nitrobenzene-d5	9.069	128	54192	10.433	ng/u1	0.00
24) 2-Nitrophenol-d4	9.799	143	56359	10.052	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.340	165	112308	10.414	ng/u1	0.00
31) 4-Chloroaniline-d4	10.852	131	152396	10.800	ng/u1	0.00
46) Dimethylphthalate-d6	13.951	166	297833	10.592	ng/u1	0.00
49) Acenaphthylene-d8	14.240	160	393489	10.219	ng/u1	0.00
54) 4-Nitrophenol-d4	14.722	143	45162	9.534	ng/u1	0.00
60) Fluorene-d10	15.540	176	266103	11.191	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	37516	8.329	ng/u1	0.00
73) Anthracene-d10	17.386	188	379675	10.423	ng/u1	0.00
81) Pyrene-d10	19.675	212	417172	12.291	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.704	264	376436	9.954	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.322	88	22613	4.259	ng/uL	90
5) Pyridine	3.722	79	147954	10.759	ng/u1	97
6) Benzaldehyde	7.040	77	98695	11.932	ng/u1	96
8) Phenol	7.093	94	166484	11.128	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.328	93	136404	11.063	ng/u1	98
12) 2-Chlorophenol	7.475	128	135614	11.218	ng/u1	95
13) 2-Methylphenol	8.352	108	117406	10.578	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.446	45	182724	11.052	ng/u1	99
16) Acetophenone	8.734	105	197167	11.531	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.716	70	92197	10.597	ng/u1#	94
18) 4-Methylphenol	8.681	108	129237	10.764	ng/u1	96
19) Hexachloroethane	9.005	117	57146	11.280	ng/u1	91
22) Nitrobenzene	9.116	77	145935	11.125	ng/u1	98
23) Isophorone	9.640	82	238040	10.594	ng/u1	95
25) 2-Nitrophenol	9.828	139	65133	10.525	ng/u1	94
26) 2,4-Dimethylphenol	9.887	107	144595	11.040	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.128	93	164206	10.871	ng/u1	99
29) 2,4-Dichlorophenol	10.363	162	117878	10.505	ng/u1	97
30) Naphthalene	10.775	128	411596	10.804	ng/u1	100
32) 4-Chloroaniline	10.875	127	161525	11.101	ng/u1	98
33) Hexachlorobutadiene	11.069	225	86378	10.989	ng/u1	97
34) Caprolactam	11.610	113	23321	7.957	ng/u1	91
35) 4-Chloro-3-methylphenol	11.993	107	111730	10.037	ng/u1	98
36) 2-Methylnaphthalene	12.381	142	268699	11.082	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021622\
 Data File : BN018619.D
 Acq On : 16 Feb 2022 13:38
 Operator : CG/JU
 Sample : SST01032
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST010232

Quant Time: Feb 16 15:21:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 16 15:18:48 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.598	142	273523	10.678	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.751	216	146562	10.361	ng/ul	94
40) Hexachlorocyclopentadiene	12.734	237	88782	10.216	ng/ul	99
41) 2,4,6-Trichlorophenol	12.981	196	77925	9.975	ng/ul	95
42) 2,4,5-Trichlorophenol	13.051	196	85755	10.230	ng/ul	98
43) 1,1'-Biphenyl	13.387	154	353020	10.532	ng/ul	97
44) 2-Chloronaphthalene	13.428	162	272046	10.500	ng/ul	96
45) 2-Nitroaniline	13.622	65	59090	9.939	ng/ul	92
47) Dimethylphthalate	13.998	163	302232	10.873	ng/ul	99
48) 2,6-Dinitrotoluene	14.110	165	51867	10.220	ng/ul	94
50) Acenaphthylene	14.269	152	419721	11.199	ng/ul	99
51) 3-Nitroaniline	14.440	138	52510	10.406	ng/ul	91
52) Acenaphthene	14.610	153	275192	11.354	ng/ul	99
53) 2,4-Dinitrophenol	14.645	184	20712	6.395	ng/ul#	88
55) 4-Nitrophenol	14.734	109	41251	9.549	ng/ul	91
56) Dibenzofuran	14.945	168	392637	11.200	ng/ul	95
57) 2,4-Dinitrotoluene	14.898	165	77527	10.350	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.169	232	65748	10.528	ng/ul#	92
59) Diethylphthalate	15.357	149	288163	10.195	ng/ul	98
61) Fluorene	15.592	166	303162	10.562	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.587	204	155264	10.489	ng/ul	94
63) 4-Nitroaniline	15.592	138	51953	12.723	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.657	198	41862	8.545	ng/ul#	97
67) N-Nitrosodiphenylamine	15.792	169	252247	9.652	ng/ul	98
68) 4-Bromophenyl-phenylether	16.475	248	88011	11.057	ng/ul#	86
69) Hexachlorobenzene	16.598	284	100243	9.410	ng/ul#	89
70) Atrazine	16.739	200	77676	8.547	ng/ul	98
71) Pentachlorophenol	16.939	266	42941	7.066	ng/ul	96
72) Phenanthrene	17.328	178	463517	10.270	ng/ul	99
74) Anthracene	17.416	178	465795	10.447	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.351	216	156073	10.199	ng/ul	99
76) Pentachlorobenzene	14.869	250	127985	9.554	ng/ul	96
77) Carbazole	17.681	167	394994	10.136	ng/ul	98
78) Di-n-butylphthalate	18.251	149	404032	9.717	ng/ul	99
80) Fluoranthene	19.339	202	496004	12.260	ng/ul	99
82) Pyrene	19.704	202	531680	11.950	ng/ul	100
83) Butylbenzylphthalate	20.598	149	148384	13.499	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.380	252	138372	11.270	ng/ul	100
85) Benzo(a)anthracene	21.445	228	506505	11.877	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.380	149	246757	9.738	ng/ul	96
87) Chrysene	21.498	228	504033	13.980	ng/ul	99
89) Di-n-octyl phthalate	22.304	149	384864	7.787	ng/ul	100
90) Benzo(b)fluoranthene	23.127	252	486922	9.607	ng/ul	97
91) Benzo(k)fluoranthene	23.174	252	494787	11.031	ng/ul	98
93) Benzo(a)pyrene	23.751	252	493195	10.305	ng/ul	94
94) Indeno(1,2,3-cd)pyrene	26.351	276	552565	10.166	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.362	278	485439	10.542	ng/ul	99
96) Benzo(g,h,i)perylene	27.109	276	486546	10.500	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021622\
 Data File : BN018619.D
 Acq On : 16 Feb 2022 13:38
 Operator : CG/JU
 Sample : SST001032
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST001032

Quant Time: Feb 16 15:21:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 16 15:18:48 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 4.259 ng/uL
RT: 3.322 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion: 88 Resp: 22613
Ion Ratio Lower Upper
88 100
43 30.4 21.9 32.9
58 79.9 56.3 84.5



#5
Pyridine
Concen: 10.759 ng/uL
RT: 3.722 min Scan# 125
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion: 79 Resp: 147954
Ion Ratio Lower Upper
79 100
52 76.5 62.9 94.3
51 37.7 27.8 41.8









#6
Benzaldehyde
Concen: 11.932 ng/ul
RT: 7.040 min Scan# 690
Delta R.T. -0.006 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion: 77 Resp: 98695
Ion Ratio Lower Upper
77 100
105 96.4 73.5 110.3
106 90.4 69.9 104.9



#8
Phenol
Concen: 11.128 ng/ul
RT: 7.093 min Scan# 698
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion: 94 Resp: 166484
Ion Ratio Lower Upper
94 100
65 29.5 21.8 32.6
66 38.9 30.2 45.4





#10

Bis(2-Chloroethyl)ether

Concen: 11.063 ng/ul

RT: 7.328 min Scan# 710

Delta R.T. -0.000 min

Lab File: BN018619.D

Acq: 16 Feb 2022 13:38

Instrument :





BNA_N

ClientSampleId :

SSTD010232

Tgt Ion: 93 Resp: 136404

Ion	Ratio	Lower	Upper
93	100		
63	76.9	63.4	95.0
95	31.6	25.9	38.9



#12

2-Chlorophenol

Concen: 11.218 ng/ul

RT: 7.475 min Scan# 763




Delta R.T. -0.000 min


Lab File: BN018619.D

Acq: 16 Feb 2022 13:38

Tgt Ion:128 Resp: 135614

Ion	Ratio	Lower	Upper
128	100		
64	49.3	44.2	66.2
130	31.5	25.0	37.6





#13

2-Methylphenol

Concen: 10.578 ng/ul

RT: 8.352 min Scan# 912

Delta R.T. -0.000 min

Lab File: BN018619.D

Acq: 16 Feb 2022 13:38

Instrument :

BNA_N

ClientSampleId :





SSTD010232

Tgt Ion:108 Resp: 117406

Ion Ratio Lower Upper

108 100

107 86.4 69.5 104.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 11.052 ng/ul

RT: 8.446 min Scan# 928

Delta R.T. -0.000 min

Lab File: BN018619.D

Acq: 16 Feb 2022 13:38




Tgt Ion: 45 Resp: 182724


Ion Ratio Lower Upper

45 100

77 14.1 11.9 17.9

79 11.0 8.6 12.8









#16
 Acetophenone
 Concen: 11.531 ng/ul
 RT: 8.734 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010232


Tgt Ion:105 Resp: 197167
 Ion Ratio Lower Upper
 105 100
 77 77.7 64.8 97.2
 51 27.1 22.8 34.2



#17
 N-Nitroso-di-n-propylamine
 Concen: 10.597 ng/ul
 RT: 8.716 min Scan# 974
 Delta R.T. -0.006 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38

Tgt Ion: 70 Resp: 92197
 Ion Ratio Lower Upper
 70 100
 42 57.3 42.4 63.6
 101 10.1 6.6 10.0#
 130 19.0 13.6 20.4









#18
4-Methylphenol
Concen: 10.764 ng/ul
RT: 8.681 min Scan# 968
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:108 Resp: 129237
Ion Ratio Lower Upper
108 100
107 107.7 89.4 134.0



#19
Hexachloroethane
Concen: 11.280 ng/ul
RT: 9.005 min Scan# 1023
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:117 Resp: 57146
Ion Ratio Lower Upper
117 100
201 98.5 85.9 128.9
199 60.0 54.5 81.7









#22
Nitrobenzene
Concen: 11.125 ng/ul
RT: 9.116 min Scan# 1042
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion: 77 Resp: 145935
Ion Ratio Lower Upper
77 100
123 41.8 32.0 48.0
65 13.5 10.7 16.1



#23
Isophorone
Concen: 10.594 ng/ul
RT: 9.640 min Scan# 1131
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion: 82 Resp: 238040
Ion Ratio Lower Upper
82 100
95 7.6 5.7 8.5
138 17.5 11.9 17.9









#25
2-Nitrophenol
Concen: 10.525 ng/ul
RT: 9.828 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:139 Resp: 65133
Ion Ratio Lower Upper
139 100
65 53.4 47.7 71.5
109 31.8 27.0 40.4



#26
2,4-Dimethylphenol
Concen: 11.040 ng/ul
RT: 9.887 min Scan# 1173
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:107 Resp: 144595
Ion Ratio Lower Upper
107 100
121 47.8 38.9 58.3
122 81.9 68.6 103.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 10.871 ng/ul
 RT: 10.128 min Scan# 1214
 Delta R.T. -0.000 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010232


Tgt Ion: 93 Resp: 164206
 Ion Ratio Lower Upper
 93 100
 95 32.7 25.7 38.5
 123 11.2 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 10.505 ng/ul
 RT: 10.363 min Scan# 1254
 Delta R.T. -0.000 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38

Tgt Ion: 162 Resp: 117878
 Ion Ratio Lower Upper
 162 100
 164 66.5 51.8 77.6
 98 36.9 31.0 46.6









#30
Naphthalene
Concen: 10.804 ng/ul
RT: 10.775 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:128 Resp: 411596
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 13.1 10.6 16.0



#32
4-Chloroaniline
Concen: 11.101 ng/ul
RT: 10.875 min Scan# 1341
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:127 Resp: 161525
Ion Ratio Lower Upper
127 100
129 32.3 25.0 37.6









#33
Hexachlorobutadiene
Concen: 10.989 ng/ul
RT: 11.069 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:225 Resp: 86378
Ion Ratio Lower Upper
225 100
223 60.6 49.7 74.5
227 65.3 49.5 74.3



#34
Caprolactam
Concen: 7.957 ng/ul
RT: 11.610 min Scan# 1466
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:113 Resp: 23321
Ion Ratio Lower Upper
113 100
55 166.8 147.4 221.2
56 129.5 107.6 161.4









#35
4-Chloro-3-methylphenol
Concen: 10.037 ng/ul
RT: 11.993 min Scan# 1531
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:107 Resp: 111730
Ion Ratio Lower Upper
107 100
144 25.9 19.4 29.0
142 77.1 62.3 93.5



#36
2-Methylnaphthalene
Concen: 11.082 ng/ul
RT: 12.381 min Scan# 1597
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:142 Resp: 268699
Ion Ratio Lower Upper
142 100
141 91.4 70.2 105.4









#37
1-Methylnaphthalene
Concen: 10.678 ng/ul
RT: 12.598 min Scan# 1
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:142 Resp: 273523
Ion Ratio Lower Upper
142 100
141 92.7 72.2 108.4
116 4.2 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 10.361 ng/ul
RT: 12.751 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:216 Resp: 146562
Ion Ratio Lower Upper
216 100
214 74.6 64.4 96.6
179 21.8 15.1 22.7
108 16.9 12.0 18.0





#40

Hexachlorocyclopentadiene

Concen: 10.216 ng/ul

RT: 12.734 min Scan# 1699

Delta R.T. -0.006 min

Lab File: BN018619.D

Acq: 16 Feb 2022 13:38

Instrument :

BNA_N

ClientSampleId :

SSTD010232





Tgt Ion:237 Resp: 88782

Ion Ratio Lower Upper

237 100

235 61.8 49.1 73.7

272 12.2 10.6 16.0



#41

2,4,6-Trichlorophenol

Concen: 9.975 ng/ul

RT: 12.981 min Scan# 1699

Delta R.T. -0.006 min

Lab File: BN018619.D

Acq: 16 Feb 2022 13:38




Tgt Ion:196 Resp: 77925


Ion Ratio Lower Upper

196 100

198 99.1 74.5 111.7

200 30.5 24.8 37.2









#42
 2,4,5-Trichlorophenol
 Concen: 10.230 ng/ul
 RT: 13.051 min Scan# 1711
 Delta R.T. -0.000 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010232


Tgt Ion:196 Resp: 85755
 Ion Ratio Lower Upper
 196 100
 198 92.6 75.3 112.9
 200 31.5 24.2 36.2



#43
 1,1'-Biphenyl
 Concen: 10.532 ng/ul
 RT: 13.387 min Scan# 1768
 Delta R.T. -0.000 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38

Tgt Ion:154 Resp: 353020
 Ion Ratio Lower Upper
 154 100
 153 41.5 32.4 48.6
 76 12.8 12.0 18.0









#44
2-Chloronaphthalene
Concen: 10.500 ng/ul
RT: 13.428 min Scan# 1775
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:162 Resp: 272046
Ion Ratio Lower Upper
162 100
164 33.3 26.8 40.2
127 40.6 28.8 43.2



#45
2-Nitroaniline
Concen: 9.939 ng/ul
RT: 13.622 min Scan# 1808
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion: 65 Resp: 59090
Ion Ratio Lower Upper
65 100
92 71.6 53.0 79.6
138 100.2 73.8 110.8









#47
 Dimethylphthalate
 Concen: 10.873 ng/ul
 RT: 13.998 min Scan# 1872
 Delta R.T. -0.000 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38




Instrument :
 BNA_N
 ClientSampleId :
 SST010232


Tgt Ion:163 Resp: 302232
 Ion Ratio Lower Upper
 163 100
 194 4.3 3.6 5.4
 164 10.1 8.2 12.4



#48
 2,6-Dinitrotoluene
 Concen: 10.220 ng/ul
 RT: 14.110 min Scan# 1891
 Delta R.T. -0.006 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38

Tgt Ion:165 Resp: 51867
 Ion Ratio Lower Upper
 165 100
 89 66.4 49.3 73.9
 121 23.3 16.8 25.2









#50
Acenaphthylene
Concen: 11.199 ng/ul
RT: 14.269 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:152 Resp: 419721
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2
153 12.6 10.6 15.8



#51
3-Nitroaniline
Concen: 10.406 ng/ul
RT: 14.440 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:138 Resp: 52510
Ion Ratio Lower Upper
138 100
108 12.1 10.6 15.8
92 123.6 107.9 161.9









#52
Acenaphthene
Concen: 11.354 ng/ul
RT: 14.610 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:153 Resp: 275192
Ion Ratio Lower Upper
153 100
152 46.9 37.5 56.3
154 88.6 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 6.395 ng/ul
RT: 14.645 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:184 Resp: 20712
Ion Ratio Lower Upper
184 100
63 69.6 58.6 87.8
154 73.5 46.2 69.2#









#55
4-Nitrophenol
Concen: 9.549 ng/ul
RT: 14.734 min Scan# 1997
Delta R.T. -0.006 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:109 Resp: 41251
Ion Ratio Lower Upper
109 100
139 123.1 104.0 156.0
65 131.4 117.5 176.3



#56
Dibenzofuran
Concen: 11.200 ng/ul
RT: 14.945 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:168 Resp: 392637
Ion Ratio Lower Upper
168 100
139 37.6 27.8 41.6









#57
2,4-Dinitrotoluene
Concen: 10.350 ng/ul
RT: 14.898 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:165 Resp: 77527
Ion Ratio Lower Upper
165 100
63 43.6 36.0 54.0
182 3.1 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 10.528 ng/ul
RT: 15.169 min Scan# 2071
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:232 Resp: 65748
Ion Ratio Lower Upper
232 100
131 45.7 31.4 47.0
130 2.7 1.8 2.6#
166 29.3 21.8 32.8









#59
Diethylphthalate
Concen: 10.195 ng/ul
RT: 15.357 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:149 Resp: 288163
Ion Ratio Lower Upper
149 100
177 21.8 18.6 28.0
150 12.2 10.0 15.0



#61
Fluorene
Concen: 10.562 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:166 Resp: 303162
Ion Ratio Lower Upper
166 100
165 98.1 79.0 118.4
167 14.1 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 10.489 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:204 Resp: 155264
Ion Ratio Lower Upper
204 100
206 33.4 26.8 40.2
141 61.1 43.1 64.7



#63
4-Nitroaniline
Concen: 12.723 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:138 Resp: 51953
Ion Ratio Lower Upper
138 100
92 55.9 45.6 68.4
108 78.6 68.0 102.0









#66
4,6-Dinitro-2-methylphenol
Concen: 8.545 ng/ul
RT: 15.657 min Scan# 2154
Delta R.T. -0.006 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:198 Resp: 41862
Ion Ratio Lower Upper
198 100
182 5.3 3.4 5.2#
77 25.0 21.1 31.7



#67
N-Nitrosodiphenylamine
Concen: 9.652 ng/ul
RT: 15.792 min Scan# 2177
Delta R.T. -0.006 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:169 Resp: 252247
Ion Ratio Lower Upper
169 100
168 64.8 52.2 78.4
167 37.0 27.4 41.2









#68
4-Bromophenyl-phenylether
Concen: 11.057 ng/ul
RT: 16.475 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:248 Resp: 88011
Ion Ratio Lower Upper
248 100
250 98.9 72.3 108.5
141 82.8 53.7 80.5#



#69
Hexachlorobenzene
Concen: 9.410 ng/ul
RT: 16.598 min Scan# 2314
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:284 Resp: 100243
Ion Ratio Lower Upper
284 100
142 37.9 24.3 36.5#
249 33.6 23.8 35.6









#70
 Atrazine
 Concen: 8.547 ng/ul
 RT: 16.739 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010232


Tgt Ion:200 Resp: 77676
 Ion Ratio Lower Upper
 200 100
 173 27.1 20.2 30.2
 215 50.0 39.7 59.5



#71
 Pentachlorophenol
 Concen: 7.066 ng/ul
 RT: 16.939 min Scan# 2372
 Delta R.T. -0.000 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38

Tgt Ion:266 Resp: 42941
 Ion Ratio Lower Upper
 266 100
 264 67.9 52.0 78.0
 268 69.9 53.5 80.3









#72
 Phenanthrene
 Concen: 10.270 ng/ul
 RT: 17.328 min Scan# 2438
 Delta R.T. -0.000 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010232

Tgt Ion:178 Resp: 463517




Ion	Ratio	Lower	Upper
178	100		
179	15.0	12.2	18.4
176	18.6	14.6	21.8




#74
 Anthracene
 Concen: 10.447 ng/ul
 RT: 17.416 min Scan# 2453
 Delta R.T. -0.006 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38

Tgt Ion:178 Resp: 465795

Ion	Ratio	Lower	Upper
178	100		
179	15.9	12.3	18.5
176	18.3	14.7	22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 10.199 ng/uL
RT: 13.351 min Scan# 1762
Delta R.T. -0.006 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:216 Resp: 156073
Ion Ratio Lower Upper
216 100
214 78.0 63.1 94.7
218 48.7 39.8 59.6



#76
Pentachlorobenzene
Concen: 9.554 ng/uL
RT: 14.869 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:250 Resp: 127985
Ion Ratio Lower Upper
250 100
248 65.4 49.5 74.3
252 62.9 52.7 79.1









#77
 Carbazole
 Concen: 10.136 ng/ul
 RT: 17.681 min Scan# 2498
 Delta R.T. -0.006 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38




Instrument :
 BNA_N
 ClientSampleId :
 SST010232


Tgt Ion:167 Resp: 394994
 Ion Ratio Lower Upper
 167 100
 166 21.4 16.9 25.3
 139 13.6 9.6 14.4



#78
 Di-n-butylphthalate
 Concen: 9.717 ng/ul
 RT: 18.251 min Scan# 2595
 Delta R.T. -0.006 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38

Tgt Ion:149 Resp: 404032
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.3 10.9
 104 5.6 4.2 6.2









#80
Fluoranthene
Concen: 12.260 ng/ul
RT: 19.339 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:202 Resp: 496004
Ion Ratio Lower Upper
202 100
101 10.9 8.5 12.7
100 8.1 6.4 9.6



#82
Pyrene
Concen: 11.950 ng/ul
RT: 19.704 min Scan# 2842
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:202 Resp: 531680
Ion Ratio Lower Upper
202 100
101 12.2 9.6 14.4
100 9.9 7.8 11.8









#83
Butylbenzylphthalate
Concen: 13.499 ng/ul
RT: 20.598 min Scan# 2994
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:149 Resp: 148384
Ion Ratio Lower Upper
149 100
91 73.0 63.7 95.5
206 20.9 19.1 28.7



#84
3,3'-Dichlorobenzidine
Concen: 11.270 ng/ul
RT: 21.380 min Scan# 3127
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:252 Resp: 138372
Ion Ratio Lower Upper
252 100
254 64.3 51.5 77.3
126 10.8 7.9 11.9









#85
Benzo(a)anthracene
Concen: 11.877 ng/ul
RT: 21.445 min Scan# 3139
Delta R.T. -0.006 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:228 Resp: 506505
Ion Ratio Lower Upper
228 100
229 19.2 15.4 23.2
226 26.9 21.0 31.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 9.738 ng/ul
RT: 21.380 min Scan# 3127
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:149 Resp: 246757
Ion Ratio Lower Upper
149 100
167 27.2 23.4 35.0
279 4.3 4.1 6.1









#87
Chrysene
Concen: 13.980 ng/ul
RT: 21.498 min Scan# 3147
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:228 Resp: 504033
Ion Ratio Lower Upper
228 100
226 29.3 23.5 35.3
229 19.3 15.0 22.4



#89
Di-n-octyl phthalate
Concen: 7.787 ng/ul
RT: 22.304 min Scan# 3284
Delta R.T. -0.000 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:149 Resp: 384864









#90
Benzo(b)fluoranthene
Concen: 9.607 ng/ul
RT: 23.127 min Scan# 3425
Delta R.T. -0.006 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38




Instrument :
BNA_N
ClientSampleId :
SSTD010232


Tgt Ion:252 Resp: 486922
Ion Ratio Lower Upper
252 100
253 23.6 17.8 26.6
125 10.7 7.2 10.8



#91
Benzo(k)fluoranthene
Concen: 11.031 ng/ul
RT: 23.174 min Scan# 3432
Delta R.T. -0.006 min
Lab File: BN018619.D
Acq: 16 Feb 2022 13:38

Tgt Ion:252 Resp: 494787
Ion Ratio Lower Upper
252 100
253 21.9 17.3 25.9
125 10.0 6.9 10.3





#93

Benzo(a)pyrene

Concen: 10.305 ng/ul

RT: 23.751 min Scan# 3531

Delta R.T. -0.006 min

Lab File: BN018619.D

Acq: 16 Feb 2022 13:38

Instrument :

BNA_N

ClientSampleId :

SSTD010232





Tgt Ion:252 Resp: 493195

Ion Ratio Lower Upper

252 100

253 19.7 19.3 28.9

125 10.3 8.0 12.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 10.166 ng/ul

RT: 26.351 min Scan# 3972

Delta R.T. -0.006 min

Lab File: BN018619.D

Acq: 16 Feb 2022 13:38




Tgt Ion:276 Resp: 552565


Ion Ratio Lower Upper

276 100

138 25.5 16.6 24.8#

277 25.1 19.1 28.7









#95
 Dibenzo(a,h)anthracene
 Concen: 10.542 ng/ul
 RT: 26.362 min Scan# 3975
 Delta R.T. -0.006 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010232

Tgt Ion:278 Resp: 485439
 Ion Ratio Lower Upper
 278 100
 139 15.6 11.7 17.5
 279 23.5 19.1 28.7



#96
 Benzo(g,h,i)perylene
 Concen: 10.500 ng/ul
 RT: 27.109 min Scan# 4101
 Delta R.T. -0.006 min
 Lab File: BN018619.D
 Acq: 16 Feb 2022 13:38

Tgt Ion:276 Resp: 486546
 Ion Ratio Lower Upper
 276 100
 138 21.3 15.6 23.4
 277 23.1 17.8 26.6

